

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL01 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103794.D Time Analyzed: 19:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.618	5.518	5.718	499.570	500.000	-0.1
Aroclor-1016-2	5.641	5.541	5.741	493.540	500.000	-1.3
Aroclor-1016-3	5.703	5.603	5.803	498.310	500.000	-0.3
Aroclor-1016-4	5.801	5.701	5.901	507.050	500.000	1.4
Aroclor-1016-5	6.095	5.995	6.195	495.920	500.000	-0.8
Aroclor-1260-1	7.219	7.118	7.318	510.260	500.000	2.1
Aroclor-1260-2	7.475	7.373	7.573	522.010	500.000	4.4
Aroclor-1260-3	7.834	7.732	7.932	537.320	500.000	7.5
Aroclor-1260-4	8.059	7.957	8.157	516.360	500.000	3.3
Aroclor-1260-5	8.380	8.278	8.478	532.810	500.000	6.6
Decachlorobiphenyl	10.208	10.102	10.302	56.450	50.000	12.9
Tetrachloro-m-xylene	4.449	4.351	4.551	50.990	50.000	2.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL01 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103794.D Time Analyzed: 19:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.660	4.571	4.771	503.540	500.000	0.7
Aroclor-1016-2	4.678	4.589	4.789	486.180	500.000	-2.8
Aroclor-1016-3	4.853	4.763	4.963	476.660	500.000	-4.7
Aroclor-1016-4	4.894	4.805	5.005	468.820	500.000	-6.2
Aroclor-1016-5	5.105	5.016	5.216	494.670	500.000	-1.1
Aroclor-1260-1	6.128	6.041	6.241	462.920	500.000	-7.4
Aroclor-1260-2	6.314	6.226	6.426	462.840	500.000	-7.4
Aroclor-1260-3	6.466	6.379	6.579	456.650	500.000	-8.7
Aroclor-1260-4	6.934	6.848	7.048	471.710	500.000	-5.7
Aroclor-1260-5	7.174	7.087	7.287	477.090	500.000	-4.6
Decachlorobiphenyl	8.539	8.456	8.656	47.600	50.000	-4.8
Tetrachloro-m-xylene	3.583	3.491	3.691	51.870	50.000	3.7