

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103805.D Time Analyzed: 23:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.619	5.518	5.718	522.790	500.000	4.6
Aroclor-1016-2	5.641	5.541	5.741	523.040	500.000	4.6
Aroclor-1016-3	5.704	5.603	5.803	557.060	500.000	11.4
Aroclor-1016-4	5.802	5.701	5.901	573.420	500.000	14.7
Aroclor-1016-5	6.096	5.995	6.195	562.340	500.000	12.5
Aroclor-1260-1	7.220	7.118	7.318	544.110	500.000	8.8
Aroclor-1260-2	7.476	7.373	7.573	544.380	500.000	8.9
Aroclor-1260-3	7.835	7.732	7.932	546.960	500.000	9.4
Aroclor-1260-4	8.060	7.957	8.157	525.810	500.000	5.2
Aroclor-1260-5	8.382	8.278	8.478	542.250	500.000	8.5
Decachlorobiphenyl	10.208	10.102	10.302	57.270	50.000	14.5
Tetrachloro-m-xylene	4.450	4.351	4.551	51.950	50.000	3.9

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PO103805.D Time Analyzed: 23:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.660	4.571	4.771	477.810	500.000	-4.4
Aroclor-1016-2	4.679	4.589	4.789	492.970	500.000	-1.4
Aroclor-1016-3	4.853	4.763	4.963	494.290	500.000	-1.1
Aroclor-1016-4	4.894	4.805	5.005	472.360	500.000	-5.5
Aroclor-1016-5	5.106	5.016	5.216	550.280	500.000	10.1
Aroclor-1260-1	6.129	6.041	6.241	489.630	500.000	-2.1
Aroclor-1260-2	6.314	6.226	6.426	474.850	500.000	-5.0
Aroclor-1260-3	6.467	6.379	6.579	465.710	500.000	-6.9
Aroclor-1260-4	6.935	6.848	7.048	477.060	500.000	-4.6
Aroclor-1260-5	7.175	7.087	7.287	482.990	500.000	-3.4
Decachlorobiphenyl	8.540	8.456	8.656	47.960	50.000	-4.1
Tetrachloro-m-xylene	3.584	3.491	3.691	52.550	50.000	5.1