

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO103892.D Time Analyzed: 11:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.617	5.518	5.718	458.260	500.000	-8.3
Aroclor-1016-2	5.640	5.541	5.741	458.200	500.000	-8.4
Aroclor-1016-3	5.702	5.603	5.803	473.080	500.000	-5.4
Aroclor-1016-4	5.800	5.701	5.901	486.130	500.000	-2.8
Aroclor-1016-5	6.094	5.995	6.195	474.970	500.000	-5.0
Aroclor-1260-1	7.218	7.118	7.318	488.440	500.000	-2.3
Aroclor-1260-2	7.473	7.373	7.573	491.070	500.000	-1.8
Aroclor-1260-3	7.832	7.732	7.932	496.450	500.000	-0.7
Aroclor-1260-4	8.057	7.957	8.157	476.030	500.000	-4.8
Aroclor-1260-5	8.378	8.278	8.478	489.940	500.000	-2.0
Decachlorobiphenyl	10.204	10.102	10.302	51.630	50.000	3.3
Tetrachloro-m-xylene	4.448	4.351	4.551	46.100	50.000	-7.8

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO103892.D Time Analyzed: 11:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.659	4.571	4.771	447.770	500.000	-10.4
Aroclor-1016-2	4.677	4.589	4.789	440.900	500.000	-11.8
Aroclor-1016-3	4.851	4.763	4.963	435.940	500.000	-12.8
Aroclor-1016-4	4.893	4.805	5.005	426.850	500.000	-14.6
Aroclor-1016-5	5.103	5.016	5.216	477.750	500.000	-4.5
Aroclor-1260-1	6.126	6.041	6.241	432.240	500.000	-13.6
Aroclor-1260-2	6.311	6.226	6.426	425.410	500.000	-14.9
Aroclor-1260-3	6.464	6.379	6.579	413.620	500.000	-17.3
Aroclor-1260-4	6.932	6.848	7.048	426.280	500.000	-14.7
Aroclor-1260-5	7.171	7.087	7.287	433.750	500.000	-13.3
Decachlorobiphenyl	8.536	8.456	8.656	43.150	50.000	-13.7
Tetrachloro-m-xylene	3.583	3.491	3.691	46.950	50.000	-6.1