

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO103902.D Time Analyzed: 14:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.618	5.518	5.718	483.140	500.000	-3.4
Aroclor-1016-2	5.640	5.541	5.741	485.880	500.000	-2.8
Aroclor-1016-3	5.703	5.603	5.803	486.460	500.000	-2.7
Aroclor-1016-4	5.801	5.701	5.901	488.500	500.000	-2.3
Aroclor-1016-5	6.095	5.995	6.195	478.040	500.000	-4.4
Aroclor-1260-1	7.218	7.118	7.318	505.950	500.000	1.2
Aroclor-1260-2	7.473	7.373	7.573	527.000	500.000	5.4
Aroclor-1260-3	7.833	7.732	7.932	546.110	500.000	9.2
Aroclor-1260-4	8.058	7.957	8.157	524.330	500.000	4.9
Aroclor-1260-5	8.379	8.278	8.478	539.030	500.000	7.8
Decachlorobiphenyl	10.205	10.102	10.302	56.710	50.000	13.4
Tetrachloro-m-xylene	4.449	4.351	4.551	49.460	50.000	-1.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/06/2024 05/06/2024

Client Sample No.: CCAL02 Date Analyzed: 05/17/2024

Lab Sample No.: AR1660CCC500 Data File : PO103902.D Time Analyzed: 14:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.659	4.571	4.771	462.980	500.000	-7.4
Aroclor-1016-2	4.677	4.589	4.789	471.480	500.000	-5.7
Aroclor-1016-3	4.851	4.763	4.963	454.940	500.000	-9.0
Aroclor-1016-4	4.893	4.805	5.005	450.490	500.000	-9.9
Aroclor-1016-5	5.103	5.016	5.216	461.100	500.000	-7.8
Aroclor-1260-1	6.126	6.041	6.241	447.690	500.000	-10.5
Aroclor-1260-2	6.312	6.226	6.426	452.300	500.000	-9.5
Aroclor-1260-3	6.464	6.379	6.579	446.160	500.000	-10.8
Aroclor-1260-4	6.932	6.848	7.048	462.110	500.000	-7.6
Aroclor-1260-5	7.171	7.087	7.287	470.830	500.000	-5.8
Decachlorobiphenyl	8.536	8.456	8.656	46.650	50.000	-6.7
Tetrachloro-m-xylene	3.583	3.491	3.691	49.990	50.000	0.0