

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2024 05/13/2024

Client Sample No.: CCAL01 Date Analyzed: 05/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP064909.D Time Analyzed: 15:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.418	5.329	5.529	500.630	500.000	0.1
Aroclor-1016-2	5.440	5.351	5.551	500.160	500.000	0.0
Aroclor-1016-3	5.502	5.414	5.614	502.290	500.000	0.5
Aroclor-1016-4	5.601	5.512	5.712	528.780	500.000	5.8
Aroclor-1016-5	5.897	5.808	6.008	550.840	500.000	10.2
Aroclor-1260-1	7.029	6.936	7.136	562.040	500.000	12.4
Aroclor-1260-2	7.287	7.193	7.393	561.180	500.000	12.2
Aroclor-1260-3	7.648	7.553	7.753	556.840	500.000	11.4
Aroclor-1260-4	7.873	7.777	7.977	554.650	500.000	10.9
Aroclor-1260-5	8.182	8.086	8.286	536.380	500.000	7.3
Decachlorobiphenyl	9.907	9.810	10.010	48.890	50.000	-2.2
Tetrachloro-m-xylene	4.254	4.160	4.360	46.470	50.000	-7.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2024 05/13/2024

Client Sample No.: CCAL01 Date Analyzed: 05/14/2024

Lab Sample No.: AR1660CCC500 Data File : PP064909.D Time Analyzed: 15:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.425	4.332	4.532	500.340	500.000	0.1
Aroclor-1016-2	4.443	4.350	4.550	492.970	500.000	-1.4
Aroclor-1016-3	4.614	4.521	4.721	500.100	500.000	0.0
Aroclor-1016-4	4.657	4.565	4.765	490.260	500.000	-1.9
Aroclor-1016-5	4.863	4.773	4.973	493.590	500.000	-1.3
Aroclor-1260-1	5.882	5.792	5.992	497.150	500.000	-0.6
Aroclor-1260-2	6.070	5.979	6.179	501.130	500.000	0.2
Aroclor-1260-3	6.220	6.130	6.330	491.920	500.000	-1.6
Aroclor-1260-4	6.689	6.596	6.796	498.460	500.000	-0.3
Aroclor-1260-5	6.931	6.838	7.038	507.080	500.000	1.4
Decachlorobiphenyl	8.264	8.167	8.367	49.050	50.000	-1.9
Tetrachloro-m-xylene	3.370	3.272	3.472	54.790	50.000	9.6