

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2024 05/13/2024

Client Sample No.: CCAL03 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP064954.D Time Analyzed: 10:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.443	5.329	5.529	481.710	500.000	-3.7
Aroclor-1016-2	5.464	5.351	5.551	466.230	500.000	-6.8
Aroclor-1016-3	5.527	5.414	5.614	466.260	500.000	-6.7
Aroclor-1016-4	5.625	5.512	5.712	505.850	500.000	1.2
Aroclor-1016-5	5.920	5.808	6.008	539.850	500.000	8.0
Aroclor-1260-1	7.044	6.936	7.136	558.000	500.000	11.6
Aroclor-1260-2	7.302	7.193	7.393	547.470	500.000	9.5
Aroclor-1260-3	7.660	7.553	7.753	530.150	500.000	6.0
Aroclor-1260-4	7.884	7.777	7.977	525.440	500.000	5.1
Aroclor-1260-5	8.194	8.086	8.286	543.870	500.000	8.8
Decachlorobiphenyl	9.920	9.810	10.010	50.360	50.000	0.7
Tetrachloro-m-xylene	4.278	4.160	4.360	43.770	50.000	-12.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2024 05/13/2024

Client Sample No.: CCAL03 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP064954.D Time Analyzed: 10:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.441	4.332	4.532	464.700	500.000	-7.1
Aroclor-1016-2	4.458	4.350	4.550	448.840	500.000	-10.2
Aroclor-1016-3	4.630	4.521	4.721	473.520	500.000	-5.3
Aroclor-1016-4	4.673	4.565	4.765	449.460	500.000	-10.1
Aroclor-1016-5	4.881	4.773	4.973	443.720	500.000	-11.3
Aroclor-1260-1	5.896	5.792	5.992	479.830	500.000	-4.0
Aroclor-1260-2	6.083	5.979	6.179	482.840	500.000	-3.4
Aroclor-1260-3	6.232	6.130	6.330	479.800	500.000	-4.0
Aroclor-1260-4	6.698	6.596	6.796	485.230	500.000	-3.0
Aroclor-1260-5	6.939	6.838	7.038	511.400	500.000	2.3
Decachlorobiphenyl	8.268	8.167	8.367	52.760	50.000	5.5
Tetrachloro-m-xylene	3.379	3.272	3.472	50.160	50.000	0.3