

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2024 05/13/2024

Client Sample No.: CCAL04 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP064966.D Time Analyzed: 16:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.425	5.329	5.529	486.350	500.000	-2.7
Aroclor-1016-2	5.447	5.351	5.551	480.330	500.000	-3.9
Aroclor-1016-3	5.510	5.414	5.614	492.410	500.000	-1.5
Aroclor-1016-4	5.609	5.512	5.712	514.850	500.000	3.0
Aroclor-1016-5	5.905	5.808	6.008	554.850	500.000	11.0
Aroclor-1260-1	7.034	6.936	7.136	554.650	500.000	10.9
Aroclor-1260-2	7.291	7.193	7.393	551.170	500.000	10.2
Aroclor-1260-3	7.651	7.553	7.753	544.620	500.000	8.9
Aroclor-1260-4	7.876	7.777	7.977	499.610	500.000	-0.1
Aroclor-1260-5	8.186	8.086	8.286	533.970	500.000	6.8
Decachlorobiphenyl	9.912	9.810	10.010	49.730	50.000	-0.5
Tetrachloro-m-xylene	4.258	4.160	4.360	43.560	50.000	-12.9

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/13/2024 05/13/2024

Client Sample No.: CCAL04 Date Analyzed: 05/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP064966.D Time Analyzed: 16:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.428	4.332	4.532	458.240	500.000	-8.4
Aroclor-1016-2	4.445	4.350	4.550	458.530	500.000	-8.3
Aroclor-1016-3	4.617	4.521	4.721	471.700	500.000	-5.7
Aroclor-1016-4	4.660	4.565	4.765	454.990	500.000	-9.0
Aroclor-1016-5	4.868	4.773	4.973	469.950	500.000	-6.0
Aroclor-1260-1	5.887	5.792	5.992	479.260	500.000	-4.1
Aroclor-1260-2	6.075	5.979	6.179	479.390	500.000	-4.1
Aroclor-1260-3	6.225	6.130	6.330	473.160	500.000	-5.4
Aroclor-1260-4	6.693	6.596	6.796	484.840	500.000	-3.0
Aroclor-1260-5	6.934	6.838	7.038	495.830	500.000	-0.8
Decachlorobiphenyl	8.265	8.167	8.367	48.880	50.000	-2.2
Tetrachloro-m-xylene	3.371	3.272	3.472	49.630	50.000	-0.7