

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2024 04/23/2024

Client Sample No.: CCAL01 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PQ066418.D Time Analyzed: 13:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.552	4.450	4.650	539.590	500.000	7.9
Aroclor-1016-2	4.571	4.470	4.670	548.210	500.000	9.6
Aroclor-1016-3	4.629	4.528	4.728	545.520	500.000	9.1
Aroclor-1016-4	4.722	4.620	4.820	546.270	500.000	9.3
Aroclor-1016-5	5.010	4.909	5.109	540.000	500.000	8.0
Aroclor-1260-1	6.116	6.015	6.215	551.910	500.000	10.4
Aroclor-1260-2	6.376	6.276	6.476	539.120	500.000	7.8
Aroclor-1260-3	6.731	6.630	6.830	546.750	500.000	9.4
Aroclor-1260-4	6.949	6.847	7.047	537.820	500.000	7.6
Aroclor-1260-5	7.261	7.160	7.360	533.790	500.000	6.8
Decachlorobiphenyl	8.650	8.547	8.747	54.820	50.000	9.6
Tetrachloro-m-xylene	3.441	3.340	3.540	52.700	50.000	5.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2024 04/23/2024

Client Sample No.: CCAL01 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PQ066418.D Time Analyzed: 13:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.730	3.631	3.831	515.490	500.000	3.1
Aroclor-1016-2	3.745	3.646	3.846	512.410	500.000	2.5
Aroclor-1016-3	3.906	3.806	4.006	511.500	500.000	2.3
Aroclor-1016-4	3.954	3.855	4.055	504.490	500.000	0.9
Aroclor-1016-5	4.149	4.050	4.250	506.400	500.000	1.3
Aroclor-1260-1	5.141	5.042	5.242	500.830	500.000	0.2
Aroclor-1260-2	5.332	5.232	5.432	503.600	500.000	0.7
Aroclor-1260-3	5.472	5.374	5.574	508.930	500.000	1.8
Aroclor-1260-4	5.934	5.835	6.035	491.440	500.000	-1.7
Aroclor-1260-5	6.182	6.082	6.282	513.560	500.000	2.7
Decachlorobiphenyl	7.498	7.398	7.598	54.690	50.000	9.4
Tetrachloro-m-xylene	2.733	2.633	2.833	52.420	50.000	4.8