

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2024 04/23/2024

Client Sample No.: CCAL02 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PQ066429.D Time Analyzed: 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.550	4.450	4.650	546.300	500.000	9.3
Aroclor-1016-2	4.569	4.470	4.670	543.740	500.000	8.7
Aroclor-1016-3	4.627	4.528	4.728	546.700	500.000	9.3
Aroclor-1016-4	4.720	4.620	4.820	549.410	500.000	9.9
Aroclor-1016-5	5.008	4.909	5.109	550.170	500.000	10.0
Aroclor-1260-1	6.115	6.015	6.215	552.710	500.000	10.5
Aroclor-1260-2	6.375	6.276	6.476	540.800	500.000	8.2
Aroclor-1260-3	6.730	6.630	6.830	543.350	500.000	8.7
Aroclor-1260-4	6.948	6.847	7.047	539.140	500.000	7.8
Aroclor-1260-5	7.260	7.160	7.360	536.040	500.000	7.2
Decachlorobiphenyl	8.648	8.547	8.747	54.860	50.000	9.7
Tetrachloro-m-xylene	3.439	3.340	3.540	52.750	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2024 04/23/2024

Client Sample No.: CCAL02 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PQ066429.D Time Analyzed: 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.730	3.631	3.831	514.320	500.000	2.9
Aroclor-1016-2	3.745	3.646	3.846	520.010	500.000	4.0
Aroclor-1016-3	3.905	3.806	4.006	517.410	500.000	3.5
Aroclor-1016-4	3.955	3.855	4.055	513.320	500.000	2.7
Aroclor-1016-5	4.149	4.050	4.250	508.100	500.000	1.6
Aroclor-1260-1	5.141	5.042	5.242	504.120	500.000	0.8
Aroclor-1260-2	5.331	5.232	5.432	505.380	500.000	1.1
Aroclor-1260-3	5.472	5.374	5.574	511.830	500.000	2.4
Aroclor-1260-4	5.934	5.835	6.035	494.890	500.000	-1.0
Aroclor-1260-5	6.181	6.082	6.282	512.630	500.000	2.5
Decachlorobiphenyl	7.496	7.398	7.598	56.290	50.000	12.6
Tetrachloro-m-xylene	2.733	2.633	2.833	52.700	50.000	5.4