

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/24/2024 04/24/2024

Client Sample No.: CCAL01 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR066586.D Time Analyzed: 13:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.872	4.782	4.982	537.310	500.000	7.5
Aroclor-1016-2	4.894	4.804	5.004	542.050	500.000	8.4
Aroclor-1016-3	4.958	4.868	5.068	535.040	500.000	7.0
Aroclor-1016-4	5.060	4.969	5.169	540.570	500.000	8.1
Aroclor-1016-5	5.374	5.283	5.483	547.720	500.000	9.5
Aroclor-1260-1	6.583	6.493	6.693	534.550	500.000	6.9
Aroclor-1260-2	6.865	6.775	6.975	534.980	500.000	7.0
Aroclor-1260-3	7.253	7.166	7.366	539.690	500.000	7.9
Aroclor-1260-4	7.495	7.410	7.610	538.570	500.000	7.7
Aroclor-1260-5	7.833	7.748	7.948	524.430	500.000	4.9
Decachlorobiphenyl	9.536	9.470	9.670	47.790	50.000	-4.4
Tetrachloro-m-xylene	3.653	3.562	3.762	54.270	50.000	8.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/24/2024 04/24/2024

Client Sample No.: CCAL01 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR066586.D Time Analyzed: 13:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.004	3.908	4.108	480.340	500.000	-3.9
Aroclor-1016-2	4.021	3.925	4.125	493.200	500.000	-1.4
Aroclor-1016-3	4.199	4.103	4.303	494.330	500.000	-1.1
Aroclor-1016-4	4.250	4.154	4.354	491.810	500.000	-1.6
Aroclor-1016-5	4.466	4.369	4.569	498.940	500.000	-0.2
Aroclor-1260-1	5.552	5.455	5.655	480.310	500.000	-3.9
Aroclor-1260-2	5.756	5.660	5.860	475.020	500.000	-5.0
Aroclor-1260-3	5.913	5.817	6.017	476.300	500.000	-4.7
Aroclor-1260-4	6.418	6.323	6.523	466.170	500.000	-6.8
Aroclor-1260-5	6.683	6.587	6.787	463.260	500.000	-7.3
Decachlorobiphenyl	8.125	8.032	8.232	45.220	50.000	-9.6
Tetrachloro-m-xylene	2.894	2.797	2.997	49.850	50.000	-0.3