

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/24/2024 04/24/2024

Client Sample No.: CCAL02 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR066597.D Time Analyzed: 16:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.873	4.782	4.982	530.150	500.000	6.0
Aroclor-1016-2	4.894	4.804	5.004	531.590	500.000	6.3
Aroclor-1016-3	4.959	4.868	5.068	526.860	500.000	5.4
Aroclor-1016-4	5.060	4.969	5.169	529.980	500.000	6.0
Aroclor-1016-5	5.375	5.283	5.483	541.150	500.000	8.2
Aroclor-1260-1	6.582	6.493	6.693	538.410	500.000	7.7
Aroclor-1260-2	6.865	6.775	6.975	524.710	500.000	4.9
Aroclor-1260-3	7.253	7.166	7.366	532.880	500.000	6.6
Aroclor-1260-4	7.495	7.410	7.610	528.040	500.000	5.6
Aroclor-1260-5	7.832	7.748	7.948	515.290	500.000	3.1
Decachlorobiphenyl	9.536	9.470	9.670	46.870	50.000	-6.3
Tetrachloro-m-xylene	3.653	3.562	3.762	53.110	50.000	6.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/24/2024 04/24/2024

Client Sample No.: CCAL02 Date Analyzed: 05/09/2024

Lab Sample No.: AR1660CCC500 Data File : PR066597.D Time Analyzed: 16:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.005	3.908	4.108	470.990	500.000	-5.8
Aroclor-1016-2	4.022	3.925	4.125	480.090	500.000	-4.0
Aroclor-1016-3	4.200	4.103	4.303	484.610	500.000	-3.1
Aroclor-1016-4	4.251	4.154	4.354	484.450	500.000	-3.1
Aroclor-1016-5	4.466	4.369	4.569	486.890	500.000	-2.6
Aroclor-1260-1	5.552	5.455	5.655	460.170	500.000	-8.0
Aroclor-1260-2	5.756	5.660	5.860	455.310	500.000	-8.9
Aroclor-1260-3	5.914	5.817	6.017	459.050	500.000	-8.2
Aroclor-1260-4	6.418	6.323	6.523	455.470	500.000	-8.9
Aroclor-1260-5	6.682	6.587	6.787	449.510	500.000	-10.1
Decachlorobiphenyl	8.125	8.032	8.232	44.450	50.000	-11.1
Tetrachloro-m-xylene	2.894	2.797	2.997	48.030	50.000	-3.9