

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: PP051524

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified	✓		
2. PEM Meet Criteria.	✓		
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.	✓		
3(a).Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. COMMENTS:	✓		
3(b).Continuous calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. COMMENTS: AR1660CCC500(PP064986.D)-failing high -associated samples have no any hit. no further corrective action required.	✓		
4. Blank Contamination - If yes, list compounds and concentrations in each blank: COMMENTS:	✓		
5. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. COMMENTS: DCB high in 1st column,--P2475-01 ,03,04,05,06			✓

6(a).Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria ✓

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

6(b).Blank Spike/Blank Spike Duplicate Recoveries Meet Criteria ✓

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

7. Retention Time Shift Meet Criteria (if applicable) ✓

COMMENTS:

8. Extraction Holding Time Met ✓

If not met, list number of days exceeded for each sample:

COMMENTS:

9. Analysis Holding Time Met ✓

If not met, list those compounds and their recoveries which fall outside the acceptable range.

COMMENTS:

ADDITIONAL COMMENTS:

P2475-01,02,03,04,05,06 all sample were PILC in matrix, therefore as a precautionary measure lab has analyzed these sample with straight 10x dilution. Due to sample nature, it was not plausible for lab to run these samples undiluted.

Nilesh

QA REVIEW

05/16/2024

Date