



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Instrument ID: ECD_D Date(s) Analyzed: 05/13/2024 05/13/2024

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.89	5.79	5.99	8074030
		2	6.68	6.58	6.78	15170200
		3	7.10	7.00	7.20	60731700
		4	7.19	7.09	7.29	45259600
		5	7.62	7.52	7.72	34046800



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Instrument ID: ECD_D Date(s) Analyzed: 05/13/2024 05/13/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.00	4.90	5.10	29321200
		2	5.68	5.58	5.78	32517000
		3	6.72	6.62	6.82	139801000
		4	7.04	6.94	7.14	53169400
		5	7.17	7.07	7.27	42256200