



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Instrument ID: ECD_L Date(s) Analyzed: 04/26/2024 04/26/2024

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.86	5.76	5.96	10470200
		2	6.45	6.35	6.55	12927300
		3	7.07	6.97	7.17	63848600
		4	7.16	7.06	7.26	48198400
		5	7.94	7.84	8.04	36188200



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Instrument ID: ECD_L Date(s) Analyzed: 04/26/2024 04/26/2024

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.36	5.26	5.46	10206000
		2	5.72	5.62	5.82	10727800
		3	6.63	6.53	6.73	35349700
		4	6.76	6.66	6.86	48366000
		5	7.07	6.97	7.17	20498500