

Report of Analysis

Client:	Chemtech Consult	ing Group			Date Collected:	05/07/24		
Project:	LOD-LOQ Study				Date Received:	05/07/24		
Client Sample ID:	LOD-MDL-WATE	ER-01-QT2-2024			SDG No.:	P2403		
Lab Sample ID:	P2403-07				Matrix:	WATER		
Analytical Method	: SW8081				% Solid:	0	Decanted:	
-		T						
Sample Wt/Vol:	1000 Units:	mL			Final Vol:	10000	uL	
Soil Aliquot Vol:		uL			Test:	Pesticide-TCI		
Extraction Type:					Injection Volume :			
GPC Factor :	1.0	PH :						
Prep Method :	3510C							
	- 4 .	_	_			_		
File ID/Qc Batch:	Dilution:	Dilution: Prep Date			Date Analyzed	Prep Batch ID		
PL089424.D	1	05/09	0/24 08:51		05/09/24 15:31	PB160816		
CAS Number	Parameter	Conc.	Qualifier	MDL		LOQ / CF	RQ L	Units
TARGETS								
319-84-6	alpha-BHC	0.014	JP	0.0061		0	0.050	ug/L
319-85-7	beta-BHC	0.015	J	0.014			0.050	ug/L
319-86-8	delta-BHC	0.050	U	0.015			0.050	ug/L
58-89-9	gamma-BHC (Lindane)	0.015	JP	0.0049		0	0.050	ug/L
76-44-8	Heptachlor	0.015	ЛЬ	0.0054		0	.050	ug/L
309-00-2	Aldrin	0.014	JP	0.0044		0	0.050	ug/L
1024-57-3	Heptachlor epoxide	0.015	ЛЬ	0.0090		0	.050	ug/L
959-98-8	Endosulfan I	0.016	JP	0.0050		0	0.050	ug/L
60-57-1	Dieldrin	0.015	JP	0.0047		0	0.050	ug/L
72-55-9	4,4-DDE	0.015	JP	0.0045		0	.050	ug/L
72-20-8	Endrin	0.013	J	0.0043		0	0.050	ug/L
33213-65-9	Endosulfan II	0.014	J	0.0075		0	0.050	ug/L
72-54-8	4,4-DDD	0.015	J	0.0092		0	.050	ug/L
1031-07-8	Endosulfan Sulfate	0.015	J	0.0035			0.050	ug/L
50-29-3	4,4-DDT	0.013	JP	0.0044		0	0.050	ug/L
72-43-5	Methoxychlor	0.013	J	0.011			0.050	ug/L
53494-70-5	Endrin ketone	0.015	JP	0.0097			0.050	ug/L
7421-93-4	Endrin aldehyde	0.015	J	0.0099			0.050	ug/L
	alpha-Chlordane	0.015	J	0.0060			0.050	ug/L
5103-71-9		0.015	J	0.0060			.050	ug/L
5103-71-9 5103-74-2	gamma-Chlordane	0.013						
	gamma-Chlordane Toxaphene	1.00	U	0.15		1	.00	ug/L
5103-74-2 8001-35-2				0.15		1	.00	ug/L
5103-74-2				0.15 27 - 142			.00	ug/L SPK: 2



Report of Analysis

Client:	Chemtech Cons	sulting Group		Date Collected:	05/07/24	
Project:	LOD-LOQ Stud			Date Received:	05/07/24	
Client Sample ID:	LOD-MDL-WA	ATER-01-QT2-2024		SDG No.:	P2403	
Lab Sample ID:	P2403-07			Matrix:	WATER	
Analytical Method:	SW8081			% Solid:	0	Decanted:
Sample Wt/Vol:	1000 Uni	ts: mL		Final Vol:	10000	uL
Soil Aliquot Vol:		uL		Test:	Pesticide-TCL	
Extraction Type:				Injection Volume :		
GPC Factor :	1.0	PH :				
Prep Method :	3510C					
File ID/Qc Batch:	Dilution:	Prep	Date	Date Analyzed	Prep l	Batch ID
PL089424.D	1	05/09/24 08:51		05/09/24 15:31	PB160816	
CAS Number Pa	arameter	Conc.	Qualifier MDI	L	LOQ / CR	QL Units

Comments:

U = Not Detected

- LOQ = Limit of Quantitation
- MDL = Method Detection Limit
- LOD = Limit of Detection
- E = Value Exceeds Calibration Range
- P = Indicates > 25% difference for detected
- concentrations between the two GC columns
- Q = indicates LCS control criteria did not meet requirements
- M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound
- * = Values outside of QC limits
- D = Dilution
- S = Indicates estimated value where valid five-point calibration
- was not performed prior to analyte detection in sample.
- () = Laboratory InHouse Limit