

Hit Summary Sheet
SW-846

SDG No.: P2403

Order ID: P2403

Client: Chemtech Consulting Group

Project ID: LOD-LOQ Study

Sample ID	Client ID	Parameter	Concentration	C	MDL	RDL	Units
Client ID : LOD-MDL-WATER-01-QT2-2024							
P2403-07	LOD-MDL-WATER-0 WATER	alpha-BHC	0.014	JP	0.0061	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	alpha-BHC	0.032	JP	0.0061	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	beta-BHC	0.033	J	0.014	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	beta-BHC	0.015	J	0.014	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	delta-BHC	0.033	JP	0.015	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	gamma-BHC (Lindane)	0.015	JP	0.0049	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	gamma-BHC (Lindane)	0.033	JP	0.0049	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Heptachlor	0.015	JP	0.0054	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Heptachlor	0.032	JP	0.0054	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Aldrin	0.032	J	0.0044	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Aldrin	0.014	JP	0.0044	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Heptachlor epoxide	0.034	J	0.0090	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Heptachlor epoxide	0.015	JP	0.0090	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endosulfan I	0.034	J	0.0050	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endosulfan I	0.016	JP	0.0050	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Dieldrin	0.034	J	0.0047	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Dieldrin	0.015	JP	0.0047	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	4,4-DDE	0.015	JP	0.0045	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	4,4-DDE	0.033	JP	0.0045	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endrin	0.013	J	0.0043	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endrin	0.031	J	0.0043	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endosulfan II	0.014	J	0.0075	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endosulfan II	0.033	J	0.0075	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	4,4-DDD	0.015	J	0.0092	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	4,4-DDD	0.035	J	0.0092	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endosulfan Sulfate	0.015	J	0.0035	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endosulfan Sulfate	0.034	J	0.0035	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	4,4-DDT	0.013	JP	0.0044	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	4,4-DDT	0.033	JP	0.0044	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Methoxychlor	0.029	J	0.011	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Methoxychlor	0.013	J	0.011	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endrin ketone	0.034	J	0.0097	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endrin ketone	0.015	JP	0.0097	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endrin aldehyde	0.015	J	0.0099	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	Endrin aldehyde	0.034	J	0.0099	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	alpha-Chlordane	0.015	J	0.0060	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	alpha-Chlordane	0.034	J	0.0060	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	gamma-Chlordane	0.015	J	0.0060	0.050	ug/L
P2403-07	LOD-MDL-WATER-0 WATER	gamma-Chlordane	0.035	J	0.0060	0.050	ug/L



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P2403

Order ID: P2403

Client: Chemtech Consulting Group

Project ID: LOD-LOQ Study

Sample ID	Client ID	Parameter	Concentration	C	MDL	RDL	Units
Total Concentration:			0.939				