

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG No.: P2403

Calibration Sequence : FF041124		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	198427453	116722	FF014100.D
850	96597132	113644	FF014101.D
340	38099294	112057	FF014102.D
170	19475563	114562	FF014103.D
85	8955262	105356	FF014104.D
AVG RF : 112468		% RSD : 3.84	AVG RT : 14.9444