

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG No.: P2403

Calibration Sequence : FG041124		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	211079640	124164	FG013679.D
850	103027394	121209	FG013680.D
340	40734131	119806	FG013681.D
170	21809557	128292	FG013682.D
85	10722721	126150	FG013683.D
AVG RF : 123924		% RSD : 2.808	AVG RT : 15.0676