

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG No.: P2403

Calibration Sequence : fe050924		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
3900	406029203	104110	FE048049.D
850	106521741	125320	FE048050.D
340	44631547	131269	FE048051.D
170	24435104	143736	FE048052.D
85	12302657	144737	FE048053.D
AVG RF : 129834		% RSD : 12.77	AVG RT : 15.0528