

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Instrument ID: ECD_R

Calibration Date(s): 05/20/2024 05/20/2024

Calibration Times: 13:03 13:41

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:							CF 0.5 =	<u>PR066687.D</u>	CF 0.25 =	<u>PR066688.D</u>	
CF 0.1 =							<u>PR066689.D</u>	CF 0.05 =	<u>PR066690.D</u>	CF 0.025 =	<u>PR066691.D</u>
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD				
DBCP	130872000000	128652000000	141933000000	143961000000	137395000000	136563000000	5				
EDB	62605400000	55597100000	63579400000	59519000000	67017200000	61663600000	7				

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Instrument ID: ECD_R

Calibration Date(s): 05/20/2024 05/20/2024

Calibration Times: 13:03 13:41

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR066687.D</u>		CF 0.25 = <u>PR066688.D</u>			
CF 0.1 = <u>PR066689.D</u>		CF 0.05 = <u>PR066690.D</u>		CF 0.025 = <u>PR066691.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	91423800000	92214900000	101373000000	103548000000	102462000000	98204600000	6
EDB	45764100000	42896400000	48826300000	50008200000	47656800000	47030300000	6