



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/20/2024 Initial Calibration Date(s): 05/20/2024 05/20/2024

Continuing Calib Time: 14:11 Initial Calibration Time(s): 13:03 13:41

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DBCP	6.58	6.58	6.48	6.68	0.00
EDB	2.84	2.84	2.74	2.94	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

Continuing Calib Date: 05/20/2024 Initial Calibration Date(s): 05/20/2024 05/20/2024

Continuing Calib Time: 14:11 Initial Calibration Time(s): 13:03 13:41

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.40	5.40	5.30	5.50	0.00
EDB	2.14	2.14	2.04	2.24	0.00