



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 892

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/20/2024 05/20/2024

Client Sample No.: CCAL02 Date Analyzed: 05/20/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR066699.D Time Analyzed: 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	6.583	6.484	6.684	0.110	0.100	0.0
EDB	2.842	2.742	2.942	0.100	0.100	0.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 892

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P2403 SAS No.: P2403 SDG NO.: P2403

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/20/2024 05/20/2024

Client Sample No.: CCAL02 Date Analyzed: 05/20/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR066699.D Time Analyzed: 14:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.396	5.296	5.496	0.100	0.100	0.0
EDB	2.142	2.042	2.242	0.100	0.100	0.0