



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: PSEG **SDG No.:** P2687
Contract: PSEG03 **Lab Code:** CHEM **Case No.:** P2687 **SAS No.:** P2687
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	17.1	20.0	85	40 - 160	P	06/03/2024	12:28	LB131042
	Barium	99.1	100	99	40 - 160	P	06/03/2024	12:28	LB131042
	Cadmium	6.33	6.0	106	40 - 160	P	06/03/2024	12:28	LB131042
	Chromium	9.51	10.0	95	40 - 160	P	06/03/2024	12:28	LB131042
	Copper	21.1	20.0	105	40 - 160	P	06/03/2024	12:28	LB131042
	Lead	12.2	12.0	102	40 - 160	P	06/03/2024	12:28	LB131042
	Nickel	40.3	40.0	101	40 - 160	P	06/03/2024	12:28	LB131042
	Selenium	19.6	20.0	98	40 - 160	P	06/03/2024	12:28	LB131042
	Silver	10.6	10.0	106	40 - 160	P	06/03/2024	12:28	LB131042
	Zinc	44.9	40.0	112	40 - 160	P	06/03/2024	12:28	LB131042
CRA	Mercury	0.15	0.2	76	40 - 160	CV	06/05/2024	12:38	LB131063