



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>PSEG</u>				SDG No.:	<u>P2687</u>			
Contract:	<u>PSEG03</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>P2687</u>	SAS No.:	<u>P2687</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB75	Mercury	0.20	+/-0.20	U	0.20	CV	06/05/2024	12:14	LB131063

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: PSEG

SDG No.: P2687

Contract: PSEG03

Lab Code: CHEM

Case No.: P2687

SAS No.: P2687

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB84	Mercury	0.20	+/-0.20	U	0.20	CV	06/05/2024	12:23	LB131063
CCB85	Mercury	0.20	+/-0.20	U	0.20	CV	06/05/2024	13:05	LB131063
CCB86	Mercury	0.20	+/-0.20	U	0.20	CV	06/05/2024	13:43	LB131063
CCB87	Mercury	0.20	+/-0.20	U	0.20	CV	06/05/2024	14:21	LB131063

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>PSEG</u>		SDG No.: <u>P2687</u>							
Contract: <u>PSEG03</u>		Lab Code: <u>CHEM</u>		Case No.: <u>P2687</u>		SAS No.: <u>P2687</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	12:23	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	12:23	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	12:23	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	12:23	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	12:23	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	12:23	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	12:23	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	12:23	LB131042
	Silver	10.0	+/-10.0	U	10.0	P	06/03/2024	12:23	LB131042
	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	12:23	LB131042



- 3a -

Client:	PSEG		SDG No.:	P2687	
Contract:	PSEG03	Lab Code:	CHEM	Case No.:	P2687
				SAS No.:	P2687

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	12:50	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	12:50	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	12:50	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	12:50	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	12:50	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	12:50	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	12:50	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	12:50	LB131042
	Silver	10.0	+/-10.0	U	10.0	P	06/03/2024	12:50	LB131042
	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	12:50	LB131042
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	13:48	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	13:48	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	13:48	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	13:48	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	13:48	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	13:48	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	13:48	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	13:48	LB131042
	Silver	10.0	+/-10.0	U	10.0	P	06/03/2024	13:48	LB131042
	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	13:48	LB131042
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	14:47	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	14:47	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	14:47	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	14:47	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	14:47	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	14:47	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	14:47	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	14:47	LB131042
	Silver	10.0	+/-10.0	U	10.0	P	06/03/2024	14:47	LB131042
	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	14:47	LB131042
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	15:38	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	15:38	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	15:38	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	15:38	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	15:38	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	15:38	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	15:38	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	15:38	LB131042
	Silver	10.0	+/-10.0	U	10.0	P	06/03/2024	15:38	LB131042

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>PSEG</u>		SDG No.: <u>P2687</u>							
Contract: <u>PSEG03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P2687</u>	SAS No.: <u>P2687</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	15:38	LB131042
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	16:24	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	16:24	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	16:24	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	16:24	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	16:24	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	16:24	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	16:24	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	16:24	LB131042
	Silver	10.0	+/-10.0	U	10.0	P	06/03/2024	16:24	LB131042
	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	16:24	LB131042
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	17:31	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	17:31	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	17:31	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	17:31	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	17:31	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	17:31	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	17:31	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	17:31	LB131042
	Silver	1.31	+/-10.0	J	10.0	P	06/03/2024	17:31	LB131042
	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	17:31	LB131042
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	06/03/2024	18:39	LB131042
	Barium	100	+/-100	U	100	P	06/03/2024	18:39	LB131042
	Cadmium	6.00	+/-6.00	U	6.00	P	06/03/2024	18:39	LB131042
	Chromium	10.0	+/-10.0	U	10.0	P	06/03/2024	18:39	LB131042
	Copper	20.0	+/-20.0	U	20.0	P	06/03/2024	18:39	LB131042
	Lead	12.0	+/-12.0	U	12.0	P	06/03/2024	18:39	LB131042
	Nickel	40.0	+/-40.0	U	40.0	P	06/03/2024	18:39	LB131042
	Selenium	20.0	+/-20.0	U	20.0	P	06/03/2024	18:39	LB131042
	Silver	1.33	+/-10.0	J	10.0	P	06/03/2024	18:39	LB131042
	Zinc	40.0	+/-40.0	U	40.0	P	06/03/2024	18:39	LB131042

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>PSEG</u>	SDG No.:	<u>P2687</u>						
Contract:	<u>PSEG03</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P2687</u>	SAS No.:	<u>P2687</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number