

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: PSEG

SDG No.: P2687

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB161253TB		WATER		Batch Number:	PB161315		Prep Date:	06/04/2024	
	Mercury	2.00	<2.00	U	2.00	CV	06/05/2024	14:07	LB131063
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB161315BL		WATER		Batch Number:	PB161315		Prep Date:	06/04/2024	
	Mercury	0.20	<0.20	U	0.20	CV	06/05/2024	12:41	LB131063

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: PSEG

SDG No.: P2687

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB161253TB	WATER			Batch Number:	PB161289		Prep Date:	06/03/2024	
	Arsenic	100	<100	U	100	P	06/03/2024	18:06	LB131042
	Barium	500	<500	U	500	P	06/03/2024	18:06	LB131042
	Cadmium	30.0	<30.0	U	30.0	P	06/03/2024	18:06	LB131042
	Chromium	22.7	<50.0	J	50.0	P	06/03/2024	18:06	LB131042
	Copper	100	<100	U	100	P	06/03/2024	18:06	LB131042
	Lead	60.0	<60.0	U	60.0	P	06/03/2024	18:06	LB131042
	Nickel	13.9	<200	J	200	P	06/03/2024	18:06	LB131042
	Selenium	100	<100	U	100	P	06/03/2024	18:06	LB131042
	Silver	6.98	<50.0	J	50.0	P	06/03/2024	18:06	LB131042
	Zinc	67.5	<200	J	200	P	06/03/2024	18:06	LB131042
Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB161289BL	WATER			Batch Number:	PB161289		Prep Date:	06/03/2024	
	Arsenic	100	<100	U	100	P	06/03/2024	18:11	LB131042
	Barium	500	<500	U	500	P	06/03/2024	18:11	LB131042
	Cadmium	30.0	<30.0	U	30.0	P	06/03/2024	18:11	LB131042
	Chromium	50.0	<50.0	U	50.0	P	06/03/2024	18:11	LB131042
	Copper	100	<100	U	100	P	06/03/2024	18:11	LB131042
	Lead	60.0	<60.0	U	60.0	P	06/03/2024	18:11	LB131042
	Nickel	200	<200	U	200	P	06/03/2024	18:11	LB131042
	Selenium	100	<100	U	100	P	06/03/2024	18:11	LB131042
	Silver	6.51	<50.0	J	50.0	P	06/03/2024	18:11	LB131042
	Zinc	200	<200	U	200	P	06/03/2024	18:11	LB131042