

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PSEG **Level:** LOW **SDG No.:** P2687
Contract: PSEG03 **Lab Code:** CHEM **Case No.:** P2687 **SAS No.:** P2687
Matrix: Water **Sample ID:** P2687-09 **Client ID:** 72-11998DUP
Percent Solids for Sample: NA **Duplicate ID** P2687-09DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	2560		2830		10		P
Cadmium	ug/L	20	3.66	J	3.81	J	4		P
Chromium	ug/L	20	50.0	U	50.0	U			P
Copper	ug/L	20	100	U	100	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	65.8	J	64.4	J	2		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	7.91	J	50.0	U	200.0		P
Zinc	ug/L	20	547		549		0		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: PSEG **Level:** LOW **SDG No.:** P2687
Contract: PSEG03 **Lab Code:** CHEM **Case No.:** P2687 **SAS No.:** P2687
Matrix: Water **Sample ID:** P2687-09MS **Client ID:** 72-11998MSD
Percent Solids for Sample: NA **Duplicate ID** P2687-09MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	4060		4130		2		P
Barium	ug/L	20	3630		3570		2		P
Cadmium	ug/L	20	952		964		1		P
Chromium	ug/L	20	1710		1670		2		P
Copper	ug/L	20	1370		1380		1		P
Lead	ug/L	20	4870		4930		1		P
Mercury	ug/L	20	46.5		45.5		2		CV
Nickel	ug/L	20	2370		2400		1		P
Selenium	ug/L	20	10400		10500		1		P
Silver	ug/L	20	343		333		3		P
Zinc	ug/L	20	1590		1570		1		P