

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: PSEG **SDG No.:** P2687
Contract: PSEG03 **Lab Code:** CHEM **Case No.:** P2687 **SAS No.:** P2687

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB161289BS							
Arsenic	ug/L	4000	3980		100	80 - 120	P
Barium	ug/L	1000	960		96	80 - 120	P
Cadmium	ug/L	1000	941		94	80 - 120	P
Chromium	ug/L	2000	1850		92	80 - 120	P
Copper	ug/L	1500	1450		97	80 - 120	P
Lead	ug/L	5000	4900		98	80 - 120	P
Nickel	ug/L	2500	2320		93	80 - 120	P
Selenium	ug/L	10000	9740		97	80 - 120	P
Silver	ug/L	380	344		90	80 - 120	P
Zinc	ug/L	1000	1020		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: PSEG

SDG No.: P2687

Contract: PSEG03

Lab Code: CHEM

Case No.: P2687

SAS No.: P2687

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB161315BS Mercury	ug/L	4.0	3.98		100	80 - 120	CV