

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P2687

Client: PSEG

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P2668-04MS	Client Sample ID:	WC-TP-1MS					DataFile:	BP020517.D		
Pyridine	500	0	310	ug/L	62				10	109	
1,4-Dichlorobenzene	500	0	400	ug/L	80				55	125	
2-Methylphenol	500	0	410	ug/L	82				23	132	
3+4-Methylphenols	500	0	410	ug/L	82				17	136	
Hexachloroethane	500	0	390	ug/L	78				35	137	
Nitrobenzene	500	0	410	ug/L	82				40	134	
Hexachlorobutadiene	500	0	400	ug/L	80				52	125	
2,4,6-Trichlorophenol	500	0	490	ug/L	98				46	139	
2,4,5-Trichlorophenol	500	0	460	ug/L	92				42	129	
2,4-Dinitrotoluene	500	0	460	ug/L	92				40	151	
Hexachlorobenzene	500	0	440	ug/L	88				60	129	
Pentachlorophenol	1000	0	770	ug/L	77				15	137	