

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: P2687

Client: PSEG

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	P2668-04MSD	Client Sample ID:	WC-TP-1MSD					DataFile:	BP020518.D		
Pyridine	500	0	310	ug/L	62	0			10	109	20
1,4-Dichlorobenzene	500	0	400	ug/L	80	0			55	125	20
2-Methylphenol	500	0	420	ug/L	84	2			23	132	20
3+4-Methylphenols	500	0	420	ug/L	84	2			17	136	20
Hexachloroethane	500	0	390	ug/L	78	0			35	137	20
Nitrobenzene	500	0	410	ug/L	82	0			40	134	20
Hexachlorobutadiene	500	0	410	ug/L	82	2			52	125	20
2,4,6-Trichlorophenol	500	0	510	ug/L	102	4			46	139	20
2,4,5-Trichlorophenol	500	0	470	ug/L	94	2			42	129	20
2,4-Dinitrotoluene	500	0	470	ug/L	94	2			40	151	20
Hexachlorobenzene	500	0	440	ug/L	88	0			60	129	20
Pentachlorophenol	1000	0	760	ug/L	76	1			15	137	20