

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P2687 SAS No.: P2687 SDG NO.: P2687

EPA Sample No.: SSTDCCC040 Date Analyzed: 06/04/2024

Lab File ID: BP020494.D Time Analyzed: 12:27

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	393653	7.769	1640780	10.55	1066330	14.40
UPPER LIMIT	787306	8.269	3281560	11.046	2132660	14.898
LOWER LIMIT	196827	7.269	820390	10.046	533165	13.898
EPA SAMPLE NO.						
01 VNJ-223	255014	7.75	987390	10.52	590413	14.38
02 72-11977	274125	7.76	1058340	10.54	665506	14.38
03 72-11998	267600	7.77	1037480	10.53	656828	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.