

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX052924

Review By	John Carlone	Review On	5/30/2024 9:55:17 AM
Supervise By	Semsettin Yesilyurt	Supervise On	5/30/2024 5:31:32 PM
SubDirectory	VX052924	HP Acquire Method	HP Processing Method 82X052924W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP128050		
Initial Calibration Stds	VP128065,VP128066,VP128067,VP128068,VP128069,VP128070		
CCC	VP128051		
Internal Standard/PEM			
ICV/I.BLK	VP128071		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX041661.D	29 May 2024 08:43	JC/MD	Ok
2	VSTDICC001	VX041662.D	29 May 2024 09:34	JC/MD	Ok,M
3	VSTDICC005	VX041663.D	29 May 2024 09:56	JC/MD	Ok,M
4	VSTDICC020	VX041664.D	29 May 2024 10:19	JC/MD	Ok,M
5	VSTDICCC050	VX041665.D	29 May 2024 10:42	JC/MD	Ok,M
6	VSTDICC100	VX041666.D	29 May 2024 11:06	JC/MD	Ok,M
7	VSTDICC150	VX041667.D	29 May 2024 11:29	JC/MD	Ok,M
8	IBLK	VX041668.D	29 May 2024 11:52	JC/MD	Ok
9	VSTDICC001	VX041669.D	29 May 2024 13:09	JC/MD	Not Ok
10	VSTDICV050	VX041670.D	29 May 2024 13:32	JC/MD	Ok,M
11	VX0529MBL01	VX041671.D	29 May 2024 14:06	JC/MD	Ok
12	VX0529WBL01	VX041672.D	29 May 2024 14:29	JC/MD	Ok
13	VX0529WBS01	VX041673.D	29 May 2024 14:54	JC/MD	Ok,M
14	VX0529WBSD01	VX041674.D	29 May 2024 15:20	JC/MD	Ok,M
15	PB161143ZHE#01	VX041675.D	29 May 2024 15:43	JC/MD	Ok
16	PB161143ZHE#02	VX041676.D	29 May 2024 16:06	JC/MD	Ok
17	PB161143ZHE#03	VX041677.D	29 May 2024 16:30	JC/MD	Ok
18	PB161143ZHE#04	VX041678.D	29 May 2024 16:53	JC/MD	Ok
19	PB161143ZHE#05	VX041679.D	29 May 2024 17:16	JC/MD	Ok
20	VSTDCCC050	VX041680.D	29 May 2024 17:41	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX060324

Review By	John Carlone	Review On	6/4/2024 10:01:04 AM	
Supervise By	Maresh Dadoda	Supervise On	6/4/2024 6:06:46 PM	
SubDirectory	VX060324	HP Acquire Method	MSVOA_X	HP Processing Method
STD. NAME		STD REF.#		
Tune/Reschk Initial Calibration Stds		VP128155		
CCC		VP128156,VP128157		
Internal Standard/PEM		VP125083		
ICV/I.BLK				
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX041692.D	03 Jun 2024 09:42	JC/MD	Ok
2	VSTDCCC050	VX041693.D	03 Jun 2024 10:12	JC/MD	Ok,M
3	VX0603MBL01	VX041694.D	03 Jun 2024 10:42	JC/MD	Ok
4	VX0603WBL01	VX041695.D	03 Jun 2024 11:33	JC/MD	Ok
5	VX0603WBS01	VX041696.D	03 Jun 2024 11:58	JC/MD	Ok,M
6	VX0603WBSD01	VX041697.D	03 Jun 2024 12:23	JC/MD	Ok,M
7	P2694-01	VX041698.D	03 Jun 2024 12:47	JC/MD	Not Ok
8	P2694-02	VX041699.D	03 Jun 2024 13:10	JC/MD	Ok
9	IBLK	VX041700.D	03 Jun 2024 13:33	JC/MD	Ok
10	P2694-01	VX041701.D	03 Jun 2024 13:57	JC/MD	Ok
11	P2660-02	VX041702.D	03 Jun 2024 14:20	JC/MD	Ok
12	P2691-04	VX041703.D	03 Jun 2024 14:43	JC/MD	Ok
13	P2691-02	VX041704.D	03 Jun 2024 15:07	JC/MD	Ok
14	P2691-03	VX041705.D	03 Jun 2024 15:30	JC/MD	Ok
15	P2660-01	VX041706.D	03 Jun 2024 15:53	JC/MD	Ok
16	P2660-03	VX041707.D	03 Jun 2024 16:17	JC/MD	Ok
17	P2660-04	VX041708.D	03 Jun 2024 16:41	JC/MD	Ok
18	P2660-05	VX041709.D	03 Jun 2024 17:04	JC/MD	Ok
19	P2660-06	VX041710.D	03 Jun 2024 17:27	JC/MD	Dilution
20	P2687-03	VX041711.D	03 Jun 2024 17:51	JC/MD	Ok
21	P2687-06	VX041712.D	03 Jun 2024 18:14	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX060324

Review By	John Carlone	Review On	6/4/2024 10:01:04 AM	
Supervise By	Mahesh Dadoda	Supervise On	6/4/2024 6:06:46 PM	
SubDirectory	VX060324	HP Acquire Method	MSVOA_X	HP Processing Method
STD. NAME	STD REF.#			
Tune/Reschk Initial Calibration Stds	VP128155			
CCC	VP128156,VP128157			
Internal Standard/PEM	VP125083			
ICV/I.BLK				
Surrogate Standard				
MS/MSD Standard				
LCS Standard				

22	P2687-09	VX041713.D	03 Jun 2024 18:38	JC/MD	Ok
23	VX0603MBS01	VX041714.D	03 Jun 2024 19:01	JC/MD	Ok,M
24	P2657-18	VX041715.D	03 Jun 2024 19:25	JC/MD	Not Ok
25	P2657-14	VX041716.D	03 Jun 2024 19:48	JC/MD	Not Ok
26	P2657-10	VX041717.D	03 Jun 2024 20:12	JC/MD	Not Ok
27	P2657-06	VX041718.D	03 Jun 2024 20:35	JC/MD	Ok
28	P2657-02	VX041719.D	03 Jun 2024 20:59	JC/MD	Ok
29	VSTDCCC050	VX041720.D	03 Jun 2024 21:22	JC/MD	Ok,M

M : Manual Integration