

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX052924

SequenceID : VX052924 NA NO YES

1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)                       ✓

2. GC/MS Tuning Specifications. BFB Meet Criteria                       ✓  
(NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)

3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series                       ✓

4. GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series                       ✓

5. GC/MS Calibration Requirements                       ✓

5a. Initial Calibration Meet Criteria                       ✓  
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.  
The %RSD is greater than 20% in the Initial Calibration method (82X052924W.M) for Acetone this compound is passing on Quadratic Regression.

5b. Continuous Calibration(CCC) Meet Criteria                       ✓  
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

6. Blank Contamination - If yes, list compounds and concentrations in each blank:            ✓           

7. Surrogate Recoveries Meet Criteria                       ✓  
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

8a. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria                       ✓  
If not met, list those compounds and their recoveries which fall outside the acceptable range.

8b. Blank Spike/Blank spike Duplicate Recoveries Meet Criteria                       ✓  
If not met, list those compounds and their recoveries which fall outside the acceptable range.

9. Internal Standard Area/Retention Time Shift Meet Criteria                       ✓

Comments:

jalpa

Analyst

05/30/2024

Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX052924

SequenceID : VX052924

NA NO YES

10. Analysis Holding Time Met

\_\_\_\_ \_ ✓

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

jalpa

Analyst

05/30/2024

Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: VX052924

SequenceID : VX052924 NA NO YES

jalpa  
Analyst

05/30/2024  
Date