

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: VX060324

SequenceID : <u>VX060324</u>	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	<u>✓</u>	<u> </u>	<u> </u>
2. GC/MS Tuning Specifications. BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	<u>✓</u>	<u> </u>	<u> </u>
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	<u>✓</u>	<u> </u>	<u> </u>
4. GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	<u>✓</u>	<u> </u>	<u> </u>
5. GC/MS Calibration Requirements	<u>✓</u>	<u> </u>	<u> </u>
5a. Initial Calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	<u> </u>	<u> </u>
5b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	<u> </u>	<u> </u>
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	<u>✓</u>	<u> </u>	<u> </u>
7. Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges.	<u>✓</u>	<u> </u>	<u> </u>
8a. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>✓</u>	<u> </u>	<u> </u>
8b. Blank Spike/Blank spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	<u>✓</u>	<u> </u>	<u> </u>
9. Internal Standard Area/Retention Time Shift Meet Criteria	<u>✓</u>	<u> </u>	<u> </u>
Comments:			

JOHN
Analyst

06/04/2024
Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: VX060324

SequenceID : VX060324

NA NO YES

10. Analysis Holding Time Met

✓

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

P2694-01,02. Analyst ran these samples at straight dilution after checking past history of these samples containing high amounts of compound Trichloroethene.

P2691-02,03,04. Analyst ran these samples at straight dilution due to all samples having green/yellow color and turbid, not allowing straight run analysis.

P2660-02. Analyst ran this sample at straight high dilution after checking past history of this sample containing high amounts of compounds ethyl benzene, o xylene, mp xylenes

JOHN
Analyst

06/04/2024
Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: VX060324

SequenceID : VX060324 NA NO YES

JOHN
Analyst

06/04/2024
Date