

Surrogate Summary

SDG No.: P2687

Client: PSEG

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P2687-03	VNJ-223	1,2-Dichloroethane-d4	50	45.0	90	74	125
		Dibromofluoromethane	50	44.4	89	75	124
		Toluene-d8	50	49.3	99	86	113
		4-Bromofluorobenzene	50	51.9	104	64	133
P2687-06	72-11977	1,2-Dichloroethane-d4	50	47.0	94	74	125
		Dibromofluoromethane	50	45.5	91	75	124
		Toluene-d8	50	49.6	99	86	113
		4-Bromofluorobenzene	50	50.9	102	64	133
P2687-09	72-11998	1,2-Dichloroethane-d4	50	46.0	92	74	125
		Dibromofluoromethane	50	45.2	90	75	124
		Toluene-d8	50	49.9	100	86	113
		4-Bromofluorobenzene	50	51.1	102	64	133

Surrogate Summary

SDG No.: P2687

Client: PSEG

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0603WBL01	VX0603WBL01	1,2-Dichloroethane-d4	50	46.7	93	74	125
		Dibromofluoromethane	50	46.2	92	75	124
		Toluene-d8	50	49.7	99	86	113
		4-Bromofluorobenzene	50	50.8	102	64	133
VX0603WBS01	VX0603WBS01	1,2-Dichloroethane-d4	50	45.9	92	74	125
		Dibromofluoromethane	50	45.4	91	75	124
		Toluene-d8	50	47.2	94	86	113
		4-Bromofluorobenzene	50	46.7	93	64	133
VX0603WBSD0	VX0603WBSD01	1,2-Dichloroethane-d4	50	47.0	94	74	125
		Dibromofluoromethane	50	45.7	91	75	124
		Toluene-d8	50	46.7	93	86	113
		4-Bromofluorobenzene	50	47.2	94	64	133