

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P2687

Client: PSEG

Analytical Method: SW8260-Low

Datafile : VX041696.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0603WBS01	Vinyl chloride	20	19.2	ug/L	96			65	117	
	1,1-Dichloroethene	20	18.5	ug/L	93			74	110	
	2-Butanone	100	98.6	ug/L	99			65	122	
	Carbon Tetrachloride	20	19.1	ug/L	96			77	113	
	Chloroform	20	19.2	ug/L	96			79	113	
	Benzene	20	19.2	ug/L	96			82	109	
	1,2-Dichloroethane	20	19.2	ug/L	96			80	115	
	Trichloroethene	20	19.0	ug/L	95			77	113	
	Tetrachloroethene	20	18.9	ug/L	95			67	123	
	Chlorobenzene	20	19.2	ug/L	96			82	109	