

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P2687

Client: PSEG

Analytical Method: SW8260-Low

Datafile : VX041697.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0603WBSD01	Vinyl chloride	20	18.8	ug/L	94	2		65	117	20
	1,1-Dichloroethene	20	18.4	ug/L	92	1		74	110	20
	2-Butanone	100	110	ug/L	110	11		65	122	20
	Carbon Tetrachloride	20	19.1	ug/L	96	0		77	113	20
	Chloroform	20	19.4	ug/L	97	1		79	113	20
	Benzene	20	19.2	ug/L	96	0		82	109	20
	1,2-Dichloroethane	20	19.8	ug/L	99	3		80	115	20
	Trichloroethene	20	19.2	ug/L	96	1		77	113	20
	Tetrachloroethene	20	19.1	ug/L	96	1		67	123	20
	Chlorobenzene	20	19.2	ug/L	96	0		82	109	20