

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PSEG03
 Lab Code: CHEM Case No.: P2687 SAS No.: P2687 SDG NO.: P2687
 Lab File ID: VX041693.D Date Analyzed: 06/03/2024
 Instrument ID: MSVOA_X Time Analyzed: 10:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	142787	12.018				
UPPER LIMIT	285574	12.518				
LOWER LIMIT	71393.5	11.518				
EPA SAMPLE NO.						
VNJ-223	114199	12.02				
72-11977	101855	12.02				
72-11998	102157	12.02				
VX0603WBL01	111449	12.02				
VX0603WBS01	131481	12.02				
VX0603WBSD01	123243	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.