

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PSEG03  
Lab Code: CHEM Case No.: P2687 SAS No.: P2687 SDG NO.: P2687  
Lab File ID: VX041661.D BFB Injection Date: 05/29/2024  
Instrument ID: MSVOA\_X BFB Injection Time: 08:43  
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

| m/e | ION ABUNDANCE CRITERIA             | % RELATIVE ABUNDANCE |
|-----|------------------------------------|----------------------|
| 50  | 15.0 - 40.0% of mass 95            | 19.7                 |
| 75  | 30.0 - 60.0% of mass 95            | 52.5                 |
| 95  | Base Peak, 100% relative abundance | 100                  |
| 96  | 5.0 - 9.0% of mass 95              | 6.9                  |
| 173 | Less than 2.0% of mass 174         | 0.8 ( 0.8 ) 1        |
| 174 | 50.0 - 100.0% of mass 95           | 96.8                 |
| 175 | 5.0 - 9.0% of mass 174             | 7.1 ( 7.4 ) 1        |
| 176 | 95.0 - 101.0% of mass 174          | 92.1 ( 95.1 ) 1      |
| 177 | 5.0 - 9.0% of mass 176             | 6.3 ( 6.8 ) 2        |

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

| EPA<br>SAMPLE NO. | LAB<br>SAMPLE ID | LAB<br>FILE ID | DATE<br>ANALYZED | TIME<br>ANALYZED |
|-------------------|------------------|----------------|------------------|------------------|
| VSTDIC001         | VSTDIC001        | VX041662.D     | 05/29/2024       | 09:34            |
| VSTDIC005         | VSTDIC005        | VX041663.D     | 05/29/2024       | 09:56            |
| VSTDIC020         | VSTDIC020        | VX041664.D     | 05/29/2024       | 10:19            |
| VSTDIC050         | VSTDIC050        | VX041665.D     | 05/29/2024       | 10:42            |
| VSTDIC100         | VSTDIC100        | VX041666.D     | 05/29/2024       | 11:06            |
| VSTDIC150         | VSTDIC150        | VX041667.D     | 05/29/2024       | 11:29            |