

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD3S6
 Instrument ID: P5 Start Date: 06/14/2024
 Analytical Method: ICP-AES Run Batch: LB131248
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.020	0	60	56	-7	1250	1000	-20
Arsenic	0	0.0	0	10	7.4	-26	1250	1000	-20
Barium	0	0.030	0	200	190	-5	2500	2200	-12
Beryllium	0	0.0	0	5	4.9	-2	125	110	-12
Cadmium	0	0.0	0	5	4.7	-6	625	560	-10
Chromium	0	0.0	0	10	11	10	3875	3600	-7
Cobalt	0	0.010	0	50	45	-10	625	540	-14
Copper	0	-0	0	25	29	16	3750	3300	-12
Lead	0	0.0	0	10	10	0	6250	5500	-12
Manganese	0	0.0	0	15	15	0	3750	3500	-7
Molybdenum	0	0.0	0	10	9.7	-3	1250	1100	-12
Nickel	0	0.010	0	40	38	-5	625	550	-12
Selenium	0	0.010	0	35	31	-11	1250	1100	-12
Silver	0	0.0	0	10	9.6	-4	312.5	270	-14
Thallium	0	0.0	0	25	24	-4	1250	1200	-4
Vanadium	0	0.010	0	50	44	-12	625	560	-10
Zinc	0	0.020	0	60	54	-10	3750	3100	-17

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD3S6
 Instrument ID: P5 Start Date: 06/14/2024
 Analytical Method: ICP-AES Run Batch: LB131248
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	2500	2300	-8	5000	5000	0	10000	10000	0
Arsenic	2500	2300	-8	5000	5000	0	10000	10000	0
Barium	5000	4900	-2	10000	10000	0	20000	20000	0
Beryllium	250	250	0	500	510	2	1000	1000	0
Cadmium	1250	1200	-4	2500	2600	4	5000	5000	0
Chromium	7750	7800	1	15500	16000	3	31000	31000	0
Cobalt	1250	1200	-4	2500	2600	4	5000	5000	0
Copper	7500	7300	-3	15000	15000	0	30000	30000	0
Lead	12500	12000	-4	25000	26000	4	50000	51000	2
Manganese	7500	7600	1	15000	15000	0	30000	30000	0
Molybdenum	2500	2400	-4	5000	5100	2	10000	10000	0
Nickel	1250	1200	-4	2500	2600	4	5000	5000	0
Selenium	2500	2400	-4	5000	5000	0	10000	10000	0
Silver	625	600	-4	1250	1300	4	2500	2600	4
Thallium	2500	2500	0	5000	5200	4	10000	9900	-1
Vanadium	1250	1200	-4	2500	2600	4	5000	5000	0
Zinc	7500	7000	-7	15000	15000	0	30000	31000	3

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD3S6
 Instrument ID: P5 Start Date: 06/18/2024
 Analytical Method: ICP-AES Run Batch: LB131291
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.010	0	60	59	-2	1250	1200	-4
Arsenic	0	0.0	0	10	8.3	-17	1250	1100	-12
Barium	0	0.020	0	200	200	0	2500	2300	-8
Beryllium	0	-0	0	5	5.3	6	125	130	4
Cadmium	0	-0	0	5	5.2	4	625	630	1
Chromium	0	-0	0	10	12	20	3875	4000	3
Cobalt	0	0.0	0	50	48	-4	625	620	-1
Copper	0	0.0	0	25	26	4	3750	3700	-1
Lead	0	0.0	0	10	9.5	-5	6250	6200	-1
Manganese	0	-0	0	15	17	13	3750	3800	1
Molybdenum	0	-0	0	10	11	10	1250	1200	-4
Nickel	0	-0	0	40	40	0	625	630	1
Selenium	0	0.0	0	35	35	0	1250	1200	-4
Silver	0	0.0	0	10	10	0	312.5	300	-4
Thallium	0	-0	0	25	26	4	1250	1300	4
Vanadium	0	0.0	0	50	51	2	625	590	-6
Zinc	0	0.010	0	60	58	-3	3750	3400	-9

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD3S6
 Instrument ID: P5 Start Date: 06/18/2024
 Analytical Method: ICP-AES Run Batch: LB131291
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	2500	2300	-8	5000	4900	-2	10000	10000	0
Arsenic	2500	2300	-8	5000	5000	0	10000	10000	0
Barium	5000	4700	-6	10000	9900	-1	20000	21000	5
Beryllium	250	240	-4	500	500	0	1000	1000	0
Cadmium	1250	1200	-4	2500	2500	0	5000	5000	0
Chromium	7750	7700	-1	15500	16000	3	31000	31000	0
Cobalt	1250	1200	-4	2500	2500	0	5000	5000	0
Copper	7500	7200	-4	15000	15000	0	30000	31000	3
Lead	12500	12000	-4	25000	25000	0	50000	50000	0
Manganese	7500	7400	-1	15000	15000	0	30000	30000	0
Molybdenum	2500	2400	-4	5000	5000	0	10000	10000	0
Nickel	1250	1200	-4	2500	2500	0	5000	5000	0
Selenium	2500	2400	-4	5000	4900	-2	10000	10000	0
Silver	625	600	-4	1250	1300	4	2500	2500	0
Thallium	2500	2500	0	5000	5100	2	10000	9900	-1
Vanadium	1250	1200	-4	2500	2500	0	5000	5200	4
Zinc	7500	6900	-8	15000	15000	0	30000	31000	3