

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/27/24 11:22		Kareem	OK
2	S1	S01	CAL2	06/27/24 11:27		Kareem	OK
3	S2	S02	CAL3	06/27/24 11:31		Kareem	OK
4	S3	S03	CAL4	06/27/24 11:36		Kareem	OK
5	S4	S04	CAL5	06/27/24 11:40		Kareem	OK
6	S5	S05	CAL6	06/27/24 11:44		Kareem	OK
7	S6	S06	CAL7	06/27/24 11:49		Kareem	OK
8	ICV011	ICV011	ICV	06/27/24 11:54		Kareem	OK
9	ICB011	ICB011	ICB	06/27/24 12:01	Inst. paused for review	Kareem	OK
10	ICSA011	ICSA011	ICSA	06/27/24 12:06		Kareem	OK
11	ICSAB011	ICSAB011	ICSAB	06/27/24 12:10		Kareem	OK
12	CCV065	CCV065	CCV	06/27/24 12:25	Inst. paused for review	Kareem	OK
13	CCB065	CCB065	CCB	06/27/24 12:29		Kareem	OK
14	P2809-07	MYD4R4	SAM	06/27/24 12:34		Kareem	OK
15	P2809-08	MYD4R5	SAM	06/27/24 12:38		Kareem	OK
16	P2809-09	MYD4R6	SAM	06/27/24 12:43		Kareem	OK
17	P2809-10	MYD4R7	SAM	06/27/24 12:47		Kareem	OK
18	P2809-11	MYD4R7D	DUP	06/27/24 12:51		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

19	P2809-10L	MYD4R7L	SD	06/27/24 12:56		Kareem	OK
20	P2809-12	MYD4R7S	MS	06/27/24 13:00	MS fail for Zn (2789.87 ppb) & Sb,Se,Ag,Tl - Below RL	Kareem	OK
21	P2809-13	MYD4R8	SAM	06/27/24 13:04		Kareem	OK
22	P2809-14	MYD4R9	SAM	06/27/24 13:09		Kareem	OK
23	P2809-15	MYD4S0	SAM	06/27/24 13:13		Kareem	OK
24	P2809-16	MYD4S1	SAM	06/27/24 13:17		Kareem	OK
25	P2809-17	MYD4S2	SAM	06/27/24 13:22		Kareem	OK
26	P2809-18	MYD4S3	SAM	06/27/24 13:26		Kareem	OK
27	P2809-19	MYD4S4	SAM	06/27/24 13:30		Kareem	OK
28	P2809-20	MYD4S5	SAM	06/27/24 13:35		Kareem	OK
29	P2809-21	MYD4S6	SAM	06/27/24 13:39		Kareem	OK
30	P2809-22	MYD4S7	SAM	06/27/24 13:43		Kareem	OK
31	P2810-01	MYD4G6	SAM	06/27/24 13:48		Kareem	OK
32	P2810-02	MYD4G6D	DUP	06/27/24 13:52		Kareem	OK
33	P2810-01L	MYD4G6L	SD	06/27/24 13:56		Kareem	OK
34	CCV066	CCV066	CCV	06/27/24 14:01		Kareem	OK
35	CCB066	CCB066	CCB	06/27/24 14:05		Kareem	OK
36	P2810-03	MYD4G6S	MS	06/27/24 14:10	MS fail for Sb,Se,Ag,Tl - Below RL	Kareem	OK
37	P2810-04	MYD4S8	SAM	06/27/24 14:14		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

38	P2810-05	MYD4S9	SAM	06/27/24 14:18		Kareem	OK
39	P2810-06	MYD4T0	SAM	06/27/24 14:23		Kareem	OK
40	P2810-07	MYD4T1	SAM	06/27/24 14:27		Kareem	OK
41	P2810-08	MYD4T2	SAM	06/27/24 14:31		Kareem	OK
42	P2810-09	MYD4T3	SAM	06/27/24 14:36		Kareem	OK
43	P2810-10	MYD4T4	SAM	06/27/24 14:40		Kareem	OK
44	P2810-11	MYD4T5	SAM	06/27/24 14:45		Kareem	OK
45	P2810-12	MYD4T6	SAM	06/27/24 14:49		Kareem	OK
46	P2810-13	MYD4T7	SAM	06/27/24 14:53		Kareem	OK
47	P2810-14	MYD4T8	SAM	06/27/24 14:58		Kareem	OK
48	P2810-15	MYD4T9	SAM	06/27/24 15:02		Kareem	OK
49	P2810-16	MYD4W0	SAM	06/27/24 15:07		Kareem	OK
50	P2810-17	MYD4W1	SAM	06/27/24 15:11		Kareem	OK
51	P2810-18	MYD4W2	SAM	06/27/24 15:15		Kareem	OK
52	P2810-19	MYD4W3	SAM	06/27/24 15:20		Kareem	OK
53	P2810-20	MYD4W4	SAM	06/27/24 15:24		Kareem	OK
54	P2810-21	MYD4W5	SAM	06/27/24 15:28		Kareem	OK
55	P2810-22	MYD4W6	SAM	06/27/24 15:33		Kareem	OK
56	CCV067	CCV067	CCV	06/27/24 15:37		Kareem	OK
57	CCB067	CCB067	CCB	06/27/24 15:42		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

58	P2811-01	MYD4W7	SAM	06/27/24 15:46		Kareem	OK
59	P2811-02	MYD4W8	SAM	06/27/24 15:51		Kareem	OK
60	P2811-03	MYD4W9	SAM	06/27/24 15:55		Kareem	OK
61	P2811-04	MYD4X0	SAM	06/27/24 15:59		Kareem	OK
62	P2811-05	MYD4X1	SAM	06/27/24 16:04		Kareem	OK
63	P2811-06	MYD4X2	SAM	06/27/24 16:08		Kareem	OK
64	P2811-07	MYD4X3	SAM	06/27/24 16:12		Kareem	OK
65	P2811-08	MYD4X4	SAM	06/27/24 16:17		Kareem	OK
66	P2811-09	MYD4X5	SAM	06/27/24 16:21		Kareem	OK
67	P2811-10	MYD4X6	SAM	06/27/24 16:26		Kareem	OK
68	P2811-11	MYD4X7	SAM	06/27/24 16:30		Kareem	OK
69	P2811-12	MYD4X8	SAM	06/27/24 16:34		Kareem	OK
70	P2811-13	MYD4X9	SAM	06/27/24 16:39		Kareem	OK
71	P2811-14	MYD4Y0	SAM	06/27/24 16:43		Kareem	OK
72	P2811-15	MYD4Y1	SAM	06/27/24 16:47		Kareem	OK
73	P2811-16	MYD4Y2	SAM	06/27/24 16:52		Kareem	OK
74	P2811-17	MYD4Y3	SAM	06/27/24 16:56		Kareem	OK
75	P2811-18	MYD4Y3D	DUP	06/27/24 17:01		Kareem	OK
76	P2811-17L	MYD4Y3L	SD	06/27/24 17:05		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

77	P2811-19	MYD4Y3S	MS	06/27/24 17:09	MS fail for Zn (1597.37 ppb) & Sb,Se,Tl - Below RL	Kareem	OK
78	CCV068	CCV068	CCV	06/27/24 17:14		Kareem	OK
79	CCB068	CCB068	CCB	06/27/24 17:18		Kareem	OK
80	P2811-20	MYD4Y5	SAM	06/27/24 17:23		Kareem	OK
81	P2811-21	MYD4Y6	SAM	06/27/24 17:27		Kareem	OK
82	P2811-22	MYD4Y7	SAM	06/27/24 17:31		Kareem	OK
83	P2813-01	MYD4Y4	SAM	06/27/24 17:36		Kareem	OK
84	P2813-02	MYD4Y4D	DUP	06/27/24 17:40		Kareem	OK
85	P2813-01L	MYD4Y4L	SD	06/27/24 17:44		Kareem	OK
86	P2813-03	MYD4Y4S	MS	06/27/24 17:49	MS fail for Sb,Se,Ag,Tl - Below RL	Kareem	OK
87	P2813-04	MYD4Y8	SAM	06/27/24 17:53		Kareem	OK
88	P2813-05	MYD4Y9	SAM	06/27/24 17:58		Kareem	OK
89	P2813-06	MYD4Z0	SAM	06/27/24 18:02		Kareem	OK
90	P2813-07	MYD4Z1	SAM	06/27/24 18:06		Kareem	OK
91	P2813-08	MYD4Z2	SAM	06/27/24 18:11		Kareem	OK
92	P2813-09	MYD4Z3	SAM	06/27/24 18:15		Kareem	OK
93	P2813-10	MYD4Z4	SAM	06/27/24 18:19		Kareem	OK
94	P2813-11	MYD4Z5	SAM	06/27/24 18:24		Kareem	OK
95	P2813-12	MYD4Z6	SAM	06/27/24 18:28		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

96	P2813-13	MYD4Z7	SAM	06/27/24 18:32		Kareem	OK
97	P2813-14	MYD4Z8	SAM	06/27/24 18:37		Kareem	OK
98	P2813-15	MYD4Z9	SAM	06/27/24 18:41		Kareem	OK
99	P2813-16	MYD500	SAM	06/27/24 18:45		Kareem	OK
100	CCV069	CCV069	CCV	06/27/24 18:50		Kareem	OK
101	CCB069	CCB069	CCB	06/27/24 18:54		Kareem	OK
102	P2813-17	MYD501	SAM	06/27/24 18:59		Kareem	OK
103	P2813-18	MYD502	SAM	06/27/24 19:03		Kareem	OK
104	P2813-19	MYD503	SAM	06/27/24 19:07		Kareem	OK
105	P2813-20	MYD504	SAM	06/27/24 19:12		Kareem	OK
106	P2813-21	MYD505	SAM	06/27/24 19:16		Kareem	OK
107	P2813-22	MYD506	SAM	06/27/24 19:20		Kareem	OK
108	P2814-01	MYD4H3	SAM	06/27/24 19:25		Kareem	OK
109	P2814-02	MYD4H3D	DUP	06/27/24 19:29		Kareem	OK
110	P2814-01L	MYD4H3L	SD	06/27/24 19:33		Kareem	OK
111	P2814-03	MYD4H3S	MS	06/27/24 19:38	MS fail for Sb,Se,Ag,Tl - Below RL	Kareem	OK
112	P2814-04	MYD507	SAM	06/27/24 19:42		Kareem	OK
113	P2814-05	MYD508	SAM	06/27/24 19:47		Kareem	OK
114	P2814-06	MYD509	SAM	06/27/24 19:51		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

115	P2814-07	MYD510	SAM	06/27/24 19:55		Kareem	OK
116	P2814-08	MYD511	SAM	06/27/24 20:00		Kareem	OK
117	P2814-09	MYD512	SAM	06/27/24 20:04		Kareem	OK
118	P2814-10	MYD513	SAM	06/27/24 20:08		Kareem	OK
119	P2814-11	MYD514	SAM	06/27/24 20:13		Kareem	OK
120	P2814-12	MYD515	SAM	06/27/24 20:17		Kareem	OK
121	P2814-13	MYD516	SAM	06/27/24 20:21		Kareem	OK
122	CCV070	CCV070	CCV	06/27/24 20:26		Kareem	OK
123	CCB070	CCB070	CCB	06/27/24 20:30		Kareem	OK
124	P2814-14	MYD517	SAM	06/27/24 20:35		Kareem	OK
125	P2814-15	MYD518	SAM	06/27/24 20:39		Kareem	OK
126	P2814-16	MYD519	SAM	06/27/24 20:43		Kareem	OK
127	P2814-17	MYD520	SAM	06/27/24 20:48		Kareem	OK
128	P2814-18	MYD521	SAM	06/27/24 20:52		Kareem	OK
129	P2814-19	MYD522	SAM	06/27/24 20:56	As high	Kareem	Dilution
130	P2814-20	MYD523	SAM	06/27/24 21:01		Kareem	OK
131	P2814-21	MYD524	SAM	06/27/24 21:05		Kareem	OK
132	P2814-22	MYD525	SAM	06/27/24 21:09		Kareem	OK
133	P2815-01	MYD526	SAM	06/27/24 21:14		Kareem	OK
134	P2815-02	MYD527	SAM	06/27/24 21:18		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

135	P2815-03	MYD528	SAM	06/27/24 21:22		Kareem	OK
136	P2815-04	MYD529	SAM	06/27/24 21:27		Kareem	OK
137	P2815-05	MYD530	SAM	06/27/24 21:31		Kareem	OK
138	P2815-06	MYD531	SAM	06/27/24 21:35		Kareem	OK
139	P2815-07	MYD532	SAM	06/27/24 21:40		Kareem	OK
140	P2815-08	MYD533	SAM	06/27/24 21:44		Kareem	OK
141	P2815-09	MYD534	SAM	06/27/24 21:48		Kareem	OK
142	P2815-10	MYD535	SAM	06/27/24 21:53		Kareem	OK
143	P2815-11	MYD536	SAM	06/27/24 21:57		Kareem	OK
144	CCV071	CCV071	CCV	06/27/24 22:01		Kareem	OK
145	CCB071	CCB071	CCB	06/27/24 22:06		Kareem	OK
146	P2815-12	MYD536D	DUP	06/27/24 22:10		Kareem	OK
147	P2815-11L	MYD536L	SD	06/27/24 22:15		Kareem	OK
148	P2815-13	MYD536S	MS	06/27/24 22:19	MS fail for Zn (1335.09 ppb)& Sb,Se,Ag,Tl - Below RL	Kareem	OK
149	P2815-14	MYD537	SAM	06/27/24 22:23		Kareem	OK
150	P2815-15	MYD538	SAM	06/27/24 22:28		Kareem	OK
151	P2815-16	MYD539	SAM	06/27/24 22:32		Kareem	OK
152	P2815-17	MYD540	SAM	06/27/24 22:36		Kareem	OK
153	P2815-18	MYD541	SAM	06/27/24 22:41		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

154	P2815-19	MYD543	SAM	06/27/24 22:45		Kareem	OK
155	P2815-20	MYD544	SAM	06/27/24 22:50		Kareem	OK
156	P2815-21	MYD545	SAM	06/27/24 22:54		Kareem	OK
157	P2815-22	MYD547	SAM	06/27/24 22:58		Kareem	OK
158	PB161696BL	PBS696	MB	06/27/24 23:03		Kareem	OK
159	PB161696BS	LCS696	LCS	06/27/24 23:07		Kareem	OK
160	PB161697BL	PBS697	MB	06/27/24 23:12		Kareem	OK
161	PB161697BS	LCS697	LCS	06/27/24 23:16		Kareem	OK
162	PB161698BL	PBS698	MB	06/27/24 23:21		Kareem	OK
163	PB161698BS	LCS698	LCS	06/27/24 23:25		Kareem	OK
164	PB161694BL	PBS694	MB	06/27/24 23:30		Kareem	OK
165	PB161694BS	LCS694	LCS	06/27/24 23:34		Kareem	OK
166	CCV072	CCV072	CCV	06/27/24 23:39		Kareem	OK
167	CCB072	CCB072	CCB	06/27/24 23:43		Kareem	OK
168	P2789-02A	MYD425A	PS	06/27/24 23:48	PS for Se - 0.07 mp81115 - 10ml sample	Kareem	OK
169	P2790-03A	MYD447A	PS	06/27/24 23:52	PS for Se - 0.07 mp81115 , TI - 0.05 - mp81116 - 10ml sample	Kareem	OK
170	P2791-01A	MYD430A	PS	06/27/24 23:56	PS for Cr - 0.003 - m5658 , Se - 0.07 mp81115 - 10ml sample	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

171	P2792-01A	MYD438A	PS	06/28/24 00:01	PS for AS-0.0061ml M5801, Ba-0.028ml M5390, Cr-0.0032ml m5658, Cu-0.0096ml M5521, Zn-0.019ml M5819, Se-0.07 MP81115 - 10ml sample	Kareem	OK
172	P2793-01A	MYD442A	PS	06/28/24 00:05	PS for Zn-0.032ml M5819 , Se 0.07 ml MP81115, Ag 0.02 MP80855, TI-0.05 MP81116 - 10ml sample	Kareem	OK
173	P2797-12A	MYD4P9A	PS	06/28/24 00:10	PS for Se, TI(0.07ml mp81115,0.05ml mp81116-10ml sample)	Kareem	OK
174	P2748-01A	MYD3M4A	PS	06/28/24 00:14	PS for Se - 0.07 ml MP81115, TI - 0.05 ml MP81115 - 10ml sample	Kareem	OK
175	P2788-01A	MYD404A	PS	06/28/24 00:18	PS for Se - 0.07 ml MP81115 - 10ml sample	Kareem	OK
176	PB161699BL	PBS699	MB	06/28/24 00:23		Kareem	OK
177	PB161699BS	LCS699	LCS	06/28/24 00:27		Kareem	OK
178	PB161700BL	PBS700	MB	06/28/24 00:32		Kareem	OK
179	PB161700BS	LCS700	LCS	06/28/24 00:36		Kareem	OK
180	PB161701BL	PBS701	MB	06/28/24 00:41		Kareem	OK
181	PB161701BS	LCS701	LCS	06/28/24 00:45		Kareem	OK
182	PB161702BL	PBS702	MB	06/28/24 00:50		Kareem	OK
183	PB161702BS	LCS702	LCS	06/28/24 00:54		Kareem	OK
184	PB161703BL	PBS703	MB	06/28/24 00:59		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

185	PB161703BS	LCS703	LCS	06/28/24 01:03		Kareem	OK
186	PB161736BL	PBW736	MB	06/28/24 01:08	NOT USE	Kareem	Not Ok
187	PB161736BS	LCS736	LCS	06/28/24 01:12	NOT USE	Kareem	Not Ok
188	CCV073	CCV073	CCV	06/28/24 01:17		Kareem	OK
189	CCB073	CCB073	CCB	06/28/24 01:21		Kareem	OK
190	P2816-01	MYD4J8	SAM	06/28/24 01:26		Kareem	OK
191	P2816-02	MYD4J8D	DUP	06/28/24 01:30		Kareem	OK
192	P2816-01L	MYD4J8L	SD	06/28/24 01:35		Kareem	OK
193	P2816-03	MYD4J8S	MS	06/28/24 01:39	MS fail for Zn (535.18 ppb) & Sb,Se,Ag - Below RL	Kareem	OK
194	P2816-04	MYD548	SAM	06/28/24 01:43		Kareem	OK
195	P2816-05	MYD549	SAM	06/28/24 01:48		Kareem	OK
196	P2816-06	MYD550	SAM	06/28/24 01:52		Kareem	OK
197	P2816-07	MYD551	SAM	06/28/24 01:57		Kareem	OK
198	P2816-08	MYD552	SAM	06/28/24 02:01		Kareem	OK
199	P2816-09	MYD553	SAM	06/28/24 02:05		Kareem	OK
200	P2816-10	MYD554	SAM	06/28/24 02:10		Kareem	OK
201	P2816-11	MYD555	SAM	06/28/24 02:14		Kareem	OK
202	P2816-12	MYD556	SAM	06/28/24 02:18		Kareem	OK
203	P2816-13	MYD557	SAM	06/28/24 02:23		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

204	P2816-14	MYD558	SAM	06/28/24 02:27		Kareem	OK
205	P2816-15	MYD559	SAM	06/28/24 02:32		Kareem	OK
206	P2816-16	MYD560	SAM	06/28/24 02:36		Kareem	OK
207	P2816-17	MYD561	SAM	06/28/24 02:40		Kareem	OK
208	P2816-18	MYD562	SAM	06/28/24 02:45		Kareem	OK
209	P2816-19	MYD563	SAM	06/28/24 02:49		Kareem	OK
210	CCV074	CCV074	CCV	06/28/24 02:54		Kareem	OK
211	CCB074	CCB074	CCB	06/28/24 02:58		Kareem	OK
212	P2816-20	MYD564	SAM	06/28/24 03:02		Kareem	OK
213	P2816-21	MYD565	SAM	06/28/24 03:07		Kareem	OK
214	P2816-22	MYD566	SAM	06/28/24 03:11		Kareem	OK
215	P2817-01	MYD567	SAM	06/28/24 03:16		Kareem	OK
216	P2817-02	MYD568	SAM	06/28/24 03:20		Kareem	OK
217	P2817-03	MYD569	SAM	06/28/24 03:24		Kareem	OK
218	P2817-04	MYD570	SAM	06/28/24 03:29		Kareem	OK
219	P2817-05	MYD570D	DUP	06/28/24 03:33		Kareem	OK
220	P2817-04L	MYD570L	SD	06/28/24 03:38		Kareem	OK
221	P2817-06	MYD570S	MS	06/28/24 03:42	MS fail for As (154.43 ppb) & Sb,Se,Ag - Below RL	Kareem	OK
222	P2817-07	MYD571	SAM	06/28/24 03:46		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

223	P2817-08	MYD572	SAM	06/28/24 03:51		Kareem	OK
224	P2817-09	MYD574	SAM	06/28/24 03:55		Kareem	OK
225	P2817-10	MYD575	SAM	06/28/24 03:59		Kareem	OK
226	P2817-11	MYD576	SAM	06/28/24 04:04		Kareem	OK
227	P2817-12	MYD577	SAM	06/28/24 04:08		Kareem	OK
228	P2817-13	MYD578	SAM	06/28/24 04:13		Kareem	OK
229	P2817-14	MYD579	SAM	06/28/24 04:17		Kareem	OK
230	P2817-15	MYD580	SAM	06/28/24 04:21		Kareem	OK
231	P2817-16	MYD582	SAM	06/28/24 04:26		Kareem	OK
232	CCV075	CCV075	CCV	06/28/24 04:30		Kareem	OK
233	CCB075	CCB075	CCB	06/28/24 04:35		Kareem	OK
234	P2817-17	MYD583	SAM	06/28/24 04:39		Kareem	OK
235	P2817-18	MYD584	SAM	06/28/24 04:43		Kareem	OK
236	P2817-19	MYD585	SAM	06/28/24 04:48		Kareem	OK
237	P2817-20	MYD586	SAM	06/28/24 04:52		Kareem	OK
238	P2817-21	MYD587	SAM	06/28/24 04:56		Kareem	OK
239	P2817-22	MYD588	SAM	06/28/24 05:01		Kareem	OK
240	P2818-01	MYD581	SAM	06/28/24 05:05		Kareem	OK
241	P2818-02	MYD581D	DUP	06/28/24 05:10		Kareem	OK
242	P2818-01L	MYD581L	SD	06/28/24 05:14		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

243	P2818-03	MYD581S	MS	06/28/24 05:18	MS fail for Zn (1781.92 ppb) & Sb,Se - Below RL	Kareem	OK
244	P2818-04	MYD589	SAM	06/28/24 05:23		Kareem	OK
245	P2818-05	MYD590	SAM	06/28/24 05:27		Kareem	OK
246	P2818-06	MYD591	SAM	06/28/24 05:31		Kareem	OK
247	P2818-07	MYD592	SAM	06/28/24 05:36		Kareem	OK
248	P2818-08	MYD593	SAM	06/28/24 05:40		Kareem	OK
249	P2818-09	MYD594	SAM	06/28/24 05:44		Kareem	OK
250	P2818-10	MYD595	SAM	06/28/24 05:49		Kareem	OK
251	P2818-11	MYD596	SAM	06/28/24 05:53		Kareem	OK
252	P2818-12	MYD597	SAM	06/28/24 05:57		Kareem	OK
253	P2818-13	MYD598	SAM	06/28/24 06:02		Kareem	OK
254	CCV076	CCV076	CCV	06/28/24 06:06		Kareem	OK
255	CCB076	CCB076	CCB	06/28/24 06:11		Kareem	OK
256	P2818-14	MYD5A0	SAM	06/28/24 06:15		Kareem	OK
257	P2818-15	MYD5A1	SAM	06/28/24 06:19		Kareem	OK
258	P2818-16	MYD5A2	SAM	06/28/24 06:24		Kareem	OK
259	P2818-17	MYD5A3	SAM	06/28/24 06:28		Kareem	OK
260	P2818-18	MYD5A4	SAM	06/28/24 06:33		Kareem	OK
261	P2818-19	MYD5A5	SAM	06/28/24 06:37		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

262	P2818-20	MYD5A6	SAM	06/28/24 06:41		Kareem	OK
263	P2818-21	MYD5A7	SAM	06/28/24 06:46		Kareem	OK
264	P2818-22	MYD5A8	SAM	06/28/24 06:50		Kareem	OK
265	P2819-01	MYD5A9	SAM	06/28/24 06:54		Kareem	OK
266	P2819-02	MYD5B0	SAM	06/28/24 06:59		Kareem	OK
267	P2819-03	MYD5B1	SAM	06/28/24 07:03		Kareem	OK
268	P2819-04	MYD5B2	SAM	06/28/24 07:07		Kareem	OK
269	P2819-05	MYD5B3	SAM	06/28/24 07:12		Kareem	OK
270	P2819-06	MYD5B3D	DUP	06/28/24 07:16		Kareem	OK
271	P2819-05L	MYD5B3L	SD	06/28/24 07:20		Kareem	OK
272	P2819-07	MYD5B3S	MS	06/28/24 07:25	MS fail for Cr (830.18 ppb), Cu (520.67 ppb), Zn (607.25 ppb) & Sb,Ag - Below RL	Kareem	OK
273	P2819-08	MYD5B4	SAM	06/28/24 07:29		Kareem	OK
274	P2819-09	MYD5B5	SAM	06/28/24 07:33		Kareem	OK
275	P2819-10	MYD5B6	SAM	06/28/24 07:38		Kareem	OK
276	CCV077	CCV077	CCV	06/28/24 07:42		Kareem	OK
277	CCB077	CCB077	CCB	06/28/24 07:47		Kareem	OK
278	P2819-11	MYD5B7	SAM	06/28/24 07:51		Kareem	OK
279	P2819-12	MYD5B8	SAM	06/28/24 07:55		Kareem	OK
280	P2819-13	MYD5B9	SAM	06/28/24 08:00		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

281	P2819-14	MYD5C0	SAM	06/28/24 08:04		Kareem	OK
282	P2819-15	MYD5C1	SAM	06/28/24 08:08		Kareem	OK
283	P2819-16	MYD5C2	SAM	06/28/24 08:13		Kareem	OK
284	P2819-17	MYD5C3	SAM	06/28/24 08:17		Kareem	OK
285	P2819-18	MYD5C4	SAM	06/28/24 08:21		Kareem	OK
286	P2819-19	MYD5C6	SAM	06/28/24 08:26		Kareem	OK
287	P2819-20	MYD5C7	SAM	06/28/24 08:30		Kareem	OK
288	P2819-21	MYD5C8	SAM	06/28/24 08:35		Kareem	OK
289	P2819-22	MYD5C9	SAM	06/28/24 08:39		Kareem	OK
290	P2820-01	MYD5D0	SAM	06/28/24 08:43		Kareem	OK
291	P2820-02	MYD5D1	SAM	06/28/24 08:48		Kareem	OK
292	P2820-03	MYD5D2	SAM	06/28/24 08:52		Kareem	OK
293	P2820-04	MYD5D3	SAM	06/28/24 08:56		Kareem	OK
294	P2820-05	MYD5D4	SAM	06/28/24 09:01		Kareem	OK
295	P2820-06	MYD5D4D	DUP	06/28/24 09:05		Kareem	OK
296	P2820-05L	MYD5D4L	SD	06/28/24 09:09		Kareem	OK
297	P2820-07	MYD5D4S	MS	06/28/24 09:14	MS fail for Zn (801.79 ppb) & Sb,Se,Ag,Tl - Below RL	Kareem	OK
298	CCV078	CCV078	CCV	06/28/24 09:18		Kareem	OK
299	CCB078	CCB078	CCB	06/28/24 09:22		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131438

Review By	Sarabjit Jaswal	Review On	6/28/2024 12:19:21 PM
Supervise By	Mohan Bera	Supervise On	6/28/2024 12:22:21 PM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81115,M5819,M5521,M5658,M5390,M5801
LCS Standard	
Chk Standard	MP80942,MP80944

300	P2820-08	MYD5D5	SAM	06/28/24 09:27		Kareem	OK
301	P2820-09	MYD5D6	SAM	06/28/24 09:31		Kareem	OK
302	P2820-10	MYD5D7	SAM	06/28/24 09:35		Kareem	OK
303	P2820-11	MYD5D8	SAM	06/28/24 09:40		Kareem	OK
304	P2820-12	MYD5D9	SAM	06/28/24 09:44		Kareem	OK
305	P2820-13	MYD5E0	SAM	06/28/24 09:48		Kareem	OK
306	P2820-14	MYD5E1	SAM	06/28/24 09:53		Kareem	OK
307	P2820-15	MYD5E2	SAM	06/28/24 09:57		Kareem	OK
308	P2820-16	MYD5E4	SAM	06/28/24 10:01		Kareem	OK
309	P2820-17	MYD5E5	SAM	06/28/24 10:06		Kareem	OK
310	P2820-18	MYD5E6	SAM	06/28/24 10:10		Kareem	OK
311	P2820-19	MYD5E7	SAM	06/28/24 10:15	Mn high	Kareem	Dilution
312	P2820-20	MYD5E8	SAM	06/28/24 10:19		Kareem	OK
313	CCV079	CCV079	CCV	06/28/24 10:23		Kareem	OK
314	CCB079	CCB079	CCB	06/28/24 10:28		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/28/24 14:22		Kareem	OK
2	S1	S01	CAL2	06/28/24 14:26		Kareem	OK
3	S2	S02	CAL3	06/28/24 14:31		Kareem	OK
4	S3	S03	CAL4	06/28/24 14:35		Kareem	OK
5	S4	S04	CAL5	06/28/24 14:40		Kareem	OK
6	S5	S05	CAL6	06/28/24 14:44		Kareem	OK
7	S6	S06	CAL7	06/28/24 14:48		Kareem	OK
8	ICV012	ICV012	ICV	06/28/24 15:04		Kareem	OK
9	ICB012	ICB012	ICB	06/28/24 15:08		Kareem	OK
10	ICSA012	ICSA012	ICSA	06/28/24 15:15		Kareem	OK
11	ICSAB012	ICSAB012	ICSAB	06/28/24 15:20		Kareem	OK
12	CCV080	CCV080	CCV	06/28/24 15:39		Kareem	OK
13	CCB080	CCB080	CCB	06/28/24 15:46		Kareem	OK
14	PB161736BL	PBW736	MB	06/28/24 15:50		Kareem	OK
15	PB161736BS	LCS736	LCS	06/28/24 15:55		Kareem	OK
16	PB161704BL	PBS704	MB	06/28/24 15:59		Kareem	OK
17	PB161704BS	LCS704	LCS	06/28/24 16:04		Kareem	OK
18	PB161705BL	PBS705	MB	06/28/24 16:08		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

19	PB161705BS	LCS705	LCS	06/28/24 16:13		Kareem	OK
20	PB161706BL	PBS706	MB	06/28/24 16:18		Kareem	OK
21	PB161706BS	LCS706	LCS	06/28/24 16:22		Kareem	OK
22	PB161707BL	PBS707	MB	06/28/24 16:27		Kareem	OK
23	PB161707BS	LCS707	LCS	06/28/24 16:31		Kareem	OK
24	PB161708BL	PBS708	MB	06/28/24 16:36		Kareem	OK
25	PB161708BS	LCS708	LCS	06/28/24 16:40		Kareem	OK
26	PB161709BL	PBS709	MB	06/28/24 16:45		Kareem	OK
27	PB161709BS	LCS709	LCS	06/28/24 16:49		Kareem	OK
28	P2975-01	E10R8	SAM	06/28/24 16:54		Kareem	OK
29	P2975-02	E10R9	SAM	06/28/24 16:58		Kareem	OK
30	P2975-03	E10S0	SAM	06/28/24 17:03		Kareem	OK
31	P2975-04	E10S1	SAM	06/28/24 17:07		Kareem	OK
32	P2975-05	E10S2	SAM	06/28/24 17:12		Kareem	OK
33	P2975-06	E10S3	SAM	06/28/24 17:16		Kareem	OK
34	CCV081	CCV081	CCV	06/28/24 17:21		Kareem	OK
35	CCB081	CCB081	CCB	06/28/24 17:25		Kareem	OK
36	P2975-07	E1179	SAM	06/28/24 17:30		Kareem	OK
37	P2975-08	E1180	SAM	06/28/24 17:34		Kareem	OK
38	P3016-18	E11R4	SAM	06/28/24 17:39		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

39	P3016-19	E11R5	SAM	06/28/24 17:43		Kareem	OK
40	P3016-20	E11R6	SAM	06/28/24 17:48		Kareem	OK
41	P2821-01	MYD4L1	SAM	06/28/24 17:52		Kareem	OK
42	P2821-02	MYD4L1D	DUP	06/28/24 17:57		Kareem	OK
43	P2821-01L	MYD4L1L	SD	06/28/24 18:01		Kareem	OK
44	P2821-03	MYD4L1S	MS	06/28/24 18:06	MS fail for Cu(0.3605673PPM), Pb(0.1279198PPM), Se(Below RL), Tl(Below RL), Zn(0.7395081PPM)	Kareem	OK
45	P2821-04	MYD5E9	SAM	06/28/24 18:10		Kareem	OK
46	P2821-05	MYD5F0	SAM	06/28/24 18:14		Kareem	OK
47	P2821-06	MYD5F1	SAM	06/28/24 18:19		Kareem	OK
48	P2821-07	MYD5F2	SAM	06/28/24 18:23		Kareem	OK
49	P2821-08	MYD5F3	SAM	06/28/24 18:27		Kareem	OK
50	P2821-09	MYD5F4	SAM	06/28/24 18:32		Kareem	OK
51	P2821-10	MYD5F5	SAM	06/28/24 18:36		Kareem	OK
52	P2821-11	MYD5F6	SAM	06/28/24 18:40		Kareem	OK
53	P2821-12	MYD5F7	SAM	06/28/24 18:45		Kareem	OK
54	P2821-13	MYD5F8	SAM	06/28/24 18:49		Kareem	OK
55	P2821-14	MYD5F9	SAM	06/28/24 18:54		Kareem	OK
56	CCV082	CCV082	CCV	06/28/24 18:58		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

57	CCB082	CCB082	CCB	06/28/24 19:02		Kareem	OK
58	P2821-15	MYD5G0	SAM	06/28/24 19:07		Kareem	OK
59	P2821-16	MYD5G1	SAM	06/28/24 19:11		Kareem	OK
60	P2821-17	MYD5G2	SAM	06/28/24 19:15		Kareem	OK
61	P2821-18	MYD5G3	SAM	06/28/24 19:20		Kareem	OK
62	P2821-19	MYD5G4	SAM	06/28/24 19:24		Kareem	OK
63	P2821-20	MYD5G5	SAM	06/28/24 19:29		Kareem	OK
64	P2821-21	MYD5G6	SAM	06/28/24 19:33		Kareem	OK
65	P2821-22	MYD5G7	SAM	06/28/24 19:37		Kareem	OK
66	P2809-10A	MYD4R7A	PS	06/28/24 19:42	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
67	P2810-01A	MYD4G6A	PS	06/28/24 19:46	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
68	P2811-17A	MYD4Y3A	PS	06/28/24 19:50	PS For Se- 0.007-MP81024,TI-0.005-M P81025,Zn- 0.032 - M5819 - 10ml sample	Kareem	OK
69	P2813-01A	MYD4Y4A	PS	06/28/24 19:55	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
70	P2814-01A	MYD4H3A	PS	06/28/24 19:59	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

71	P2815-11A	MYD536A	PS	06/28/24 20:03	PS For Se-0.007-MP81024,TI-0.005-M P81025,Zn-0.027-M5819 - 10ml sample	Kareem	OK
72	P2816-01A	MYD4J8A	PS	06/28/24 20:08	PS For Se-0.007-MP81024,TI-0.005-M P81025,Zn-0.01-M5819 - 10ml sample	Kareem	OK
73	P2817-04A	MYD570A	PS	06/28/24 20:12	PS For As-0.031-MP81248,Cr-0.16-MP 81247,Se-0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
74	P2818-01A	MYD581A	PS	06/28/24 20:17	PS For Se-0.007-MP81024,Zn-0.027-M 5819 - 10ml sample	Kareem	OK
75	P2819-05A	MYD5B3A	PS	06/28/24 20:21	PS For Cr-0.17-MP81246,Cu-0.1-MP81247,Se-0.007-MP81024,Zn-0.012-M 5819 - 10ml sample	Kareem	OK
76	P2820-05A	MYD5D4A	PS	06/28/24 20:25	PS For Se-0.007-MP81024,TI-0.005-M P81025,Zn- 0.016 - M5819 - 10ml sample	Kareem	OK
77	P2814-19DL	MYD522	SAM	06/28/24 20:29	20X for As	Kareem	Confirms
78	CCV083	CCV083	CCV	06/28/24 20:34		Kareem	OK
79	CCB083	CCB083	CCB	06/28/24 20:38		Kareem	OK
80	P2820-19DL	MYD5E7	SAM	06/28/24 20:43	Not use	Kareem	Not Ok
81	P2822-01	MYD5H2	SAM	06/28/24 20:47		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

82	P2822-02	MYD5H2D	DUP	06/28/24 20:52		Kareem	OK
83	P2822-01L	MYD5H2L	SD	06/28/24 20:56		Kareem	OK
84	P2822-03	MYD5H2S	MS	06/28/24 21:01	MS fail for As(0.1722545PPM), Cu(0.3878602PPM), Se(Below RL), TI(Below RL), V(0.5143108PPM)	Kareem	OK
85	P2822-04	MYD5H4	SAM	06/28/24 21:05		Kareem	OK
86	P2822-05	MYD5H5	SAM	06/28/24 21:09		Kareem	OK
87	P2822-06	MYD5H6	SAM	06/28/24 21:13		Kareem	OK
88	P2822-07	MYD5H7	SAM	06/28/24 21:18		Kareem	OK
89	P2822-08	MYD5H8	SAM	06/28/24 21:22		Kareem	OK
90	P2822-09	MYD5H9	SAM	06/28/24 21:27		Kareem	OK
91	P2822-10	MYD5J0	SAM	06/28/24 21:31		Kareem	OK
92	P2822-11	MYD5J1	SAM	06/28/24 21:35		Kareem	OK
93	P2822-12	MYD5J2	SAM	06/28/24 21:40		Kareem	OK
94	P2822-13	MYD5J3	SAM	06/28/24 21:44		Kareem	OK
95	P2822-14	MYD5J4	SAM	06/28/24 21:48	Mn high	Kareem	Dilution
96	P2822-15	MYD5J5	SAM	06/28/24 21:53	Mn high	Kareem	Dilution
97	P2822-16	MYD5J6	SAM	06/28/24 21:57	Mn high	Kareem	Dilution
98	P2822-17	MYD5J7	SAM	06/28/24 22:02	Mn high	Kareem	Dilution
99	P2822-18	MYD5J8	SAM	06/28/24 22:06	Mn high	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

100	CCV084	CCV084	CCV	06/28/24 22:10		Kareem	OK
101	CCB084	CCB084	CCB	06/28/24 22:15		Kareem	OK
102	P2822-19	MYD5K0	SAM	06/28/24 22:19	Mn high	Kareem	Dilution
103	P2822-20	MYD5K1	SAM	06/28/24 22:23	Mn high	Kareem	Dilution
104	P2822-21	MYD5K2	SAM	06/28/24 22:28	Mn high	Kareem	Dilution
105	P2822-22	MYD5K4	SAM	06/28/24 22:32	Mn high	Kareem	Dilution
106	P2823-01	MYD5X1	SAM	06/28/24 22:36		Kareem	OK
107	P2823-02	MYD5X1D	DUP	06/28/24 22:41		Kareem	OK
108	P2823-01L	MYD5X1L	SD	06/28/24 22:45		Kareem	OK
109	P2823-03	MYD5X1S	MS	06/28/24 22:50	MS fail for As(0.0417517PPM),Cr(0.60 59198PPM),Pb(0.0818535P PM),Mn(1.991713PPM),Ni(0 .2536192PPM),Se(Below RL)	Kareem	OK
110	P2823-04	MYD5L1	SAM	06/28/24 22:54	Mn high	Kareem	Dilution
111	P2823-05	MYD5L2	SAM	06/28/24 22:58		Kareem	OK
112	P2823-06	MYD5L3	SAM	06/28/24 23:03	Mn high	Kareem	Dilution
113	P2823-07	MYD5L4	SAM	06/28/24 23:07	Mn high	Kareem	Dilution
114	P2823-08	MYD5L5	SAM	06/28/24 23:11	Mn high	Kareem	Dilution
115	P2823-09	MYD5L6	SAM	06/28/24 23:16	Mn high	Kareem	Dilution
116	P2823-10	MYD5L7	SAM	06/28/24 23:20	Mn high	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

117	P2823-11	MYD5L8	SAM	06/28/24 23:24	Mn high	Kareem	Dilution
118	P2823-12	MYD5L9	SAM	06/28/24 23:29		Kareem	OK
119	P2823-13	MYD5M0	SAM	06/28/24 23:33	Mn high	Kareem	Dilution
120	P2823-14	MYD5M1	SAM	06/28/24 23:37	Mn high	Kareem	Dilution
121	P2823-15	MYD5M2	SAM	06/28/24 23:42	Mn high	Kareem	Dilution
122	CCV085	CCV085	CCV	06/28/24 23:46		Kareem	OK
123	CCB085	CCB085	CCB	06/28/24 23:50		Kareem	OK
124	P2823-16	MYD5M3	SAM	06/28/24 23:55	Mn high	Kareem	Dilution
125	P2823-17	MYD5M4	SAM	06/28/24 23:59	Mn high	Kareem	Dilution
126	P2823-18	MYD5M5	SAM	06/29/24 00:04	Mn high	Kareem	Dilution
127	P2823-19	MYD5M6	SAM	06/29/24 00:08	Mn high	Kareem	Dilution
128	P2823-20	MYD5M7	SAM	06/29/24 00:12	Mn high	Kareem	Dilution
129	P2823-21	MYD5M8	SAM	06/29/24 00:17	Mn high	Kareem	Dilution
130	P2823-22	MYD5M9	SAM	06/29/24 00:21	Mn high	Kareem	Dilution
131	P2824-01	MYD5K5	SAM	06/29/24 00:25	Mn high	Kareem	Dilution
132	P2824-02	MYD5K5D	DUP	06/29/24 00:30	Mn high	Kareem	Dilution
133	P2824-01L	MYD5K5L	SD	06/29/24 00:34		Kareem	OK
134	P2824-03	MYD5K5S	MS	06/29/24 00:38	Mn high,MS fail for Cr(0.6045513PPM),Se(Below RL),Ti(Below RL)	Kareem	Dilution
135	P2824-04	MYD5N0	SAM	06/29/24 00:43		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

136	P2824-05	MYD5N1	SAM	06/29/24 00:47	Mn high	Kareem	Dilution
137	P2824-06	MYD5N2	SAM	06/29/24 00:51	Mn high	Kareem	Dilution
138	P2824-07	MYD5N3	SAM	06/29/24 00:56	Mn high	Kareem	Dilution
139	P2824-08	MYD5N4	SAM	06/29/24 01:00	Mn high	Kareem	Dilution
140	P2824-09	MYD5N5	SAM	06/29/24 01:04	Mn high	Kareem	Dilution
141	P2824-10	MYD5N6	SAM	06/29/24 01:09	Mn high	Kareem	Dilution
142	P2824-11	MYD5N7	SAM	06/29/24 01:13	Mn high	Kareem	Dilution
143	P2824-12	MYD5N8	SAM	06/29/24 01:17	Mn high	Kareem	Dilution
144	CCV086	CCV086	CCV	06/29/24 01:22		Kareem	OK
145	CCB086	CCB086	CCB	06/29/24 01:26		Kareem	OK
146	P2824-13	MYD5N9	SAM	06/29/24 01:31	Mn high	Kareem	Dilution
147	P2824-14	MYD5P0	SAM	06/29/24 01:35		Kareem	OK
148	P2824-15	MYD5P1	SAM	06/29/24 01:40		Kareem	OK
149	P2824-16	MYD5P2	SAM	06/29/24 01:44		Kareem	OK
150	P2824-17	MYD5P3	SAM	06/29/24 01:48		Kareem	OK
151	P2824-18	MYD5P4	SAM	06/29/24 01:53		Kareem	OK
152	P2824-19	MYD5P5	SAM	06/29/24 01:57		Kareem	OK
153	P2824-20	MYD5P6	SAM	06/29/24 02:01	Mn high	Kareem	Dilution
154	P2824-21	MYD5P7	SAM	06/29/24 02:06		Kareem	OK
155	P2824-22	MYD5P8	SAM	06/29/24 02:10		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

156	P2825-01	MYD542	SAM	06/29/24 02:14		Kareem	OK
157	P2825-02	MYD542D	DUP	06/29/24 02:19		Kareem	OK
158	P2825-01L	MYD542L	SD	06/29/24 02:23		Kareem	OK
159	P2825-03	MYD542S	MS	06/29/24 02:28	MS fail for As(0.2705676PPM), Cu(0.3912561PPM), Se(Below RL), Zn(1.091943PPM)	Kareem	OK
160	P2825-04	MYD5P9	SAM	06/29/24 02:32		Kareem	OK
161	P2825-05	MYD5Q0	SAM	06/29/24 02:36		Kareem	OK
162	P2825-06	MYD5Q1	SAM	06/29/24 02:41		Kareem	OK
163	P2825-07	MYD5Q2	SAM	06/29/24 02:45		Kareem	OK
164	P2825-08	MYD5Q3	SAM	06/29/24 02:49		Kareem	OK
165	P2825-09	MYD5Q4	SAM	06/29/24 02:54		Kareem	OK
166	P2825-10	MYD5Q5	SAM	06/29/24 03:07		Kareem	OK
167	P2825-11	MYD5Q6	SAM	06/29/24 03:11	Mn high	Kareem	Dilution
168	CCV087	CCV087	CCV	06/29/24 03:16		Kareem	OK
169	CCB087	CCB087	CCB	06/29/24 03:25		Kareem	OK
170	P2825-12	MYD5Q7	SAM	06/29/24 03:30		Kareem	OK
171	P2825-13	MYD5Q8	SAM	06/29/24 03:34	Mn high	Kareem	Dilution
172	P2825-14	MYD5Q9	SAM	06/29/24 03:39		Kareem	OK
173	P2825-15	MYD5R0	SAM	06/29/24 03:43		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

174	P2825-16	MYD5R1	SAM	06/29/24 03:47		Kareem	OK
175	P2825-17	MYD5R2	SAM	06/29/24 03:52	Mn high	Kareem	Dilution
176	P2825-18	MYD5G8	SAM	06/29/24 03:56		Kareem	OK
177	P2825-19	MYD5G9	SAM	06/29/24 04:00		Kareem	OK
178	P2825-20	MYD5H0	SAM	06/29/24 04:05		Kareem	OK
179	P2825-21	MYD5H1	SAM	06/29/24 04:09		Kareem	OK
180	P2826-01	MYD546	SAM	06/29/24 04:13		Kareem	OK
181	P2826-01L	MYD546L	SD	06/29/24 04:18		Kareem	OK
182	P2826-02	MYD546D	DUP	06/29/24 04:22		Kareem	OK
183	P2826-03	MYD546S	MS	06/29/24 04:27	MS fail for As(0.2951106PPM),Se(Below RL)	Kareem	OK
184	P2826-04	MYD5R3	SAM	06/29/24 04:31	Mn high	Kareem	Dilution
185	P2826-05	MYD5R4	SAM	06/29/24 04:35		Kareem	OK
186	P2826-06	MYD5R5	SAM	06/29/24 04:40		Kareem	OK
187	P2826-07	MYD5R6	SAM	06/29/24 04:44		Kareem	OK
188	CCV088	CCV088	CCV	06/29/24 04:48		Kareem	OK
189	CCB088	CCB088	CCB	06/29/24 04:53		Kareem	OK
190	P2826-08	MYD5R7	SAM	06/29/24 04:57		Kareem	OK
191	P2826-09	MYD5R8	SAM	06/29/24 05:02		Kareem	OK
192	P2826-10	MYD5R9	SAM	06/29/24 05:06		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	
STD. NAME	STD REF.#		
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935		
ICV Standard	MP80931		
CCV Standard	MP80934		
ICSA Standard	MP80932,MP80933		
CRI Standard			
LCS Standard			
Chk Standard	MP80942,MP80944		

193	P2826-11	MYD5S0	SAM	06/29/24 05:10	Mn high	Kareem	Dilution
194	P2826-12	MYD5S1	SAM	06/29/24 05:15		Kareem	OK
195	P2826-13	MYD5S2	SAM	06/29/24 05:19		Kareem	OK
196	P2826-14	MYD5S3	SAM	06/29/24 05:23		Kareem	OK
197	P2826-15	MYD5S4	SAM	06/29/24 05:28	Mn high	Kareem	Dilution
198	P2826-16	MYD5S5	SAM	06/29/24 05:32		Kareem	OK
199	P2826-17	MYD5S6	SAM	06/29/24 05:37	Mn high	Kareem	Dilution
200	P2826-18	MYD5S7	SAM	06/29/24 05:41		Kareem	OK
201	P2826-19	MYD5S8	SAM	06/29/24 05:45		Kareem	OK
202	P2826-20	MYD5S9	SAM	06/29/24 05:50	Mn high	Kareem	Dilution
203	P2826-21	MYD5T0	SAM	06/29/24 05:54		Kareem	OK
204	P2826-22	MYD5T1	SAM	06/29/24 05:58		Kareem	OK
205	P2828-01	MYD5H3	SAM	06/29/24 06:03		Kareem	OK
206	P2828-01L	MYD5H3L	SD	06/29/24 06:07		Kareem	OK
207	P2828-02	MYD5H3D	DUP	06/29/24 06:11		Kareem	OK
208	P2828-03	MYD5H3S	MS	06/29/24 06:16	MS fail for Cr(0.3328370PPM),Cu(0.65 30863ppm).Se(Below RL),Tl(Below RL),Zn(1.610189PPM)	Kareem	OK
209	P2828-04	MYD5K6	SAM	06/29/24 06:20	Mn high	Kareem	Dilution
210	CCV089	CCV089	CCV	06/29/24 06:24		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

211	CCB089	CCB089	CCB	06/29/24 06:29		Kareem	OK
212	P2828-05	MYD5K7	SAM	06/29/24 06:33	Mn high	Kareem	Dilution
213	P2828-06	MYD5K8	SAM	06/29/24 06:38	Mn high	Kareem	Dilution
214	P2828-07	MYD5K9	SAM	06/29/24 06:42	Mn high	Kareem	Dilution
215	P2828-08	MYD5L0	SAM	06/29/24 06:46	Mn high	Kareem	Dilution
216	P2828-09	MYD4E6	SAM	06/29/24 06:50		Kareem	OK
217	P2828-10	MYD4E7	SAM	06/29/24 06:55		Kareem	OK
218	P2828-11	MYD4E8	SAM	06/29/24 06:59		Kareem	OK
219	P2828-12	MYD4E9	SAM	06/29/24 07:04		Kareem	OK
220	P2828-13	MYD4F0	SAM	06/29/24 07:08		Kareem	OK
221	P2828-14	MYD4F1	SAM	06/29/24 07:12		Kareem	OK
222	P2828-15	MYD4F2	SAM	06/29/24 07:17		Kareem	OK
223	P2828-16	MYD4F3	SAM	06/29/24 07:21		Kareem	OK
224	P2828-17	MYD4F4	SAM	06/29/24 07:25		Kareem	OK
225	P2828-18	MYD4F5	SAM	06/29/24 07:29		Kareem	OK
226	P2828-19	MYD4F6	SAM	06/29/24 07:34		Kareem	OK
227	P2820-19DL2	MYD5E7	SAM	06/29/24 07:38	2X for Mn	Kareem	Confirms
228	P2829-01	MYD5T2	SAM	06/29/24 07:43		Kareem	OK
229	P2829-02	MYD5T3	SAM	06/29/24 07:47		Kareem	OK
230	P2829-03	MYD5T4	SAM	06/29/24 07:51		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

231	P2829-04	MYD5T5	SAM	06/29/24 07:56		Kareem	OK
232	CCV090	CCV090	CCV	06/29/24 08:00		Kareem	OK
233	CCB090	CCB090	CCB	06/29/24 08:04		Kareem	OK
234	PB161710BL	PBS710	MB	06/29/24 08:09		Kareem	OK
235	PB161710BS	LCS710	LCS	06/29/24 08:13		Kareem	OK
236	PB161711BL	PBS711	MB	06/29/24 08:18		Kareem	OK
237	PB161711BS	LCS711	LCS	06/29/24 08:22		Kareem	OK
238	PB161712BL	PBS712	MB	06/29/24 08:27		Kareem	OK
239	PB161712BS	LCS712	LCS	06/29/24 08:31		Kareem	OK
240	PB161713BL	PBS713	MB	06/29/24 08:36		Kareem	OK
241	PB161713BS	LCS713	LCS	06/29/24 08:40		Kareem	OK
242	P2829-05	MYD5T6	SAM	06/29/24 08:45		Kareem	OK
243	P2829-06	MYD5T7	SAM	06/29/24 08:49		Kareem	OK
244	P2829-07	MYD5T8	SAM	06/29/24 08:53		Kareem	OK
245	P2829-08	MYD5T9	SAM	06/29/24 08:58		Kareem	OK
246	P2829-09	MYD5T9D	DUP	06/29/24 09:02		Kareem	OK
247	P2829-08L	MYD5T9L	SD	06/29/24 09:06		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

248	P2829-10	MYD5T9S	MS	06/29/24 09:11	MS fail for Ba(3.039342ppm),Cr(0.5076730ppm),Cu(1.570988ppm),Se(Below RL),Ti(Below RL),V(1.385149ppm),Zn(1.300744ppm)	Kareem	OK
249	P2829-11	MYD5W0	SAM	06/29/24 09:15		Kareem	OK
250	P2829-12	MYD5W1	SAM	06/29/24 09:19	Mn high	Kareem	Dilution
251	P2829-13	MYD5W2	SAM	06/29/24 09:24	Mn high	Kareem	Dilution
252	P2829-14	MYD5W3	SAM	06/29/24 09:28	Mn high	Kareem	Dilution
253	P2829-15	MYD5W4	SAM	06/29/24 09:32		Kareem	OK
254	CCV091	CCV091	CCV	06/29/24 09:37		Kareem	OK
255	CCB091	CCB091	CCB	06/29/24 09:41		Kareem	OK
256	P2829-16	MYD5W5	SAM	06/29/24 09:46		Kareem	OK
257	P2829-17	MYD5W6	SAM	06/29/24 09:50	Mn high	Kareem	Dilution
258	P2829-18	MYD5X0	SAM	06/29/24 09:54		Kareem	OK
259	P2829-19	MYD5X2	SAM	06/29/24 09:59		Kareem	OK
260	P2829-20	MYD5X3	SAM	06/29/24 10:03		Kareem	OK
261	P2830-01	MYD573	SAM	06/29/24 10:07		Kareem	OK
262	P2830-02	MYD573D	DUP	06/29/24 10:12		Kareem	OK
263	P2830-01L	MYD573L	SD	06/29/24 10:16		Kareem	OK
264	P2830-03	MYD573S	MS	06/29/24 10:21	MS fail for Se(Below RL),Ti(Below RL),Zn(1.235142ppm)	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

265	P2831-01	MYD5C5	SAM	06/29/24 10:25		Kareem	OK
266	P2831-02	MYD5C5D	DUP	06/29/24 10:29		Kareem	OK
267	P2831-01L	MYD5C5L	SD	06/29/24 10:34		Kareem	OK
268	P2831-03	MYD5C5S	MS	06/29/24 10:38	MS fail for Cu(0.7983432ppm),Pb(0.10 47544ppm),Se(Below RL),Tl(Below RL)	Kareem	OK
269	P2832-01	MYD5E3	SAM	06/29/24 10:42		Kareem	OK
270	P2832-02	MYD5E3D	DUP	06/29/24 10:47		Kareem	OK
271	P2832-01L	MYD5E3L	SD	06/29/24 10:51		Kareem	OK
272	P2832-03	MYD5E3S	MS	06/29/24 10:55	MS fail for Cu(1.167239ppm),Se(Below RL),Zn(1.444952ppm)	Kareem	OK
273	P2833-01	MYD5J9	SAM	06/29/24 11:00	Mn high	Kareem	Dilution
274	P2833-02	MYD5J9D	DUP	06/29/24 11:04	Mn high	Kareem	Dilution
275	P2833-01L	MYD5J9L	SD	06/29/24 11:08		Kareem	OK
276	CCV092	CCV092	CCV	06/29/24 11:13		Kareem	OK
277	CCB092	CCB092	CCB	06/29/24 11:17		Kareem	OK
278	P2833-03	MYD5J9S	MS	06/29/24 11:22	Mn high,MS fail for Ba(4.687471ppm),Se(Below RL),Tl(Below RL),Zn(1.234995ppm)	Kareem	Dilution
279	P2834-01	MYD5K3	SAM	06/29/24 11:26		Kareem	OK
280	P2834-02	MYD5K3D	DUP	06/29/24 11:30	Mn high	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

281	P2834-01L	MYD5K3L	SD	06/29/24 11:35		Kareem	OK
282	P2834-03	MYD5K3S	MS	06/29/24 11:39	Mn high,MS fail for Ba(1.897607ppm),Se(Below RL),Ti(Below RL)	Kareem	Dilution
283	P3031-01	E11M7	SAM	06/29/24 11:43		Kareem	OK
284	P3031-02	E11M7D	DUP	06/29/24 11:48		Kareem	OK
285	P3031-01L	E11M7L	SD	06/29/24 11:52		Kareem	OK
286	P3031-03	E11M7S	MS	06/29/24 11:56	MS fail for As(0.1258610ppm)	Kareem	OK
287	P3032-01	E11E8	SAM	06/29/24 12:01		Kareem	OK
288	P3032-02	E11E8D	DUP	06/29/24 12:05		Kareem	OK
289	P3032-01L	E11E8L	SD	06/29/24 12:10		Kareem	OK
290	P3032-03	E11E8S	MS	06/29/24 12:14		Kareem	OK
291	P2835-01	MYD4A6	SAM	06/29/24 12:18		Kareem	OK
292	P2835-02	MYD4A7	SAM	06/29/24 12:23		Kareem	OK
293	P2835-03	MYD4A8	SAM	06/29/24 12:27		Kareem	OK
294	P2835-04	MYD4A9	SAM	06/29/24 12:31		Kareem	OK
295	P2835-05	MYD4A9D	DUP	06/29/24 12:36		Kareem	OK
296	P2835-04L	MYD4A9L	SD	06/29/24 12:40		Kareem	OK
297	P2835-06	MYD4A9S	MS	06/29/24 12:45	P2835 : MS fail for As(0.2225569ppm),Ba(1.530131ppm),Se(Below RL)	Kareem	OK
298	CCV093	CCV093	CCV	06/29/24 12:49		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By		Supervise On	

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	
LCS Standard	
Chk Standard	MP80942,MP80944

299	CCB093	CCB093	CCB	06/29/24 12:53		Kareem	OK
300	P2835-07	MYD4B0	SAM	06/29/24 12:58		Kareem	OK
301	P2835-08	MYD4B1	SAM	06/29/24 13:02		Kareem	OK
302	P2835-09	MYD4B2	SAM	06/29/24 13:07		Kareem	OK
303	P2835-10	MYD4B3	SAM	06/29/24 13:11		Kareem	OK
304	P2835-11	MYD4B4	SAM	06/29/24 13:15		Kareem	OK
305	P2835-12	MYD4B5	SAM	06/29/24 13:20		Kareem	OK
306	P2835-13	MYD4B6	SAM	06/29/24 13:24		Kareem	OK
307	P2835-14	MYD4B7	SAM	06/29/24 13:29		Kareem	OK
308	P2835-15	MYD4B8	SAM	06/29/24 13:33		Kareem	OK
309	P2835-16	MYD4B9	SAM	06/29/24 13:37		Kareem	OK
310	P2835-17	MYD4C0	SAM	06/29/24 13:42		Kareem	OK
311	P2835-18	MYD4C1	SAM	06/29/24 13:46		Kareem	OK
312	P2835-19	MYD4C2	SAM	06/29/24 13:50		Kareem	OK
313	P2835-20	MYD4C3	SAM	06/29/24 13:55		Kareem	OK
314	P2835-21	MYD4C4	SAM	06/29/24 13:59		Kareem	OK
315	P2835-22	MYD4C5	SAM	06/29/24 14:04		Kareem	OK
316	CCV094	CCV094	CCV	06/29/24 14:08		Kareem	OK
317	CCB094	CCB094	CCB	06/29/24 14:12		Kareem	OK