

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No. : SDG No.: MYD4Y8
 Initial Calibration Verification Source : MP82578
 Continuing Calibration Verification Source : MP82579
 Run Batch: LB132751 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	200	102	1	500	510	102	2	500	99	1
Arsenic	200	210	104	1	500	480	96	0	480	97	1
Barium	100	100	102	1	2500	2600	104	1	2500	102	1
Beryllium	100	110	107	0	500	500	101	1	490	97	0
Cadmium	100	100	105	1	500	500	100	1	480	97	1
Chromium	100	97	97	0	500	500	99	1	510	102	1
Cobalt	100	100	105	1	500	470	94	0	490	98	2
Copper	100	100	102	3	5000	4500	91	1	4700	95	1
Lead	200	190	97	0	2500	2600	102	1	2600	102	1
Nickel	110	100	91	0	500	470	94	0	490	97	1
Selenium	200	210	107	1	500	480	97	0	480	96	1
Silver	50.0	54	109	2	500	490	99	1	470	95	2
Thallium	210	190	91	1	500	520	104	1	520	103	1
Vanadium	100	93	93	1	500	500	101	0	510	103	1
Zinc	200	200	98	0	5000	4600	92	1	4800	95	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51495 MA No. : SDG No.: MYD4Y8Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB132751 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID: CCV004		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	480	96	2	490	98	2
Arsenic					500	480	97	1	480	96	1
Barium					2500	2500	99	2	2500	100	0
Beryllium					500	460	92	2	480	96	1
Cadmium					500	470	94	1	500	101	1
Chromium					500	510	103	0	510	102	1
Cobalt					500	490	98	1	490	97	1
Copper					5000	4800	97	1	4800	95	1
Lead					2500	2600	102	2	2500	102	1
Nickel					500	490	98	2	490	98	2
Selenium					500	470	95	2	480	95	1
Silver					500	460	92	2	490	97	1
Thallium					500	520	104	1	510	103	0
Vanadium					500	520	105	1	510	103	1
Zinc					5000	4800	96	2	4800	96	1