


Batch Folder: D:\Agilent\ICPMH\1\DATA\P7101824-MS.b\
Analysis File: P7101824-MS.batch.bin
DA Date-Time: 2024-10-18 16:42:04
Calibration Title:
Calibration Method: External Calibration
VIS Interpolation Fit:

Level	Standard Data File	Sample Name	Acq. Date-Time
1	004CALB.d	S00	2024-10-18 11:46:13
2	005CAL.S.d	S02	2024-10-18 11:49:35
3	006CAL.S.d	S03	2024-10-18 11:53:00
4	007CAL.S.d	S04	2024-10-18 11:56:17
5	008CAL.S.d	S05	2024-10-18 11:59:27
6	009CAL.S.d	S06	2024-10-18 12:02:28
7	010CAL.S.d	S07	2024-10-18 12:05:26
8	011CAL.S.d	S08	2024-10-18 12:08:18

9 Be [No Gas] ISTD:6 Li [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	35.56	0.0001	P	48.1
2	☐	1.000	1.074	853.36	0.0012	P	7.9
3	☐	50.000	52.955	40734.31	0.0568	P	0.4
4	☐	125.000	131.204	98467.94	0.1407	P	0.9
5	☐	250.000	258.343	189928.74	0.2769	P	1.1
6	☐	500.000	507.773	365397.15	0.5443	P	1.6
7	☐	1000.000	993.105	712753.69	1.0645	P	1.2
8	☐			421.12	0.0006	P	12.1

$y = 0.0011 * x + 5.1797E-005$

$R = 0.9999$


$DL = 0.06971$

$BEC = 0.04833$

Weight: <None>


Min Conc: 0

12 C [No Gas] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

12 C [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

23 Na [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	7388.50	0.0544	P	2.9
2	☐	500.000	482.836	129794.45	0.9562	P	0.4
3	☐	5000.000	4717.655	1199439.84	8.8656	A	1.5
4	☐	12500.000	12108.326	2938633.57	22.6691	A	0.2
5	☐	25000.000	24017.612	5578423.25	44.9120	A	1.2
6	☐	50000.000	48935.018	11114437.06	91.4502	A	0.9
7	☐	100000.00	96533.420	22096994.39	180.3495	A	0.5
8	☐	500000.00	500861.56	107863426.1	935.5115	A	2.6

$y = 0.0019 * x + 0.0544$


$R = 1.0000$

$DL = 2.529$

$BEC = 29.14$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	1458.98	0.0108	P	6.6
2	☐	500.000	473.471	77398.78	0.5702	P	1.1
3	☐	5000.000	4619.071	739918.65	5.4687	P	0.6
4	☐	12500.000	11375.553	1743715.97	13.4523	A	1.2
5	☐	25000.000	22641.174	3324378.94	26.7639	A	0.7
6	☐	50000.000	45469.758	6531189.90	53.7385	A	1.3
7	☐	100000.00	89989.295	13029361.61	106.3435	A	0.9
8	☐	500000.00	502605.05	68465590.09	593.8967	A	3.3

$$y = 0.0012 * x + 0.0108$$


$$R = 0.9998$$

$$DL = 1.79$$

$$BEC = 9.1$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	111.11	0.0008	P	10.3
2	☐	20.000	18.035	1840.13	0.0136	P	8.8
3	☐	1000.000	970.326	92845.24	0.6862	P	0.3
4	☐	2500.000	2439.935	223506.12	1.7243	P	1.3
5	☐	5000.000	4898.559	429879.23	3.4609	P	0.7
6	☐	10000.000	9758.888	837848.97	6.8940	P	1.3
7	☐	20000.000	18857.821	1632135.59	13.3211	A	0.4
8	☐	100000.00	100259.41	8163990.78	70.8193	A	3.4

$$y = 7.0635E-004 * x + 8.1699E-004$$

$$R = 0.9999$$


$$DL = 0.358$$

$$BEC = 1.157$$

Weight: <None>

Min Conc: 0

28 Si [No Gas] ISTD :45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	185791.17	0.1212	P	6.8
2	☐	10.000	3.505	197679.45	0.1257	P	0.1
3	☐	50.000	52.424	293438.98	0.1883	P	0.7
4	☐	125.000	131.422	436027.23	0.2895	P	0.7
5	☐	250.000	263.519	655675.57	0.4587	P	2.1
6	☐	500.000	517.461	1078765.84	0.7839	P	3.2
7	☐	1000.000	987.031	1915483.96	1.3852	A	2.1
8	☐			181047.22	0.1359	P	2.0

$$y = 0.0013 * x + 0.1212$$

R = 0.9996


DL = 19.35

BEC = 94.62

Weight: <None>

Min Conc: 0

31 P [He] ISTD :45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	-16.247	87.78	0.0006	P	6.6
2	☐	0.000	16.247	231.12	0.0017	P	5.9
3	☐	1000.000	1019.153	4642.98	0.0343	P	2.3
4	☐	2500.000	2573.387	10999.59	0.0849	P	1.8
5	☐	5000.000	5025.243	20444.85	0.1646	P	0.8
6	☐	10000.000	10177.957	40370.95	0.3322	P	1.0
7	☐	20000.000	19894.580	79411.11	0.6482	P	1.5
8	☐			134.45	0.0012	P	13.2

$$y = 3.2521E-005 * x + 0.0012$$

R = 0.9999


DL = 6.565

BEC = 36.11

Weight: <None>

Min Conc: 0

34 S [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	104.685	117288.92	0.0765	P	5.7
2	☐	0.000	-104.685	117080.66	0.0744	P	0.6
3	☐	1000.000	925.692	131582.77	0.0844	P	0.1
4	☐	2500.000	2300.609	147315.20	0.0978	P	0.7
5	☐	5000.000	4810.879	174691.19	0.1222	P	1.6
6	☐	10000.000	9809.716	235025.26	0.1708	P	2.5
7	☐	20000.000	20171.062	375345.68	0.2714	P	0.9
8	☐			97537.63	0.0732	P	1.5

$y = 9.7155E-006 * x + 0.0754$

R = 0.9998


DL = 743.1

BEC = 7766

Weight: <None>


Min Conc: 0

35 Cl [No Gas] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

35 Cl [He] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐						
2	☐						
3	☐						
4	☐						
5	☐						
6	☐						
7	☐						
8	☐						

39 K [He] ISTD :45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	34901.06	0.2571	P	3.3
2	☐	500.000	498.429	129178.91	0.9517	P	0.5
3	☐	2500.000	2599.147	524831.91	3.8791	P	1.1
4	☐	6250.000	6576.153	1221227.25	9.4213	P	1.0
5	☐	12500.000	12307.219	2162273.37	17.4079	A	0.7
6	☐	25000.000	24259.401	4140132.71	34.0639	A	0.4
7	☐	50000.000	47752.077	8184733.63	66.8023	A	0.6
8	☐	250000.00	250524.14	40269401.07	349.3765	A	4.3

$y = 0.0014 * x + 0.2571$

$R = 0.9999$


$DL = 18.2$

$BEC = 184.5$

Weight: <None>

Min Conc: 0

43 Ca [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	191.12	0.0001	P	7.4
2	☐	500.000	495.291	5853.40	0.0037	P	2.1
3	☐	5000.000	5715.929	64872.22	0.0416	P	1.6
4	☐	12500.000	14232.469	155867.26	0.1035	P	0.3
5	☐	25000.000	28333.310	294306.63	0.2059	P	2.4
6	☐	50000.000	56337.689	563244.16	0.4092	P	2.9
7	☐	100000.00	110392.78	1108793.56	0.8018	P	1.0
8	☐	500000.00	497070.54	4809122.53	3.6098	A	2.5

$$y = 7.2619E-006 * x + 1.2467E-004$$

$$R = 0.9997$$


$$DL = 3.811$$

$$BEC = 17.17$$

Weight: <None>

Min Conc: 0

44 Ca [No Gas] ISTD:45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	4287.35	0.0028	P	3.1
2	☐	500.000	514.871	97682.75	0.0621	P	0.6
3	☐	5000.000	5560.104	1002352.45	0.6433	P	0.6
4	☐	12500.000	12948.047	2250792.95	1.4943	A	0.5
5	☐	25000.000	25741.427	4243105.87	2.9680	A	1.7
6	☐	50000.000	51927.781	8235692.38	5.9844	A	3.3
7	☐	100000.00	102069.65	16263010.45	11.7602	A	1.2
8	☐	500000.00	499339.40	76635896.08	57.5219	A	1.7

$$y = 1.1519E-004 * x + 0.0028$$

$$R = 1.0000$$


$$DL = 2.252$$

$$BEC = 24.24$$

Weight: <None>

Min Conc: 0

47 Ti [No Gas] ISTD : 45 Sc [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	33.33	0.0000	P	35.5
2	☐	5.000	4.655	2560.25	0.0016	P	4.1
3	☐	500.000	519.149	279236.67	0.1792	P	0.9
4	☐	1250.000	1298.902	675285.78	0.4483	P	0.4
5	☐	2500.000	2540.026	1253206.12	0.8767	M	3.1
6	☐	5000.000	5075.249	2410902.41	1.7517	A	2.9
7	☐	10000.000	9945.299	4746846.67	3.4326	A	1.2
8	☐			768.92	0.0006	P	7.2

$$y = 3.4514\text{E-}004 * x + 2.1406\text{E-}005$$

$$R = 0.9999$$


$$DL = 0.066$$

$$BEC = 0.06202$$

Weight: <None>

Min Conc: 0

51 V [He] ISTD : 45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	6.67	0.0000	P	2.2
2	☐	5.000	4.970	6794.93	0.0501	P	0.8
3	☐	50.000	50.437	68666.23	0.5076	P	2.0
4	☐	125.000	127.464	166263.35	1.2826	P	0.7
5	☐	250.000	254.033	317507.51	2.5562	P	0.6
6	☐	500.000	505.259	617907.06	5.0841	P	1.0
7	☐	1000.000	996.033	1227959.39	10.0225	P	0.6
8	☐			386.68	0.0034	P	10.1

$$y = 0.0101 * x + 4.9130\text{E-}005$$

$$R = 1.0000$$


$$DL = 0.000326$$

$$BEC = 0.004883$$

Weight: <None>

Min Conc: 0

52 Cr [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	280.01	0.0021	P	21.8
2	☐	2.000	1.952	3204.82	0.0236	P	5.7
3	☐	50.000	51.263	76849.94	0.5680	P	1.4
4	☐	125.000	128.175	183697.99	1.4171	P	0.8
5	☐	250.000	255.961	351252.27	2.8279	P	1.0
6	☐	500.000	506.864	680323.83	5.5980	P	1.5
7	☐	1000.000	994.618	1345632.93	10.9829	P	0.6
8	☐			4725.24	0.0410	P	5.9

$$y = 0.0110 * x + 0.0021$$

$$R = 0.9999$$


$$DL = 0.1223$$

$$BEC = 0.1867$$

Weight: <None>

Min Conc: 0

54 Fe [He] ISTD:45 Sc [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det	RSD
1	☐	0.000	0.000	876.70	0.0065	P	7.8
2	☐	50.000	55.334	4864.16	0.0358	P	1.7
3	☐	2500.000	2733.435	197171.94	1.4574	P	1.1
4	☐	6250.000	6910.559	476315.78	3.6745	P	0.8
5	☐	12500.000	13790.073	909964.23	7.3261	P	1.1
6	☐	25000.000	26069.472	1682396.28	13.8440	A	2.4
7	☐	50000.000	51236.816	3332869.05	27.2026	A	0.9
8	☐	250000.00	249562.33	15270920.61	132.4725	A	3.5

$$y = 5.3079E-004 * x + 0.0065$$


$$R = 1.0000$$

$$DL = 2.838$$

$$BEC = 12.18$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	92.22	0.0007	P	19.0
2	☐	1.000	1.068	1333.41	0.0098	P	4.3
3	☐	500.000	518.254	600441.65	4.4379	P	0.7
4	☐	1250.000	1267.258	1406505.36	10.8508	A	1.2
5	☐	2500.000	2554.499	2716595.27	21.8721	A	1.8
6	☐	5000.000	5071.712	5277496.90	43.4242	A	1.2
7	☐	10000.000	9947.449	10435196.86	85.1698	A	0.6
8	☐			6709.33	0.0582	P	2.6

$$y = 0.0086 * x + 6.7814E-004$$


$$R = 0.9999$$

$$DL = 0.04514$$

$$BEC = 0.0792$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	5355.45	0.0394	P	1.9
2	☐	50.000	52.445	75405.44	0.5555	P	0.6
3	☐	2500.000	2575.550	3434442.83	25.3836	A	0.3
4	☐	6250.000	6430.343	8207160.50	63.3160	A	1.5
5	☐	12500.000	12818.787	15672712.54	126.1802	A	0.9
6	☐	25000.000	25578.762	30596574.54	251.7423	A	0.4
7	☐	50000.000	50007.452	60295883.55	492.1281	A	0.7
8	☐	250000.00	249919.42	283533082.4	2.459.323	A	2.9

$$y = 0.0098 * x + 0.0394$$


$$R = 1.0000$$

$$DL = 0.2244$$

$$BEC = 4.008$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	628.91	0.0046	P	11.9
2	☐	50.000	55.697	2468.00	0.0182	P	6.7
3	☐	2500.000	2707.293	89709.87	0.6631	P	1.3
4	☐	6250.000	6839.835	216236.23	1.6682	P	0.8
5	☐	12500.000	13650.232	412930.46	3.3245	P	1.2
6	☐	25000.000	27062.533	800521.87	6.5865	P	0.4
7	☐	50000.000	50829.047	1515186.59	12.3668	A	1.1
8	☐	250000.00	249553.60	6998069.55	60.6986	A	2.9

$$y = 2.4321\text{E-}004 * x + 0.0046$$


R = 1.0000

DL = 6.815

BEC = 19.05

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	26.67	0.0002	P	38.1
2	☐	1.000	1.090	2450.22	0.0181	P	2.2
3	☐	50.000	51.637	114506.63	0.8463	P	1.2
4	☐	125.000	130.403	277031.21	2.1370	P	0.4
5	☐	250.000	257.959	525051.26	4.2272	P	1.2
6	☐	500.000	509.722	1015159.64	8.3527	P	0.8
7	☐	1000.000	992.392	1992421.46	16.2620	A	0.8
8	☐			5199.83	0.0451	P	0.8

$$y = 0.0164 * x + 1.9697\text{E-}004$$


R = 0.9999

DL = 0.01373

BEC = 0.01202

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	67.78	0.0005	P	21.0
2	☐	1.000	1.203	718.91	0.0053	P	12.1
3	☐	50.000	53.589	28968.04	0.2141	P	1.3
4	☐	125.000	134.020	69308.06	0.5347	P	1.9
5	☐	250.000	261.747	129653.08	1.0439	P	1.5
6	☐	500.000	514.201	249165.71	2.0502	P	1.1
7	☐	1000.000	988.656	482908.53	3.9414	P	0.2
8	☐			1984.60	0.0172	P	10.1

$$y = 0.0040 * x + 5.0011E-004$$


R = 0.9998

DL = 0.07905

BEC = 0.1255

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	771.14	0.0057	P	7.9
2	☐	2.000	2.076	3997.24	0.0294	P	4.8
3	☐	500.000	524.272	813128.41	6.0098	P	0.5
4	☐	1250.000	1327.819	1971893.77	15.2122	A	0.8
5	☐	2500.000	2602.032	3702071.44	29.8048	A	1.2
6	☐	5000.000	5145.994	7162861.14	58.9389	A	1.5
7	☐	10000.000	9890.554	13878808.13	113.2747	A	0.7
8	☐			4570.74	0.0397	P	4.1

$$y = 0.0115 * x + 0.0057$$


R = 0.9998

DL = 0.1174

BEC = 0.4955

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	85.55	0.0006	P	17.4
2	☐	2.000	5.061	1403.42	0.0103	P	6.0
3	☐	500.000	515.876	133985.65	0.9903	P	0.6
4	☐	1250.000	1292.831	321583.18	2.4808	P	0.7
5	☐	2500.000	2576.920	614105.80	4.9442	P	1.2
6	☐	5000.000	5098.535	1188808.14	9.7817	P	1.1
7	☐	10000.000	9925.354	2332925.45	19.0415	A	1.4
8	☐			8270.14	0.0717	P	4.7

$$y = 0.0019 * x + 6.3110E-004$$


R = 0.9999

DL = 0.1719

BEC = 0.329

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1.11	0.0000	P	173.
2	☐	1.000	1.135	212.23	0.0002	P	19.3
3	☐	50.000	51.379	9573.18	0.0077	P	1.3
4	☐	125.000	128.064	23029.05	0.0193	P	0.6
5	☐	250.000	256.612	44378.35	0.0386	P	0.9
6	☐	500.000	503.633	85715.37	0.0758	P	1.0
7	☐	1000.000	996.078	169645.52	0.1500	P	1.1
8	☐			74.45	0.0001	P	17.8

$$y = 1.5056E-004 * x + 8.9667E-007$$


R = 1.0000

DL = 0.03095

BEC = 0.005955

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	2.22	0.0000	P	86.6
2	☐	5.000	5.867	35.56	0.0000	P	39.0
3	☐	50.000	52.711	302.23	0.0002	P	18.4
4	☐	125.000	124.026	683.35	0.0006	P	6.4
5	☐	250.000	270.229	1428.97	0.0012	P	2.7
6	☐	500.000	502.374	2612.47	0.0023	P	4.2
7	☐	1000.000	993.738	5169.84	0.0046	P	1.4
8	☐			2.22	0.0000	P	86.6

$$y = 4.5974\text{E-}006 * x + 1.7929\text{E-}006$$


R = 0.9998

DL = 1.013

BEC = 0.39

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	108.89	0.0001	P	23.2
2	☐	5.000	4.427	190.00	0.0002	P	4.5
3	☐	50.000	53.462	1104.50	0.0009	P	6.0
4	☐	125.000	129.418	2433.56	0.0020	P	8.1
5	☐	250.000	259.883	4599.64	0.0040	P	1.9
6	☐	500.000	505.576	8713.75	0.0077	P	3.3
7	☐	1000.000	994.019	17048.80	0.0151	P	1.2
8	☐			103.34	0.0001	P	30.0

$$y = 1.5075\text{E-}005 * x + 8.6989\text{E-}005$$


R = 0.9999

DL = 4.018


BEC = 5.77

Weight: <None>

Min Conc: 0


81 Br [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			7865.50		P	0.7
2	☐			7815.48		P	2.1
3	☐			7301.87		P	1.3
4	☐			6781.62		P	1.3
5	☐			6492.60		P	2.8
6	☐			6163.56		P	1.9
7	☐			6238.04		P	3.0
8	☐			6348.10		P	1.3

81 Br [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			34.45		P	39.1
2	☐			43.34		P	13.3
3	☐			12.22		P	31.5
4	☐			22.22		P	8.7
5	☐			21.11		P	39.7
6	☐			21.11		P	71.2
7	☐			15.56		P	75.3
8	☐			23.34		P	24.7

82 Se [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	120.62	0.0000	P	18.5
2	☐	5.000	4.879	405.31	0.0001	P	10.2
3	☐	50.000	52.066	3144.10	0.0006	P	2.6
4	☐	125.000	131.652	7490.14	0.0016	P	0.7
5	☐	250.000	260.029	14125.45	0.0031	P	1.6
6	☐	500.000	509.042	27040.65	0.0060	P	3.5
7	☐	1000.000	992.037	52582.63	0.0116	P	1.2
8	☐			141.89	0.0000	P	9.6

$y = 1.1712E-005 * x + 2.5107E-005$

$R = 0.9999$


$DL = 1.187$

$BEC = 2.144$

Weight: <None>


Min Conc: 0

83 Kr [No Gas] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			63.34		P	18.2
2	☐			70.00		P	21.8
3	☐			67.78		P	28.4
4	☐			78.89		P	28.1
5	☐			67.78		P	27.1
6	☐			91.11		P	2.1
7	☐			66.67		P	18.0
8	☐			95.56		P	10.1

86 Sr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	136.67	0.0000	P	11.2
2	☐	1.000	1.007	1486.76	0.0003	P	7.1
3	☐	1250.000	1270.922	1707825.62	0.3449	A	1.1
4	☐	3125.000	3214.612	4169045.94	0.8722	A	1.0
5	☐	6250.000	6341.609	7914713.28	1.7206	A	2.3
6	☐	12500.000	12578.819	15418109.07	3.4129	A	2.0
7	☐	25000.000	24925.441	30539624.27	6.7627	A	2.0
8	☐			5199.85	0.0012	P	1.0

$$y = 2.7132E-004 * x + 2.8658E-005$$

R = 1.0000


DL = 0.03556

BEC = 0.1056

Weight: <None>

Min Conc: 0

88 Sr [No Gas] ISTD: 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	167.78	0.0000	P	9.3
2	☐	1.000	0.981	11736.98	0.0024	P	1.8
3	☐	1250.000	1254.428	14859798.11	3.0005	A	0.5
4	☐	3125.000	3175.559	36305130.02	7.5955	A	0.2
5	☐	6250.000	6343.291	69797016.18	15.1723	A	1.7
6	☐	12500.000	12645.892	136638035.7	30.2473	A	2.4
7	☐	25000.000	24897.190	268932924.8	59.5507	A	0.8
8	☐			44341.12	0.0105	P	4.3

$$y = 0.0024 * x + 3.4991E-005$$

R = 1.0000

DL = 0.004098


BEC = 0.01463

Weight: <None>

Min Conc: 0

Calibration for 010CAL5.d

90 Zr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	176.67	0.0000	P	19.7
2	☐	1.000	0.967	6902.78	0.0014	P	2.7
3	☐	50.000	50.210	350803.25	0.0708	P	1.4
4	☐	125.000	127.590	860100.55	0.1799	P	0.5
5	☐	250.000	253.053	1641544.21	0.3569	A	2.2
6	☐	500.000	506.653	3227592.42	0.7145	A	2.1
7	☐	1000.000	995.576	6339859.56	1.4039	A	1.3
8	☐			3791.64	0.0009	P	1.1

$$y = 0.0014 * x + 3.6769E-005$$

$$R = 1.0000$$


$$DL = 0.01541$$

$$BEC = 0.02608$$

Weight: <None>

Min Conc: 0

91 Zr [No Gas] ISTD : 89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	31.11	0.0000	P	13.3
2	☐	1.000	0.998	1581.21	0.0003	P	5.2
3	☐	50.000	50.074	78187.73	0.0158	P	0.5
4	☐	125.000	127.605	192248.75	0.0402	P	0.5
5	☐	250.000	254.640	369192.55	0.0803	P	1.9
6	☐	500.000	506.741	721411.99	0.1597	P	2.9
7	☐	1000.000	995.140	1416342.53	0.3136	M	1.3
8	☐			805.59	0.0002	P	0.5

$$y = 3.1515E-004 * x + 6.4867E-006$$

$$R = 1.0000$$


$$DL = 0.008231$$

$$BEC = 0.02058$$

Weight: <None>

Min Conc: 0

94 Mo [No Gas] ISTD:89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	86.66	0.0000	P	5.6
2	☐	5.000	5.253	8202.36	0.0017	P	3.2
3	☐	500.000	501.638	778872.46	0.1573	P	0.9
4	☐	1250.000	1273.245	1907842.09	0.3992	A	1.0
5	☐	2500.000	2522.164	3637324.53	0.7907	A	1.5
6	☐	5000.000	5015.439	7101864.96	1.5723	A	3.2
7	☐	10000.000	9983.752	14133879.37	3.1297	A	1.0
8	☐			3707.17	0.0009	P	1.1

$$y = 3.1348E-004 * x + 1.8117E-005$$

$$R = 1.0000$$


$$DL = 0.009767$$

$$BEC = 0.05779$$

Weight: <None>

Min Conc: 0

95 Mo [No Gas] ISTD:89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	51.11	0.0000	P	27.5
2	☐	5.000	4.386	9989.05	0.0020	P	0.6
3	☐	500.000	497.570	1132965.71	0.2288	P	1.2
4	☐	1250.000	1263.868	2777460.34	0.5811	A	0.5
5	☐	2500.000	2550.911	5394877.70	1.1728	A	2.4
6	☐	5000.000	5052.967	10494544.29	2.3232	A	2.4
7	☐	10000.000	9959.177	20678088.03	4.5789	A	1.1
8	☐			3570.47	0.0008	P	2.2

$$y = 4.5976E-004 * x + 1.0766E-005$$

$$R = 1.0000$$


$$DL = 0.0193$$

$$BEC = 0.02342$$

Weight: <None>

Min Conc: 0

96 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	51.11	0.0000	P	3.1
2	☐	5.000	4.486	11191.04	0.0023	P	2.0
3	☐	500.000	498.026	1242857.12	0.2510	P	1.0
4	☐	1250.000	1272.638	3065225.44	0.6413	A	0.2
5	☐	2500.000	2517.082	5834991.09	1.2683	A	1.2
6	☐	5000.000	5071.147	11543593.30	2.5553	A	2.3
7	☐	10000.000	9957.425	22658499.94	5.0175	A	1.4
8	☐			7145.15	0.0017	P	4.7

$$y = 5.0389\text{E-}004 * x + 1.0671\text{E-}005$$

R = 1.0000


DL = 0.001954

BEC = 0.02118

Weight: <None>

Min Conc: 0

97 Mo [No Gas] ISTD:89 Y [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	26.66	0.0000	P	40.4
2	☐	5.000	4.440	6357.00	0.0013	P	2.4
3	☐	500.000	500.354	716966.26	0.1448	P	1.1
4	☐	1250.000	1266.350	1751274.77	0.3664	A	0.6
5	☐	2500.000	2520.641	3354973.77	0.7293	A	1.6
6	☐	5000.000	5056.740	6609473.30	1.4631	A	2.1
7	☐	10000.000	9964.408	13019250.92	2.8830	A	1.7
8	☐			2245.75	0.0005	P	5.8

$$y = 2.8933\text{E-}004 * x + 5.5048\text{E-}006$$

R = 1.0000

DL = 0.02303


BEC = 0.01903

Weight: <None>

Min Conc: 0

Calibration for 010CAL5.d

98 Mo [No Gas] ISTD:89 Y [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	54.44	0.0000	P	20.3
2	☐	5.000	4.406	16273.62	0.0033	P	1.4
3	☐	500.000	500.534	1851696.61	0.3739	A	0.8
4	☐	1250.000	1270.914	4537630.00	0.9493	A	0.3
5	☐	2500.000	2537.151	8717513.76	1.8951	A	2.4
6	☐	5000.000	5044.712	17023750.44	3.7682	A	1.9
7	☐	10000.000	9965.715	33615790.33	7.4439	A	1.9
8	☐			5421.05	0.0013	P	2.2

$$y = 7.4695E-004 * x + 1.1426E-005$$

$$R = 1.0000$$


$$DL = 0.009309$$

$$BEC = 0.0153$$

Weight: <None>

Min Conc: 0

106 Cd [No Gas] ISTD:159 Tb [No Gas]



	R _j ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	1422.31	0.0004	P	3.8
2	☐	1.000	1.501	1605.66	0.0004	P	9.2
3	☐	50.000	52.595	6919.49	0.0017	P	3.0
4	☐	125.000	134.421	15117.94	0.0038	P	2.4
5	☐	250.000	262.795	28062.70	0.0071	P	0.8
6	☐	500.000	514.428	53101.01	0.0135	P	2.5
7	☐	1000.000	988.279	102309.15	0.0255	P	2.3
8	☐			1268.96	0.0003	P	7.3

$$y = 2.5465E-005 * x + 3.5789E-004$$


$$R = 0.9997$$

$$DL = 1.585$$

$$BEC = 14.05$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	30.00	0.0000	P	56.6
2	☐	1.000	1.026	4946.45	0.0012	P	2.2
3	☐	50.000	54.447	262481.41	0.0644	P	0.9
4	☐	125.000	134.833	637467.79	0.1594	P	0.7
5	☐	250.000	260.065	1223791.42	0.3075	P	1.3
6	☐	500.000	517.354	2413527.20	0.6116	A	2.2
7	☐	1000.000	987.355	4678451.39	1.1673	A	3.0
8	☐			595.57	0.0002	P	9.5

$$y = 0.0012 * x + 7.7103E-006$$


$$R = 0.9997$$

$$DL = 0.01107$$

$$BEC = 0.006522$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	6.67	0.0000	P	46.6
2	☐	1.000	1.026	81.11	0.0000	P	6.2
3	☐	50.000	56.140	4100.62	0.0010	P	2.5
4	☐	125.000	135.823	9721.12	0.0024	P	1.2
5	☐	250.000	267.679	19061.47	0.0048	P	1.8
6	☐	500.000	520.391	36732.03	0.0093	P	1.7
7	☐	1000.000	983.725	70527.92	0.0176	P	2.0
8	☐			35.56	0.0000	P	26.1

$$y = 1.7885E-005 * x + 1.6546E-006$$


$$R = 0.9995$$

$$DL = 0.1294$$

$$BEC = 0.09252$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	12.22	0.0000	P	53.9
2	☐	1.000	0.993	4504.08	0.0011	P	1.1
3	☐	50.000	54.761	249233.05	0.0611	P	1.5
4	☐	125.000	135.498	604824.17	0.1513	P	0.8
5	☐	250.000	259.524	1153078.23	0.2897	P	0.9
6	☐	500.000	525.962	2316654.46	0.5871	A	2.0
7	☐	1000.000	983.088	4398498.65	1.0974	A	2.3
8	☐			540.01	0.0001	P	8.2

$$y = 0.0011 * x + 3.0644E-006$$


$$R = 0.9994$$

$$DL = 0.004443$$

$$BEC = 0.002745$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	992.06	0.0002	P	3.6
2	☐	1.000	1.196	2171.68	0.0005	P	4.4
3	☐	50.000	54.737	54403.32	0.0133	P	1.2
4	☐	125.000	137.005	132050.76	0.0330	P	1.1
5	☐	250.000	263.653	252020.26	0.0633	P	1.5
6	☐	500.000	518.098	490021.92	0.1242	P	2.0
7	☐	1000.000	985.800	946233.49	0.2361	P	1.8
8	☐			1082.28	0.0003	P	8.7

$$y = 2.3921E-004 * x + 2.4966E-004$$


$$R = 0.9996$$

$$DL = 0.1123$$

$$BEC = 1.044$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	746.69	0.0002	P	10.1
2	☐	5.000	4.832	17722.10	0.0044	P	1.0
3	☐	50.000	51.580	182858.48	0.0449	P	2.1
4	☐	125.000	128.861	446924.57	0.1118	P	1.1
5	☐	250.000	250.209	863064.71	0.2168	P	1.4
6	☐	500.000	516.267	1764514.28	0.4472	A	2.8
7	☐	1000.000	991.253	3441048.25	0.8585	A	2.6
8	☐			2139.07	0.0006	P	7.7

$$y = 8.6588E-004 * x + 1.8875E-004$$


R = 0.9998

DL = 0.06589

BEC = 0.218

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	11.11	0.0000	P	65.5
2	☐	2.000	2.013	8339.15	0.0021	P	1.5
3	☐	50.000	50.874	211751.09	0.0519	P	1.6
4	☐	125.000	128.013	522564.30	0.1307	P	1.0
5	☐	250.000	250.096	1016179.64	0.2553	P	1.4
6	☐	500.000	516.638	2080961.43	0.5274	A	2.6
7	☐	1000.000	991.236	4056211.33	1.0119	A	1.8
8	☐			2916.99	0.0008	P	3.8

$$y = 0.0010 * x + 2.8512E-006$$


R = 0.9998

DL = 0.005491

BEC = 0.002793

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	27.78	0.0000	P	47.5
2	☐	10.000	9.670	11222.28	0.0028	P	3.1
3	☐	250.000	249.688	290806.00	0.0713	P	1.1
4	☐	625.000	624.816	713642.83	0.1785	P	1.3
5	☐	1250.000	1222.649	1390013.00	0.3492	P	1.2
6	☐	2500.000	2599.541	2929845.24	0.7425	A	2.0
7	☐	5000.000	4957.107	5674869.36	1.4159	A	2.5
8	☐			2854.76	0.0008	P	5.0

$$y = 2.8563E-004 * x + 7.1093E-006$$


$$R = 0.9997$$

$$DL = 0.03547$$

$$BEC = 0.02489$$

Weight: <None>

Min Conc: 0



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	44.44	0.0000	P	25.8
2	☐	10.000	9.451	18913.75	0.0047	P	1.4
3	☐	250.000	249.558	501226.32	0.1229	P	1.3
4	☐	625.000	627.983	1236936.53	0.3093	P	1.0
5	☐	1250.000	1305.680	2559871.16	0.6431	A	1.1
6	☐	2500.000	2573.464	5001546.00	1.2676	A	2.9
7	☐	5000.000	4948.999	9771489.78	2.4377	A	1.5
8	☐			5029.83	0.0013	P	2.5

$$y = 4.9256E-004 * x + 1.1212E-005$$

$$R = 0.9998$$


$$DL = 0.01765$$

$$BEC = 0.02276$$

Weight: <None>


Min Conc: 0

150 Nd [No Gas] No available calibration points



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			3.33		P	173.
2	☐			2.22		P	86.6
3	☐			24.44		P	41.7
4	☐			54.44		P	27.6
5	☐			110.00		P	29.8
6	☐			258.90		P	7.4
7	☐			445.57		P	2.3
8	☐			103.34		P	22.6

150 Nd [He] No available calibration points




	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			0.00		P	
2	☐			1.11		P	173.
3	☐			3.33		P	173.
4	☐			8.89		P	21.6
5	☐			24.45		P	43.8
6	☐			46.67		P	51.5
7	☐			86.67		P	44.4
8	☐			62.22		P	35.7

156 Dy [No Gas] No available calibration points


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			7.78		P	99.0
2	☐			12.22		P	56.8
3	☐			18.89		P	53.9
4	☐			23.33		P	28.6
5	☐			38.89		P	35.7
6	☐			77.78		P	13.1
7	☐			124.44		P	15.5
8	☐			165.56		P	17.4

156 Dy [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			1.11		P	173.
2	☐			5.55		P	69.3
3	☐			18.89		P	27.0
4	☐			28.89		P	17.6
5	☐			62.22		P	8.2
6	☐			121.11		P	26.0
7	☐			276.67		P	7.5
8	☐			114.45		P	37.5

160 Gd [No Gas] No available calibration poin...

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			11.11		P	17.3
2	☐			20.00		P	101.
3	☐			11.11		P	17.3
4	☐			22.22		P	43.3
5	☐			22.22		P	34.7
6	☐			35.55		P	30.1
7	☐			61.11		P	17.5
8	☐			133.33		P	34.7

160 Gd [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			105.56		P	20.5
2	☐			110.00		P	31.6
3	☐			105.56		P	14.9
4	☐			114.45		P	13.8
5	☐			121.11		P	15.7
6	☐			130.00		P	11.2
7	☐			121.12		P	11.1
8	☐			146.67		P	17.2


164 Er [No Gas] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			16.67		P	20.0
2	☐			12.22		P	41.7
3	☐			17.78		P	10.8
4	☐			18.89		P	20.4
5	☐			28.89		P	13.3
6	☐			35.55		P	37.9
7	☐			56.66		P	40.8
8	☐			183.34		P	25.3

164 Er [He] No available calibration points

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐			8.89		P	57.3
2	☐			4.44		P	43.4
3	☐			10.00		P	66.7
4	☐			8.89		P	43.3
5	☐			14.44		P	81.1
6	☐			24.44		P	31.5
7	☐			38.89		P	17.8
8	☐			121.11		P	34.9

203 Tl [No Gas] ISTD : 209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	32.22	0.0000	P	10.5
2	☐	1.000	0.839	3660.54	0.0015	P	4.2
3	☐	50.000	44.832	195514.08	0.0815	P	1.1
4	☐	125.000	112.771	486539.45	0.2049	P	1.2
5	☐	250.000	223.454	963045.34	0.4061	P	1.5
6	☐	500.000	506.215	2162027.21	0.9199	A	2.8
7	☐	1000.000	1005.316	4293133.44	1.8269	A	2.1
8	☐			444.46	0.0002	P	8.1

$$y = 0.0018 * x + 1.3886E-005$$

$$R = 0.9996$$


$$DL = 0.00241$$

$$BEC = 0.007641$$

Weight: <None>

Min Conc: 0

205 Tl [No Gas] ISTD : 209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	0.000	0.000	65.56	0.0000	P	49.7
2	☐	1.000	0.855	8930.79	0.0038	P	2.9
3	☐	50.000	44.083	460877.89	0.1921	P	1.2
4	☐	125.000	111.882	1157242.15	0.4875	P	1.4
5	☐	250.000	249.699	2580072.46	1.0879	A	1.4
6	☐	500.000	506.644	5188388.01	2.2074	A	1.5
7	☐	1000.000	998.689	10224531.10	4.3511	A	2.2
8	☐			1092.28	0.0005	P	13.8

$$y = 0.0044 * x + 2.8765E-005$$

$$R = 0.9999$$


$$DL = 0.00984$$

$$BEC = 0.006602$$

Weight: <None>

Min Conc: 0

206 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	395.57	0.0002	P	19.6
2	☐	1.000	0.797	3207.07	0.0013	P	1.1
3	☐	250.000	223.691	793352.21	0.3307	P	1.4
4	☐	625.000	619.680	2173769.67	0.9157	A	1.7
5	☐	1250.000	1236.028	4330978.93	1.8263	A	2.1
6	☐	2500.000	2531.391	8790153.83	3.7401	A	2.5
7	☐	5000.000	4989.778	17324041.41	7.3722	A	1.9
8	☐			3436.02	0.0017	P	5.1

$$y = 0.0015 * x + 1.7148E-004$$

$$R = 1.0000$$


$$DL = 0.06831$$

$$BEC = 0.1161$$

Weight: <None>

Min Conc: 0

207 [Pb] [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	356.68	0.0002	P	6.8
2	☐	1.000	0.788	2794.76	0.0012	P	3.9
3	☐	250.000	220.604	686458.32	0.2861	P	1.7
4	☐	625.000	627.458	1931383.57	0.8135	A	0.5
5	☐	1250.000	1263.469	3884207.13	1.6379	A	2.1
6	☐	2500.000	2526.587	7696978.22	3.2753	A	3.2
7	☐	5000.000	4984.502	15184062.41	6.4613	A	1.4
8	☐			3062.60	0.0015	P	8.0

$$y = 0.0013 * x + 1.5406E-004$$

$$R = 1.0000$$

$$DL = 0.02433$$


$$BEC = 0.1189$$

Weight: <None>

Min Conc: 0

Calibration for 010CAL5.d

208 Pb [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	1680.06	0.0007	P	7.0
2	☐	1.000	0.787	12833.60	0.0054	P	0.3
3	☐	250.000	235.168	3348564.49	1.3957	A	1.8
4	☐	625.000	622.010	8761213.75	3.6904	A	0.9
5	☐	1250.000	1248.108	17559309.11	7.4043	A	1.6
6	☐	2500.000	2528.048	35243427.46	14.9967	A	3.0
7	☐	5000.000	4987.565	69527116.05	29.5862	A	1.3
8	☐			13931.83	0.0068	P	4.6

$$y = 0.0059 * x + 7.2560E-004$$

$$R = 1.0000$$


$$DL = 0.02582$$

$$BEC = 0.1223$$

Weight: <None>

Min Conc: 0

238 U [No Gas] ISTD :209 Bi [No Gas]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	0.000	0.000	10.00	0.0000	P	28.9
2	☐	1.000	0.792	11356.03	0.0048	P	1.4
3	☐	50.000	41.829	604736.29	0.2520	P	0.9
4	☐	125.000	107.456	1537083.92	0.6475	P	1.2
5	☐	250.000	249.359	3562977.45	1.5025	A	2.4
6	☐	500.000	501.910	7107465.38	3.0242	A	2.8
7	☐	1000.000	1001.807	14185154.65	6.0363	A	1.4
8	☐			692.25	0.0003	P	17.0

$$y = 0.0060 * x + 4.2679E-006$$


$$R = 0.9998$$

$$DL = 0.000615$$


$$BEC = 0.0007083$$

Weight: <None>


Min Conc: 0

6 Li (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	1.000		689626.08		P	4.4
2	☐	1.000		709518.52		P	0.9
3	☐	1.000		717028.03		P	0.2
4	☐	1.000		699973.28		P	0.6
5	☐	1.000		685831.78		P	0.8
6	☐	1.000		671393.70		P	0.9
7	☐	1.000		669582.18		P	0.9
8	☐	1.000		652468.93		P	1.0

45 Sc (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det.	RSD
1	☐	1.000		1537426.78		A	6.1
2	☐	1.000		1573033.83		A	0.8
3	☐	1.000		1558245.97		A	0.4
4	☐	1.000		1506247.63		A	0.5
5	☐	1.000		1429936.16		A	1.8
6	☐	1.000		1377006.99		A	2.7
7	☐	1.000		1382939.02		A	0.8
8	☐	1.000		1332378.07		A	0.9

45 Sc (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		135807.75		P	2.2
2	☐	1.000		135738.99		P	0.4
3	☐	1.000		135304.21		P	1.0
4	☐	1.000		129632.94		P	1.0
5	☐	1.000		124212.87		P	0.6
6	☐	1.000		121542.41		P	1.0
7	☐	1.000		122522.31		P	0.4
8	☐	1.000		115352.34		P	2.8

89 Y (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		4785223.85		A	4.5
2	☐	1.000		4927398.99		A	0.2
3	☐	1.000		4952656.14		A	0.9
4	☐	1.000		4779822.46		A	0.6
5	☐	1.000		4601054.66		A	1.6
6	☐	1.000		4519012.09		A	2.3
7	☐	1.000		4516164.72		A	0.6
8	☐	1.000		4209149.17		A	0.5

89 Y (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1251877.37		P	1.9
2	☐	1.000		1235919.86		P	0.1
3	☐	1.000		1237304.16		P	0.8
4	☐	1.000		1194282.22		P	0.8
5	☐	1.000		1148590.78		P	0.2
6	☐	1.000		1130392.14		P	0.5
7	☐	1.000		1131180.17		P	0.2
8	☐	1.000		1031929.43		P	3.8

103 Rh (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2640539.98		A	4.5
2	☐	1.000		2717977.84		A	1.4
3	☐	1.000		2691414.72		A	0.7
4	☐	1.000		2562043.66		A	1.5
5	☐	1.000		2486477.36		A	1.6
6	☐	1.000		2392686.95		A	2.5
7	☐	1.000		2337037.22		A	0.3
8	☐	1.000		2064237.45		A	1.5

103 Rh (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1489929.91		P	1.9
2	☐	1.000		1491820.19		M	0.3
3	☐	1.000		1472222.95		P	1.0
4	☐	1.000		1409642.06		P	0.6
5	☐	1.000		1352701.16		P	0.5
6	☐	1.000		1315784.01		P	1.1
7	☐	1.000		1298525.89		P	0.2
8	☐	1.000		1106398.84		P	3.6

115 In (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2921663.09		A	4.7
2	☐	1.000		3016675.06		A	1.4
3	☐	1.000		2997082.59		A	1.2
4	☐	1.000		2900838.68		A	1.1
5	☐	1.000		2794426.48		A	2.3
6	☐	1.000		2689771.90		A	2.5
7	☐	1.000		2661996.12		A	1.3
8	☐	1.000		2506454.08		A	1.3

115 In (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		974296.63		P	2.1
2	☐	1.000		967494.76		P	0.3
3	☐	1.000		962205.47		P	0.5
4	☐	1.000		928696.20		P	0.8
5	☐	1.000		894866.26		P	0.7
6	☐	1.000		865274.37		P	0.6
7	☐	1.000		850363.61		P	0.2
8	☐	1.000		780312.28		P	3.3

159 Tb (ISTD) [No Gas]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3970396.57		A	5.4
2	☐	1.000		4053059.24		A	0.3
3	☐	1.000		4077400.46		A	0.7
4	☐	1.000		3998962.68		A	0.9
5	☐	1.000		3980498.58		A	0.8
6	☐	1.000		3946700.05		A	1.6
7	☐	1.000		4009227.44		A	1.9
8	☐	1.000		3742550.22		A	2.6

159 Tb (ISTD) [He]


	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2642343.26		A	2.8
2	☐	1.000		2647131.37		A	0.5
3	☐	1.000		2644014.37		A	0.5
4	☐	1.000		2607281.83		A	1.4
5	☐	1.000		2545439.09		A	0.7
6	☐	1.000		2575772.52		A	1.9
7	☐	1.000		2581242.84		A	1.1
8	☐	1.000		2336873.49		A	4.2

165 Ho (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		3849416.96		A	6.2
2	☐	1.000		3947299.07		A	0.6
3	☐	1.000		4005517.48		A	0.5
4	☐	1.000		3938021.75		A	0.8
5	☐	1.000		3939629.42		A	1.0
6	☐	1.000		3910456.71		A	0.6
7	☐	1.000		3929824.73		A	0.2
8	☐	1.000		3681198.69		A	0.5


165 Ho (ISTD) [He]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2699428.80		A	3.3
2	☐	1.000		2687380.93		A	0.8
3	☐	1.000		2665877.91		A	0.9
4	☐	1.000		2647221.60		A	0.6
5	☐	1.000		2618055.59		A	1.5
6	☐	1.000		2605263.93		A	1.9
7	☐	1.000		2643841.14		A	0.7
8	☐	1.000		2405889.18		A	3.7

209 Bi (ISTD) [No Gas]

	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		2320958.84		A	5.6
2	☐	1.000		2378389.24		A	0.2
3	☐	1.000		2399497.26		A	1.1
4	☐	1.000		2374122.97		A	1.0
5	☐	1.000		2371791.63		A	1.2
6	☐	1.000		2350881.03		A	1.7
7	☐	1.000		2350089.93		A	0.7
8	☐	1.000		2047546.97		A	0.2

209 Bi (ISTD) [He]



	Rj ct	Conc.	Calc Conc.	CPS	Ratio	Det .	RSD
1	☐	1.000		1909624.42		A	2.7
2	☐	1.000		1900812.02		A	1.5
3	☐	1.000		1884802.18		A	1.8
4	☐	1.000		1866549.63		A	0.7
5	☐	1.000		1772714.02		M	5.1
6	☐	1.000		1761909.54		M	6.2
7	☐	1.000		1755972.47		M	6.1
8	☐	1.000		1385872.88		P	3.3

US EPA Tune Check Report

Reviewed By: Mohan
On: 10/19/2024 2:12:32 AM
Inst Id : P7
Client Sample Id : TUNE025

Operator Name Jaswal
Acq/Data Batch D:\Agilent\ICPMH\1\DATA\IP7101824-MS.b
Acq. Date-Time 2024-10-18 10:30:27
Report Comment ---
Instrument Name G8403A JP14410463

[No Gas]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
9		1794	17939.60			0.829	5.000
24		61289	612885.04			0.962	5.000
25		7659	76587.01			1.279	5.000
26		8583	85834.47			1.306	5.000
59		15373	153726.72			0.593	5.000
113		1386	13858.94			0.699	5.000
115		30397	303971.49			1.095	5.000
206		6656	66562.89			0.800	5.000
207		5581	55811.48			0.641	5.000
208		13563	135634.19			0.884	5.000
220		0	2.40			75.691	

Mass	RSD% (Flag)
9	
24	
25	
26	
59	
113	
115	
206	
207	
208	
220	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
9	1778	1805	1797	1779	1811
24	60267	61676	61525	61310	61664
25	7505	7707	7636	7680	7765
26	8433	8627	8563	8554	8740
59	15273	15510	15379	15309	15391
113	1383	1399	1390	1373	1384
115	30052	30808	30647	30093	30385
206	6563	6681	6675	6694	6669
207	5519	5609	5588	5593	5598
208	13356	13601	13601	13590	13668


US EPA Tune Check Report

Reviewed By:Mohan
On:10/19/2024 2:12:32
AM
Inst Id :P7
ClientSampleId
:TUNE025

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
220	0	0	0	0	0

Integration Time [sec] 0.1

Resolution/Axis



Mass	Peak Height	Axis	Axis (Required)	Axis (Flag)
9	3289.38	9.00	8.90 - 9.10	
24	108988.78	24.00	23.90 - 24.10	
25	13959.05	25.00	24.90 - 25.10	
26	15433.47	26.00	25.90 - 26.10	
59	26680.75	58.95	58.90 - 59.10	
113	2577.55	113.00	112.90 - 113.10	
115	57192.05	115.00	114.90 - 115.10	
206	13048.20	206.00	205.90 - 206.10	
207	10614.02	207.00	206.90 - 207.10	
208	26333.04	208.00	207.90 - 208.10	
220	0.55	219.90	-	

Mass	W-50%	W-5%	W-5% (Required)	W-5% (Flag)
9	0.58	0.731	0.900	
24	0.59	0.740	0.900	
25	0.58	0.736	0.900	
26	0.60	0.738	0.900	
59	0.60	0.754	0.900	
113	0.56	0.738	0.900	
115	0.54	0.743	0.900	
206	0.53	0.741	0.900	
207	0.54	0.761	0.900	
208	0.52	0.768	0.900	
220	0.31	0.535		

Integration Time [sec] 0.1

Acquisition Time [sec] 256.770000000002

Y Axis Linear

Tune Parameters

Plasma Parameters

Plasma Mode --- Nebulizer Gas 0.82 L/min Dilution Gas 0.40 L/min

US EPA Tune Check Report

Reviewed By: Mohan
On: 10/19/2024 2:12:32 AM
Inst Id : P7
Client Sample Id
TUNE025

RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.9 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	7.8 V	Deflect	17.0 V
Extract 2	-105.0 V	Cell Entrance	-30 V	Plate Bias	-35 V
Omega Bias	-65 V	Cell Exit	-50 V		

Cell Parameters

Use Gas	No	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	0.0 mL/min	OctP Bias	-8.0 V		
H2 Flow	---	OctP RF	160 V		

QP Parameters

Mass Gain	135	Axis Gain	0.9976	QP Bias	-3.0 V
Mass Offset	131	Axis Offset	0.11		

Hardware Settings

Torch

Torch H	-0.5 mm	Torch V	0.6 mm
---------	---------	---------	--------

EM

Discriminator	4.8 mV	Analog HV	2399 V	Pulse HV	1788 V
---------------	--------	-----------	--------	----------	--------

[He]

Sensitivity

Mass	Conc. [ug/l]	Count	CPS	Resp (Required) [cps/ug/l]	Resp (Flag)	RSD%	RSD% (Required)
59		5090	50903.01			0.935	
89		50854	508536.90			0.732	
205		16333	163330.88			0.537	

Mass	RSD% (Flag)
59	
89	
205	

Mass	Rep#1 Count	Rep#2 Count	Rep#3 Count	Rep#4 Count	Rep#5 Count
59	5174	5066	5081	5076	5055
89	51466	50919	50603	50741	50539
205	16447	16348	16218	16281	16372

Integration Time [sec] 0.1

Tune Parameters

Plasma Parameters

Plasma Mode	---	Nebulizer Gas	0.82 L/min	Dilution Gas	0.40 L/min
RF Power	1600 W	Option Gas	---	Auxiliary Gas	0.90 L/min
RF Matching	1.80 V	Nebulizer Pump	0.10 rps	Plasma Gas	15.0 L/min
Sample Depth	9.0 mm	S/C Temp	2 °C		

Lens Parameters

Extract 1	0.0 V	Omega Lens	8.3 V	Deflect	3.2 V
-----------	-------	------------	-------	---------	-------

US EPA Tune Check Report

Reviewed By:Mohan
On:10/19/2024 2:12:32
AM
Inst Id :P7
ClientSampleId
TUNE025

Extract 2	-90.0 V	Cell Entrance	-40 V	Plate Bias	-60 V
Omega Bias	-55 V	Cell Exit	-60 V		
Cell Parameters					
Use Gas	Yes	3rd Gas Flow	---	Energy Discrimination	5.0 V
He Flow	3.9 mL/min	OctP Bias	-18.0 V		
H2 Flow	---	OctP RF	190 V		
QP Parameters					
Mass Gain	135	Axis Gain	0.9976	QP Bias	-13.0 V
Mass Offset	131	Axis Offset	0.11		
Hardware Settings					
Torch					
Torch H	-0.5 mm	Torch V	0.6 mm		
EM					
Discriminator	4.8 mV	Analog HV	2399 V	Pulse HV	1788 V

LB Number :LB133015Operator :Jaswal

Lab Sample ID :S00Instrumnet Name :P7

Client Sample ID :S0Dilution Factor :1

Date & Time Acquired :2024-10-18 11:46:13DataFile Name :004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	0.14	-0.04	-0.09	0.00	N/A	ppb
Antimony	121-1	0.00	0.00	0.00	0.00	N/A	ppb
Arsenic	75-2	-0.01	0.01	-0.01	0.00	N/A	ppb
Barium	135-1	0.00	-0.01	0.01	0.00	N/A	ppb
Barium	137-1	-0.01	0.01	0.00	0.00	N/A	ppb
Beryllium	9-1	0.02	-0.03	0.01	0.00	N/A	ppb
Bismuth	209-1				100		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.04	0.00	-0.04	0.00	N/A	ppb
Cadmium	106-1	-0.15	0.59	-0.44	0.00	N/A	ppb
Cadmium	111-1	-0.01	0.04	-0.03	0.00	N/A	ppb
Calcium	43-1	-0.92	-0.53	1.45	0.00	N/A	ppb
Calcium	44-1	-0.10	-0.69	0.80	0.00	N/A	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.03	-0.05	0.02	0.00	N/A	ppb
Cobalt	59-2	0.00	0.00	0.00	0.00	N/A	ppb
Copper	63-2	0.04	-0.04	0.00	0.00	N/A	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				100		%
Indium	115-1				100		%
Indium	115-2				100		%
Iron	56-2	-0.09	0.04	0.05	0.00	N/A	ppb
Iron	57-2	0.30	-2.41	2.11	0.00	N/A	ppb
Iron	54-2	-1.00	0.12	0.88	0.00	N/A	ppb
Krypton	83-1						cps

LB Number :	LB133015	Operator :	Jaswal
Lab Sample ID :	S00	Instrumnet Name :	P7
Client Sample ID :	S0	Dilution Factor :	1
Date & Time Acquired :	2024-10-18 11:46:13	DataFile Name :	004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.02	0.00	0.02	0.00	N/A	ppb
Lead	207-1	0.00	-0.01	0.01	0.00	N/A	ppb
Lead	208-1	-0.01	0.00	0.01	0.00	N/A	ppb
Lithium	6-1				100		%
Magnesium	24-2	-0.69	0.29	0.40	0.00	N/A	ppb
Manganese	55-2	0.00	0.02	-0.01	0.00	N/A	ppb
Molybdenum	94-1	0.00	0.00	0.00	0.00	N/A	ppb
Molybdenum	95-1	0.00	-0.01	0.01	0.00	N/A	ppb
Molybdenum	96-1	0.00	0.00	0.00	0.00	N/A	ppb
Molybdenum	97-1	0.00	0.00	-0.01	0.00	N/A	ppb
Molybdenum	98-1	0.00	0.00	0.00	0.00	N/A	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	-0.03	0.02	0.01	0.00	N/A	ppb
Phosphorus	31-2	-16.17	-14.98	-17.58	-16.25		ppb
Potassium	39-2	-6.99	3.88	3.11	0.00	N/A	ppb
Rhodium	103-1				100		%
Rhodium	103-2				100		%
Scandium	45-1				100		%
Scandium	45-2				100		%
Selenium	82-1	0.44	-0.10	-0.33	0.00	N/A	ppb
Selenium	77-2	-0.39	0.20	0.19	0.00	N/A	ppb
Selenium	78-2	-0.07	-1.31	1.37	0.00	N/A	ppb
Silicon	28-1	-3.60	-3.84	7.44	0.00	N/A	ppb
Silver	107-1	0.00	0.00	0.00	0.00	N/A	ppb
Silver	109-1	0.00	0.00	0.00	0.00	N/A	ppb
Sodium	23-2	-0.95	0.30	0.65	0.00	N/A	ppb
Strontium	86-1	-0.01	-0.01	0.01	0.00	N/A	ppb
Strontium	88-1	0.00	0.00	0.00	0.00	N/A	ppb
Sulfur	34-1	-115.00	-193.06	622.11	104.68	429.67	ppb
Terbium	159-1				100		%
Terbium	159-2				100		%
Thallium	203-1	0.00	0.00	0.00	0.00	N/A	ppb
Thallium	205-1	0.00	0.00	0.00	0.00	N/A	ppb
Tin	118-1	-0.01	-0.02	0.03	0.00	N/A	ppb
Titanium	47-1	0.00	0.02	-0.02	0.00	N/A	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S00

Instrumnet Name :

P7

Client Sample ID :

S0

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:46:13

DataFile Name :

004CALB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	N/A	ppb
Vanadium	51-2	0.00	0.00	0.00	0.00	N/A	ppb
Yttrium	89-1				100		%
Yttrium	89-2				100		%
Zinc	66-2	-0.02	-0.05	0.06	0.00	N/A	ppb
Zirconium	90-1	0.01	0.00	0.00	0.00	N/A	ppb
Zirconium	91-1	0.00	0.00	0.00	0.00	N/A	ppb

LB Number :LB133015Operator :Jaswal

Lab Sample ID :S02Instrumnet Name :P7

Client Sample ID :S02Dilution Factor :1

Date & Time Acquired :2024-10-18 11:49:35DataFile Name :005CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	17.95	19.77	16.39	18.03	9.40	ppb
Antimony	121-1	1.98	2.03	2.03	2.01	1.50	ppb
Arsenic	75-2	0.94	1.37	1.09	1.13	19.36	ppb
Barium	135-1	9.52	10.02	9.47	9.67	3.12	ppb
Barium	137-1	9.31	9.57	9.47	9.45	1.36	ppb
Beryllium	9-1	1.05	1.17	1.00	1.07	8.24	ppb
Bismuth	209-1				102		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.96	1.01	1.10	1.03	6.76	ppb
Cadmium	106-1	3.04	0.20	1.26	1.50	95.73	ppb
Cadmium	111-1	1.28	1.22	1.09	1.20	8.19	ppb
Calcium	43-1	500.52	482.66	502.69	495.29	2.22	ppb
Calcium	44-1	518.65	512.27	513.70	514.87	0.65	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1.82	1.97	2.07	1.95	6.25	ppb
Cobalt	59-2	1.12	1.07	1.08	1.09	2.21	ppb
Copper	63-2	2.22	2.00	2.01	2.08	5.95	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				100		%
Indium	115-1				103		%
Indium	115-2				99		%
Iron	56-2	52.27	52.84	52.22	52.45	0.65	ppb
Iron	57-2	59.35	50.02	57.72	55.70	8.95	ppb
Iron	54-2	56.32	54.10	55.59	55.33	2.05	ppb
Krypton	83-1						cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S02

Instrumnet Name :

P7

Client Sample ID :

S02

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:49:35

DataFile Name :

005CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.80	0.80	0.79	0.80	1.22	ppb
Lead	207-1	0.81	0.75	0.81	0.79	4.44	ppb
Lead	208-1	0.79	0.79	0.79	0.79	0.37	ppb
Lithium	6-1				103		%
Magnesium	24-2	468.81	479.26	472.35	473.47	1.12	ppb
Manganese	55-2	1.01	1.11	1.08	1.07	4.59	ppb
Molybdenum	94-1	5.06	5.33	5.37	5.25	3.25	ppb
Molybdenum	95-1	4.36	4.40	4.39	4.39	0.56	ppb
Molybdenum	96-1	4.42	4.59	4.45	4.49	2.03	ppb
Molybdenum	97-1	4.32	4.47	4.53	4.44	2.40	ppb
Molybdenum	98-1	4.44	4.33	4.45	4.41	1.42	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	1.36	1.20	1.04	1.20	13.40	ppb
Phosphorus	31-2	16.51	13.05	19.18	16.25	18.93	ppb
Potassium	39-2	495.08	501.33	498.87	498.43	0.63	ppb
Rhodium	103-1				103		%
Rhodium	103-2				100		%
Scandium	45-1				102		%
Scandium	45-2				100		%
Selenium	82-1	5.44	5.12	4.07	4.88	14.73	ppb
Selenium	77-2	7.81	3.13	6.66	5.87	41.56	ppb
Selenium	78-2	4.77	4.61	3.90	4.43	10.45	ppb
Silicon	28-1	3.50	3.43	3.59	3.51	2.38	ppb
Silver	107-1	1.04	1.04	1.00	1.03	2.21	ppb
Silver	109-1	1.00	0.98	1.00	0.99	1.15	ppb
Sodium	23-2	480.37	483.78	484.36	482.84	0.45	ppb
Strontium	86-1	1.06	0.92	1.04	1.01	7.87	ppb
Strontium	88-1	0.98	0.96	1.00	0.98	1.81	ppb
Sulfur	34-1	-64.54	-154.30	-95.20	-104.68		ppb
Terbium	159-1				102		%
Terbium	159-2				100		%
Thallium	203-1	0.80	0.87	0.85	0.84	4.22	ppb
Thallium	205-1	0.88	0.84	0.84	0.86	2.88	ppb
Tin	118-1	4.81	4.89	4.79	4.83	1.05	ppb
Titanium	47-1	4.86	4.47	4.64	4.65	4.17	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S02

Instrumnet Name :

P7

Client Sample ID :

S02

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:49:35

DataFile Name :

005CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.80	0.78	0.79	0.79	1.35	ppb
Vanadium	51-2	4.93	4.99	4.99	4.97	0.78	ppb
Yttrium	89-1				103		%
Yttrium	89-2				99		%
Zinc	66-2	4.69	5.29	5.21	5.06	6.37	ppb
Zirconium	90-1	0.95	0.96	1.00	0.97	2.78	ppb
Zirconium	91-1	0.95	0.99	1.05	1.00	5.26	ppb

LB Number :	LB133015	Operator :	Jaswal
Lab Sample ID :	S03	Instrumnet Name :	P7
Client Sample ID :	S03	Dilution Factor :	1
Date & Time Acquired :	2024-10-18 11:53:00	DataFile Name :	006CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	970.55	967.43	973.00	970.33	0.29	ppb
Antimony	121-1	49.92	51.48	51.22	50.87	1.65	ppb
Arsenic	75-2	51.84	51.65	50.64	51.38	1.26	ppb
Barium	135-1	246.81	250.13	252.12	249.69	1.07	ppb
Barium	137-1	246.13	250.27	252.27	249.56	1.26	ppb
Beryllium	9-1	52.81	53.17	52.88	52.96	0.37	ppb
Bismuth	209-1				103		%
Bismuth	209-2				99		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	56.36	57.41	54.65	56.14	2.48	ppb
Cadmium	106-1	50.85	54.81	52.12	52.59	3.84	ppb
Cadmium	111-1	54.03	55.28	54.91	54.74	1.17	ppb
Calcium	43-1	5618.80	5721.41	5807.58	5715.93	1.65	ppb
Calcium	44-1	5525.41	5592.80	5562.10	5560.10	0.61	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	50.80	50.92	52.07	51.26	1.37	ppb
Cobalt	59-2	51.91	50.94	52.06	51.64	1.17	ppb
Copper	63-2	525.82	521.54	525.46	524.27	0.45	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				99		%
Indium	115-1				103		%
Indium	115-2				99		%
Iron	56-2	2578.06	2565.84	2582.75	2575.55	0.34	ppb
Iron	57-2	2714.28	2669.78	2737.82	2707.29	1.28	ppb
Iron	54-2	2727.02	2706.02	2767.26	2733.44	1.14	ppb
Krypton	83-1						cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S03

Instrumnet Name :

P7

Client Sample ID :

S03

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:53:00

DataFile Name :

006CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	221.28	222.71	227.08	223.69	1.35	ppb
Lead	207-1	217.53	219.44	224.84	220.60	1.72	ppb
Lead	208-1	232.70	232.61	240.19	235.17	1.85	ppb
Lithium	6-1				104		%
Magnesium	24-2	4640.61	4586.04	4630.56	4619.07	0.63	ppb
Manganese	55-2	518.92	514.17	521.68	518.25	0.73	ppb
Molybdenum	94-1	500.01	498.14	506.76	501.64	0.90	ppb
Molybdenum	95-1	493.05	495.51	504.15	497.57	1.17	ppb
Molybdenum	96-1	493.10	497.55	503.43	498.03	1.04	ppb
Molybdenum	97-1	495.38	499.63	506.05	500.35	1.07	ppb
Molybdenum	98-1	503.65	496.11	501.84	500.53	0.79	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	53.24	53.11	54.41	53.59	1.34	ppb
Phosphorus	31-2	995.52	1018.56	1043.38	1019.15	2.35	ppb
Potassium	39-2	2584.99	2578.72	2633.73	2599.15	1.16	ppb
Rhodium	103-1				102		%
Rhodium	103-2				99		%
Scandium	45-1				101		%
Scandium	45-2				100		%
Selenium	82-1	52.80	50.45	52.95	52.07	2.69	ppb
Selenium	77-2	62.81	51.97	43.35	52.71	18.49	ppb
Selenium	78-2	54.90	49.41	56.07	53.46	6.65	ppb
Silicon	28-1	51.27	53.16	52.85	52.42	1.93	ppb
Silver	107-1	53.89	54.81	54.65	54.45	0.90	ppb
Silver	109-1	53.81	55.37	55.10	54.76	1.52	ppb
Sodium	23-2	4702.46	4657.32	4793.18	4717.66	1.47	ppb
Strontium	86-1	1268.96	1258.63	1285.18	1270.92	1.05	ppb
Strontium	88-1	1251.16	1251.03	1261.10	1254.43	0.46	ppb
Sulfur	34-1	931.22	923.62	922.24	925.69	0.52	ppb
Terbium	159-1				103		%
Terbium	159-2				100		%
Thallium	203-1	44.37	44.79	45.33	44.83	1.07	ppb
Thallium	205-1	43.85	43.71	44.69	44.08	1.20	ppb
Tin	118-1	50.49	52.68	51.57	51.58	2.12	ppb
Titanium	47-1	514.42	523.57	519.46	519.15	0.88	ppb

LB Number :LB133015

Operator :Jaswal

Lab Sample ID :S03

Instrumnet Name :P7

Client Sample ID :S03

Dilution Factor :1

Date & Time Acquired :2024-10-18 11:53:00

DataFile Name :006CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	41.51	41.71	42.26	41.83	0.92	ppb
Vanadium	51-2	50.44	49.41	51.46	50.44	2.04	ppb
Yttrium	89-1				103		%
Yttrium	89-2				99		%
Zinc	66-2	515.92	512.95	518.76	515.88	0.56	ppb
Zirconium	90-1	49.60	50.08	50.95	50.21	1.36	ppb
Zirconium	91-1	49.82	50.06	50.35	50.07	0.53	ppb

LB Number :LB133015Operator :Jaswal

Lab Sample ID :S04Instrumnet Name :P7

Client Sample ID :S04Dilution Factor :1

Date & Time Acquired :2024-10-18 11:56:17DataFile Name :007CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2402.88	2455.53	2461.40	2439.94	1.32	ppb
Antimony	121-1	126.53	128.47	129.04	128.01	1.03	ppb
Arsenic	75-2	128.40	127.25	128.54	128.06	0.55	ppb
Barium	135-1	616.28	625.22	632.95	624.82	1.34	ppb
Barium	137-1	622.03	627.51	634.40	627.98	0.99	ppb
Beryllium	9-1	130.22	130.94	132.45	131.20	0.87	ppb
Bismuth	209-1				102		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	136.49	137.01	133.96	135.82	1.20	ppb
Cadmium	106-1	130.31	135.76	137.19	134.42	2.70	ppb
Cadmium	111-1	136.18	136.13	138.71	137.01	1.08	ppb
Calcium	43-1	14233.86	14277.90	14185.65	14232.47	0.32	ppb
Calcium	44-1	12982.77	12986.06	12875.31	12948.05	0.49	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	127.26	127.94	129.32	128.17	0.82	ppb
Cobalt	59-2	130.81	129.75	130.64	130.40	0.44	ppb
Copper	63-2	1315.73	1329.87	1337.86	1327.82	0.84	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				98		%
Indium	115-1				99		%
Indium	115-2				95		%
Iron	56-2	6323.10	6470.94	6496.99	6430.34	1.46	ppb
Iron	57-2	6810.94	6805.11	6903.46	6839.83	0.81	ppb
Iron	54-2	6847.45	6920.56	6963.67	6910.56	0.85	ppb
Krypton	83-1						cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S04

Instrumnet Name :

P7

Client Sample ID :

S04

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:56:17

DataFile Name :

007CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	616.44	611.12	631.48	619.68	1.70	ppb
Lead	207-1	624.17	630.49	627.71	627.46	0.50	ppb
Lead	208-1	616.78	621.58	627.68	622.01	0.88	ppb
Lithium	6-1				102		%
Magnesium	24-2	11237.04	11375.00	11514.62	11375.55	1.22	ppb
Manganese	55-2	1251.80	1266.53	1283.44	1267.26	1.25	ppb
Molybdenum	94-1	1267.56	1263.69	1288.48	1273.24	1.05	ppb
Molybdenum	95-1	1256.57	1267.43	1267.60	1263.87	0.50	ppb
Molybdenum	96-1	1275.52	1270.03	1272.36	1272.64	0.22	ppb
Molybdenum	97-1	1258.55	1267.99	1272.50	1266.35	0.56	ppb
Molybdenum	98-1	1272.52	1273.63	1266.59	1270.91	0.30	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	131.59	133.68	136.78	134.02	1.95	ppb
Phosphorus	31-2	2541.21	2550.81	2628.14	2573.39	1.85	ppb
Potassium	39-2	6515.88	6559.25	6653.33	6576.15	1.07	ppb
Rhodium	103-1				97		%
Rhodium	103-2				95		%
Scandium	45-1				98		%
Scandium	45-2				95		%
Selenium	82-1	132.79	131.06	131.11	131.65	0.75	ppb
Selenium	77-2	133.11	118.63	120.34	124.03	6.38	ppb
Selenium	78-2	126.65	141.48	120.13	129.42	8.46	ppb
Silicon	28-1	129.62	131.82	132.82	131.42	1.25	ppb
Silver	107-1	133.86	135.58	135.06	134.83	0.66	ppb
Silver	109-1	134.35	135.55	136.59	135.50	0.83	ppb
Sodium	23-2	12108.17	12083.68	12133.13	12108.33	0.20	ppb
Strontium	86-1	3212.34	3246.66	3184.83	3214.61	0.96	ppb
Strontium	88-1	3169.26	3180.65	3176.77	3175.56	0.18	ppb
Sulfur	34-1	2374.87	2292.68	2234.28	2300.61	3.07	ppb
Terbium	159-1				101		%
Terbium	159-2				99		%
Thallium	203-1	112.32	111.75	114.24	112.77	1.15	ppb
Thallium	205-1	110.83	111.20	113.62	111.88	1.35	ppb
Tin	118-1	127.60	128.51	130.47	128.86	1.14	ppb
Titanium	47-1	1298.13	1294.73	1303.84	1298.90	0.35	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S04

Instrumnet Name :

P7

Client Sample ID :

S04

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:56:17

DataFile Name :

007CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	106.82	106.62	108.93	107.46	1.19	ppb
Vanadium	51-2	126.47	127.55	128.37	127.46	0.75	ppb
Yttrium	89-1				100		%
Yttrium	89-2				95		%
Zinc	66-2	1282.55	1298.01	1297.94	1292.83	0.69	ppb
Zirconium	90-1	126.85	127.80	128.12	127.59	0.52	ppb
Zirconium	91-1	126.95	127.78	128.08	127.60	0.46	ppb

LB Number :LB133015

Operator :Jaswal

Lab Sample ID :S05

Instrumnet Name :P7

Client Sample ID :S05

Dilution Factor :1

Date & Time Acquired :2024-10-18 11:59:27

DataFile Name :008CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	4916.71	4920.64	4858.33	4898.56	0.71	ppb
Antimony	121-1	246.47	250.32	253.50	250.10	1.41	ppb
Arsenic	75-2	256.59	258.98	254.26	256.61	0.92	ppb
Barium	135-1	1207.67	1223.41	1236.86	1222.65	1.19	ppb
Barium	137-1	1288.38	1314.25	1314.41	1305.68	1.15	ppb
Beryllium	9-1	255.46	261.25	258.32	258.34	1.12	ppb
Bismuth	209-1				102		%
Bismuth	209-2				93		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	266.19	263.86	272.98	267.68	1.77	ppb
Cadmium	106-1	261.64	265.26	261.49	262.79	0.81	ppb
Cadmium	111-1	259.19	265.06	266.71	263.65	1.50	ppb
Calcium	43-1	27541.19	28696.83	28761.90	28333.31	2.42	ppb
Calcium	44-1	25244.80	25917.52	26061.95	25741.43	1.69	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	257.74	257.00	253.14	255.96	0.97	ppb
Cobalt	59-2	260.85	258.50	254.52	257.96	1.24	ppb
Copper	63-2	2637.19	2575.78	2593.13	2602.03	1.22	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				97		%
Indium	115-1				96		%
Indium	115-2				92		%
Iron	56-2	12928.88	12817.09	12710.39	12818.79	0.85	ppb
Iron	57-2	13821.17	13621.24	13508.29	13650.23	1.16	ppb
Iron	54-2	13949.28	13779.61	13641.33	13790.07	1.12	ppb
Krypton	83-1						cps

LB Number :	LB133015	Operator :	Jaswal
Lab Sample ID :	S05	Instrumnet Name :	P7
Client Sample ID :	S05	Dilution Factor :	1
Date & Time Acquired :	2024-10-18 11:59:27	DataFile Name :	008CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1208.98	1239.66	1259.44	1236.03	2.06	ppb
Lead	207-1	1235.03	1266.88	1288.49	1263.47	2.13	ppb
Lead	208-1	1225.82	1253.85	1264.65	1248.11	1.61	ppb
Lithium	6-1				99		%
Magnesium	24-2	22811.83	22535.17	22576.52	22641.17	0.66	ppb
Manganese	55-2	2586.68	2574.29	2502.53	2554.50	1.78	ppb
Molybdenum	94-1	2486.77	2564.00	2515.72	2522.16	1.55	ppb
Molybdenum	95-1	2487.52	2608.37	2556.84	2550.91	2.38	ppb
Molybdenum	96-1	2491.43	2551.63	2508.19	2517.08	1.23	ppb
Molybdenum	97-1	2480.41	2561.77	2519.75	2520.64	1.61	ppb
Molybdenum	98-1	2468.76	2585.33	2557.36	2537.15	2.40	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	265.94	261.39	257.91	261.75	1.54	ppb
Phosphorus	31-2	5069.81	4999.63	5006.29	5025.24	0.77	ppb
Potassium	39-2	12239.85	12404.11	12277.69	12307.22	0.70	ppb
Rhodium	103-1				94		%
Rhodium	103-2				91		%
Scandium	45-1				93		%
Scandium	45-2				91		%
Selenium	82-1	255.25	262.97	261.88	260.03	1.61	ppb
Selenium	77-2	261.82	275.47	273.40	270.23	2.72	ppb
Selenium	78-2	254.53	260.70	264.42	259.88	1.92	ppb
Silicon	28-1	254.89	266.49	269.18	263.52	2.88	ppb
Silver	107-1	256.08	261.71	262.40	260.07	1.33	ppb
Silver	109-1	256.96	259.89	261.72	259.52	0.92	ppb
Sodium	23-2	24042.26	24293.53	23717.05	24017.61	1.20	ppb
Strontium	86-1	6203.16	6494.77	6326.90	6341.61	2.31	ppb
Strontium	88-1	6224.01	6422.67	6383.19	6343.29	1.66	ppb
Sulfur	34-1	4584.96	4961.09	4886.59	4810.88	4.14	ppb
Terbium	159-1				100		%
Terbium	159-2				96		%
Thallium	203-1	219.47	225.22	225.67	223.45	1.55	ppb
Thallium	205-1	246.82	248.75	253.52	249.70	1.38	ppb
Tin	118-1	246.18	251.37	253.08	250.21	1.44	ppb
Titanium	47-1	2472.00	2521.11	2626.97	2540.03	3.12	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:59:27 DataFile Name : 008CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	243.12	249.88	255.07	249.36	2.40	ppb
Vanadium	51-2	255.31	254.28	252.51	254.03	0.56	ppb
Yttrium	89-1				96		%
Yttrium	89-2				92		%
Zinc	66-2	2600.54	2587.02	2543.20	2576.92	1.16	ppb
Zirconium	90-1	247.39	258.55	253.22	253.05	2.21	ppb
Zirconium	91-1	249.72	259.57	254.63	254.64	1.94	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:02:28 DataFile Name : 009CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	9863.10	9621.12	9792.44	9758.89	1.28	ppb
Antimony	121-1	503.75	515.23	530.93	516.64	2.64	ppb
Arsenic	75-2	502.62	499.34	508.94	503.63	0.97	ppb
Barium	135-1	2546.12	2602.23	2650.27	2599.54	2.01	ppb
Barium	137-1	2489.74	2634.46	2596.20	2573.46	2.91	ppb
Beryllium	9-1	499.01	514.79	509.51	507.77	1.58	ppb
Bismuth	209-1				101		%
Bismuth	209-2				92		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	511.64	519.76	529.77	520.39	1.75	ppb
Cadmium	106-1	499.03	520.40	523.85	514.43	2.61	ppb
Cadmium	111-1	506.44	522.44	525.41	518.10	1.97	ppb
Calcium	43-1	54694.82	57908.05	56410.20	56337.69	2.85	ppb
Calcium	44-1	50213.48	53611.22	51958.64	51927.78	3.27	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	514.35	498.72	507.52	506.86	1.55	ppb
Cobalt	59-2	512.97	504.81	511.39	509.72	0.85	ppb
Copper	63-2	5222.85	5063.92	5151.21	5145.99	1.55	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				97		%
Indium	115-1				92		%
Indium	115-2				89		%
Iron	56-2	25668.41	25465.19	25602.68	25578.76	0.41	ppb
Iron	57-2	27123.90	26928.70	27135.00	27062.53	0.43	ppb
Iron	54-2	26431.11	25356.64	26420.66	26069.47	2.37	ppb
Krypton	83-1						cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S06

Instrumnet Name :

P7

Client Sample ID :

S06

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 12:02:28

DataFile Name :

009CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2458.43	2569.12	2566.62	2531.39	2.50	ppb
Lead	207-1	2436.21	2590.36	2553.19	2526.59	3.18	ppb
Lead	208-1	2439.49	2579.19	2565.47	2528.05	3.05	ppb
Lithium	6-1				97		%
Magnesium	24-2	45487.15	44879.13	46042.99	45469.76	1.28	ppb
Manganese	55-2	5115.27	5003.25	5096.62	5071.71	1.18	ppb
Molybdenum	94-1	4832.69	5143.42	5070.21	5015.44	3.24	ppb
Molybdenum	95-1	4940.47	5184.29	5034.14	5052.97	2.43	ppb
Molybdenum	96-1	4945.36	5175.58	5092.50	5071.15	2.30	ppb
Molybdenum	97-1	4954.88	5167.95	5047.39	5056.74	2.11	ppb
Molybdenum	98-1	4939.36	5125.58	5069.19	5044.71	1.89	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	518.54	507.96	516.10	514.20	1.08	ppb
Phosphorus	31-2	10174.06	10076.09	10283.72	10177.96	1.02	ppb
Potassium	39-2	24281.31	24150.13	24346.76	24259.40	0.41	ppb
Rhodium	103-1				91		%
Rhodium	103-2				88		%
Scandium	45-1				90		%
Scandium	45-2				89		%
Selenium	82-1	494.35	528.87	503.91	509.04	3.50	ppb
Selenium	77-2	512.30	478.27	516.56	502.37	4.18	ppb
Selenium	78-2	489.71	503.31	523.71	505.58	3.38	ppb
Silicon	28-1	496.56	535.69	520.14	517.46	3.81	ppb
Silver	107-1	504.85	527.22	519.99	517.35	2.21	ppb
Silver	109-1	513.83	530.36	533.69	525.96	2.02	ppb
Sodium	23-2	49397.23	48517.21	48890.61	48935.02	0.90	ppb
Strontium	86-1	12332.37	12834.30	12569.78	12578.82	2.00	ppb
Strontium	88-1	12355.41	12956.09	12626.18	12645.89	2.38	ppb
Sulfur	34-1	9375.68	10265.51	9787.95	9809.72	4.54	ppb
Terbium	159-1				99		%
Terbium	159-2				97		%
Thallium	203-1	490.11	512.27	516.26	506.21	2.78	ppb
Thallium	205-1	498.06	512.56	509.31	506.64	1.50	ppb
Tin	118-1	499.36	524.06	525.38	516.27	2.84	ppb
Titanium	47-1	4951.68	5236.65	5037.41	5075.25	2.88	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S06

Instrumnet Name :

P7

Client Sample ID :

S06

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 12:02:28

DataFile Name :

009CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	486.02	508.32	511.39	501.91	2.76	ppb
Vanadium	51-2	507.00	499.66	509.12	505.26	0.98	ppb
Yttrium	89-1				94		%
Yttrium	89-2				90		%
Zinc	66-2	5149.73	5040.29	5105.58	5098.54	1.08	ppb
Zirconium	90-1	496.27	517.33	506.36	506.65	2.08	ppb
Zirconium	91-1	491.04	520.09	509.10	506.74	2.90	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S07

Instrumnet Name :

P7

Client Sample ID :

S07

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 12:05:26

DataFile Name :

010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	18836.67	18794.98	18941.82	18857.82	0.40	ppb
Antimony	121-1	978.32	1011.82	983.57	991.24	1.82	ppb
Arsenic	75-2	1009.03	989.08	990.13	996.08	1.13	ppb
Barium	135-1	4941.03	5085.90	4844.39	4957.11	2.45	ppb
Barium	137-1	4931.24	5031.73	4884.02	4949.00	1.52	ppb
Beryllium	9-1	980.03	1004.10	995.19	993.10	1.23	ppb
Bismuth	209-1				101		%
Bismuth	209-2				92		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	974.21	1005.86	971.10	983.72	1.96	ppb
Cadmium	106-1	977.51	1014.67	972.66	988.28	2.33	ppb
Cadmium	111-1	981.30	1005.20	970.90	985.80	1.78	ppb
Calcium	43-1	109278.28	111500.01	110400.08	110392.79	1.01	ppb
Calcium	44-1	101038.63	103375.94	101794.40	102069.66	1.17	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1000.68	988.89	994.28	994.62	0.59	ppb
Cobalt	59-2	1000.65	983.81	992.71	992.39	0.85	ppb
Copper	63-2	9859.31	9846.67	9965.67	9890.55	0.66	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				98		%
Indium	115-1				91		%
Indium	115-2				87		%
Iron	56-2	50314.04	49643.89	50064.42	50007.45	0.68	ppb
Iron	57-2	51167.86	50183.58	51135.70	50829.05	1.10	ppb
Iron	54-2	51598.03	51373.65	50738.77	51236.82	0.87	ppb
Krypton	83-1						cps

LB Number :LB133015Operator :Jaswal

Lab Sample ID :S07Instrumnet Name :P7

Client Sample ID :S07Dilution Factor :1

Date & Time Acquired :2024-10-18 12:05:26DataFile Name :010CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	4949.97	5100.48	4918.88	4989.78	1.95	ppb
Lead	207-1	4930.26	5061.44	4961.80	4984.50	1.37	ppb
Lead	208-1	4938.75	5064.06	4959.89	4987.56	1.35	ppb
Lithium	6-1				97		%
Magnesium	24-2	90339.40	89081.61	90546.87	89989.29	0.88	ppb
Manganese	55-2	9942.68	10012.25	9887.41	9947.45	0.63	ppb
Molybdenum	94-1	9871.04	10050.49	10029.72	9983.75	0.98	ppb
Molybdenum	95-1	9835.82	10039.69	10002.02	9959.18	1.09	ppb
Molybdenum	96-1	9797.59	10051.04	10023.64	9957.43	1.40	ppb
Molybdenum	97-1	9775.92	10013.92	10103.39	9964.41	1.70	ppb
Molybdenum	98-1	9779.92	9955.86	10161.36	9965.72	1.92	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	990.19	988.96	986.83	988.66	0.17	ppb
Phosphorus	31-2	20189.77	19603.83	19890.14	19894.58	1.47	ppb
Potassium	39-2	47874.48	47408.97	47972.78	47752.08	0.63	ppb
Rhodium	103-1				89		%
Rhodium	103-2				87		%
Scandium	45-1				90		%
Scandium	45-2				90		%
Selenium	82-1	980.05	992.54	1003.52	992.04	1.18	ppb
Selenium	77-2	994.00	979.78	1007.44	993.74	1.39	ppb
Selenium	78-2	980.50	1001.26	1000.30	994.02	1.18	ppb
Silicon	28-1	976.09	1012.69	972.32	987.03	2.26	ppb
Silver	107-1	972.45	1021.52	968.10	987.36	3.00	ppb
Silver	109-1	972.54	1009.03	967.69	983.09	2.30	ppb
Sodium	23-2	96293.35	96165.12	97141.80	96533.42	0.55	ppb
Strontium	86-1	24476.56	24854.60	25445.17	24925.44	1.96	ppb
Strontium	88-1	24684.88	24939.12	25067.57	24897.19	0.78	ppb
Sulfur	34-1	19971.69	20464.76	20076.74	20171.06	1.29	ppb
Terbium	159-1				101		%
Terbium	159-2				98		%
Thallium	203-1	994.72	1029.82	991.42	1005.32	2.12	ppb
Thallium	205-1	993.30	1022.61	980.16	998.69	2.18	ppb
Tin	118-1	973.39	1021.30	979.07	991.25	2.64	ppb
Titanium	47-1	9830.16	10071.94	9933.80	9945.30	1.22	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S07

Instrumnet Name :

P7

Client Sample ID :

S07

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 12:05:26

DataFile Name :

010CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	993.94	1018.29	993.19	1001.81	1.43	ppb
Vanadium	51-2	1002.27	991.94	993.89	996.03	0.55	ppb
Yttrium	89-1				94		%
Yttrium	89-2				90		%
Zinc	66-2	10070.99	9901.53	9803.54	9925.35	1.36	ppb
Zirconium	90-1	985.59	991.11	1010.02	995.58	1.29	ppb
Zirconium	91-1	984.15	992.49	1008.78	995.14	1.26	ppb

LB Number :
Lab Sample ID :
Client Sample ID :
Date & Time Acquired :

LB133015
S08
S08
2024-10-18 12:08:18

Operator :
Instrumnet Name :
Dilution Factor :
DataFile Name :

Jaswal
P7
1
011CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	103726.18	100062.42	96989.66	100259.42	3.36	ppb
Antimony	121-1	0.76	0.74	0.79	0.76	3.81	ppb
Arsenic	75-2	0.37	0.50	0.53	0.47	18.04	ppb
Barium	135-1	2.65	2.51	2.78	2.65	5.10	ppb
Barium	137-1	2.63	2.73	2.76	2.71	2.56	ppb
Beryllium	9-1	0.61	0.59	0.47	0.55	13.18	ppb
Bismuth	209-1				88		%
Bismuth	209-2				73		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.60	0.35	0.37	0.44	31.64	ppb
Cadmium	106-1	0.21	-1.73	-0.69	-0.74		ppb
Cadmium	111-1	0.28	0.08	0.14	0.17	63.51	ppb
Calcium	43-1	491540.99	488339.90	511330.73	497070.54	2.51	ppb
Calcium	44-1	495101.42	494054.87	508861.92	499339.40	1.65	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3.69	3.27	3.62	3.52	6.25	ppb
Cobalt	59-2	2.71	2.76	2.74	2.74	0.81	ppb
Copper	63-2	3.13	2.89	2.88	2.97	4.78	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				96		%
Holmium	165-2				89		%
Indium	115-1				86		%
Indium	115-2				80		%
Iron	56-2	257980.72	248084.22	243693.35	249919.43	2.93	ppb
Iron	57-2	257822.43	246357.38	244481.01	249553.61	2.89	ppb
Iron	54-2	259025.99	247912.41	241748.60	249562.34	3.51	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P7
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:08:18 DataFile Name : 011CAL.S.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.09	1.00	0.98	1.02	5.69	ppb
Lead	207-1	1.07	1.10	0.93	1.03	8.93	ppb
Lead	208-1	1.08	1.03	0.97	1.02	5.16	ppb
Lithium	6-1				95		%
Magnesium	24-2	521717.41	495872.78	490224.96	502605.05	3.34	ppb
Manganese	55-2	6.90	6.71	6.54	6.72	2.63	ppb
Molybdenum	94-1	2.79	2.73	2.74	2.75	1.11	ppb
Molybdenum	95-1	1.87	1.79	1.81	1.82	2.24	ppb
Molybdenum	96-1	3.53	3.23	3.29	3.35	4.72	ppb
Molybdenum	97-1	1.92	1.84	1.71	1.83	5.84	ppb
Molybdenum	98-1	1.75	1.70	1.68	1.71	2.21	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	4.01	4.69	3.88	4.19	10.37	ppb
Phosphorus	31-2	5.10	-1.57	-4.08	-0.18		ppb
Potassium	39-2	262598.39	246961.26	242012.78	250524.14	4.29	ppb
Rhodium	103-1				78		%
Rhodium	103-2				74		%
Scandium	45-1				87		%
Scandium	45-2				85		%
Selenium	82-1	0.45	0.76	1.00	0.73	37.67	ppb
Selenium	77-2	-0.39	0.31	0.29	0.07	578.81	ppb
Selenium	78-2	3.17	-0.67	0.24	0.91	219.06	ppb
Silicon	28-1	13.41	9.19	11.88	11.49	18.57	ppb
Silver	107-1	0.14	0.11	0.13	0.13	9.97	ppb
Silver	109-1	0.13	0.14	0.12	0.13	8.34	ppb
Sodium	23-2	516075.23	493975.17	492534.30	500861.57	2.63	ppb
Strontium	86-1	4.50	4.42	4.42	4.45	1.00	ppb
Strontium	88-1	4.59	4.37	4.21	4.39	4.34	ppb
Sulfur	34-1	-330.70	-256.66	-104.72	-230.69		ppb
Terbium	159-1				94		%
Terbium	159-2				88		%
Thallium	203-1	0.12	0.11	0.10	0.11	8.61	ppb
Thallium	205-1	0.13	0.12	0.10	0.12	14.59	ppb
Tin	118-1	0.41	0.42	0.50	0.44	11.43	ppb
Titanium	47-1	1.47	1.66	1.70	1.61	7.43	ppb

LB Number :LB133015

Operator :Jaswal

Lab Sample ID :S08

Instrumnet Name :P7

Client Sample ID :S08

Dilution Factor :1

Date & Time Acquired :2024-10-18 12:08:18

DataFile Name :011CALS.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.07	0.05	0.05	0.06	17.22	ppb
Vanadium	51-2	0.32	0.36	0.30	0.33	10.30	ppb
Yttrium	89-1				88		%
Yttrium	89-2				82		%
Zinc	66-2	39.09	35.81	36.32	37.07	4.76	ppb
Zirconium	90-1	0.61	0.61	0.62	0.61	1.15	ppb
Zirconium	91-1	0.59	0.58	0.59	0.59	0.52	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

ICV025

Instrumnet Name :

P7

Client Sample ID :

ICV025

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 12:59:06

DataFile Name :

012ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	467.90	462.81	463.23	464.65	0.61	ppb
Antimony	121-1	186.90	190.72	195.06	190.89	2.14	ppb
Arsenic	75-2	199.41	200.93	201.65	200.66	0.57	ppb
Barium	135-1	91.03	91.81	94.03	92.29	1.69	ppb
Barium	137-1	90.93	92.29	94.34	92.52	1.86	ppb
Beryllium	9-1	101.39	103.31	101.60	102.10	1.03	ppb
Bismuth	209-1				106		%
Bismuth	209-2				99		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	148.03	150.23	156.93	151.73	3.06	ppb
Cadmium	106-1	168.18	165.74	169.63	167.85	1.17	ppb
Cadmium	111-1	106.26	107.41	108.58	107.42	1.08	ppb
Calcium	43-1	2159.52	2124.30	2155.83	2146.55	0.90	ppb
Calcium	44-1	2104.63	2072.89	2098.76	2092.09	0.81	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	101.45	100.95	100.10	100.83	0.68	ppb
Cobalt	59-2	103.05	103.21	100.77	102.34	1.33	ppb
Copper	63-2	102.94	102.93	102.29	102.72	0.36	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				104		%
Holmium	165-2				97		%
Indium	115-1				96		%
Indium	115-2				91		%
Iron	56-2	2056.21	2005.72	1976.85	2012.93	2.00	ppb
Iron	57-2	2115.37	2094.95	2130.47	2113.59	0.84	ppb
Iron	54-2	2209.45	2177.28	2158.35	2181.69	1.18	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICV025 Instrumnet Name : P7
Client Sample ID : ICV025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:59:06 DataFile Name : 012ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	185.69	186.10	190.08	187.29	1.29	ppb
Lead	207-1	177.14	176.49	181.57	178.40	1.55	ppb
Lead	208-1	181.00	181.03	185.89	182.64	1.54	ppb
Lithium	6-1				101		%
Magnesium	24-2	1125.02	1124.55	1099.84	1116.47	1.29	ppb
Manganese	55-2	103.13	103.20	100.02	102.12	1.78	ppb
Molybdenum	94-1	4612.10	4491.86	4613.56	4572.51	1.53	ppb
Molybdenum	95-1	4508.17	4462.32	4506.85	4492.44	0.58	ppb
Molybdenum	96-1	4490.07	4472.26	4550.34	4504.22	0.91	ppb
Molybdenum	97-1	4480.36	4520.14	4551.98	4517.50	0.79	ppb
Molybdenum	98-1	4532.71	4480.20	4471.11	4494.67	0.74	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	108.95	108.67	104.90	107.51	2.10	ppb
Phosphorus	31-2	-4.19	-10.71	-6.82	-7.24		ppb
Potassium	39-2	1943.11	1947.84	1935.14	1942.03	0.33	ppb
Rhodium	103-1				96		%
Rhodium	103-2				93		%
Scandium	45-1				95		%
Scandium	45-2				92		%
Selenium	82-1	203.25	205.27	211.78	206.76	2.16	ppb
Selenium	77-2	192.58	208.11	191.19	197.29	4.76	ppb
Selenium	78-2	203.41	219.22	202.26	208.30	4.55	ppb
Silicon	28-1	517.08	511.58	526.05	518.23	1.41	ppb
Silver	107-1	48.22	48.92	49.89	49.01	1.71	ppb
Silver	109-1	47.79	48.35	49.90	48.68	2.24	ppb
Sodium	23-2	1951.30	1951.07	1938.66	1947.01	0.37	ppb
Strontium	86-1	497.05	495.00	501.47	497.84	0.66	ppb
Strontium	88-1	506.00	509.52	511.75	509.09	0.57	ppb
Sulfur	34-1	-7.84	-281.43	-130.39	-139.89		ppb
Terbium	159-1				102		%
Terbium	159-2				97		%
Thallium	203-1	183.95	184.52	190.02	186.16	1.80	ppb
Thallium	205-1	208.44	206.74	208.74	207.98	0.52	ppb
Tin	118-1	540.09	544.78	548.03	544.30	0.73	ppb
Titanium	47-1	4474.81	4381.77	4468.78	4441.79	1.17	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICV025 Instrumnet Name : P7
Client Sample ID : ICV025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:59:06 DataFile Name : 012ICV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	499.31	497.55	508.09	501.65	1.13	ppb
Vanadium	51-2	96.97	97.42	95.64	96.68	0.96	ppb
Yttrium	89-1				97		%
Yttrium	89-2				92		%
Zinc	66-2	198.53	195.78	197.12	197.15	0.70	ppb
Zirconium	90-1	501.93	507.20	511.18	506.77	0.92	ppb
Zirconium	91-1	497.99	501.62	503.41	501.00	0.55	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICB025 Instrumnet Name : P7
Client Sample ID : ICB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:16:30 DataFile Name : 014CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	-0.39	-0.64	-0.36	-0.46		ppb
Antimony	121-1	0.01	0.01	0.01	0.01	40.65	ppb
Arsenic	75-2	-0.01	-0.01	-0.01	-0.01		ppb
Barium	135-1	0.01	-0.01	-0.01	0.00		ppb
Barium	137-1	-0.01	-0.01	0.00	0.00		ppb
Beryllium	9-1	0.02	0.02	0.02	0.02	11.15	ppb
Bismuth	209-1				107		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.00	-0.09	-0.09	-0.06		ppb
Cadmium	106-1	-1.51	-2.46	-3.32	-2.43		ppb
Cadmium	111-1	-0.11	-0.18	-0.24	-0.18		ppb
Calcium	43-1	-1.68	-4.21	-7.27	-4.39		ppb
Calcium	44-1	-5.29	-5.38	-5.93	-5.53		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	-0.01	-0.02	0.04	0.00	848.50	ppb
Cobalt	59-2	-0.01	-0.01	-0.01	-0.01		ppb
Copper	63-2	-0.03	-0.05	-0.06	-0.04		ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				99		%
Indium	115-1				101		%
Indium	115-2				95		%
Iron	56-2	-0.43	-0.24	-0.32	-0.33		ppb
Iron	57-2	-3.39	-0.65	-1.62	-1.89		ppb
Iron	54-2	-0.88	-1.06	-0.31	-0.75		ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICB025 Instrumnet Name : P7
Client Sample ID : ICB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:16:30 DataFile Name : 014CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	-0.06	-0.05	-0.05	-0.05		ppb
Lead	207-1	-0.06	-0.04	-0.04	-0.05		ppb
Lead	208-1	-0.07	-0.05	-0.06	-0.06		ppb
Lithium	6-1				102		%
Magnesium	24-2	-0.76	-0.65	-0.92	-0.78		ppb
Manganese	55-2	0.01	-0.01	-0.01	0.00		ppb
Molybdenum	94-1	0.02	0.02	0.02	0.02	21.19	ppb
Molybdenum	95-1	0.01	0.00	0.01	0.01	89.53	ppb
Molybdenum	96-1	0.00	0.01	0.00	0.01	120.40	ppb
Molybdenum	97-1	0.00	0.00	0.01	0.01	113.72	ppb
Molybdenum	98-1	0.01	0.00	0.02	0.01	64.92	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.00	-0.02	0.00	-0.01		ppb
Phosphorus	31-2	-12.37	-10.65	-17.15	-13.39		ppb
Potassium	39-2	-14.98	-11.01	-14.09	-13.36		ppb
Rhodium	103-1				97		%
Rhodium	103-2				96		%
Scandium	45-1				97		%
Scandium	45-2				95		%
Selenium	82-1	-0.57	-0.43	0.54	-0.15		ppb
Selenium	77-2	-0.39	0.22	-0.39	-0.19		ppb
Selenium	78-2	-0.60	0.05	-0.19	-0.25		ppb
Silicon	28-1	-7.54	-6.78	-7.29	-7.20		ppb
Silver	107-1	0.00	0.00	0.00	0.00	56.42	ppb
Silver	109-1	0.00	0.00	0.00	0.00	62.80	ppb
Sodium	23-2	4.91	4.11	4.77	4.60	9.39	ppb
Strontium	86-1	0.01	-0.01	-0.03	-0.01		ppb
Strontium	88-1	0.00	0.00	0.00	0.00		ppb
Sulfur	34-1	-91.89	-41.70	-15.15	-49.58		ppb
Terbium	159-1				103		%
Terbium	159-2				99		%
Thallium	203-1	0.01	0.00	0.00	0.00	72.36	ppb
Thallium	205-1	0.00	0.01	0.01	0.00	63.97	ppb
Tin	118-1	-0.02	0.02	-0.02	-0.01		ppb
Titanium	47-1	0.02	-0.02	0.00	0.00		ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

ICB025

Instrumnet Name :

P7

Client Sample ID :

ICB025

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 13:16:30

DataFile Name :

014CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	144.77	ppb
Vanadium	51-2	0.00	0.00	0.00	0.00		ppb
Yttrium	89-1				98		%
Yttrium	89-2				95		%
Zinc	66-2	0.13	0.04	0.11	0.09	53.92	ppb
Zirconium	90-1	0.02	0.01	0.01	0.02	16.73	ppb
Zirconium	91-1	0.01	0.02	0.03	0.02	45.18	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSA025 Instrumnet Name : P7
Client Sample ID : ICSA025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:19:53 DataFile Name : 015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	85032.07	85923.34	85163.75	85373.06	0.56	ppb
Antimony	121-1	1.07	0.97	1.08	1.04	5.83	ppb
Arsenic	75-2	0.40	0.25	0.34	0.33	23.20	ppb
Barium	135-1	1.17	1.22	1.23	1.21	2.85	ppb
Barium	137-1	1.25	1.32	1.28	1.28	2.54	ppb
Beryllium	9-1	0.33	0.33	0.24	0.30	15.81	ppb
Bismuth	209-1				100		%
Bismuth	209-2				92		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	8.23	7.62	7.34	7.73	5.92	ppb
Cadmium	106-1	-2.13	-3.35	-2.33	-2.60		ppb
Cadmium	111-1	0.09	0.19	0.07	0.12	56.18	ppb
Calcium	43-1	99179.43	99639.24	101427.14	100081.93	1.19	ppb
Calcium	44-1	93987.89	94543.06	96242.26	94924.40	1.24	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	19.46	19.93	19.42	19.60	1.45	ppb
Cobalt	59-2	1.22	1.29	1.28	1.26	3.26	ppb
Copper	63-2	7.08	7.30	7.70	7.36	4.30	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				99		%
Indium	115-1				93		%
Indium	115-2				90		%
Iron	56-2	97244.12	99929.14	97897.48	98356.92	1.42	ppb
Iron	57-2	97221.73	99537.63	97853.47	98204.28	1.22	ppb
Iron	54-2	97701.92	99083.13	98632.71	98472.59	0.72	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSA025 Instrumnet Name : P7
Client Sample ID : ICSA025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:19:53 DataFile Name : 015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	4.00	4.11	4.11	4.08	1.59	ppb
Lead	207-1	3.60	3.61	3.75	3.65	2.35	ppb
Lead	208-1	3.71	3.78	3.81	3.77	1.29	ppb
Lithium	6-1				97		%
Magnesium	24-2	86040.60	88634.32	86712.74	87129.22	1.54	ppb
Manganese	55-2	8.15	8.24	7.77	8.06	3.12	ppb
Molybdenum	94-1	1428.78	1440.55	1423.21	1430.85	0.62	ppb
Molybdenum	95-1	1740.16	1738.98	1710.65	1729.93	0.97	ppb
Molybdenum	96-1	1699.72	1706.07	1681.68	1695.82	0.75	ppb
Molybdenum	97-1	1726.25	1728.06	1729.86	1728.06	0.10	ppb
Molybdenum	98-1	1730.11	1752.47	1742.53	1741.70	0.64	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	5.53	5.19	5.40	5.37	3.16	ppb
Phosphorus	31-2	96239.18	97763.34	96795.38	96932.64	0.80	ppb
Potassium	39-2	91475.51	94392.25	92292.53	92720.10	1.62	ppb
Rhodium	103-1				88		%
Rhodium	103-2				88		%
Scandium	45-1				88		%
Scandium	45-2				90		%
Selenium	82-1	-0.04	-0.62	-0.68	-0.45		ppb
Selenium	77-2	-0.39	-0.39	-0.39	-0.39		ppb
Selenium	78-2	0.19	-0.52	-0.38	-0.24		ppb
Silicon	28-1	16.81	15.87	17.11	16.60	3.90	ppb
Silver	107-1	0.05	0.04	0.06	0.05	16.88	ppb
Silver	109-1	0.05	0.05	0.06	0.05	16.39	ppb
Sodium	23-2	92603.52	93593.49	92631.68	92942.90	0.61	ppb
Strontium	86-1	32.80	33.36	33.11	33.09	0.85	ppb
Strontium	88-1	32.73	33.41	33.37	33.17	1.14	ppb
Sulfur	34-1	98445.51	97853.96	100411.56	98903.68	1.35	ppb
Terbium	159-1				102		%
Terbium	159-2				98		%
Thallium	203-1	0.06	0.06	0.07	0.06	8.60	ppb
Thallium	205-1	0.06	0.06	0.07	0.06	10.08	ppb
Tin	118-1	0.22	0.25	0.24	0.24	5.83	ppb
Titanium	47-1	1839.41	1839.21	1861.64	1846.75	0.70	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

ICSA025

Instrumnet Name :

P7

Client Sample ID :

ICSA025

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 13:19:53

DataFile Name :

015ICSA.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.01	0.02	0.02	8.49	ppb
Vanadium	51-2	0.21	0.20	0.21	0.21	2.80	ppb
Yttrium	89-1				94		%
Yttrium	89-2				92		%
Zinc	66-2	10.69	11.43	10.78	10.97	3.65	ppb
Zirconium	90-1	0.04	0.04	0.04	0.04	6.39	ppb
Zirconium	91-1	0.05	0.06	0.06	0.05	11.09	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSAB025 Instrumnet Name : P7
Client Sample ID : ICSAB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:23:05 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	84722.42	84497.09	84046.19	84421.90	0.41	ppb
Antimony	121-1	19.44	19.14	19.72	19.44	1.48	ppb
Arsenic	75-2	21.03	18.62	19.42	19.69	6.23	ppb
Barium	135-1	18.91	18.64	19.17	18.90	1.40	ppb
Barium	137-1	18.63	19.29	18.75	18.89	1.88	ppb
Beryllium	9-1	19.77	19.89	20.27	19.98	1.31	ppb
Bismuth	209-1				101		%
Bismuth	209-2				96		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	25.40	27.45	26.93	26.59	4.00	ppb
Cadmium	106-1	21.25	21.89	21.10	21.41	1.95	ppb
Cadmium	111-1	20.03	19.25	19.03	19.44	2.68	ppb
Calcium	43-1	98126.52	96844.02	100946.06	98638.87	2.13	ppb
Calcium	44-1	93998.50	92073.02	95909.83	93993.79	2.04	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	38.76	38.20	38.46	38.47	0.73	ppb
Cobalt	59-2	20.80	20.53	20.53	20.62	0.75	ppb
Copper	63-2	25.44	25.56	25.10	25.37	0.92	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				106		%
Holmium	165-2				100		%
Indium	115-1				95		%
Indium	115-2				91		%
Iron	56-2	95850.26	96037.29	94960.46	95616.00	0.60	ppb
Iron	57-2	96713.96	97035.84	95582.78	96444.19	0.79	ppb
Iron	54-2	97979.23	97389.20	97805.10	97724.51	0.31	ppb
Krypton	83-1						cps

LB Number :LB133015Operator :Jaswal

Lab Sample ID :ICSAB025Instrumnet Name :P7

Client Sample ID :ICSAB025Dilution Factor :1

Date & Time Acquired :2024-10-18 13:23:05DataFile Name :016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	23.89	23.72	24.78	24.13	2.37	ppb
Lead	207-1	23.02	22.28	23.16	22.82	2.06	ppb
Lead	208-1	23.17	22.76	23.53	23.15	1.67	ppb
Lithium	6-1				101		%
Magnesium	24-2	87514.75	86332.42	85810.84	86552.67	1.01	ppb
Manganese	55-2	27.33	27.15	27.29	27.26	0.33	ppb
Molybdenum	94-1	1541.37	1540.40	1581.16	1554.31	1.50	ppb
Molybdenum	95-1	1752.55	1756.69	1769.12	1759.45	0.49	ppb
Molybdenum	96-1	1722.62	1741.68	1748.01	1737.44	0.76	ppb
Molybdenum	97-1	1775.04	1756.18	1793.73	1774.98	1.06	ppb
Molybdenum	98-1	1770.70	1723.94	1786.42	1760.35	1.85	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	24.28	24.79	23.94	24.34	1.76	ppb
Phosphorus	31-2	96157.32	96072.45	95822.69	96017.49	0.18	ppb
Potassium	39-2	91459.49	90225.41	90450.73	90711.88	0.72	ppb
Rhodium	103-1				91		%
Rhodium	103-2				90		%
Scandium	45-1				94		%
Scandium	45-2				93		%
Selenium	82-1	20.16	19.87	20.77	20.27	2.27	ppb
Selenium	77-2	15.77	21.92	16.80	18.17	18.13	ppb
Selenium	78-2	20.01	22.96	18.58	20.52	10.90	ppb
Silicon	28-1	66.36	65.17	71.95	67.83	5.34	ppb
Silver	107-1	17.36	17.44	17.63	17.48	0.80	ppb
Silver	109-1	17.45	17.42	17.53	17.47	0.33	ppb
Sodium	23-2	92652.28	92001.74	92219.34	92291.12	0.36	ppb
Strontium	86-1	81.94	81.28	82.06	81.76	0.51	ppb
Strontium	88-1	81.22	81.14	81.65	81.34	0.33	ppb
Sulfur	34-1	97089.12	96505.13	99218.72	97604.32	1.46	ppb
Terbium	159-1				106		%
Terbium	159-2				100		%
Thallium	203-1	22.24	21.80	22.70	22.25	2.04	ppb
Thallium	205-1	22.13	21.50	22.58	22.07	2.46	ppb
Tin	118-1	48.45	47.99	48.60	48.35	0.65	ppb
Titanium	47-1	1854.88	1834.87	1902.60	1864.12	1.87	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSAB025 Instrumnet Name : P7
Client Sample ID : ICSAB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:23:05 DataFile Name : 016ICSB.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	44.32	43.37	44.92	44.21	1.78	ppb
Vanadium	51-2	19.51	19.15	19.14	19.27	1.11	ppb
Yttrium	89-1				97		%
Yttrium	89-2				94		%
Zinc	66-2	30.85	30.22	29.62	30.23	2.04	ppb
Zirconium	90-1	48.82	48.99	49.39	49.07	0.59	ppb
Zirconium	91-1	49.31	48.62	49.66	49.20	1.08	ppb

LB Number :LB133015Operator :Jaswal

Lab Sample ID :CCV094Instrumnet Name :P7

Client Sample ID :CCV094Dilution Factor :1

Date & Time Acquired :2024-10-18 13:35:33DataFile Name :017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	48494.52	47437.81	47683.60	47871.98	1.16	ppb
Antimony	121-1	498.95	503.40	495.05	499.14	0.84	ppb
Arsenic	75-2	498.26	499.37	501.93	499.85	0.38	ppb
Barium	135-1	2559.05	2520.41	2549.55	2543.00	0.79	ppb
Barium	137-1	2532.44	2509.16	2541.69	2527.76	0.66	ppb
Beryllium	9-1	512.74	523.72	516.35	517.60	1.08	ppb
Bismuth	209-1				95		%
Bismuth	209-2				80		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	496.73	481.48	476.45	484.89	2.18	ppb
Cadmium	106-1	491.64	483.61	486.97	487.41	0.83	ppb
Cadmium	111-1	481.14	479.66	476.36	479.05	0.51	ppb
Calcium	43-1	241695.22	248323.08	251560.50	247192.93	2.03	ppb
Calcium	44-1	244452.46	244410.62	250449.57	246437.55	1.41	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	505.17	504.19	506.56	505.31	0.24	ppb
Cobalt	59-2	536.16	533.40	534.48	534.68	0.26	ppb
Copper	63-2	5091.13	5173.00	5066.06	5110.06	1.09	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				98		%
Holmium	165-2				96		%
Indium	115-1				89		%
Indium	115-2				85		%
Iron	56-2	124685.90	125331.36	123769.14	124595.47	0.63	ppb
Iron	57-2	125986.07	124929.98	125972.58	125629.55	0.48	ppb
Iron	54-2	126545.80	124987.51	124855.35	125462.89	0.75	ppb
Krypton	83-1						cps

LB Number :	LB133015	Operator :	Jaswal
Lab Sample ID :	CCV094	Instrumnet Name :	P7
Client Sample ID :	CCV094	Dilution Factor :	1
Date & Time Acquired :	2024-10-18 13:35:33	DataFile Name :	017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2582.28	2602.22	2619.90	2601.47	0.72	ppb
Lead	207-1	2586.97	2620.23	2610.83	2606.01	0.66	ppb
Lead	208-1	2575.76	2632.05	2607.64	2605.15	1.08	ppb
Lithium	6-1				93		%
Magnesium	24-2	232051.31	228609.88	230042.33	230234.51	0.75	ppb
Manganese	55-2	5058.96	5047.77	5017.38	5041.37	0.43	ppb
Molybdenum	94-1	5088.83	5237.63	5193.45	5173.31	1.48	ppb
Molybdenum	95-1	5120.77	5213.33	5188.37	5174.16	0.93	ppb
Molybdenum	96-1	5101.65	5207.03	5161.47	5156.72	1.02	ppb
Molybdenum	97-1	5044.19	5214.98	5202.77	5153.98	1.85	ppb
Molybdenum	98-1	5152.88	5245.39	5168.41	5188.89	0.95	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	532.41	532.15	530.86	531.81	0.16	ppb
Phosphorus	31-2	10012.32	10034.09	10088.93	10045.11	0.39	ppb
Potassium	39-2	122293.68	120983.07	120773.70	121350.15	0.68	ppb
Rhodium	103-1				83		%
Rhodium	103-2				83		%
Scandium	45-1				88		%
Scandium	45-2				89		%
Selenium	82-1	485.55	493.47	491.06	490.03	0.83	ppb
Selenium	77-2	513.83	476.75	480.76	490.45	4.15	ppb
Selenium	78-2	492.19	496.40	494.27	494.29	0.43	ppb
Silicon	28-1	781.25	784.14	814.40	793.26	2.31	ppb
Silver	107-1	477.44	486.25	479.25	480.98	0.97	ppb
Silver	109-1	484.61	486.07	475.16	481.94	1.23	ppb
Sodium	23-2	245702.62	243909.56	247096.00	245569.39	0.65	ppb
Strontium	86-1	12789.97	12943.22	12903.62	12878.94	0.62	ppb
Strontium	88-1	12774.36	13036.55	12878.72	12896.55	1.02	ppb
Sulfur	34-1	9838.97	9885.33	9994.56	9906.29	0.81	ppb
Terbium	159-1				99		%
Terbium	159-2				94		%
Thallium	203-1	521.39	522.77	520.43	521.53	0.23	ppb
Thallium	205-1	521.15	525.96	529.51	525.54	0.80	ppb
Tin	118-1	506.02	518.83	510.10	511.65	1.28	ppb
Titanium	47-1	4967.70	5058.51	5130.05	5052.09	1.61	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCV094

Instrumnet Name :

P7

Client Sample ID :

CCV094

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 13:35:33

DataFile Name :

017CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	529.40	537.59	535.34	534.11	0.79	ppb
Vanadium	51-2	510.15	506.34	509.50	508.66	0.40	ppb
Yttrium	89-1				91		%
Yttrium	89-2				88		%
Zinc	66-2	5216.12	5190.44	5207.86	5204.80	0.25	ppb
Zirconium	90-1	518.67	524.65	521.62	521.65	0.57	ppb
Zirconium	91-1	511.12	525.10	518.28	518.17	1.35	ppb

LB Number :LB133015Operator :Jaswal

Lab Sample ID :CCB094Instrumnet Name :P7

Client Sample ID :CCB094Dilution Factor :1

Date & Time Acquired :2024-10-18 13:38:25DataFile Name :018CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	1.71	1.65	2.22	1.86	16.79	ppb
Antimony	121-1	0.32	0.33	0.35	0.33	5.81	ppb
Arsenic	75-2	0.11	0.09	-0.01	0.06	96.02	ppb
Barium	135-1	0.10	0.05	0.03	0.06	56.29	ppb
Barium	137-1	0.06	0.04	0.04	0.05	29.78	ppb
Beryllium	9-1	0.30	0.20	0.29	0.26	20.97	ppb
Bismuth	209-1				108		%
Bismuth	209-2				103		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	-0.05	-0.05	-0.05	-0.05		ppb
Cadmium	106-1	-1.16	-2.39	-1.16	-1.57		ppb
Cadmium	111-1	-0.02	-0.14	-0.04	-0.07		ppb
Calcium	43-1	8.52	1.25	0.60	3.46	127.29	ppb
Calcium	44-1	3.44	1.77	0.59	1.93	74.16	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.06	0.00	0.07	0.04	90.96	ppb
Cobalt	59-2	0.01	0.02	0.02	0.02	36.96	ppb
Copper	63-2	0.36	0.19	0.34	0.29	30.58	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				103		%
Holmium	165-2				99		%
Indium	115-1				100		%
Indium	115-2				94		%
Iron	56-2	6.21	6.66	7.59	6.82	10.28	ppb
Iron	57-2	1.45	3.51	4.72	3.23	51.32	ppb
Iron	54-2	4.45	6.38	7.16	6.00	23.27	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCB094 Instrumnet Name : P7
Client Sample ID : CCB094 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:38:25 DataFile Name : 018CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.15	0.12	0.05	0.11	51.57	ppb
Lead	207-1	0.11	0.10	0.09	0.10	10.13	ppb
Lead	208-1	0.13	0.09	0.07	0.10	29.75	ppb
Lithium	6-1				103		%
Magnesium	24-2	8.34	7.97	9.83	8.71	11.25	ppb
Manganese	55-2	0.22	0.19	0.24	0.22	11.02	ppb
Molybdenum	94-1	0.50	0.50	0.45	0.48	5.81	ppb
Molybdenum	95-1	0.30	0.27	0.28	0.28	4.73	ppb
Molybdenum	96-1	0.42	0.32	0.29	0.34	20.49	ppb
Molybdenum	97-1	0.30	0.26	0.29	0.28	7.14	ppb
Molybdenum	98-1	0.33	0.29	0.28	0.30	8.92	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.05	0.06	0.05	0.05	11.35	ppb
Phosphorus	31-2	-4.95	-4.77	-12.46	-7.39		ppb
Potassium	39-2	2.82	0.12	5.88	2.94	97.87	ppb
Rhodium	103-1				95		%
Rhodium	103-2				95		%
Scandium	45-1				93		%
Scandium	45-2				93		%
Selenium	82-1	-0.74	-0.53	0.02	-0.42		ppb
Selenium	77-2	0.86	0.23	-0.39	0.23	269.81	ppb
Selenium	78-2	0.50	-2.20	-1.00	-0.90		ppb
Silicon	28-1	-1.68	-4.46	-3.07	-3.07		ppb
Silver	107-1	0.07	0.05	0.03	0.05	34.20	ppb
Silver	109-1	0.06	0.04	0.04	0.05	14.72	ppb
Sodium	23-2	47.15	43.45	43.96	44.85	4.47	ppb
Strontium	86-1	0.50	0.34	0.37	0.40	20.99	ppb
Strontium	88-1	0.47	0.34	0.30	0.37	24.02	ppb
Sulfur	34-1	-115.28	-254.09	-139.49	-169.62		ppb
Terbium	159-1				104		%
Terbium	159-2				99		%
Thallium	203-1	0.07	0.07	0.06	0.07	14.00	ppb
Thallium	205-1	0.08	0.05	0.05	0.06	21.46	ppb
Tin	118-1	0.04	0.07	0.05	0.05	25.41	ppb
Titanium	47-1	0.31	0.24	0.20	0.25	22.55	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCB094

Instrumnet Name :

P7

Client Sample ID :

CCB094

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 13:38:25

DataFile Name :

018CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.02	0.01	0.01	0.01	17.87	ppb
Vanadium	51-2	0.02	0.02	0.02	0.02	15.57	ppb
Yttrium	89-1				97		%
Yttrium	89-2				93		%
Zinc	66-2	0.42	0.57	0.64	0.54	20.44	ppb
Zirconium	90-1	0.14	0.11	0.13	0.13	11.09	ppb
Zirconium	91-1	0.12	0.15	0.17	0.15	15.27	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2816-01A Instrumnet Name : P7
Client Sample ID : MYD4J8ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:52:29 DataFile Name : 020SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	34542.87	33842.84	34716.11	34367.28	1.35	ppb
Antimony	121-1	0.83	0.81	0.82	0.82	1.36	ppb
Arsenic	75-2	219.63	218.82	222.13	220.20	0.78	ppb
Barium	135-1	232.44	237.27	232.63	234.11	1.17	ppb
Barium	137-1	233.42	234.14	233.99	233.85	0.16	ppb
Beryllium	9-1	1.25	1.39	1.24	1.29	6.42	ppb
Bismuth	209-1				106		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.80	1.86	1.59	1.75	8.08	ppb
Cadmium	106-1	1.25	-0.27	-0.65	0.11	905.90	ppb
Cadmium	111-1	0.81	0.69	0.67	0.72	10.51	ppb
Calcium	43-1	13283.45	13272.76	13402.06	13319.42	0.54	ppb
Calcium	44-1	12853.37	13063.33	12712.07	12876.26	1.37	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	37.38	37.03	37.87	37.43	1.12	ppb
Cobalt	59-2	31.65	31.18	32.05	31.63	1.37	ppb
Copper	63-2	108.35	103.13	106.25	105.91	2.48	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				107		%
Holmium	165-2				101		%
Indium	115-1				101		%
Indium	115-2				95		%
Iron	56-2	79969.85	78307.37	79925.20	79400.81	1.19	ppb
Iron	57-2	80241.80	77122.19	80504.17	79289.39	2.37	ppb
Iron	54-2	79736.87	78613.15	80879.40	79743.14	1.42	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2816-01A Instrumnet Name : P7
Client Sample ID : MYD4J8ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:52:29 DataFile Name : 020SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	38.60	39.89	39.69	39.39	1.77	ppb
Lead	207-1	38.32	39.10	39.79	39.07	1.88	ppb
Lead	208-1	38.39	38.94	39.86	39.06	1.89	ppb
Lithium	6-1				108		%
Magnesium	24-2	12528.05	12332.60	13072.42	12644.36	3.03	ppb
Manganese	55-2	1421.94	1399.02	1420.57	1413.84	0.91	ppb
Molybdenum	94-1	14.60	14.14	14.59	14.45	1.83	ppb
Molybdenum	95-1	0.84	0.85	0.87	0.85	1.35	ppb
Molybdenum	96-1	2.82	2.77	2.84	2.81	1.25	ppb
Molybdenum	97-1	0.85	0.92	0.99	0.92	7.32	ppb
Molybdenum	98-1	0.85	0.85	0.82	0.84	1.95	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	34.03	34.57	33.96	34.18	0.97	ppb
Phosphorus	31-2	1991.44	2023.92	2080.98	2032.11	2.23	ppb
Potassium	39-2	5260.56	5154.52	5305.74	5240.28	1.48	ppb
Rhodium	103-1				97		%
Rhodium	103-2				96		%
Scandium	45-1				105		%
Scandium	45-2				104		%
Selenium	82-1	-0.50	0.06	0.02	-0.14		ppb
Selenium	77-2	11.35	15.15	8.65	11.72	27.85	ppb
Selenium	78-2	1.73	1.68	1.64	1.68	2.65	ppb
Silicon	28-1	7896.74	7861.02	7878.15	7878.63	0.23	ppb
Silver	107-1	0.29	0.30	0.32	0.30	4.87	ppb
Silver	109-1	0.28	0.30	0.30	0.30	3.50	ppb
Sodium	23-2	191.40	189.18	196.83	192.47	2.04	ppb
Strontium	86-1	54.84	55.34	56.73	55.64	1.76	ppb
Strontium	88-1	55.61	55.23	57.06	55.97	1.72	ppb
Sulfur	34-1	-297.87	-568.40	-758.41	-541.56		ppb
Terbium	159-1				105		%
Terbium	159-2				100		%
Thallium	203-1	0.26	0.25	0.23	0.25	5.71	ppb
Thallium	205-1	0.25	0.24	0.26	0.25	3.73	ppb
Tin	118-1	1.43	1.53	1.42	1.46	4.15	ppb
Titanium	47-1	1146.96	1150.46	1152.46	1149.96	0.24	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2816-01A

Instrumnet Name :

P7

Client Sample ID :

MYD4J8ARE

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 13:52:29

DataFile Name :

020SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.66	0.65	0.68	0.66	1.85	ppb
Vanadium	51-2	147.47	145.87	150.97	148.10	1.76	ppb
Yttrium	89-1				109		%
Yttrium	89-2				103		%
Zinc	66-2	200.05	196.28	205.49	200.60	2.31	ppb
Zirconium	90-1	8.06	8.08	8.29	8.14	1.56	ppb
Zirconium	91-1	7.94	8.00	8.24	8.06	1.98	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2823-01ADLX4 Instrumnet Name : P7
Client Sample ID : MYD5X1ARE Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:06:11 DataFile Name : 021SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	6941.07	6834.82	6969.52	6915.14	1.03	ppb
Antimony	121-1	0.21	0.19	0.22	0.21	7.94	ppb
Arsenic	75-2	9.36	9.70	9.84	9.63	2.58	ppb
Barium	135-1	245.40	210.32	218.20	224.64	8.19	ppb
Barium	137-1	247.33	213.60	218.64	226.52	8.03	ppb
Beryllium	9-1	0.32	0.29	0.28	0.30	6.79	ppb
Bismuth	209-1				100		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.53	0.36	0.46	0.45	19.85	ppb
Cadmium	106-1	-1.66	-1.54	-2.75	-1.98		ppb
Cadmium	111-1	0.07	0.10	-0.01	0.05	101.95	ppb
Calcium	43-1	4315.27	3695.94	3723.60	3911.61	8.94	ppb
Calcium	44-1	4410.18	3762.33	3783.94	3985.48	9.23	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	12.03	11.82	11.57	11.81	1.93	ppb
Cobalt	59-2	6.63	6.59	6.48	6.56	1.13	ppb
Copper	63-2	19.78	19.55	19.78	19.71	0.68	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				99		%
Indium	115-1				95		%
Indium	115-2				95		%
Iron	56-2	14633.07	14357.53	14636.25	14542.28	1.10	ppb
Iron	57-2	15299.90	14980.05	15360.88	15213.61	1.34	ppb
Iron	54-2	15354.13	15163.64	15430.68	15316.15	0.90	ppb
Krypton	83-1						cps

LB Number :	LB133015	Operator :	Jaswal
Lab Sample ID :	P2823-01ADLX4	Instrumnet Name :	P7
Client Sample ID :	MYD5X1ARE	Dilution Factor :	4
Date & Time Acquired :	2024-10-18 14:06:11	DataFile Name :	021SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	18.69	15.82	16.75	17.09	8.55	ppb
Lead	207-1	18.95	16.49	16.85	17.43	7.62	ppb
Lead	208-1	18.73	16.11	16.62	17.15	8.10	ppb
Lithium	6-1				99		%
Magnesium	24-2	2883.52	2856.68	2896.23	2878.81	0.70	ppb
Manganese	55-2	327.57	325.60	329.73	327.64	0.63	ppb
Molybdenum	94-1	4.30	3.55	3.64	3.83	10.64	ppb
Molybdenum	95-1	0.22	0.20	0.19	0.20	8.52	ppb
Molybdenum	96-1	0.73	0.66	0.69	0.69	5.08	ppb
Molybdenum	97-1	0.22	0.25	0.18	0.22	16.14	ppb
Molybdenum	98-1	0.20	0.20	0.21	0.20	1.84	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	11.81	11.66	12.16	11.88	2.13	ppb
Phosphorus	31-2	402.86	361.55	383.99	382.80	5.40	ppb
Potassium	39-2	1410.86	1397.77	1416.96	1408.53	0.70	ppb
Rhodium	103-1				93		%
Rhodium	103-2				96		%
Scandium	45-1				94		%
Scandium	45-2				97		%
Selenium	82-1	-0.79	-0.88	-0.15	-0.61		ppb
Selenium	77-2	1.40	3.17	2.58	2.39	37.88	ppb
Selenium	78-2	2.23	0.03	0.76	1.01	111.60	ppb
Silicon	28-1	2144.66	1845.11	1845.45	1945.07	8.89	ppb
Silver	107-1	0.11	0.09	0.11	0.10	9.92	ppb
Silver	109-1	0.12	0.09	0.11	0.10	16.24	ppb
Sodium	23-2	98.69	96.48	97.97	97.72	1.15	ppb
Strontium	86-1	21.19	18.00	18.35	19.18	9.13	ppb
Strontium	88-1	20.78	18.12	18.29	19.06	7.81	ppb
Sulfur	34-1	231.25	-849.76	-900.02	-506.18		ppb
Terbium	159-1				99		%
Terbium	159-2				100		%
Thallium	203-1	0.07	0.06	0.06	0.06	13.91	ppb
Thallium	205-1	0.08	0.06	0.06	0.07	11.67	ppb
Tin	118-1	1.27	1.07	1.05	1.13	10.73	ppb
Titanium	47-1	310.50	264.56	264.92	279.99	9.44	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2823-01ADLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5X1ARE

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:06:11

DataFile Name :

021SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.27	0.24	0.25	0.25	7.56	ppb
Vanadium	51-2	80.47	79.33	80.22	80.01	0.75	ppb
Yttrium	89-1				97		%
Yttrium	89-2				97		%
Zinc	66-2	66.43	65.35	66.04	65.94	0.83	ppb
Zirconium	90-1	2.41	2.12	2.14	2.22	7.26	ppb
Zirconium	91-1	2.35	2.13	2.13	2.20	5.64	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2824-01A Instrumnet Name : P7
Client Sample ID : MYD5K5ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:09:32 DataFile Name : 022SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	41527.31	42364.27	40855.12	41582.23	1.82	ppb
Antimony	121-1	1.24	1.32	1.28	1.28	3.20	ppb
Arsenic	75-2	167.54	168.47	167.16	167.72	0.40	ppb
Barium	135-1	431.61	450.00	446.08	442.56	2.19	ppb
Barium	137-1	440.98	444.42	445.26	443.55	0.51	ppb
Beryllium	9-1	1.23	1.13	1.18	1.18	4.03	ppb
Bismuth	209-1				104		%
Bismuth	209-2				96		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	1.95	1.70	1.04	1.56	30.23	ppb
Cadmium	106-1	1.18	1.85	0.84	1.29	40.24	ppb
Cadmium	111-1	0.81	0.88	0.88	0.86	4.38	ppb
Calcium	43-1	8589.03	8618.68	8691.13	8632.95	0.61	ppb
Calcium	44-1	8243.09	8191.73	8264.87	8233.23	0.46	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	132.52	132.82	131.13	132.16	0.68	ppb
Cobalt	59-2	75.16	75.57	73.94	74.89	1.13	ppb
Copper	63-2	183.62	186.61	181.72	183.98	1.34	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				106		%
Holmium	165-2				98		%
Indium	115-1				99		%
Indium	115-2				93		%
Iron	56-2	163364.44	166089.85	161858.17	163770.82	1.31	ppb
Iron	57-2	164595.50	164759.34	161938.52	163764.45	0.97	ppb
Iron	54-2	165819.20	167386.27	164548.44	165917.97	0.86	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2824-01A Instrumnet Name : P7
Client Sample ID : MYD5K5ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:09:32 DataFile Name : 022SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	50.23	52.11	50.44	50.92	2.02	ppb
Lead	207-1	52.25	52.52	50.94	51.90	1.63	ppb
Lead	208-1	51.05	51.65	50.62	51.11	1.02	ppb
Lithium	6-1				106		%
Magnesium	24-2	24126.35	24197.68	23890.41	24071.48	0.67	ppb
Manganese	55-2	8009.84	8207.84	8077.07	8098.25	1.24	ppb
Molybdenum	94-1	15.36	15.37	15.49	15.41	0.47	ppb
Molybdenum	95-1	1.96	2.00	1.94	1.97	1.49	ppb
Molybdenum	96-1	4.57	4.51	4.22	4.43	4.12	ppb
Molybdenum	97-1	1.87	1.94	1.82	1.88	3.02	ppb
Molybdenum	98-1	1.82	1.92	1.79	1.84	3.84	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	97.42	99.80	97.83	98.35	1.30	ppb
Phosphorus	31-2	1782.07	1776.14	1657.56	1738.59	4.04	ppb
Potassium	39-2	8569.08	8733.44	8475.46	8592.66	1.52	ppb
Rhodium	103-1				98		%
Rhodium	103-2				93		%
Scandium	45-1				110		%
Scandium	45-2				105		%
Selenium	82-1	12.15	12.03	12.13	12.11	0.53	ppb
Selenium	77-2	22.63	20.14	20.03	20.94	7.03	ppb
Selenium	78-2	13.89	18.38	12.91	15.06	19.35	ppb
Silicon	28-1	5701.93	5630.67	5693.03	5675.21	0.68	ppb
Silver	107-1	0.25	0.27	0.28	0.27	5.36	ppb
Silver	109-1	0.20	0.27	0.26	0.25	14.65	ppb
Sodium	23-2	182.87	184.51	185.49	184.29	0.72	ppb
Strontium	86-1	75.21	77.23	76.10	76.18	1.33	ppb
Strontium	88-1	75.85	75.88	75.26	75.66	0.46	ppb
Sulfur	34-1	-1301.00	-1350.35	-1405.13	-1352.16		ppb
Terbium	159-1				105		%
Terbium	159-2				99		%
Thallium	203-1	0.29	0.35	0.33	0.32	9.90	ppb
Thallium	205-1	0.31	0.32	0.32	0.32	2.75	ppb
Tin	118-1	2.28	2.28	2.31	2.29	0.87	ppb
Titanium	47-1	829.16	822.63	818.77	823.52	0.64	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2824-01A

Instrumnet Name :

P7

Client Sample ID :

MYD5K5ARE

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 14:09:32

DataFile Name :

022SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.82	0.85	0.83	0.83	1.52	ppb
Vanadium	51-2	266.74	272.43	265.73	268.30	1.35	ppb
Yttrium	89-1				107		%
Yttrium	89-2				100		%
Zinc	66-2	285.90	289.43	286.38	287.23	0.67	ppb
Zirconium	90-1	8.24	8.43	8.24	8.30	1.29	ppb
Zirconium	91-1	8.46	8.35	8.26	8.36	1.22	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-10DLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:23:04

DataFile Name :

023SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11813.86	11555.91	11789.48	11719.75	1.22	ppb
Antimony	121-1	0.07	0.06	0.06	0.06	6.52	ppb
Arsenic	75-2	7.75	8.19	8.39	8.11	4.03	ppb
Barium	135-1	136.37	135.29	135.43	135.70	0.43	ppb
Barium	137-1	136.24	136.84	135.65	136.24	0.44	ppb
Beryllium	9-1	0.51	0.43	0.58	0.51	14.73	ppb
Bismuth	209-1				102		%
Bismuth	209-2				92		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.60	0.19	0.23	0.34	67.23	ppb
Cadmium	106-1	-2.23	-1.54	-1.82	-1.86		ppb
Cadmium	111-1	-0.04	-0.03	0.00	-0.02		ppb
Calcium	43-1	4480.22	4445.82	4487.29	4471.11	0.50	ppb
Calcium	44-1	4513.19	4511.86	4531.50	4518.85	0.24	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	14.59	14.57	14.80	14.65	0.90	ppb
Cobalt	59-2	8.87	8.61	8.84	8.77	1.64	ppb
Copper	63-2	21.33	20.76	21.14	21.08	1.38	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				96		%
Indium	115-1				98		%
Indium	115-2				91		%
Iron	56-2	17833.29	17487.02	17967.03	17762.45	1.39	ppb
Iron	57-2	18783.66	18376.86	18732.86	18631.13	1.19	ppb
Iron	54-2	18902.46	18575.86	18981.27	18819.86	1.14	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3 Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:23:04 DataFile Name : 023SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	10.05	10.43	10.45	10.31	2.21	ppb
Lead	207-1	10.06	10.13	10.08	10.09	0.35	ppb
Lead	208-1	9.90	10.18	10.11	10.07	1.46	ppb
Lithium	6-1				102		%
Magnesium	24-2	4549.64	4432.06	4571.75	4517.81	1.66	ppb
Manganese	55-2	498.38	495.24	506.23	499.95	1.13	ppb
Molybdenum	94-1	3.79	3.72	3.73	3.75	1.07	ppb
Molybdenum	95-1	0.15	0.13	0.16	0.15	12.21	ppb
Molybdenum	96-1	0.65	0.64	0.63	0.64	1.34	ppb
Molybdenum	97-1	0.15	0.17	0.15	0.16	9.46	ppb
Molybdenum	98-1	0.13	0.15	0.14	0.14	6.92	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	13.43	12.83	12.84	13.04	2.65	ppb
Phosphorus	31-2	581.26	527.64	574.69	561.20	5.21	ppb
Potassium	39-2	1458.46	1432.22	1472.31	1454.33	1.40	ppb
Rhodium	103-1				96		%
Rhodium	103-2				92		%
Scandium	45-1				97		%
Scandium	45-2				94		%
Selenium	82-1	-0.09	-0.26	-0.04	-0.13		ppb
Selenium	77-2	2.69	5.72	2.69	3.70	47.29	ppb
Selenium	78-2	0.61	-1.11	-1.26	-0.59		ppb
Silicon	28-1	2119.35	2158.56	2171.54	2149.82	1.26	ppb
Silver	107-1	0.04	0.05	0.05	0.05	12.67	ppb
Silver	109-1	0.04	0.05	0.05	0.05	16.68	ppb
Sodium	23-2	60.55	56.15	59.67	58.79	3.96	ppb
Strontium	86-1	25.12	25.23	25.36	25.24	0.47	ppb
Strontium	88-1	24.97	25.21	25.30	25.16	0.67	ppb
Sulfur	34-1	-900.16	-1073.26	-1173.05	-1048.83		ppb
Terbium	159-1				101		%
Terbium	159-2				95		%
Thallium	203-1	0.09	0.07	0.07	0.08	18.83	ppb
Thallium	205-1	0.07	0.08	0.08	0.08	3.02	ppb
Tin	118-1	0.67	0.64	0.64	0.65	2.56	ppb
Titanium	47-1	103.98	107.02	107.42	106.14	1.77	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-10DLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:23:04

DataFile Name :

023SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.21	0.22	0.21	0.21	2.19	ppb
Vanadium	51-2	29.69	29.57	29.97	29.74	0.69	ppb
Yttrium	89-1				101		%
Yttrium	89-2				94		%
Zinc	66-2	48.16	47.64	47.21	47.67	1.00	ppb
Zirconium	90-1	2.15	2.18	2.17	2.16	0.73	ppb
Zirconium	91-1	2.17	2.13	2.21	2.17	1.87	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-11DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3D Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:26:25 DataFile Name : 024SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	11860.65	11828.31	11730.92	11806.63	0.57	ppb
Antimony	121-1	0.07	0.07	0.06	0.07	3.42	ppb
Arsenic	75-2	8.05	8.18	8.13	8.12	0.81	ppb
Barium	135-1	135.56	135.11	134.41	135.03	0.43	ppb
Barium	137-1	136.32	134.74	135.18	135.41	0.60	ppb
Beryllium	9-1	0.58	0.66	0.56	0.60	9.43	ppb
Bismuth	209-1				101		%
Bismuth	209-2				96		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.19	0.09	0.51	0.26	83.36	ppb
Cadmium	106-1	0.02	-0.74	-1.61	-0.77		ppb
Cadmium	111-1	0.13	0.06	-0.05	0.05	201.52	ppb
Calcium	43-1	4552.63	4534.52	4620.86	4569.34	1.00	ppb
Calcium	44-1	4609.81	4586.14	4647.29	4614.41	0.67	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	16.71	17.32	16.45	16.83	2.66	ppb
Cobalt	59-2	8.83	8.81	8.69	8.78	0.90	ppb
Copper	63-2	21.03	21.32	21.05	21.13	0.77	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				96		%
Indium	115-1				96		%
Indium	115-2				91		%
Iron	56-2	18027.93	17976.41	17621.38	17875.24	1.24	ppb
Iron	57-2	18887.60	18632.76	18554.45	18691.60	0.93	ppb
Iron	54-2	19054.64	18775.75	18782.78	18871.06	0.84	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-11DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3D Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:26:25 DataFile Name : 024SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	10.46	10.42	10.84	10.58	2.20	ppb
Lead	207-1	10.21	10.02	10.60	10.28	2.91	ppb
Lead	208-1	10.24	10.10	10.53	10.29	2.11	ppb
Lithium	6-1				101		%
Magnesium	24-2	4609.65	4574.45	4493.08	4559.06	1.31	ppb
Manganese	55-2	498.49	497.42	495.42	497.11	0.31	ppb
Molybdenum	94-1	3.90	4.00	3.96	3.95	1.33	ppb
Molybdenum	95-1	0.14	0.16	0.15	0.15	8.02	ppb
Molybdenum	96-1	0.71	0.72	0.74	0.73	2.39	ppb
Molybdenum	97-1	0.17	0.18	0.23	0.19	14.80	ppb
Molybdenum	98-1	0.17	0.17	0.17	0.17	1.47	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	15.75	13.91	14.47	14.71	6.43	ppb
Phosphorus	31-2	528.13	551.63	536.92	538.89	2.20	ppb
Potassium	39-2	1479.67	1475.20	1443.57	1466.15	1.34	ppb
Rhodium	103-1				95		%
Rhodium	103-2				91		%
Scandium	45-1				95		%
Scandium	45-2				94		%
Selenium	82-1	0.40	0.26	-0.50	0.05	905.77	ppb
Selenium	77-2	4.52	5.80	3.32	4.55	27.27	ppb
Selenium	78-2	-1.84	1.21	0.45	-0.06		ppb
Silicon	28-1	2186.52	2187.49	2195.48	2189.83	0.22	ppb
Silver	107-1	0.05	0.04	0.04	0.04	11.14	ppb
Silver	109-1	0.05	0.04	0.05	0.05	10.34	ppb
Sodium	23-2	65.69	63.33	66.19	65.07	2.35	ppb
Strontium	86-1	25.85	26.34	26.35	26.18	1.10	ppb
Strontium	88-1	25.79	26.22	26.19	26.07	0.92	ppb
Sulfur	34-1	-1402.87	-1507.61	-1388.49	-1432.99		ppb
Terbium	159-1				101		%
Terbium	159-2				96		%
Thallium	203-1	0.06	0.07	0.08	0.07	11.24	ppb
Thallium	205-1	0.07	0.08	0.08	0.07	6.36	ppb
Tin	118-1	0.63	0.62	0.59	0.61	2.82	ppb
Titanium	47-1	106.19	105.93	106.93	106.35	0.49	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-11DLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3D

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:26:25

DataFile Name :

024SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.21	0.21	0.22	0.21	2.19	ppb
Vanadium	51-2	30.55	30.07	29.93	30.18	1.07	ppb
Yttrium	89-1				98		%
Yttrium	89-2				94		%
Zinc	66-2	48.00	47.41	47.02	47.48	1.05	ppb
Zirconium	90-1	2.14	2.27	2.27	2.23	3.33	ppb
Zirconium	91-1	2.22	2.14	2.21	2.19	1.94	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10LDLX20 Instrumnet Name : P7
Client Sample ID : MYD5Y3L Dilution Factor : 20
Date & Time Acquired : 2024-10-18 14:29:50 DataFile Name : 025SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	2290.57	2275.59	2288.75	2284.97	0.36	ppb
Antimony	121-1	0.02	0.01	0.01	0.01	9.11	ppb
Arsenic	75-2	1.87	1.62	1.73	1.74	7.29	ppb
Barium	135-1	25.88	26.26	26.98	26.38	2.12	ppb
Barium	137-1	26.13	26.75	26.78	26.55	1.39	ppb
Beryllium	9-1	0.03	0.05	0.16	0.08	81.46	ppb
Bismuth	209-1				105		%
Bismuth	209-2				100		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	0.00	0.05	-0.05	0.00	9481.12	ppb
Cadmium	106-1	-2.97	-0.64	-1.72	-1.77		ppb
Cadmium	111-1	-0.19	-0.02	-0.10	-0.10		ppb
Calcium	43-1	867.30	863.69	878.12	869.70	0.86	ppb
Calcium	44-1	878.93	874.68	887.20	880.27	0.72	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	2.92	2.93	2.95	2.94	0.52	ppb
Cobalt	59-2	1.79	1.72	1.65	1.72	3.96	ppb
Copper	63-2	4.18	4.00	4.10	4.09	2.18	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				101		%
Holmium	165-2				98		%
Indium	115-1				98		%
Indium	115-2				94		%
Iron	56-2	3545.92	3502.57	3499.52	3516.01	0.74	ppb
Iron	57-2	3705.94	3687.77	3666.13	3686.61	0.54	ppb
Iron	54-2	3684.40	3651.31	3725.66	3687.12	1.01	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10LDLX20 Instrumnet Name : P7
Client Sample ID : MYD5Y3L Dilution Factor : 20
Date & Time Acquired : 2024-10-18 14:29:50 DataFile Name : 025SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	1.90	1.95	1.99	1.95	2.48	ppb
Lead	207-1	1.84	1.85	1.91	1.87	1.99	ppb
Lead	208-1	1.84	1.86	1.92	1.88	2.26	ppb
Lithium	6-1				101		%
Magnesium	24-2	881.77	879.88	890.26	883.97	0.63	ppb
Manganese	55-2	99.47	96.89	98.32	98.23	1.32	ppb
Molybdenum	94-1	0.65	0.81	0.78	0.75	11.27	ppb
Molybdenum	95-1	0.02	0.02	0.03	0.02	38.42	ppb
Molybdenum	96-1	0.13	0.14	0.14	0.13	5.20	ppb
Molybdenum	97-1	0.04	0.03	0.03	0.03	22.05	ppb
Molybdenum	98-1	0.01	0.02	0.04	0.02	50.54	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	2.74	2.94	2.79	2.82	3.58	ppb
Phosphorus	31-2	101.26	94.24	97.77	97.76	3.59	ppb
Potassium	39-2	281.55	271.40	274.54	275.83	1.88	ppb
Rhodium	103-1				95		%
Rhodium	103-2				95		%
Scandium	45-1				94		%
Scandium	45-2				94		%
Selenium	82-1	-0.23	-0.88	-0.47	-0.53		ppb
Selenium	77-2	-0.39	0.21	1.48	0.43	219.53	ppb
Selenium	78-2	0.98	-0.61	-1.02	-0.22		ppb
Silicon	28-1	412.45	415.39	420.06	415.96	0.92	ppb
Silver	107-1	0.01	0.01	0.02	0.01	33.94	ppb
Silver	109-1	0.01	0.01	0.01	0.01	23.66	ppb
Sodium	23-2	22.88	21.82	22.33	22.34	2.36	ppb
Strontium	86-1	4.90	5.09	5.19	5.06	2.87	ppb
Strontium	88-1	5.06	4.95	5.08	5.03	1.34	ppb
Sulfur	34-1	-1676.98	-1712.21	-1649.96	-1679.72		ppb
Terbium	159-1				101		%
Terbium	159-2				97		%
Thallium	203-1	0.01	0.01	0.02	0.02	31.20	ppb
Thallium	205-1	0.01	0.01	0.02	0.02	29.06	ppb
Tin	118-1	0.14	0.14	0.16	0.15	8.50	ppb
Titanium	47-1	19.95	20.81	20.85	20.54	2.47	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-10LDLX20

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3L

Dilution Factor :

20

Date & Time Acquired :

2024-10-18 14:29:50

DataFile Name :

025SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.04	0.04	0.04	0.04	4.86	ppb
Vanadium	51-2	5.69	5.66	5.92	5.75	2.51	ppb
Yttrium	89-1				97		%
Yttrium	89-2				94		%
Zinc	66-2	9.61	9.51	9.79	9.64	1.48	ppb
Zirconium	90-1	0.43	0.46	0.44	0.44	2.93	ppb
Zirconium	91-1	0.45	0.47	0.43	0.45	4.23	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-12DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3S Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:33:11 DataFile Name : 026SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	12178.48	12101.23	12231.99	12170.57	0.54	ppb
Antimony	121-1	15.81	16.10	15.81	15.91	1.04	ppb
Arsenic	75-2	15.22	15.80	15.05	15.36	2.55	ppb
Barium	135-1	437.18	445.70	441.38	441.42	0.97	ppb
Barium	137-1	436.97	444.51	441.58	441.02	0.86	ppb
Beryllium	9-1	8.69	8.72	8.90	8.77	1.32	ppb
Bismuth	209-1				104		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	8.60	7.41	7.75	7.92	7.79	ppb
Cadmium	106-1	5.93	7.07	6.59	6.53	8.81	ppb
Cadmium	111-1	8.16	8.53	8.46	8.38	2.32	ppb
Calcium	43-1	4767.43	4672.25	4883.54	4774.41	2.22	ppb
Calcium	44-1	4893.42	4844.59	4891.03	4876.34	0.56	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	48.82	49.50	49.31	49.21	0.71	ppb
Cobalt	59-2	91.59	92.01	91.89	91.83	0.24	ppb
Copper	63-2	60.92	60.82	61.20	60.98	0.32	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				102		%
Holmium	165-2				97		%
Indium	115-1				98		%
Indium	115-2				92		%
Iron	56-2	18974.74	19108.82	19049.01	19044.19	0.35	ppb
Iron	57-2	19629.84	19711.78	19869.08	19736.90	0.62	ppb
Iron	54-2	19943.08	19934.71	19957.34	19945.04	0.06	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-12DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3S Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:33:11 DataFile Name : 026SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	13.94	14.17	14.35	14.15	1.45	ppb
Lead	207-1	13.62	13.68	13.68	13.66	0.25	ppb
Lead	208-1	13.61	13.67	13.81	13.70	0.74	ppb
Lithium	6-1				100		%
Magnesium	24-2	4715.02	4739.32	4783.98	4746.11	0.74	ppb
Manganese	55-2	622.26	624.05	621.19	622.50	0.23	ppb
Molybdenum	94-1	3.87	3.49	3.98	3.78	6.79	ppb
Molybdenum	95-1	0.16	0.20	0.19	0.19	10.44	ppb
Molybdenum	96-1	0.69	0.70	0.75	0.71	4.19	ppb
Molybdenum	97-1	0.21	0.19	0.18	0.19	9.00	ppb
Molybdenum	98-1	0.19	0.17	0.21	0.19	8.94	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	98.97	99.69	99.98	99.55	0.52	ppb
Phosphorus	31-2	617.32	595.85	586.28	599.82	2.65	ppb
Potassium	39-2	1533.80	1526.54	1551.60	1537.31	0.84	ppb
Rhodium	103-1				95		%
Rhodium	103-2				93		%
Scandium	45-1				95		%
Scandium	45-2				94		%
Selenium	82-1	17.01	17.76	16.65	17.14	3.31	ppb
Selenium	77-2	24.30	23.22	15.27	20.93	23.58	ppb
Selenium	78-2	15.90	20.08	17.73	17.90	11.69	ppb
Silicon	28-1	2180.66	2158.81	2166.08	2168.52	0.51	ppb
Silver	107-1	7.86	8.08	7.90	7.95	1.51	ppb
Silver	109-1	7.84	8.00	7.95	7.93	1.05	ppb
Sodium	23-2	60.48	62.42	66.11	63.00	4.53	ppb
Strontium	86-1	26.80	26.72	27.18	26.90	0.92	ppb
Strontium	88-1	26.87	26.95	27.45	27.09	1.15	ppb
Sulfur	34-1	-1659.30	-1916.75	-1746.52	-1774.19		ppb
Terbium	159-1				103		%
Terbium	159-2				96		%
Thallium	203-1	7.02	7.09	7.21	7.11	1.37	ppb
Thallium	205-1	7.05	7.09	7.12	7.08	0.49	ppb
Tin	118-1	0.70	0.77	0.68	0.72	6.73	ppb
Titanium	47-1	114.47	113.08	116.62	114.72	1.56	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-12DLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3S

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:33:11

DataFile Name :

026SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.23	0.23	0.22	0.23	2.14	ppb
Vanadium	51-2	112.29	112.51	111.90	112.24	0.27	ppb
Yttrium	89-1				100		%
Yttrium	89-2				96		%
Zinc	66-2	131.80	131.97	134.15	132.64	0.99	ppb
Zirconium	90-1	2.13	2.12	2.21	2.15	2.25	ppb
Zirconium	91-1	2.20	2.15	2.14	2.17	1.58	ppb

LB Number :LB133015Operator :Jaswal

Lab Sample ID :P2827-10MSDLX4Instrumnet Name :P7

Client Sample ID :MYD5Y3SREDilution Factor :4

Date & Time Acquired :2024-10-18 14:36:40DataFile Name :027SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	12800.76	12761.38	12879.26	12813.80	0.47	ppb
Antimony	121-1	78.28	79.68	78.15	78.71	1.08	ppb
Arsenic	75-2	37.66	37.32	38.74	37.91	1.96	ppb
Barium	135-1	1749.26	1746.53	1691.37	1729.05	1.89	ppb
Barium	137-1	1721.84	1706.94	1726.22	1718.34	0.59	ppb
Beryllium	9-1	40.95	41.42	39.61	40.66	2.31	ppb
Bismuth	209-1				102		%
Bismuth	209-2				98		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	37.08	38.69	35.89	37.22	3.77	ppb
Cadmium	106-1	32.97	33.19	32.58	32.92	0.94	ppb
Cadmium	111-1	40.95	40.83	40.67	40.82	0.35	ppb
Calcium	43-1	4723.74	4756.12	4687.11	4722.32	0.73	ppb
Calcium	44-1	4768.33	4882.90	4751.74	4800.99	1.49	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	171.87	172.44	174.30	172.87	0.74	ppb
Cobalt	59-2	401.52	399.77	404.96	402.09	0.66	ppb
Copper	63-2	208.42	207.58	210.18	208.73	0.64	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				100		%
Holmium	165-2				96		%
Indium	115-1				97		%
Indium	115-2				92		%
Iron	56-2	18728.33	18626.53	18958.90	18771.25	0.91	ppb
Iron	57-2	19689.57	19683.41	19956.17	19776.38	0.79	ppb
Iron	54-2	19947.26	19912.92	20090.24	19983.47	0.47	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10MSDLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3SRE Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:36:40 DataFile Name : 027SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	25.10	25.27	24.47	24.95	1.68	ppb
Lead	207-1	24.59	24.68	23.85	24.37	1.87	ppb
Lead	208-1	24.66	24.94	23.88	24.49	2.25	ppb
Lithium	6-1				97		%
Magnesium	24-2	4533.49	4516.35	4574.94	4541.60	0.66	ppb
Manganese	55-2	905.52	901.84	911.96	906.44	0.57	ppb
Molybdenum	94-1	3.61	3.87	3.74	3.74	3.55	ppb
Molybdenum	95-1	0.20	0.19	0.19	0.20	2.04	ppb
Molybdenum	96-1	0.69	0.67	0.67	0.67	1.53	ppb
Molybdenum	97-1	0.18	0.21	0.15	0.18	16.17	ppb
Molybdenum	98-1	0.18	0.21	0.18	0.19	8.73	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	420.45	415.59	421.00	419.01	0.71	ppb
Phosphorus	31-2	560.37	562.12	590.40	570.96	2.95	ppb
Potassium	39-2	1457.01	1461.27	1480.96	1466.41	0.87	ppb
Rhodium	103-1				92		%
Rhodium	103-2				93		%
Scandium	45-1				92		%
Scandium	45-2				93		%
Selenium	82-1	88.35	90.68	87.96	89.00	1.65	ppb
Selenium	77-2	87.97	84.95	90.32	87.75	3.07	ppb
Selenium	78-2	87.81	82.65	79.64	83.37	4.96	ppb
Silicon	28-1	2105.72	2175.69	2096.27	2125.89	2.04	ppb
Silver	107-1	38.85	39.89	39.10	39.28	1.38	ppb
Silver	109-1	39.69	40.30	39.69	39.89	0.89	ppb
Sodium	23-2	63.13	62.54	71.29	65.65	7.45	ppb
Strontium	86-1	26.85	26.79	25.81	26.49	2.22	ppb
Strontium	88-1	26.56	26.84	26.31	26.57	1.01	ppb
Sulfur	34-1	-1654.65	-1674.21	-1875.58	-1734.81		ppb
Terbium	159-1				99		%
Terbium	159-2				97		%
Thallium	203-1	34.85	35.00	33.71	34.52	2.04	ppb
Thallium	205-1	34.46	34.75	33.41	34.21	2.05	ppb
Tin	118-1	0.76	0.77	0.74	0.76	1.79	ppb
Titanium	47-1	111.61	115.85	113.29	113.58	1.88	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-10MSDLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3SRE

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:36:40

DataFile Name :

027SMPL.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.21	0.22	0.21	0.22	3.10	ppb
Vanadium	51-2	410.69	411.07	416.03	412.59	0.72	ppb
Yttrium	89-1				97		%
Yttrium	89-2				96		%
Zinc	66-2	434.86	435.85	449.15	439.95	1.81	ppb
Zirconium	90-1	2.07	2.17	2.04	2.09	3.25	ppb
Zirconium	91-1	2.10	2.23	2.07	2.13	4.16	ppb

LB Number :LB133015Operator :Jaswal

Lab Sample ID :CCV095Instrumnet Name :P7

Client Sample ID :CCV095Dilution Factor :1

Date & Time Acquired :2024-10-18 14:44:37DataFile Name :028CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	47392.72	48682.61	47776.55	47950.62	1.38	ppb
Antimony	121-1	503.86	502.17	512.42	506.15	1.09	ppb
Arsenic	75-2	502.17	508.25	505.83	505.42	0.61	ppb
Barium	135-1	2570.00	2554.40	2590.30	2571.57	0.70	ppb
Barium	137-1	2593.21	2533.77	2570.39	2565.79	1.17	ppb
Beryllium	9-1	520.68	521.21	525.02	522.30	0.45	ppb
Bismuth	209-1				93		%
Bismuth	209-2				79		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	485.79	494.44	497.05	492.43	1.20	ppb
Cadmium	106-1	504.80	494.37	501.07	500.08	1.06	ppb
Cadmium	111-1	488.96	487.07	495.93	490.65	0.95	ppb
Calcium	43-1	244242.92	255160.73	252666.76	250690.14	2.28	ppb
Calcium	44-1	245921.34	253937.95	250533.52	250130.94	1.61	ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	509.78	514.12	511.22	511.71	0.43	ppb
Cobalt	59-2	538.41	543.35	541.28	541.01	0.46	ppb
Copper	63-2	5249.11	5259.57	5231.25	5246.64	0.27	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				97		%
Holmium	165-2				94		%
Indium	115-1				87		%
Indium	115-2				83		%
Iron	56-2	125863.99	127571.13	126607.14	126680.76	0.68	ppb
Iron	57-2	127341.44	126755.31	125511.87	126536.21	0.74	ppb
Iron	54-2	127121.59	129924.77	127692.58	128246.31	1.16	ppb
Krypton	83-1						cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCV095

Instrumnet Name :

P7

Client Sample ID :

CCV095

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 14:44:37

DataFile Name :

028CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	2570.78	2644.66	2603.72	2606.39	1.42	ppb
Lead	207-1	2557.33	2664.35	2604.61	2608.76	2.06	ppb
Lead	208-1	2553.14	2643.98	2589.75	2595.62	1.76	ppb
Lithium	6-1				89		%
Magnesium	24-2	229989.47	230000.38	231478.10	230489.32	0.37	ppb
Manganese	55-2	5105.81	5197.62	5181.49	5161.64	0.95	ppb
Molybdenum	94-1	5203.82	5227.35	5208.07	5213.08	0.24	ppb
Molybdenum	95-1	5112.47	5237.34	5151.58	5167.13	1.24	ppb
Molybdenum	96-1	5160.57	5259.32	5188.29	5202.73	0.98	ppb
Molybdenum	97-1	5108.50	5235.52	5213.69	5185.90	1.31	ppb
Molybdenum	98-1	5139.18	5289.02	5200.46	5209.55	1.45	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	537.22	541.74	538.75	539.24	0.43	ppb
Phosphorus	31-2	9982.25	10340.86	10042.90	10122.00	1.90	ppb
Potassium	39-2	121483.67	122730.97	122040.93	122085.19	0.51	ppb
Rhodium	103-1				81		%
Rhodium	103-2				82		%
Scandium	45-1				85		%
Scandium	45-2				87		%
Selenium	82-1	495.69	493.64	496.22	495.19	0.28	ppb
Selenium	77-2	485.69	502.68	534.08	507.48	4.84	ppb
Selenium	78-2	477.94	505.53	481.29	488.25	3.08	ppb
Silicon	28-1	780.06	821.68	813.32	805.02	2.74	ppb
Silver	107-1	492.35	484.95	491.81	489.70	0.84	ppb
Silver	109-1	489.79	489.47	494.30	491.19	0.55	ppb
Sodium	23-2	242913.88	246295.97	246725.66	245311.83	0.85	ppb
Strontium	86-1	12877.29	13139.83	12899.57	12972.23	1.12	ppb
Strontium	88-1	12936.77	13005.43	12926.38	12956.19	0.33	ppb
Sulfur	34-1	10077.81	10645.09	10572.00	10431.63	2.96	ppb
Terbium	159-1				96		%
Terbium	159-2				93		%
Thallium	203-1	505.56	530.24	520.84	518.88	2.40	ppb
Thallium	205-1	509.01	535.60	523.12	522.58	2.55	ppb
Tin	118-1	522.37	517.68	527.36	522.47	0.93	ppb
Titanium	47-1	4959.64	5192.86	5149.92	5100.81	2.43	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCV095

Instrumnet Name :

P7

Client Sample ID :

CCV095

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 14:44:37

DataFile Name :

028CCV.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	521.75	542.40	525.93	530.02	2.06	ppb
Vanadium	51-2	510.14	520.96	516.56	515.89	1.06	ppb
Yttrium	89-1				89		%
Yttrium	89-2				87		%
Zinc	66-2	5227.96	5342.73	5283.07	5284.58	1.09	ppb
Zirconium	90-1	522.92	524.59	526.59	524.70	0.35	ppb
Zirconium	91-1	513.40	520.72	521.22	518.45	0.84	ppb

LB Number :LB133015Operator :Jaswal

Lab Sample ID :CCB095Instrumnet Name :P7

Client Sample ID :CCB095Dilution Factor :1

Date & Time Acquired :2024-10-18 14:50:48DataFile Name :030CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Aluminium	27-2	0.28	0.61	0.49	0.46	36.69	ppb
Antimony	121-1	0.10	0.11	0.12	0.11	9.81	ppb
Arsenic	75-2	0.05	0.01	-0.01	0.02	150.09	ppb
Barium	135-1	0.03	0.02	0.01	0.02	35.31	ppb
Barium	137-1	0.03	-0.01	0.01	0.01	141.29	ppb
Beryllium	9-1	0.11	0.18	0.13	0.14	27.54	ppb
Bismuth	209-1				109		%
Bismuth	209-2				104		%
Bromine	81-1						cps
Bromine	81-2						cps
Cadmium	108-1	-0.09	0.00	-0.09	-0.06		ppb
Cadmium	106-1	-1.09	-0.03	-2.02	-1.05		ppb
Cadmium	111-1	-0.05	0.01	-0.12	-0.05		ppb
Calcium	43-1	-1.16	-6.19	-4.56	-3.97		ppb
Calcium	44-1	-6.24	-5.68	-5.61	-5.85		ppb
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	0.03	-0.04	-0.03	-0.01		ppb
Cobalt	59-2	0.01	0.02	0.01	0.01	66.73	ppb
Copper	63-2	0.06	0.02	0.03	0.04	50.22	ppb
Dysprosium	156-1						cps
Dysprosium	156-2						cps
Erbium	164-1						cps
Erbium	164-2						cps
Gadolinium	160-1						cps
Gadolinium	160-2						cps
Holmium	165-1				105		%
Holmium	165-2				101		%
Indium	115-1				101		%
Indium	115-2				94		%
Iron	56-2	2.67	2.72	2.60	2.66	2.27	ppb
Iron	57-2	-0.25	2.13	0.59	0.82	146.50	ppb
Iron	54-2	3.15	1.94	1.11	2.07	49.64	ppb
Krypton	83-1						cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCB095 Instrumnet Name : P7
Client Sample ID : CCB095 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:50:48 DataFile Name : 030CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Lead	206-1	0.02	0.00	0.00	0.01	168.85	ppb
Lead	207-1	0.00	-0.01	0.03	0.01	244.53	ppb
Lead	208-1	0.01	0.00	0.01	0.01	117.35	ppb
Lithium	6-1				104		%
Magnesium	24-2	3.24	2.90	1.94	2.69	25.03	ppb
Manganese	55-2	0.07	0.07	0.10	0.08	20.78	ppb
Molybdenum	94-1	0.22	0.16	0.14	0.17	23.78	ppb
Molybdenum	95-1	0.10	0.08	0.12	0.10	16.67	ppb
Molybdenum	96-1	0.09	0.09	0.10	0.10	8.27	ppb
Molybdenum	97-1	0.10	0.11	0.09	0.10	8.39	ppb
Molybdenum	98-1	0.08	0.09	0.08	0.08	6.38	ppb
Neodymium	150-1						cps
Neodymium	150-2						cps
Nickel	60-2	0.04	-0.02	-0.02	0.00	1408.62	ppb
Phosphorus	31-2	-7.35	-4.11	-8.31	-6.59		ppb
Potassium	39-2	-9.58	-4.83	-13.48	-9.29		ppb
Rhodium	103-1				98		%
Rhodium	103-2				96		%
Scandium	45-1				96		%
Scandium	45-2				95		%
Selenium	82-1	-0.15	-0.77	-0.32	-0.41		ppb
Selenium	77-2	-0.39	-0.39	-0.39	-0.39		ppb
Selenium	78-2	-0.36	-1.06	-0.54	-0.66		ppb
Silicon	28-1	-5.65	-6.81	-4.35	-5.61		ppb
Silver	107-1	0.01	0.02	0.01	0.01	58.09	ppb
Silver	109-1	0.02	0.02	0.02	0.02	6.28	ppb
Sodium	23-2	27.98	27.62	25.34	26.98	5.31	ppb
Strontium	86-1	0.12	0.12	0.16	0.14	15.87	ppb
Strontium	88-1	0.14	0.13	0.12	0.13	7.40	ppb
Sulfur	34-1	-109.64	-338.75	-151.65	-200.02		ppb
Terbium	159-1				104		%
Terbium	159-2				99		%
Thallium	203-1	0.04	0.03	0.04	0.03	15.94	ppb
Thallium	205-1	0.04	0.03	0.04	0.03	8.22	ppb
Tin	118-1	0.01	0.01	0.01	0.01	36.96	ppb
Titanium	47-1	0.07	0.11	0.04	0.08	45.05	ppb

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCB095

Instrumnet Name :

P7

Client Sample ID :

CCB095

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 14:50:48

DataFile Name :

030CCBE.d

Parameter	Mass	ConRep1	ConRep2	ConRep3	Avg. Conc.	ConcRSD	Units
Uranium	238-1	0.00	0.00	0.00	0.00	8.96	ppb
Vanadium	51-2	0.01	0.01	0.00	0.00	33.24	ppb
Yttrium	89-1				98		%
Yttrium	89-2				94		%
Zinc	66-2	0.12	0.09	0.10	0.10	12.89	ppb
Zirconium	90-1	0.07	0.06	0.05	0.06	13.71	ppb
Zirconium	91-1	0.05	0.05	0.06	0.05	3.81	ppb

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:46:13 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	127	107	100	111	12.49	cps
Antimony	121-1	3	13	17	11	62.48	cps
Arsenic	75-2	0	3	0	1	173.21	cps
Barium	135-1	27	17	40	28	42.13	cps
Barium	137-1	33	57	43	44	26.35	cps
Beryllium	9-1	50	17	40	36	48.10	cps
Bismuth	209-1	2384120	2406920	2171836	2320959	5.59	cps
Bismuth	209-2	1965873	1896367	1866633	1909624	2.67	cps
Bromine	81-1	7909	7886	7802	7866	0.71	cps
Bromine	81-2	20	47	37	34	39.11	cps
Cadmium	108-1	10	7	3	7	50.03	cps
Cadmium	106-1	1453	1523	1290	1422	8.42	cps
Cadmium	111-1	1010	1062	904	992	8.11	cps
Calcium	43-1	187	193	193	191	2.01	cps
Calcium	44-1	4397	4341	4124	4287	3.37	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	330	210	300	280	22.31	cps
Cobalt	59-2	17	37	27	27	37.50	cps
Copper	63-2	850	710	753	771	9.30	cps
Dysprosium	156-1	3	3	17	8	99.04	cps
Dysprosium	156-2	3	0	0	1	173.21	cps
Erbium	164-1	17	20	13	17	20.01	cps
Erbium	164-2	3	13	10	9	57.30	cps
Gadolinium	160-1	13	10	10	11	17.30	cps
Gadolinium	160-2	107	83	127	106	20.55	cps
Holmium	165-1	3995967	3976160	3576123	3849417	6.15	cps
Holmium	165-2	2791643	2689999	2616645	2699429	3.26	cps
Indium	115-1	3017088	2985096	2762805	2921663	4.74	cps
Indium	115-2	998374	962417	962099	974297	2.14	cps
Iron	56-2	5358	5408	5301	5355	1.00	cps
Iron	57-2	653	550	683	629	11.12	cps
Iron	54-2	823	887	920	877	5.60	cps
Krypton	83-1	57	57	77	63	18.23	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S0 Instrumnet Name : P7
Client Sample ID : S0 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:46:13 DataFile Name : 004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	330	410	447	396	15.08	cps
Lead	207-1	360	350	360	357	1.62	cps
Lead	208-1	1623	1720	1697	1680	3.00	cps
Lithium	6-1	707392	707247	654240	689626	4.44	cps
Magnesium	24-2	1380	1507	1490	1459	4.72	cps
Manganese	55-2	93	110	73	92	19.91	cps
Molybdenum	94-1	83	93	83	87	6.66	cps
Molybdenum	95-1	50	40	63	51	22.90	cps
Molybdenum	96-1	53	53	47	51	7.52	cps
Molybdenum	97-1	33	33	13	27	43.31	cps
Molybdenum	98-1	60	43	60	54	17.68	cps
Neodymium	150-1	0	0	10	3	173.21	cps
Neodymium	150-2	0	0	0	0	0.00	cps
Nickel	60-2	53	80	70	68	19.88	cps
Phosphorus	31-2	90	93	80	88	7.90	cps
Potassium	39-2	34335	35658	34709	34901	1.95	cps
Rhodium	103-1	2684104	2730177	2507338	2640540	4.45	cps
Rhodium	103-2	1522008	1482160	1465622	1489930	1.95	cps
Scandium	45-1	1581840	1600426	1430014	1537427	6.08	cps
Scandium	45-2	138813	135839	132772	135808	2.22	cps
Selenium	82-1	148	117	96	121	21.72	cps
Selenium	77-2	0	3	3	2	86.60	cps
Selenium	78-2	110	83	133	109	22.98	cps
Silicon	28-1	184384	186068	186922	185791	0.70	cps
Silver	107-1	17	27	47	30	50.91	cps
Silver	109-1	20	7	10	12	56.76	cps
Sodium	23-2	7308	7469	7389	7388	1.08	cps
Strontium	86-1	133	130	147	137	6.45	cps
Strontium	88-1	167	190	147	168	12.93	cps
Sulfur	34-1	117581	117749	116537	117289	0.56	cps
Terbium	159-1	4103882	4086232	3721076	3970397	5.44	cps
Terbium	159-2	2723491	2623936	2579603	2642343	2.79	cps
Thallium	203-1	37	30	30	32	11.95	cps
Thallium	205-1	60	40	97	66	43.84	cps
Tin	118-1	740	717	783	747	4.53	cps
Titanium	47-1	33	47	20	33	40.01	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S00

Instrumnet Name :

P7

Client Sample ID :

S0

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:46:13

DataFile Name :

004CALB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	10	13	7	10	33.30	cps
Vanadium	51-2	7	7	7	7	0.00	cps
Yttrium	89-1	4904375	4916846	4534450	4785224	4.54	cps
Yttrium	89-2	1279141	1237908	1238583	1251877	1.89	cps
Zinc	66-2	83	73	100	86	15.75	cps
Zirconium	90-1	220	170	140	177	22.88	cps
Zirconium	91-1	37	30	27	31	16.37	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P7
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:49:35 DataFile Name : 005CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1840	2003	1677	1840	8.88	cps
Antimony	121-1	8206	8372	8439	8339	1.44	cps
Arsenic	75-2	177	257	203	212	19.19	cps
Barium	135-1	11061	11591	11014	11222	2.86	cps
Barium	137-1	18662	19090	18989	18914	1.18	cps
Beryllium	9-1	830	927	803	853	7.60	cps
Bismuth	209-1	2376081	2383868	2375219	2378389	0.20	cps
Bismuth	209-2	1921286	1868179	1912971	1900812	1.50	cps
Bromine	81-1	7986	7652	7809	7815	2.14	cps
Bromine	81-2	37	47	47	43	13.32	cps
Cadmium	108-1	77	80	87	81	6.28	cps
Cadmium	106-1	1767	1467	1583	1606	9.42	cps
Cadmium	111-1	2256	2187	2071	2172	4.31	cps
Calcium	43-1	5868	5754	5938	5853	1.58	cps
Calcium	44-1	97608	97977	97464	97683	0.27	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	3027	3220	3367	3205	5.32	cps
Cobalt	59-2	2520	2400	2430	2450	2.55	cps
Copper	63-2	4237	3867	3887	3997	5.21	cps
Dysprosium	156-1	10	20	7	12	56.76	cps
Dysprosium	156-2	10	3	3	6	69.34	cps
Erbium	164-1	17	7	13	12	41.65	cps
Erbium	164-2	3	3	7	4	43.40	cps
Gadolinium	160-1	43	10	7	20	101.36	cps
Gadolinium	160-2	93	150	87	110	31.64	cps
Holmium	165-1	3940208	3974578	3927111	3947299	0.62	cps
Holmium	165-2	2685233	2666208	2710701	2687381	0.83	cps
Indium	115-1	3051752	3027760	2970514	3016675	1.38	cps
Indium	115-2	970695	966308	965481	967495	0.29	cps
Iron	56-2	75521	75813	74882	75405	0.63	cps
Iron	57-2	2600	2277	2527	2468	6.87	cps
Iron	54-2	4958	4767	4867	4864	1.95	cps
Krypton	83-1	57	67	87	70	21.82	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S02 Instrumnet Name : P7
Client Sample ID : S02 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:49:35 DataFile Name : 005CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	3227	3230	3164	3207	1.17	cps
Lead	207-1	2847	2677	2860	2795	3.66	cps
Lead	208-1	12806	12909	12786	12834	0.52	cps
Lithium	6-1	703115	709034	716406	709519	0.94	cps
Magnesium	24-2	77004	78207	76984	77399	0.90	cps
Manganese	55-2	1277	1380	1343	1333	3.93	cps
Molybdenum	94-1	7912	8322	8372	8202	3.08	cps
Molybdenum	95-1	9940	10040	9987	9989	0.50	cps
Molybdenum	96-1	11038	11458	11078	11191	2.07	cps
Molybdenum	97-1	6198	6401	6471	6357	2.23	cps
Molybdenum	98-1	16413	16022	16386	16274	1.34	cps
Neodymium	150-1	0	3	3	2	86.60	cps
Neodymium	150-2	0	0	3	1	173.21	cps
Nickel	60-2	810	717	630	719	12.52	cps
Phosphorus	31-2	233	217	243	231	5.83	cps
Potassium	39-2	129139	129529	128870	129179	0.26	cps
Rhodium	103-1	2693965	2696631	2763338	2717978	1.45	cps
Rhodium	103-2	1488420	1497003	1490038	1491820	0.31	cps
Scandium	45-1	1560838	1585375	1572889	1573034	0.78	cps
Scandium	45-2	136363	135529	135324	135739	0.41	cps
Selenium	82-1	438	420	358	405	10.41	cps
Selenium	77-2	47	20	40	36	39.03	cps
Selenium	78-2	197	193	180	190	4.64	cps
Silicon	28-1	196129	199071	197838	197679	0.75	cps
Silver	107-1	5031	4978	4831	4946	2.09	cps
Silver	109-1	4554	4434	4524	4504	1.39	cps
Sodium	23-2	129764	129835	129784	129794	0.03	cps
Strontium	86-1	1563	1367	1530	1487	7.08	cps
Strontium	88-1	11745	11541	11925	11737	1.64	cps
Sulfur	34-1	116785	117238	117218	117081	0.22	cps
Terbium	159-1	4058085	4041130	4059963	4053059	0.26	cps
Terbium	159-2	2662128	2645379	2633887	2647131	0.54	cps
Thallium	203-1	3490	3804	3687	3661	4.33	cps
Thallium	205-1	9217	8793	8783	8931	2.77	cps
Tin	118-1	17681	17871	17614	17722	0.75	cps
Titanium	47-1	2650	2480	2550	2560	3.34	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S02

Instrumnet Name :

P7

Client Sample ID :

S02

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:49:35

DataFile Name :

005CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	11498	11228	11342	11356	1.19	cps
Vanadium	51-2	6765	6815	6805	6795	0.39	cps
Yttrium	89-1	4934045	4931371	4916781	4927399	0.19	cps
Yttrium	89-2	1237724	1235449	1234586	1235920	0.13	cps
Zinc	66-2	1313	1460	1437	1403	5.62	cps
Zirconium	90-1	6782	6825	7102	6903	2.52	cps
Zirconium	91-1	1507	1573	1663	1581	4.97	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P7
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:53:00 DataFile Name : 006CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	92896	93435	92205	92845	0.66	cps
Antimony	121-1	209493	213159	212601	211751	0.93	cps
Arsenic	75-2	9673	9693	9353	9573	1.99	cps
Barium	135-1	289836	289776	292806	290806	0.60	cps
Barium	137-1	498434	499996	505249	501226	0.71	cps
Beryllium	9-1	40543	40944	40716	40734	0.49	cps
Bismuth	209-1	2402932	2424969	2370591	2399497	1.14	cps
Bismuth	209-2	1846920	1914176	1893310	1884802	1.83	cps
Bromine	81-1	7235	7412	7259	7302	1.31	cps
Bromine	81-2	10	17	10	12	31.50	cps
Cadmium	108-1	4151	4171	3981	4101	2.55	cps
Cadmium	106-1	6795	7112	6852	6919	2.44	cps
Cadmium	111-1	54155	54641	54414	54403	0.45	cps
Calcium	43-1	64021	64925	65671	64872	1.27	cps
Calcium	44-1	999961	1008043	999054	1002352	0.49	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	76181	77062	77307	76850	0.77	cps
Cobalt	59-2	115160	114029	114331	114507	0.51	cps
Copper	63-2	815796	816470	807119	813128	0.64	cps
Dysprosium	156-1	30	17	10	19	53.91	cps
Dysprosium	156-2	20	13	23	19	26.96	cps
Erbium	164-1	17	20	17	18	10.81	cps
Erbium	164-2	10	17	3	10	66.70	cps
Gadolinium	160-1	10	10	13	11	17.30	cps
Gadolinium	160-2	93	100	123	106	14.93	cps
Holmium	165-1	3983881	4017603	4015068	4005517	0.47	cps
Holmium	165-2	2677780	2682888	2636966	2665878	0.94	cps
Indium	115-1	2977817	3039306	2974125	2997083	1.22	cps
Indium	115-2	964507	965784	956326	962205	0.53	cps
Iron	56-2	3438879	3453569	3410880	3434443	0.63	cps
Iron	57-2	89974	89309	89846	89710	0.39	cps
Iron	54-2	196784	197042	197690	197172	0.24	cps
Krypton	83-1	57	90	57	68	28.39	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S03 Instrumnet Name : P7
Client Sample ID : S03 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:53:00 DataFile Name : 006CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	785983	798329	795745	793352	0.82	cps
Lead	207-1	677927	690169	691280	686458	1.08	cps
Lead	208-1	3318678	3347735	3379281	3348564	0.91	cps
Lithium	6-1	715609	717741	717734	717028	0.17	cps
Magnesium	24-2	743607	741526	734622	739919	0.64	cps
Manganese	55-2	601416	601303	598606	600442	0.26	cps
Molybdenum	94-1	779977	777986	778655	778872	0.13	cps
Molybdenum	95-1	1127934	1134922	1136041	1132966	0.39	cps
Molybdenum	96-1	1236328	1248954	1243289	1242857	0.51	cps
Molybdenum	97-1	713166	720132	717601	716966	0.49	cps
Molybdenum	98-1	1871880	1846053	1837157	1851697	0.97	cps
Neodymium	150-1	13	27	33	24	41.66	cps
Neodymium	150-2	10	0	0	3	173.21	cps
Nickel	60-2	28791	28981	29132	28968	0.59	cps
Phosphorus	31-2	4541	4684	4704	4643	1.92	cps
Potassium	39-2	522351	525884	526261	524832	0.41	cps
Rhodium	103-1	2712334	2685806	2676104	2691415	0.70	cps
Rhodium	103-2	1479019	1482874	1454776	1472223	1.03	cps
Scandium	45-1	1564232	1557957	1552549	1558246	0.38	cps
Scandium	45-2	135345	136569	133999	135304	0.95	cps
Selenium	82-1	3202	3068	3162	3144	2.18	cps
Selenium	77-2	360	300	247	302	18.76	cps
Selenium	78-2	1133	1037	1143	1105	5.33	cps
Silicon	28-1	292256	294851	293210	293439	0.45	cps
Silver	107-1	261933	262809	262703	262481	0.18	cps
Silver	109-1	246957	250688	250054	249233	0.80	cps
Sodium	23-2	1196067	1195372	1206881	1199440	0.54	cps
Strontium	86-1	1713175	1701247	1709054	1707826	0.35	cps
Strontium	88-1	14890027	14906131	14783237	14859798	0.45	cps
Sulfur	34-1	132172	131527	131049	131583	0.43	cps
Terbium	159-1	4110924	4055588	4065690	4077400	0.72	cps
Terbium	159-2	2631972	2658767	2641304	2644014	0.51	cps
Thallium	203-1	193798	197436	195309	195514	0.93	cps
Thallium	205-1	459150	461890	461593	460878	0.33	cps
Tin	118-1	180495	185762	182318	182858	1.46	cps
Titanium	47-1	277761	281565	278383	279237	0.73	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S03

Instrumnet Name :

P7

Client Sample ID :

S03

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:53:00

DataFile Name :

006CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	601086	609479	603644	604736	0.71	cps
Vanadium	51-2	68699	67902	69399	68666	1.09	cps
Yttrium	89-1	4975566	4981467	4900935	4952656	0.91	cps
Yttrium	89-2	1239142	1246217	1226554	1237304	0.80	cps
Zinc	66-2	134041	134476	133440	133986	0.39	cps
Zirconium	90-1	348179	351953	352278	350803	0.65	cps
Zirconium	91-1	78146	78622	77795	78188	0.53	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P7
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:56:17 DataFile Name : 007CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	222062	225532	222925	223506	0.81	cps
Antimony	121-1	521217	524159	522316	522564	0.28	cps
Arsenic	75-2	23312	22761	23015	23029	1.20	cps
Barium	135-1	710311	713748	716870	713643	0.46	cps
Barium	137-1	1236359	1235368	1239082	1236937	0.16	cps
Beryllium	9-1	97865	98788	98751	98468	0.53	cps
Bismuth	209-1	2371507	2399479	2351384	2374123	1.02	cps
Bismuth	209-2	1869362	1853227	1877060	1866550	0.65	cps
Bromine	81-1	6845	6685	6815	6782	1.25	cps
Bromine	81-2	23	20	23	22	8.65	cps
Cadmium	108-1	9857	9800	9507	9721	1.93	cps
Cadmium	106-1	14834	15248	15271	15118	1.63	cps
Cadmium	111-1	132456	131144	132553	132051	0.60	cps
Calcium	43-1	156446	156625	154530	155867	0.75	cps
Calcium	44-1	2264967	2261145	2226267	2250793	0.95	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	184005	183853	183236	183698	0.22	cps
Cobalt	59-2	280343	276365	274386	277031	1.10	cps
Copper	63-2	1971207	1980148	1964326	1971894	0.40	cps
Dysprosium	156-1	23	17	30	23	28.56	cps
Dysprosium	156-2	23	33	30	29	17.63	cps
Erbium	164-1	17	23	17	19	20.36	cps
Erbium	164-2	7	7	13	9	43.25	cps
Gadolinium	160-1	17	17	33	22	43.28	cps
Gadolinium	160-2	97	120	127	114	13.76	cps
Holmium	165-1	3955115	3957631	3901319	3938022	0.81	cps
Holmium	165-2	2658583	2655228	2627854	2647222	0.64	cps
Indium	115-1	2917406	2921930	2863180	2900839	1.13	cps
Indium	115-2	937630	925177	923282	928696	0.84	cps
Iron	56-2	8141909	8280989	8198584	8207161	0.85	cps
Iron	57-2	217227	215709	215772	216236	0.40	cps
Iron	54-2	476144	478264	474540	476316	0.39	cps
Krypton	83-1	60	73	103	79	28.14	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S04 Instrumnet Name : P7
Client Sample ID : S04 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:56:17 DataFile Name : 007CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	2160252	2166882	2194175	2173770	0.83	cps
Lead	207-1	1919124	1961407	1913620	1931384	1.35	cps
Lead	208-1	8678236	8848847	8756558	8761214	0.97	cps
Lithium	6-1	700939	703643	695339	699973	0.61	cps
Magnesium	24-2	1737769	1748284	1745095	1743716	0.31	cps
Manganese	55-2	1401672	1409452	1408392	1406505	0.30	cps
Molybdenum	94-1	1910730	1892420	1920377	1907842	0.74	cps
Molybdenum	95-1	2777974	2783628	2770779	2777460	0.23	cps
Molybdenum	96-1	3090494	3057063	3048119	3065225	0.73	cps
Molybdenum	97-1	1750926	1752511	1750388	1751275	0.06	cps
Molybdenum	98-1	4570476	4544510	4497904	4537630	0.81	cps
Neodymium	150-1	70	53	40	54	27.61	cps
Neodymium	150-2	10	10	7	9	21.63	cps
Nickel	60-2	68660	69323	69942	69308	0.93	cps
Phosphorus	31-2	10961	10934	11104	11000	0.83	cps
Potassium	39-2	1221055	1221412	1221215	1221227	0.01	cps
Rhodium	103-1	2597835	2522515	2565781	2562044	1.48	cps
Rhodium	103-2	1418827	1406798	1403301	1409642	0.58	cps
Scandium	45-1	1511709	1508776	1498258	1506248	0.47	cps
Scandium	45-2	130771	129968	128159	129633	1.03	cps
Selenium	82-1	7599	7452	7420	7490	1.27	cps
Selenium	77-2	740	650	660	683	7.22	cps
Selenium	78-2	2407	2637	2257	2434	7.87	cps
Silicon	28-1	434132	437546	436404	436027	0.40	cps
Silver	107-1	638594	640673	633136	637468	0.61	cps
Silver	109-1	605144	604752	604577	604824	0.05	cps
Sodium	23-2	2964429	2940278	2911194	2938634	0.91	cps
Strontium	86-1	4190932	4207975	4108231	4169046	1.28	cps
Strontium	88-1	36449674	36341208	36124508	36305130	0.46	cps
Sulfur	34-1	148937	147443	145565	147315	1.15	cps
Terbium	159-1	4035102	3996666	3965120	3998963	0.88	cps
Terbium	159-2	2650543	2586137	2585166	2607282	1.44	cps
Thallium	203-1	484099	487338	488181	486539	0.44	cps
Thallium	205-1	1145182	1162540	1164005	1157242	0.90	cps
Tin	118-1	446578	445488	448708	446925	0.37	cps
Titanium	47-1	677339	674253	674265	675286	0.26	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S04

Instrumnet Name :

P7

Client Sample ID :

S04

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 11:56:17

DataFile Name :

007CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	1526391	1541452	1543409	1537084	0.61	cps
Vanadium	51-2	166426	166813	165551	166263	0.39	cps
Yttrium	89-1	4808373	4776894	4754201	4779822	0.57	cps
Yttrium	89-2	1205769	1187914	1189164	1194282	0.83	cps
Zinc	66-2	321839	323716	319194	321583	0.71	cps
Zirconium	90-1	860247	860971	859084	860101	0.11	cps
Zirconium	91-1	192406	192403	191938	192249	0.14	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:59:27 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	429643	430664	429331	429879	0.16	cps
Antimony	121-1	1010304	1015409	1022825	1016180	0.62	cps
Arsenic	75-2	44443	44791	43901	44378	1.01	cps
Barium	135-1	1385103	1388594	1396342	1390013	0.41	cps
Barium	137-1	2548233	2572425	2558956	2559871	0.47	cps
Beryllium	9-1	189553	191037	189195	189929	0.51	cps
Bismuth	209-1	2403465	2360520	2351390	2371792	1.17	cps
Bismuth	209-2	1827692	1667497	1822954	1772714	5.14	cps
Bromine	81-1	6298	6652	6528	6493	2.76	cps
Bromine	81-2	13	20	30	21	39.75	cps
Cadmium	108-1	19123	18759	19303	19061	1.45	cps
Cadmium	106-1	28191	28264	27733	28063	1.03	cps
Cadmium	111-1	249961	252951	253149	252020	0.71	cps
Calcium	43-1	292173	295978	294769	294307	0.66	cps
Calcium	44-1	4249522	4241607	4238189	4243106	0.14	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	352197	351742	349818	351252	0.36	cps
Cobalt	59-2	528705	524754	521695	525051	0.67	cps
Copper	63-2	3736141	3654910	3715163	3702071	1.14	cps
Dysprosium	156-1	43	50	23	39	35.69	cps
Dysprosium	156-2	63	57	67	62	8.18	cps
Erbium	164-1	27	27	33	29	13.31	cps
Erbium	164-2	3	13	27	14	81.07	cps
Gadolinium	160-1	13	27	27	22	34.66	cps
Gadolinium	160-2	127	100	137	121	15.65	cps
Holmium	165-1	3975570	3946861	3896458	3939629	1.02	cps
Holmium	165-2	2628503	2651454	2574210	2618056	1.52	cps
Indium	115-1	2862916	2738890	2781474	2794426	2.26	cps
Indium	115-2	890312	902298	891988	894866	0.73	cps
Iron	56-2	15740336	15628873	15648929	15672713	0.38	cps
Iron	57-2	416327	410960	411505	412930	0.72	cps
Iron	54-2	916571	906859	906463	909964	0.63	cps
Krypton	83-1	67	87	50	68	27.09	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:59:27 DataFile Name : 008CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	4293462	4323751	4375723	4330979	0.96	cps
Lead	207-1	3848108	3876819	3927694	3884207	1.04	cps
Lead	208-1	17478222	17558508	17641197	17559309	0.46	cps
Lithium	6-1	692173	682115	683207	685832	0.80	cps
Magnesium	24-2	3335180	3299941	3338017	3324379	0.64	cps
Manganese	55-2	2739274	2730452	2680059	2716595	1.18	cps
Molybdenum	94-1	3644300	3640285	3627389	3637325	0.24	cps
Molybdenum	95-1	5346402	5431302	5406930	5394878	0.81	cps
Molybdenum	96-1	5868734	5823092	5813147	5834991	0.51	cps
Molybdenum	97-1	3354859	3356845	3353218	3354974	0.05	cps
Molybdenum	98-1	8620471	8745945	8786126	8717514	0.99	cps
Neodymium	150-1	137	73	120	110	29.85	cps
Neodymium	150-2	17	37	20	24	43.83	cps
Nickel	60-2	131174	129137	128649	129653	1.03	cps
Phosphorus	31-2	20537	20287	20510	20445	0.67	cps
Potassium	39-2	2141448	2173179	2172194	2162273	0.83	cps
Rhodium	103-1	2533358	2468845	2457230	2486477	1.65	cps
Rhodium	103-2	1349831	1360812	1347460	1352701	0.53	cps
Scandium	45-1	1459940	1419432	1410436	1429936	1.84	cps
Scandium	45-2	123683	123878	125078	124213	0.61	cps
Selenium	82-1	14092	14062	14222	14125	0.60	cps
Selenium	77-2	1387	1457	1443	1429	2.60	cps
Selenium	78-2	4514	4614	4671	4600	1.73	cps
Silicon	28-1	653468	656432	657127	655676	0.30	cps
Silver	107-1	1215680	1229523	1226171	1223791	0.59	cps
Silver	109-1	1151748	1152808	1154679	1153078	0.13	cps
Sodium	23-2	5560544	5627442	5547284	5578423	0.77	cps
Strontium	86-1	7867794	7980757	7895590	7914713	0.74	cps
Strontium	88-1	69592759	69574411	70223879	69797016	0.53	cps
Sulfur	34-1	175185	175511	173378	174691	0.66	cps
Terbium	159-1	4015368	3973701	3952427	3980499	0.80	cps
Terbium	159-2	2543363	2564412	2528542	2545439	0.71	cps
Thallium	203-1	958632	966165	964339	963045	0.41	cps
Thallium	205-1	2584647	2558270	2597300	2580072	0.77	cps
Tin	118-1	856676	865640	866878	863065	0.65	cps
Titanium	47-1	1245639	1235136	1278843	1253206	1.82	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S05 Instrumnet Name : P7
Client Sample ID : S05 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 11:59:27 DataFile Name : 008CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	3520893	3554117	3613923	3562977	1.32	cps
Vanadium	51-2	317747	316961	317814	317508	0.15	cps
Yttrium	89-1	4674729	4528940	4599495	4601055	1.58	cps
Yttrium	89-2	1150361	1148656	1146755	1148591	0.16	cps
Zinc	66-2	617118	614879	610320	614106	0.56	cps
Zirconium	90-1	1630899	1651296	1642438	1641544	0.62	cps
Zirconium	91-1	367926	370519	369133	369193	0.35	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:02:28 DataFile Name : 009CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	839139	834426	839982	837849	0.36	cps
Antimony	121-1	2064732	2066543	2111609	2080961	1.28	cps
Arsenic	75-2	85199	85420	86527	85715	0.83	cps
Barium	135-1	2919939	2920344	2949253	2929845	0.57	cps
Barium	137-1	4923934	5098492	4982212	5001546	1.78	cps
Beryllium	9-1	362659	367343	366189	365397	0.67	cps
Bismuth	209-1	2396416	2322569	2333658	2350881	1.69	cps
Bismuth	209-2	1635188	1834242	1816299	1761910	6.25	cps
Bromine	81-1	6285	6051	6155	6164	1.90	cps
Bromine	81-2	20	7	37	21	71.19	cps
Cadmium	108-1	36747	36529	36920	36732	0.53	cps
Cadmium	106-1	52461	53475	53367	53101	1.05	cps
Cadmium	111-1	487414	492011	490641	490022	0.48	cps
Calcium	43-1	562030	563894	563809	563244	0.19	cps
Calcium	44-1	8185996	8282214	8238867	8235692	0.59	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	684133	676223	680615	680324	0.58	cps
Cobalt	59-2	1012364	1015573	1017542	1015160	0.26	cps
Copper	63-2	7204232	7120474	7163877	7162861	0.58	cps
Dysprosium	156-1	80	87	67	78	13.09	cps
Dysprosium	156-2	110	157	97	121	26.01	cps
Erbium	164-1	23	50	33	36	37.90	cps
Erbium	164-2	20	33	20	24	31.49	cps
Gadolinium	160-1	40	43	23	36	30.14	cps
Gadolinium	160-2	137	113	140	130	11.17	cps
Holmium	165-1	3936376	3901491	3893503	3910457	0.58	cps
Holmium	165-2	2548952	2631361	2635479	2605264	1.87	cps
Indium	115-1	2744874	2616314	2708128	2689772	2.46	cps
Indium	115-2	861381	870798	863644	865274	0.57	cps
Iron	56-2	30424528	30768945	30596251	30596575	0.56	cps
Iron	57-2	795037	804624	801904	800522	0.62	cps
Iron	54-2	1690400	1653158	1703631	1682396	1.56	cps
Krypton	83-1	90	90	93	91	2.11	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S06 Instrumnet Name : P7
Client Sample ID : S06 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:02:28 DataFile Name : 009CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	8704613	8816202	8849646	8790154	0.86	cps
Lead	207-1	7568126	7799015	7723794	7696978	1.53	cps
Lead	208-1	34679541	35535512	35515230	35243427	1.39	cps
Lithium	6-1	677992	665704	670485	671394	0.92	cps
Magnesium	24-2	6474410	6511744	6607416	6531190	1.05	cps
Manganese	55-2	5274654	5259177	5298659	5277497	0.38	cps
Molybdenum	94-1	6990688	7107829	7207078	7101865	1.53	cps
Molybdenum	95-1	10481389	10507350	10494894	10494544	0.12	cps
Molybdenum	96-1	11498756	11496507	11635517	11543593	0.69	cps
Molybdenum	97-1	6615168	6591436	6621815	6609473	0.24	cps
Molybdenum	98-1	17024687	16877389	17169175	17023750	0.86	cps
Neodymium	150-1	237	270	270	259	7.43	cps
Neodymium	150-2	40	27	73	47	51.50	cps
Nickel	60-2	248994	248642	249861	249166	0.25	cps
Phosphorus	31-2	39989	40373	40751	40371	0.94	cps
Potassium	39-2	4106125	4163298	4150975	4140133	0.73	cps
Rhodium	103-1	2457704	2338418	2381940	2392687	2.52	cps
Rhodium	103-2	1306460	1331744	1309149	1315784	1.06	cps
Scandium	45-1	1414573	1340535	1375913	1377007	2.69	cps
Scandium	45-2	120434	122769	121424	121542	0.96	cps
Selenium	82-1	26832	27416	26874	27041	1.20	cps
Selenium	77-2	2654	2500	2684	2612	3.76	cps
Selenium	78-2	8409	8719	9013	8714	3.46	cps
Silicon	28-1	1070965	1082085	1083248	1078766	0.63	cps
Silver	107-1	2396448	2449014	2395120	2413527	1.27	cps
Silver	109-1	2302904	2326059	2321000	2316654	0.53	cps
Sodium	23-2	11117657	11131442	11094213	11114437	0.17	cps
Strontium	86-1	15439794	15350433	15464101	15418109	0.39	cps
Strontium	88-1	136366818	136608868	136938421	136638036	0.21	cps
Sulfur	34-1	235581	234840	234654	235025	0.21	cps
Terbium	159-1	4015067	3929031	3896002	3946700	1.56	cps
Terbium	159-2	2531816	2628836	2566666	2575773	1.91	cps
Thallium	203-1	2134445	2162186	2189450	2162027	1.27	cps
Thallium	205-1	5200172	5186593	5178400	5188388	0.21	cps
Tin	118-1	1736824	1783633	1773086	1764514	1.39	cps
Titanium	47-1	2417585	2422898	2392224	2410902	0.68	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

S06

Instrumnet Name :

P7

Client Sample ID :

S06

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 12:02:28

DataFile Name :

009CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7017876	7113640	7190880	7107465	1.22	cps
Vanadium	51-2	614410	617252	622059	617907	0.63	cps
Yttrium	89-1	4614395	4408273	4534369	4519012	2.30	cps
Yttrium	89-2	1125823	1136174	1129180	1130392	0.47	cps
Zinc	66-2	1189876	1187171	1189377	1188808	0.12	cps
Zirconium	90-1	3229226	3215855	3237696	3227592	0.34	cps
Zirconium	91-1	714114	722583	727539	721412	0.94	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:05:26 DataFile Name : 010CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1622714	1629829	1643863	1632136	0.66	cps
Antimony	121-1	3993439	4067123	4108072	4056211	1.43	cps
Arsenic	75-2	171730	168794	168413	169646	1.07	cps
Barium	135-1	5643227	5720041	5661339	5674869	0.71	cps
Barium	137-1	9712481	9759151	9842838	9771490	0.68	cps
Beryllium	9-1	700787	716164	721311	712754	1.50	cps
Bismuth	209-1	2349820	2334606	2365844	2350090	0.66	cps
Bismuth	209-2	1632995	1812965	1821957	1755972	6.07	cps
Bromine	81-1	6411	6268	6035	6238	3.05	cps
Bromine	81-2	3	27	17	16	75.27	cps
Cadmium	108-1	69676	70841	71066	70528	1.06	cps
Cadmium	106-1	100968	103153	102807	102309	1.15	cps
Cadmium	111-1	939634	947799	951268	946233	0.63	cps
Calcium	43-1	1098050	1111098	1117233	1108794	0.88	cps
Calcium	44-1	16105589	16341681	16341762	16263010	0.84	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	1347550	1340487	1348862	1345633	0.33	cps
Cobalt	59-2	1999691	1979039	1998534	1992421	0.58	cps
Copper	63-2	13770483	13843769	14022173	13878808	0.93	cps
Dysprosium	156-1	113	113	147	124	15.47	cps
Dysprosium	156-2	283	253	293	277	7.52	cps
Erbium	164-1	43	43	83	57	40.76	cps
Erbium	164-2	37	47	33	39	17.85	cps
Gadolinium	160-1	53	57	73	61	17.53	cps
Gadolinium	160-2	133	107	123	121	11.12	cps
Holmium	165-1	3919963	3939464	3930047	3929825	0.25	cps
Holmium	165-2	2662205	2623831	2645487	2643841	0.73	cps
Indium	115-1	2646468	2636718	2702803	2661996	1.34	cps
Indium	115-2	850912	851448	848731	850364	0.17	cps
Iron	56-2	60384077	59973819	60529754	60295884	0.48	cps
Iron	57-2	1518206	1498858	1528495	1515187	0.99	cps
Iron	54-2	3340805	3348275	3309527	3332869	0.62	cps
Krypton	83-1	80	57	63	67	18.03	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:05:26 DataFile Name : 010CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	17185255	17593113	17193757	17324041	1.35	cps
Lead	207-1	15017774	15317525	15216888	15184062	1.00	cps
Lead	208-1	68841874	70131495	69607979	69527116	0.93	cps
Lithium	6-1	667122	665422	676203	669582	0.87	cps
Magnesium	24-2	13019285	12922924	13145876	13029362	0.86	cps
Manganese	55-2	10381677	10523447	10400467	10435197	0.74	cps
Molybdenum	94-1	14067386	14187009	14147243	14133879	0.43	cps
Molybdenum	95-1	20558042	20784749	20691473	20678088	0.55	cps
Molybdenum	96-1	22443623	22805425	22726451	22658500	0.84	cps
Molybdenum	97-1	12858391	13046259	13153102	13019251	1.15	cps
Molybdenum	98-1	33209697	33485868	34151806	33615790	1.44	cps
Neodymium	150-1	443	437	457	446	2.29	cps
Neodymium	150-2	43	117	100	87	44.36	cps
Nickel	60-2	481408	483989	483329	482909	0.28	cps
Phosphorus	31-2	80215	78406	79612	79411	1.16	cps
Potassium	39-2	8167493	8141863	8244845	8184734	0.66	cps
Rhodium	103-1	2345591	2332713	2332808	2337037	0.32	cps
Rhodium	103-2	1297374	1301083	1297120	1298526	0.17	cps
Scandium	45-1	1383467	1372016	1393334	1382939	0.77	cps
Scandium	45-2	121952	122759	122856	122522	0.40	cps
Selenium	82-1	52295	52457	52996	52583	0.70	cps
Selenium	77-2	5168	5108	5234	5170	1.23	cps
Selenium	78-2	16806	17207	17133	17049	1.25	cps
Silicon	28-1	1897010	1945619	1903823	1915484	1.37	cps
Silver	107-1	4597111	4755400	4682843	4678451	1.69	cps
Silver	109-1	4340915	4435052	4419529	4398499	1.15	cps
Sodium	23-2	21939354	22055021	22296609	22096994	0.82	cps
Strontium	86-1	30190136	30365096	31063640	30539624	1.51	cps
Strontium	88-1	268413209	268600809	269784756	268932925	0.28	cps
Sulfur	34-1	372823	376310	376904	375346	0.59	cps
Terbium	159-1	3998618	3937593	4091471	4009227	1.93	cps
Terbium	159-2	2607938	2583903	2551888	2581243	1.09	cps
Thallium	203-1	4247746	4369144	4262510	4293133	1.54	cps
Thallium	205-1	10169168	10401412	10103013	10224531	1.53	cps
Tin	118-1	3370942	3482839	3469363	3441048	1.78	cps
Titanium	47-1	4693858	4769501	4777180	4746847	0.97	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S07 Instrumnet Name : P7
Client Sample ID : S07 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:05:26 DataFile Name : 010CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	14072902	14324401	14158161	14185155	0.90	cps
Vanadium	51-2	1229915	1225286	1228678	1227959	0.20	cps
Yttrium	89-1	4546073	4502873	4499548	4516165	0.57	cps
Yttrium	89-2	1130378	1133461	1129701	1131180	0.18	cps
Zinc	66-2	2356225	2331899	2310652	2332925	0.98	cps
Zirconium	90-1	6318057	6293098	6408424	6339860	0.96	cps
Zirconium	91-1	1410035	1408468	1430524	1416343	0.87	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P7
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:08:18 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	8199956	8179474	8112542	8163991	0.56	cps
Antimony	121-1	2927	2874	2950	2917	1.35	cps
Arsenic	75-2	57	80	87	74	21.16	cps
Barium	135-1	2894	2760	2910	2855	2.88	cps
Barium	137-1	4938	5171	4981	5030	2.47	cps
Beryllium	9-1	453	443	367	421	11.26	cps
Bismuth	209-1	2052898	2045090	2044653	2047547	0.23	cps
Bismuth	209-2	1335434	1396003	1426182	1385873	3.33	cps
Bromine	81-1	6405	6255	6385	6348	1.28	cps
Bromine	81-2	27	27	17	23	24.74	cps
Cadmium	108-1	47	30	30	36	27.07	cps
Cadmium	106-1	1373	1197	1237	1269	7.30	cps
Cadmium	111-1	1199	1023	1025	1082	9.32	cps
Calcium	43-1	4738133	4773038	4916196	4809123	1.96	cps
Calcium	44-1	75702929	76598239	77606521	76635896	1.24	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4784	4421	4971	4725	5.92	cps
Cobalt	59-2	5001	5254	5344	5200	3.42	cps
Copper	63-2	4647	4487	4577	4571	1.76	cps
Dysprosium	156-1	197	160	140	166	17.36	cps
Dysprosium	156-2	97	163	83	114	37.45	cps
Erbium	164-1	237	153	160	183	25.26	cps
Erbium	164-2	73	153	137	121	34.85	cps
Gadolinium	160-1	187	110	103	133	34.73	cps
Gadolinium	160-2	123	143	173	147	17.16	cps
Holmium	165-1	3687146	3695019	3661430	3681199	0.48	cps
Holmium	165-2	2303082	2462168	2452417	2405889	3.71	cps
Indium	115-1	2504292	2539667	2475403	2506454	1.28	cps
Indium	115-2	751710	786792	802435	780312	3.33	cps
Iron	56-2	284118582	282515876	283964789	283533082	0.31	cps
Iron	57-2	7018295	6934403	7041511	6998070	0.81	cps
Iron	54-2	15388147	15229064	15195551	15270921	0.67	cps
Krypton	83-1	107	90	90	96	10.07	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : S08 Instrumnet Name : P7
Client Sample ID : S08 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:08:18 DataFile Name : 011CAL.S.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	3644	3367	3297	3436	5.34	cps
Lead	207-1	3164	3244	2780	3063	8.09	cps
Lead	208-1	14600	13936	13259	13932	4.81	cps
Lithium	6-1	645627	652857	658922	652469	1.02	cps
Magnesium	24-2	68994756	67808241	68593774	68465590	0.88	cps
Manganese	55-2	6685	6728	6715	6709	0.33	cps
Molybdenum	94-1	3757	3657	3707	3707	1.35	cps
Molybdenum	95-1	3664	3490	3557	3570	2.45	cps
Molybdenum	96-1	7532	6848	7055	7145	4.91	cps
Molybdenum	97-1	2367	2254	2117	2246	5.58	cps
Molybdenum	98-1	5558	5361	5344	5421	2.19	cps
Neodymium	150-1	77	120	113	103	22.58	cps
Neodymium	150-2	77	37	73	62	35.67	cps
Nickel	60-2	1843	2220	1890	1985	10.35	cps
Phosphorus	31-2	150	130	123	134	10.32	cps
Potassium	39-2	40984384	39857087	39966732	40269401	1.54	cps
Rhodium	103-1	2077556	2087042	2028114	2064237	1.53	cps
Rhodium	103-2	1061358	1122341	1135497	1106399	3.58	cps
Scandium	45-1	1327336	1345879	1323919	1332378	0.89	cps
Scandium	45-2	111917	115725	118415	115352	2.83	cps
Selenium	82-1	128	142	156	142	9.79	cps
Selenium	77-2	0	3	3	2	86.60	cps
Selenium	78-2	133	80	97	103	26.41	cps
Silicon	28-1	183632	178934	180575	181047	1.32	cps
Silver	107-1	650	543	593	596	8.96	cps
Silver	109-1	547	593	480	540	10.55	cps
Sodium	23-2	107879935	106773858	108936485	107863426	1.00	cps
Strontium	86-1	5264	5141	5194	5200	1.19	cps
Strontium	88-1	46393	43915	42715	44341	4.23	cps
Sulfur	34-1	95882	98189	98542	97538	1.48	cps
Terbium	159-1	3781331	3812858	3633462	3742550	2.56	cps
Terbium	159-2	2240735	2334739	2435147	2336873	4.16	cps
Thallium	203-1	480	447	407	444	8.26	cps
Thallium	205-1	1233	1113	930	1092	13.99	cps
Tin	118-1	2040	2117	2260	2139	5.22	cps
Titanium	47-1	703	800	803	769	7.39	cps

LB Number :LB133015

Operator :Jaswal

Lab Sample ID :S08

Instrumnet Name :P7

Client Sample ID :S08

Dilution Factor :1

Date & Time Acquired :2024-10-18 12:08:18

DataFile Name :011CALS.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	823	663	590	692	17.24	cps
Vanadium	51-2	370	430	360	387	9.79	cps
Yttrium	89-1	4213790	4185392	4228266	4209149	0.52	cps
Yttrium	89-2	989594	1040022	1066172	1031929	3.77	cps
Zinc	66-2	8462	8022	8326	8270	2.72	cps
Zirconium	90-1	3771	3747	3857	3792	1.53	cps
Zirconium	91-1	810	797	810	806	0.96	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICV025 Instrumnet Name : P7
Client Sample ID : ICV025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:59:06 DataFile Name : 012ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	41489	40995	41326	41270	0.61	cps
Antimony	121-1	784850	794029	794456	791111	0.69	cps
Arsenic	75-2	34410	34614	35279	34767	1.31	cps
Barium	135-1	106984	106978	107186	107049	0.11	cps
Barium	137-1	184284	185442	185438	185055	0.36	cps
Beryllium	9-1	75783	76865	75870	76173	0.79	cps
Bismuth	209-1	2448355	2488987	2439421	2458921	1.07	cps
Bismuth	209-2	1884977	1877927	1911308	1891404	0.93	cps
Bromine	81-1	6418	6321	6241	6327	1.40	cps
Bromine	81-2	23	17	17	19	20.36	cps
Cadmium	108-1	10897	10964	11204	11022	1.46	cps
Cadmium	106-1	19089	18672	18662	18808	1.30	cps
Cadmium	111-1	105589	105804	104625	105339	0.60	cps
Calcium	43-1	22961	22914	22851	22908	0.24	cps
Calcium	44-1	356209	355940	354123	355424	0.32	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	140511	139667	139488	139888	0.39	cps
Cobalt	59-2	211484	211573	208065	210374	0.95	cps
Copper	63-2	148343	148158	148303	148268	0.07	cps
Dysprosium	156-1	13	10	10	11	17.30	cps
Dysprosium	156-2	7	3	0	3	100.05	cps
Erbium	164-1	13	20	7	13	49.99	cps
Erbium	164-2	0	0	13	4	173.21	cps
Gadolinium	160-1	13	0	10	8	89.21	cps
Gadolinium	160-2	120	120	117	119	1.62	cps
Holmium	165-1	4026792	3957517	4033045	4005784	1.05	cps
Holmium	165-2	2601852	2574401	2647504	2607919	1.42	cps
Indium	115-1	2839067	2778035	2803662	2806922	1.09	cps
Indium	115-2	884836	883199	892642	886892	0.57	cps
Iron	56-2	2538712	2473788	2455740	2489414	1.75	cps
Iron	57-2	65006	64314	65863	65061	1.19	cps
Iron	54-2	147669	145371	145148	146062	0.96	cps
Krypton	83-1	70	93	73	79	16.00	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICV025 Instrumnet Name : P7
Client Sample ID : ICV025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:59:06 DataFile Name : 012ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	672130	684772	685467	680790	1.10	cps
Lead	207-1	562559	569795	574515	568956	1.06	cps
Lead	208-1	2630491	2674607	2691714	2665604	1.19	cps
Lithium	6-1	697005	693863	696381	695750	0.24	cps
Magnesium	24-2	167814	167561	165085	166820	0.90	cps
Manganese	55-2	110653	110616	107973	109747	1.40	cps
Molybdenum	94-1	6696725	6561840	6661367	6639977	1.05	cps
Molybdenum	95-1	9600262	9560451	9543749	9568154	0.30	cps
Molybdenum	96-1	10479456	10501415	10560695	10513855	0.40	cps
Molybdenum	97-1	6004182	6094350	6066023	6054852	0.76	cps
Molybdenum	98-1	15681898	15594584	15382240	15552907	0.99	cps
Neodymium	150-1	3	17	3	8	99.04	cps
Neodymium	150-2	3	0	7	3	100.05	cps
Nickel	60-2	54445	54247	52746	53813	1.73	cps
Phosphorus	31-2	130	103	120	118	11.44	cps
Potassium	39-2	371283	371701	372138	371707	0.12	cps
Rhodium	103-1	2532200	2516041	2524866	2524369	0.32	cps
Rhodium	103-2	1379969	1374804	1390984	1381919	0.60	cps
Scandium	45-1	1452578	1473454	1448063	1458032	0.93	cps
Scandium	45-2	125225	125088	125986	125433	0.39	cps
Selenium	82-1	11142	11320	11540	11334	1.76	cps
Selenium	77-2	1017	1097	1023	1046	4.25	cps
Selenium	78-2	3614	3881	3644	3713	3.94	cps
Silicon	28-1	1137912	1143886	1151007	1144268	0.57	cps
Silver	107-1	234531	235893	235334	235253	0.29	cps
Silver	109-1	219472	220105	222223	220600	0.65	cps
Sodium	23-2	463191	462629	463029	462950	0.06	cps
Strontium	86-1	624769	625980	626785	625844	0.16	cps
Strontium	88-1	5605886	5679210	5637926	5641007	0.65	cps
Sulfur	34-1	109485	107142	107421	108016	1.19	cps
Terbium	159-1	4113590	4078252	3989699	4060514	1.57	cps
Terbium	159-2	2538894	2572580	2555625	2555700	0.66	cps
Thallium	203-1	818473	834642	842423	831846	1.47	cps
Thallium	205-1	2223483	2241993	2218611	2228029	0.55	cps
Tin	118-1	1924499	1924543	1893963	1914335	0.92	cps
Titanium	47-1	2243460	2228386	2233472	2235106	0.34	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICV025 Instrumnet Name : P7
Client Sample ID : ICV025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 12:59:06 DataFile Name : 012ICV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7366022	7461848	7468242	7432037	0.77	cps
Vanadium	51-2	122196	122626	121249	122024	0.58	cps
Yttrium	89-1	4631773	4659959	4605853	4632528	0.58	cps
Yttrium	89-2	1146032	1144144	1161932	1150703	0.85	cps
Zinc	66-2	47773	47061	47723	47519	0.84	cps
Zirconium	90-1	3278307	3332932	3320060	3310433	0.86	cps
Zirconium	91-1	726951	736709	730753	731471	0.67	cps

LB Number :LB133015Operator :Jaswal

Lab Sample ID :ICB025Instrumnet Name :P7

Client Sample ID :ICB025Dilution Factor :1

Date & Time Acquired :2024-10-18 13:16:30DataFile Name :014CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	70	47	73	63	22.94	cps
Antimony	121-1	43	37	67	49	32.22	cps
Arsenic	75-2	0	0	0	0	0.00	cps
Barium	135-1	37	20	23	27	33.08	cps
Barium	137-1	33	30	47	37	24.06	cps
Beryllium	9-1	50	53	53	52	3.68	cps
Bismuth	209-1	2503970	2453882	2458182	2472012	1.12	cps
Bismuth	209-2	1907662	1890561	1935863	1911362	1.20	cps
Bromine	81-1	6712	6895	7058	6888	2.52	cps
Bromine	81-2	10	33	47	30	61.87	cps
Cadmium	108-1	7	0	0	2	173.21	cps
Cadmium	106-1	1300	1197	1133	1210	6.95	cps
Cadmium	111-1	905	841	797	848	6.45	cps
Calcium	43-1	167	140	107	138	21.82	cps
Calcium	44-1	3234	3234	3130	3199	1.87	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	253	237	327	272	17.59	cps
Cobalt	59-2	13	13	10	12	15.73	cps
Copper	63-2	693	663	647	668	3.54	cps
Dysprosium	156-1	13	17	17	16	12.40	cps
Dysprosium	156-2	3	0	0	1	173.21	cps
Erbium	164-1	17	17	13	16	12.40	cps
Erbium	164-2	23	7	27	19	56.72	cps
Gadolinium	160-1	0	7	10	6	91.64	cps
Gadolinium	160-2	97	130	117	114	14.66	cps
Holmium	165-1	3998426	3954047	3970632	3974369	0.56	cps
Holmium	165-2	2677875	2603262	2714323	2665154	2.12	cps
Indium	115-1	2955731	2923970	2943601	2941101	0.54	cps
Indium	115-2	929362	917524	928359	925082	0.71	cps
Iron	56-2	4564	4781	4707	4684	2.35	cps
Iron	57-2	493	577	550	540	7.88	cps
Iron	54-2	777	760	817	784	3.71	cps
Krypton	83-1	90	87	43	73	35.50	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICB025 Instrumnet Name : P7
Client Sample ID : ICB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:16:30 DataFile Name : 014CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	210	237	257	234	9.99	cps
Lead	207-1	190	250	243	228	14.44	cps
Lead	208-1	837	983	953	924	8.38	cps
Lithium	6-1	702311	704504	708743	705186	0.46	cps
Magnesium	24-2	1277	1287	1253	1272	1.34	cps
Manganese	55-2	103	73	80	86	18.42	cps
Molybdenum	94-1	120	117	110	116	4.41	cps
Molybdenum	95-1	70	50	77	66	21.17	cps
Molybdenum	96-1	57	83	53	64	25.51	cps
Molybdenum	97-1	30	27	43	33	26.45	cps
Molybdenum	98-1	103	70	137	103	32.26	cps
Neodymium	150-1	7	0	0	2	173.21	cps
Neodymium	150-2	3	0	7	3	100.05	cps
Nickel	60-2	63	57	63	61	6.29	cps
Phosphorus	31-2	100	107	80	96	14.52	cps
Potassium	39-2	30597	31142	30801	30847	0.89	cps
Rhodium	103-1	2568237	2588238	2551832	2569436	0.71	cps
Rhodium	103-2	1433352	1419180	1441640	1431391	0.79	cps
Scandium	45-1	1481679	1488293	1484491	1484821	0.22	cps
Scandium	45-2	129531	128815	129709	129352	0.37	cps
Selenium	82-1	87	94	151	110	31.82	cps
Selenium	77-2	0	3	0	1	173.21	cps
Selenium	78-2	93	103	100	99	5.15	cps
Silicon	28-1	165233	167433	166028	166231	0.67	cps
Silver	107-1	37	43	37	39	9.89	cps
Silver	109-1	17	27	33	26	32.81	cps
Sodium	23-2	8239	7999	8216	8151	1.62	cps
Strontium	86-1	143	123	103	123	16.22	cps
Strontium	88-1	140	117	150	136	12.62	cps
Sulfur	34-1	110469	111687	111785	111314	0.66	cps
Terbium	159-1	4068506	4053522	4144778	4088936	1.20	cps
Terbium	159-2	2574665	2621964	2621551	2606060	1.04	cps
Thallium	203-1	67	50	40	52	25.80	cps
Thallium	205-1	83	130	127	113	22.97	cps
Tin	118-1	687	837	697	740	11.33	cps
Titanium	47-1	40	23	30	31	26.97	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICB025 Instrumnet Name : P7
Client Sample ID : ICB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:16:30 DataFile Name : 014CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7	23	20	17	52.90	cps
Vanadium	51-2	7	7	0	4	86.60	cps
Yttrium	89-1	4693101	4662388	4784376	4713289	1.35	cps
Yttrium	89-2	1196488	1178654	1188672	1187938	0.75	cps
Zinc	66-2	113	90	110	104	12.08	cps
Zirconium	90-1	303	270	277	283	6.23	cps
Zirconium	91-1	47	67	77	63	24.12	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSA025 Instrumnet Name : P7
Client Sample ID : ICSA025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:19:53 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7397515	7357206	7421524	7392082	0.44	cps
Antimony	121-1	4447	4051	4471	4323	5.46	cps
Arsenic	75-2	70	43	60	58	23.32	cps
Barium	135-1	1380	1450	1460	1430	3.05	cps
Barium	137-1	2544	2694	2604	2614	2.89	cps
Beryllium	9-1	270	263	210	248	13.27	cps
Bismuth	209-1	2303385	2320450	2325845	2316560	0.51	cps
Bismuth	209-2	1819785	1615069	1810791	1748549	6.62	cps
Bromine	81-1	6522	6455	6532	6503	0.64	cps
Bromine	81-2	50	100	87	79	32.82	cps
Cadmium	108-1	603	563	540	569	5.63	cps
Cadmium	106-1	1230	1113	1213	1186	5.32	cps
Cadmium	111-1	1102	1205	1078	1128	5.98	cps
Calcium	43-1	986934	981725	991996	986885	0.52	cps
Calcium	44-1	14836796	14777144	14932148	14848696	0.53	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	26710	26924	26710	26782	0.46	cps
Cobalt	59-2	2480	2590	2620	2564	2.88	cps
Copper	63-2	10684	10821	11585	11030	4.40	cps
Dysprosium	156-1	10	13	23	16	44.60	cps
Dysprosium	156-2	7	10	7	8	24.71	cps
Erbium	164-1	20	17	20	19	10.18	cps
Erbium	164-2	23	23	17	21	18.21	cps
Gadolinium	160-1	17	47	23	29	54.53	cps
Gadolinium	160-2	143	127	127	132	7.28	cps
Holmium	165-1	3977228	3943792	3993756	3971592	0.64	cps
Holmium	165-2	2679895	2660833	2711845	2684191	0.96	cps
Indium	115-1	2741000	2678853	2736420	2718758	1.27	cps
Indium	115-2	876373	870080	875703	874052	0.40	cps
Iron	56-2	117859828	119204545	118852695	118639023	0.59	cps
Iron	57-2	2912770	2935127	2936660	2928186	0.46	cps
Iron	54-2	6387908	6376073	6459683	6407888	0.71	cps
Krypton	83-1	77	90	93	87	10.17	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSA025 Instrumnet Name : P7
Client Sample ID : ICSA025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:19:53 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	14011	14505	14528	14348	2.04	cps
Lead	207-1	11095	11228	11675	11333	2.68	cps
Lead	208-1	52403	53678	54222	53434	1.75	cps
Lithium	6-1	672132	655709	669045	665629	1.31	cps
Magnesium	24-2	12522800	12696888	12641999	12620563	0.71	cps
Manganese	55-2	8683	8639	8292	8538	2.51	cps
Molybdenum	94-1	2016349	2005159	2008699	2010069	0.28	cps
Molybdenum	95-1	3601621	3549992	3540928	3564180	0.92	cps
Molybdenum	96-1	3855573	3817070	3815068	3829237	0.60	cps
Molybdenum	97-1	2248378	2219969	2253331	2240560	0.80	cps
Molybdenum	98-1	5817537	5812180	5859953	5829890	0.45	cps
Neodymium	150-1	13	27	7	16	65.47	cps
Neodymium	150-2	10	7	17	11	45.82	cps
Nickel	60-2	2777	2570	2717	2688	3.96	cps
Phosphorus	31-2	385610	385540	388495	386548	0.44	cps
Potassium	39-2	15731860	15976561	15898980	15869134	0.79	cps
Rhodium	103-1	2346591	2287921	2331081	2321864	1.31	cps
Rhodium	103-2	1317907	1300616	1309694	1309406	0.66	cps
Scandium	45-1	1370061	1356542	1346576	1357726	0.87	cps
Scandium	45-2	123162	121220	123370	122584	0.97	cps
Selenium	82-1	111	79	77	89	21.36	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	103	90	93	96	7.26	cps
Silicon	28-1	195516	191956	192687	193386	0.97	cps
Silver	107-1	277	247	330	284	14.84	cps
Silver	109-1	223	230	293	249	15.52	cps
Sodium	23-2	21308171	21196375	21350761	21285103	0.37	cps
Strontium	86-1	40195	40318	40576	40363	0.48	cps
Strontium	88-1	352622	354972	359484	355693	0.98	cps
Sulfur	34-1	1413765	1392018	1415252	1407012	0.92	cps
Terbium	159-1	4051551	4083115	4062930	4065865	0.39	cps
Terbium	159-2	2590719	2564093	2625344	2593385	1.18	cps
Thallium	203-1	277	290	323	297	8.10	cps
Thallium	205-1	697	663	793	718	9.41	cps
Tin	118-1	1550	1660	1610	1607	3.43	cps
Titanium	47-1	869824	861146	865246	865405	0.50	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSA025 Instrumnet Name : P7
Client Sample ID : ICSA025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:19:53 DataFile Name : 015ICSA.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	220	207	243	223	8.31	cps
Vanadium	51-2	267	250	267	261	3.69	cps
Yttrium	89-1	4501634	4440094	4502123	4481284	0.80	cps
Yttrium	89-2	1150452	1137245	1148808	1145502	0.63	cps
Zinc	66-2	2604	2734	2630	2656	2.59	cps
Zirconium	90-1	413	400	437	417	4.45	cps
Zirconium	91-1	97	107	113	106	7.94	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSAB025 Instrumnet Name : P7
Client Sample ID : ICSAB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:23:05 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	7525034	7546006	7548171	7539737	0.17	cps
Antimony	121-1	82169	83085	84855	83370	1.64	cps
Arsenic	75-2	3694	3280	3454	3476	5.97	cps
Barium	135-1	22381	22658	23105	22715	1.61	cps
Barium	137-1	38017	40437	38967	39140	3.12	cps
Beryllium	9-1	14867	15064	15137	15023	0.93	cps
Bismuth	209-1	2329516	2404403	2329683	2354534	1.83	cps
Bismuth	209-2	1831477	1848694	1845140	1841771	0.49	cps
Bromine	81-1	6962	6858	7052	6957	1.39	cps
Bromine	81-2	123	150	103	126	18.64	cps
Cadmium	108-1	1887	2094	2037	2006	5.33	cps
Cadmium	106-1	3721	3891	3774	3795	2.29	cps
Cadmium	111-1	20860	20638	20245	20581	1.51	cps
Calcium	43-1	1024150	1032245	1037005	1031133	0.63	cps
Calcium	44-1	15563201	15568415	15629864	15587160	0.24	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	54070	53581	54244	53965	0.64	cps
Cobalt	59-2	42878	42564	42794	42745	0.38	cps
Copper	63-2	37347	37721	37276	37448	0.64	cps
Dysprosium	156-1	27	17	30	24	28.38	cps
Dysprosium	156-2	23	13	13	17	34.65	cps
Erbium	164-1	33	27	30	30	11.10	cps
Erbium	164-2	30	17	7	18	65.83	cps
Gadolinium	160-1	27	33	30	30	11.10	cps
Gadolinium	160-2	87	123	113	108	17.58	cps
Holmium	165-1	4051345	4102072	4065147	4072854	0.64	cps
Holmium	165-2	2665636	2730851	2716001	2704163	1.26	cps
Indium	115-1	2755278	2809584	2801953	2788939	1.05	cps
Indium	115-2	885359	891663	893723	890248	0.49	cps
Iron	56-2	118604995	119485382	118813612	118967996	0.39	cps
Iron	57-2	2958281	2984334	2956273	2966296	0.53	cps
Iron	54-2	6540282	6536402	6601423	6559369	0.56	cps
Krypton	83-1	53	67	90	70	26.51	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSAB025 Instrumnet Name : P7
Client Sample ID : ICSAB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:23:05 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	82632	84658	85704	84331	1.85	cps
Lead	207-1	69871	69825	70300	69999	0.37	cps
Lead	208-1	321842	326371	326896	325036	0.85	cps
Lithium	6-1	699893	704766	695024	699894	0.70	cps
Magnesium	24-2	13004255	12898642	12893202	12932033	0.48	cps
Manganese	55-2	29506	29479	29790	29591	0.58	cps
Molybdenum	94-1	2228210	2263894	2301162	2264422	1.61	cps
Molybdenum	95-1	3715629	3786418	3776076	3759374	1.02	cps
Molybdenum	96-1	4002706	4114390	4089115	4068737	1.44	cps
Molybdenum	97-1	2368235	2382103	2409331	2386556	0.88	cps
Molybdenum	98-1	6099052	6036882	6194714	6110216	1.30	cps
Neodymium	150-1	20	17	23	20	16.65	cps
Neodymium	150-2	10	10	23	14	53.28	cps
Nickel	60-2	12235	12555	12195	12328	1.60	cps
Phosphorus	31-2	393356	395156	396357	394956	0.38	cps
Potassium	39-2	16058730	15928997	16058974	16015567	0.47	cps
Rhodium	103-1	2419701	2390699	2363697	2391366	1.17	cps
Rhodium	103-2	1337112	1334796	1334823	1335577	0.10	cps
Scandium	45-1	1436977	1467511	1414380	1439623	1.85	cps
Scandium	45-2	125743	126429	127144	126439	0.55	cps
Selenium	82-1	1205	1209	1246	1220	1.86	cps
Selenium	77-2	87	120	93	100	17.64	cps
Selenium	78-2	453	507	433	464	8.16	cps
Silicon	28-1	296251	300313	301717	299427	0.95	cps
Silver	107-1	84970	87682	87890	86848	1.88	cps
Silver	109-1	80629	82677	82499	81935	1.38	cps
Sodium	23-2	21766155	21731406	21905936	21801166	0.42	cps
Strontium	86-1	102647	103523	103497	103222	0.48	cps
Strontium	88-1	896024	909987	906761	904257	0.81	cps
Sulfur	34-1	1463878	1486658	1470122	1473553	0.80	cps
Terbium	159-1	4138920	4250775	4215061	4201585	1.36	cps
Terbium	159-2	2638233	2656363	2668193	2654263	0.57	cps
Thallium	203-1	94162	95273	96153	95196	1.05	cps
Thallium	205-1	224645	225241	229223	226370	1.10	cps
Tin	118-1	174405	177448	178181	176678	1.13	cps
Titanium	47-1	919977	929395	928808	926060	0.57	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : ICSAB025 Instrumnet Name : P7
Client Sample ID : ICSAB025 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:23:05 DataFile Name : 016ICSB.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	622156	628300	630635	627030	0.70	cps
Vanadium	51-2	24694	24366	24490	24517	0.67	cps
Yttrium	89-1	4611281	4688080	4642402	4647255	0.83	cps
Yttrium	89-2	1166451	1169802	1180737	1172330	0.64	cps
Zinc	66-2	7522	7409	7305	7412	1.46	cps
Zirconium	90-1	317603	324020	323459	321694	1.10	cps
Zirconium	91-1	71687	71868	72692	72082	0.74	cps

LB Number :LB133015Operator :Jaswal

Lab Sample ID :CCV094Instrumnet Name :P7

Client Sample ID :CCV094Dilution Factor :1

Date & Time Acquired :2024-10-18 13:35:33DataFile Name :017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4136849	4043443	4087157	4089149	1.14	cps
Antimony	121-1	1970457	2015878	2011294	1999210	1.25	cps
Arsenic	75-2	82940	82725	83369	83011	0.39	cps
Barium	135-1	2827685	2824037	2898252	2849991	1.47	cps
Barium	137-1	4825636	4848299	4982617	4885518	1.74	cps
Beryllium	9-1	348353	357353	361865	355857	1.93	cps
Bismuth	209-1	2198259	2197944	2250477	2215560	1.36	cps
Bismuth	209-2	1525239	1524613	1519765	1523205	0.20	cps
Bromine	81-1	6298	5748	6151	6066	4.70	cps
Bromine	81-2	33	40	23	32	26.04	cps
Cadmium	108-1	34374	33786	33920	34026	0.91	cps
Cadmium	106-1	49818	49715	50778	50104	1.17	cps
Cadmium	111-1	446217	451081	454506	450601	0.92	cps
Calcium	43-1	2352094	2447096	2471356	2423515	2.60	cps
Calcium	44-1	37736194	38206025	39029271	38323830	1.71	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	673792	671940	678878	674870	0.53	cps
Cobalt	59-2	1061049	1054730	1062786	1059521	0.40	cps
Copper	63-2	7041944	7149398	7040829	7077391	0.88	cps
Dysprosium	156-1	113	107	120	113	5.88	cps
Dysprosium	156-2	183	207	203	198	6.38	cps
Erbium	164-1	103	80	53	79	31.71	cps
Erbium	164-2	57	80	40	59	34.12	cps
Gadolinium	160-1	97	60	87	81	23.37	cps
Gadolinium	160-2	163	130	187	160	17.80	cps
Holmium	165-1	3807453	3789705	3777410	3791523	0.40	cps
Holmium	165-2	2605556	2561152	2569559	2578756	0.91	cps
Indium	115-1	2593055	2584183	2619664	2598967	0.71	cps
Indium	115-2	827457	827760	827272	827496	0.03	cps
Iron	56-2	148178514	148825394	147793531	148265813	0.35	cps
Iron	57-2	3700968	3666988	3718296	3695417	0.71	cps
Iron	54-2	8112606	8006252	8042580	8053813	0.67	cps
Krypton	83-1	90	87	70	82	13.03	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCV094 Instrumnet Name : P7
Client Sample ID : CCV094 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:35:33 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	8387059	8450624	8711361	8516348	2.02	cps
Lead	207-1	7371909	7465620	7616629	7484719	1.65	cps
Lead	208-1	33588920	34317970	34812331	34239740	1.80	cps
Lithium	6-1	633814	636553	653799	641389	1.69	cps
Magnesium	24-2	33114830	32597418	32985379	32899209	0.82	cps
Manganese	55-2	5230994	5215207	5212839	5219680	0.19	cps
Molybdenum	94-1	6900302	7069843	7108122	7026089	1.57	cps
Molybdenum	95-1	10183680	10320717	10414747	10306381	1.13	cps
Molybdenum	96-1	11119442	11297610	11355184	11257412	1.09	cps
Molybdenum	97-1	6312744	6496871	6572185	6460600	2.07	cps
Molybdenum	98-1	16648579	16870569	16855154	16791434	0.74	cps
Neodymium	150-1	290	280	260	277	5.52	cps
Neodymium	150-2	63	80	70	71	11.80	cps
Nickel	60-2	256359	256023	256833	256405	0.16	cps
Phosphorus	31-2	39464	39518	39955	39646	0.68	cps
Potassium	39-2	20612389	20375368	20454082	20480613	0.59	cps
Rhodium	103-1	2185094	2190844	2198213	2191384	0.30	cps
Rhodium	103-2	1229383	1236442	1234496	1233440	0.30	cps
Scandium	45-1	1339999	1356914	1352732	1349882	0.65	cps
Scandium	45-2	120766	120669	121345	120926	0.30	cps
Selenium	82-1	24706	24994	25219	24973	1.03	cps
Selenium	77-2	2614	2414	2440	2489	4.36	cps
Selenium	78-2	8299	8329	8316	8315	0.18	cps
Silicon	28-1	1503063	1527058	1574762	1534961	2.38	cps
Silver	107-1	2183646	2255109	2254985	2231247	1.85	cps
Silver	109-1	2092676	2128404	2110944	2110674	0.85	cps
Sodium	23-2	55425898	54977071	56007144	55470038	0.93	cps
Strontium	86-1	15010085	15121019	15285296	15138800	0.91	cps
Strontium	88-1	132162928	134263521	134490545	133638998	0.96	cps
Sulfur	34-1	229193	232698	233416	231769	0.97	cps
Terbium	159-1	3868569	3922797	3979879	3923748	1.42	cps
Terbium	159-2	2495471	2469119	2508983	2491191	0.81	cps
Thallium	203-1	2082900	2088120	2128459	2099826	1.19	cps
Thallium	205-1	4991334	5036681	5191874	5073296	2.07	cps
Tin	118-1	1695768	1763021	1758611	1739133	2.16	cps
Titanium	47-1	2297540	2369074	2395171	2353928	2.15	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCV094 Instrumnet Name : P7
Client Sample ID : CCV094 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:35:33 DataFile Name : 017CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	7012202	7119582	7259239	7130341	1.74	cps
Vanadium	51-2	619935	614815	622105	618952	0.60	cps
Yttrium	89-1	4325475	4305847	4365985	4332436	0.71	cps
Yttrium	89-2	1105556	1100263	1103165	1102995	0.24	cps
Zinc	66-2	1208539	1201620	1212404	1207521	0.45	cps
Zirconium	90-1	3163640	3185578	3211452	3186890	0.75	cps
Zirconium	91-1	696776	712583	713164	707508	1.31	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCB094 Instrumnet Name : P7
Client Sample ID : CCB094 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:38:25 DataFile Name : 018CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	253	253	300	269	10.02	cps
Antimony	121-1	1333	1417	1507	1419	6.11	cps
Arsenic	75-2	20	17	0	12	87.67	cps
Barium	135-1	147	93	67	102	39.85	cps
Barium	137-1	170	127	120	139	19.55	cps
Beryllium	9-1	263	190	260	238	17.42	cps
Bismuth	209-1	2518866	2508907	2523671	2517148	0.30	cps
Bismuth	209-2	1960260	2010579	1941007	1970615	1.82	cps
Bromine	81-1	6475	6458	6408	6447	0.54	cps
Bromine	81-2	27	23	23	24	7.89	cps
Cadmium	108-1	3	3	3	3	0.00	cps
Cadmium	106-1	1347	1223	1357	1309	5.67	cps
Cadmium	111-1	1004	891	991	962	6.43	cps
Calcium	43-1	263	193	183	213	20.43	cps
Calcium	44-1	4501	4331	4064	4298	5.12	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	343	260	350	318	15.78	cps
Cobalt	59-2	47	70	73	63	22.94	cps
Copper	63-2	1220	1003	1197	1140	10.43	cps
Dysprosium	156-1	3	10	3	6	69.34	cps
Dysprosium	156-2	3	3	0	2	86.60	cps
Erbium	164-1	10	23	13	16	44.60	cps
Erbium	164-2	10	7	0	6	91.64	cps
Gadolinium	160-1	3	10	10	8	49.52	cps
Gadolinium	160-2	113	97	107	106	7.94	cps
Holmium	165-1	3924023	3960872	4033731	3972875	1.41	cps
Holmium	165-2	2667085	2730317	2625490	2674297	1.97	cps
Indium	115-1	2922673	2905943	2914591	2914402	0.29	cps
Indium	115-2	917127	922628	907115	915623	0.86	cps
Iron	56-2	12572	13389	14337	13433	6.58	cps
Iron	57-2	623	700	727	683	7.85	cps
Iron	54-2	1103	1257	1290	1217	8.18	cps
Krypton	83-1	63	63	70	66	5.87	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCB094 Instrumnet Name : P7
Client Sample ID : CCB094 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:38:25 DataFile Name : 018CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	1007	867	607	827	24.55	cps
Lead	207-1	750	703	687	713	4.60	cps
Lead	208-1	3747	3137	2920	3268	13.12	cps
Lithium	6-1	703805	712124	712855	709595	0.71	cps
Magnesium	24-2	2577	2574	2810	2654	5.11	cps
Manganese	55-2	323	297	343	321	7.29	cps
Molybdenum	94-1	810	807	743	787	4.78	cps
Molybdenum	95-1	680	623	650	651	4.35	cps
Molybdenum	96-1	1033	790	727	850	19.05	cps
Molybdenum	97-1	423	373	420	406	6.89	cps
Molybdenum	98-1	1197	1060	1027	1094	8.23	cps
Neodymium	150-1	7	3	3	4	43.40	cps
Neodymium	150-2	3	0	3	2	86.60	cps
Nickel	60-2	87	93	87	89	4.33	cps
Phosphorus	31-2	127	130	97	118	15.59	cps
Potassium	39-2	32632	32819	33337	32929	1.11	cps
Rhodium	103-1	2526919	2507005	2516128	2516684	0.40	cps
Rhodium	103-2	1401945	1434733	1414436	1417038	1.17	cps
Scandium	45-1	1411611	1445784	1420983	1426126	1.24	cps
Scandium	45-2	125014	127568	125665	126082	1.05	cps
Selenium	82-1	76	87	118	94	22.91	cps
Selenium	77-2	7	3	0	3	100.05	cps
Selenium	78-2	110	63	83	86	27.37	cps
Silicon	28-1	168012	166948	166608	167189	0.44	cps
Silver	107-1	357	297	190	281	30.03	cps
Silver	109-1	267	207	220	231	13.63	cps
Sodium	23-2	17814	17297	17156	17422	1.99	cps
Strontium	86-1	757	557	600	638	16.50	cps
Strontium	88-1	5341	3897	3497	4245	22.85	cps
Sulfur	34-1	104924	105514	105286	105241	0.28	cps
Terbium	159-1	4103019	4117176	4132029	4117408	0.35	cps
Terbium	159-2	2548095	2665095	2614153	2609114	2.25	cps
Thallium	203-1	377	333	293	334	12.46	cps
Thallium	205-1	907	670	647	741	19.41	cps
Tin	118-1	917	1017	970	968	5.17	cps
Titanium	47-1	183	150	130	154	17.45	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCB094

Instrumnet Name :

P7

Client Sample ID :

CCB094

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 13:38:25

DataFile Name :

018CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	283	210	217	237	17.14	cps
Vanadium	51-2	27	33	27	29	13.31	cps
Yttrium	89-1	4618047	4614770	4650651	4627823	0.43	cps
Yttrium	89-2	1163501	1175427	1158100	1165676	0.76	cps
Zinc	66-2	180	220	233	211	13.15	cps
Zirconium	90-1	1083	900	1013	999	9.26	cps
Zirconium	91-1	210	253	277	247	13.72	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2816-01A Instrumnet Name : P7
Client Sample ID : MYD4J8ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:52:29 DataFile Name : 020SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3430224	3433901	3418050	3427392	0.24	cps
Antimony	121-1	3520	3477	3550	3516	1.05	cps
Arsenic	75-2	42888	43039	42898	42942	0.20	cps
Barium	135-1	274864	284568	279609	279680	1.74	cps
Barium	137-1	475989	484266	485013	481756	1.04	cps
Beryllium	9-1	1013	1153	1030	1066	7.18	cps
Bismuth	209-1	2484661	2490242	2430613	2468505	1.33	cps
Bismuth	209-2	1875251	1904709	1859399	1879786	1.22	cps
Bromine	81-1	7295	7219	7319	7277	0.72	cps
Bromine	81-2	130	150	163	148	11.36	cps
Cadmium	108-1	140	147	127	138	7.39	cps
Cadmium	106-1	1613	1473	1437	1508	6.18	cps
Cadmium	111-1	1835	1745	1720	1767	3.40	cps
Calcium	43-1	153430	157103	158872	156468	1.77	cps
Calcium	44-1	2356335	2454066	2391821	2400741	2.06	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	58305	59025	58556	58629	0.62	cps
Cobalt	59-2	72943	73428	73228	73199	0.33	cps
Copper	63-2	175239	170465	170397	172034	1.61	cps
Dysprosium	156-1	33223	34205	34101	33843	1.59	cps
Dysprosium	156-2	25356	25577	24892	25275	1.38	cps
Erbium	164-1	29267	29919	29548	29578	1.10	cps
Erbium	164-2	18582	18856	18796	18745	0.77	cps
Gadolinium	160-1	31051	32117	32030	31733	1.86	cps
Gadolinium	160-2	21944	22251	22017	22071	0.73	cps
Holmium	165-1	4069989	4189552	4078529	4112690	1.62	cps
Holmium	165-2	2717431	2758225	2694422	2723359	1.19	cps
Indium	115-1	2919037	2962232	3004964	2962078	1.45	cps
Indium	115-2	924280	937808	924608	928898	0.83	cps
Iron	56-2	110633058	110692522	109629052	110318211	0.54	cps
Iron	57-2	2744188	2694966	2729697	2722950	0.93	cps
Iron	54-2	5950854	5994774	5984672	5976767	0.38	cps
Krypton	83-1	117	117	137	123	9.36	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2816-01A Instrumnet Name : P7
Client Sample ID : MYD4J8ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 13:52:29 DataFile Name : 020SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	142110	147199	142931	144080	1.90	cps
Lead	207-1	123804	126609	125746	125386	1.15	cps
Lead	208-1	567667	577038	576401	573702	0.91	cps
Lithium	6-1	729611	749386	745517	741505	1.41	cps
Magnesium	24-2	2082587	2094768	2154497	2110617	1.82	cps
Manganese	55-2	1711605	1720700	1695384	1709229	0.75	cps
Molybdenum	94-1	23769	23716	23833	23773	0.25	cps
Molybdenum	95-1	2064	2140	2124	2109	1.91	cps
Molybdenum	96-1	7392	7502	7489	7461	0.81	cps
Molybdenum	97-1	1307	1447	1513	1422	7.42	cps
Molybdenum	98-1	3354	3437	3247	3346	2.85	cps
Neodymium	150-1	54428	54893	54391	54571	0.51	cps
Neodymium	150-2	29765	29234	29975	29658	1.29	cps
Nickel	60-2	19139	19863	18939	19314	2.52	cps
Phosphorus	31-2	9270	9623	9596	9496	2.07	cps
Potassium	39-2	1066730	1068735	1066415	1067293	0.12	cps
Rhodium	103-1	2517620	2578764	2561324	2552569	1.23	cps
Rhodium	103-2	1417644	1442538	1408982	1423055	1.22	cps
Scandium	45-1	1588495	1627837	1630305	1615546	1.45	cps
Scandium	45-2	140582	143643	139384	141203	1.56	cps
Selenium	82-1	99	138	131	123	16.70	cps
Selenium	77-2	70	93	53	72	27.82	cps
Selenium	78-2	147	147	143	146	1.32	cps
Silicon	28-1	16256841	16585008	16645908	16495919	1.27	cps
Silver	107-1	1437	1533	1607	1526	5.59	cps
Silver	109-1	1327	1420	1440	1396	4.33	cps
Sodium	23-2	57905	58571	58825	58434	0.81	cps
Strontium	86-1	77100	80157	80019	79092	2.18	cps
Strontium	88-1	688083	704105	708315	700168	1.52	cps
Sulfur	34-1	115254	113830	110992	113359	1.91	cps
Terbium	159-1	4139653	4198556	4207700	4181970	0.88	cps
Terbium	159-2	2652195	2654711	2626039	2644315	0.60	cps
Thallium	203-1	1190	1167	1047	1135	6.78	cps
Thallium	205-1	2777	2657	2787	2740	2.64	cps
Tin	118-1	5925	6355	5958	6079	3.94	cps
Titanium	47-1	628863	646402	648510	641259	1.68	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2816-01A

Instrumnet Name :

P7

Client Sample ID :

MYD4J8ARE

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 13:52:29

DataFile Name :

020SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	9837	9810	9907	9852	0.51	cps
Vanadium	51-2	208618	210844	211751	210405	0.77	cps
Yttrium	89-1	5171386	5328341	5188910	5229546	1.64	cps
Yttrium	89-2	1296913	1306285	1282627	1295275	0.92	cps
Zinc	66-2	54041	54178	55034	54418	0.99	cps
Zirconium	90-1	58964	60912	60842	60240	1.83	cps
Zirconium	91-1	12969	13476	13510	13318	2.27	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2823-01ADLX4 Instrumnet Name : P7
Client Sample ID : MYD5X1ARE Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:06:11 DataFile Name : 021SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	649270	644741	645161	646391	0.39	cps
Antimony	121-1	763	820	937	840	10.52	cps
Arsenic	75-2	1713	1783	1807	1768	2.75	cps
Barium	135-1	250182	249475	252734	250797	0.68	cps
Barium	137-1	434829	436921	436717	436156	0.26	cps
Beryllium	9-1	247	263	253	254	3.30	cps
Bismuth	209-1	2103273	2456860	2386011	2315381	8.08	cps
Bismuth	209-2	1915519	1929111	1908316	1917649	0.55	cps
Bromine	81-1	6728	7285	6425	6813	6.41	cps
Bromine	81-2	50	83	77	70	25.20	cps
Cadmium	108-1	40	33	40	38	10.19	cps
Cadmium	106-1	1127	1323	1167	1206	8.62	cps
Cadmium	111-1	951	1136	1005	1031	9.26	cps
Calcium	43-1	40581	40858	41092	40844	0.63	cps
Calcium	44-1	658854	660926	663559	661113	0.36	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	17854	17704	17010	17522	2.57	cps
Cobalt	59-2	14400	14437	13943	14260	1.93	cps
Copper	63-2	30752	30655	30424	30610	0.55	cps
Dysprosium	156-1	7769	7485	7385	7547	2.64	cps
Dysprosium	156-2	5935	5788	5701	5808	2.03	cps
Erbium	164-1	6051	6271	6235	6186	1.91	cps
Erbium	164-2	3924	4154	3991	4023	2.94	cps
Gadolinium	160-1	6775	7029	6692	6832	2.57	cps
Gadolinium	160-2	4641	4694	4931	4755	3.25	cps
Holmium	165-1	3459191	4142199	3981593	3860995	9.25	cps
Holmium	165-2	2648098	2696175	2696766	2680347	1.04	cps
Indium	115-1	2543140	2951274	2867614	2787343	7.73	cps
Indium	115-2	930493	931969	922415	928292	0.55	cps
Iron	56-2	19070773	18870051	18876844	18939222	0.60	cps
Iron	57-2	493306	487092	490127	490175	0.63	cps
Iron	54-2	1079941	1075576	1074051	1076523	0.28	cps
Krypton	83-1	130	83	77	97	30.06	cps

LB Number :	LB133015	Operator :	Jaswal
Lab Sample ID :	P2823-01ADLX4	Instrumnet Name :	P7
Client Sample ID :	MYD5X1ARE	Dilution Factor :	4
Date & Time Acquired :	2024-10-18 14:06:11	DataFile Name :	021SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	58425	57852	59473	58583	1.40	cps
Lead	207-1	51979	52882	52481	52447	0.86	cps
Lead	208-1	235221	236608	236952	236260	0.39	cps
Lithium	6-1	619551	716524	711967	682680	8.02	cps
Magnesium	24-2	452556	452150	449825	451510	0.33	cps
Manganese	55-2	371435	372329	370008	371257	0.32	cps
Molybdenum	94-1	5715	5544	5581	5613	1.59	cps
Molybdenum	95-1	467	493	463	474	3.47	cps
Molybdenum	96-1	1583	1680	1727	1663	4.40	cps
Molybdenum	97-1	293	380	277	317	17.52	cps
Molybdenum	98-1	667	793	793	751	9.74	cps
Neodymium	150-1	12779	12679	12559	12672	0.87	cps
Neodymium	150-2	6895	6915	6885	6898	0.22	cps
Nickel	60-2	6298	6275	6415	6329	1.19	cps
Phosphorus	31-2	1890	1727	1790	1802	4.57	cps
Potassium	39-2	294364	294419	292422	293735	0.39	cps
Rhodium	103-1	2226566	2613573	2557868	2466002	8.48	cps
Rhodium	103-2	1437635	1442893	1427772	1436100	0.53	cps
Scandium	45-1	1289842	1515273	1512676	1439264	8.99	cps
Scandium	45-2	132405	133525	131030	132320	0.94	cps
Selenium	82-1	66	72	112	84	29.88	cps
Selenium	77-2	10	20	17	16	32.73	cps
Selenium	78-2	147	107	120	124	16.37	cps
Silicon	28-1	3698923	3764117	3758317	3740452	0.96	cps
Silver	107-1	490	480	550	507	7.47	cps
Silver	109-1	493	417	490	467	9.29	cps
Sodium	23-2	31612	31328	31108	31350	0.81	cps
Strontium	86-1	24173	24063	24086	24108	0.24	cps
Strontium	88-1	208056	212414	210640	210370	1.04	cps
Sulfur	34-1	100215	101816	100903	100978	0.80	cps
Terbium	159-1	3569001	4152421	4054799	3925407	7.96	cps
Terbium	159-2	2639708	2640854	2654561	2645041	0.31	cps
Thallium	203-1	307	307	273	296	6.51	cps
Thallium	205-1	770	743	740	751	2.19	cps
Tin	118-1	4597	4644	4444	4562	2.29	cps
Titanium	47-1	138258	138392	138342	138330	0.05	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2823-01ADLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5X1ARE

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:06:11

DataFile Name :

021SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	3490	3531	3567	3529	1.09	cps
Vanadium	51-2	107214	106593	105774	106527	0.68	cps
Yttrium	89-1	4183484	4897724	4811209	4630806	8.42	cps
Yttrium	89-2	1215606	1220588	1218837	1218343	0.21	cps
Zinc	66-2	16956	16823	16683	16821	0.81	cps
Zirconium	90-1	14350	14791	14701	14614	1.59	cps
Zirconium	91-1	3120	3327	3257	3235	3.25	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2824-01A Instrumnet Name : P7
Client Sample ID : MYD5K5ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:09:32 DataFile Name : 022SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	4221389	4225733	4142092	4196405	1.12	cps
Antimony	121-1	5318	5604	5414	5446	2.68	cps
Arsenic	75-2	31777	31356	31276	31470	0.86	cps
Barium	135-1	518165	534129	526497	526263	1.52	cps
Barium	137-1	912972	909676	906278	909642	0.37	cps
Beryllium	9-1	997	927	973	966	3.69	cps
Bismuth	209-1	2393561	2388947	2430857	2404455	0.96	cps
Bismuth	209-2	1824855	1818763	1833815	1825811	0.41	cps
Bromine	81-1	8783	8639	8869	8764	1.33	cps
Bromine	81-2	407	393	353	384	7.22	cps
Cadmium	108-1	153	133	83	123	29.24	cps
Cadmium	106-1	1630	1683	1567	1627	3.59	cps
Cadmium	111-1	1867	1912	1898	1892	1.21	cps
Calcium	43-1	104493	107504	107020	106339	1.52	cps
Calcium	44-1	1592231	1622338	1615867	1610145	0.98	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	210850	207362	208082	208765	0.88	cps
Cobalt	59-2	177267	174886	173938	175364	0.98	cps
Copper	63-2	303442	302577	299507	301842	0.69	cps
Dysprosium	156-1	31455	32805	31933	32064	2.13	cps
Dysprosium	156-2	23981	24448	24077	24169	1.02	cps
Erbium	164-1	23817	23857	24264	23980	1.03	cps
Erbium	164-2	15702	15242	15335	15426	1.58	cps
Gadolinium	160-1	28115	28068	28128	28104	0.11	cps
Gadolinium	160-2	19327	18686	19163	19059	1.75	cps
Holmium	165-1	4047532	4159737	4061143	4089471	1.50	cps
Holmium	165-2	2660909	2588241	2650955	2633368	1.50	cps
Indium	115-1	2886207	2895933	2859213	2880451	0.66	cps
Indium	115-2	912231	889881	902576	901563	1.24	cps
Iron	56-2	231347617	230797643	228609163	230251474	0.63	cps
Iron	57-2	5761536	5659151	5653569	5691418	1.07	cps
Iron	54-2	12667196	12547180	12536926	12583767	0.58	cps
Krypton	83-1	113	103	87	101	13.32	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2824-01A Instrumnet Name : P7
Client Sample ID : MYD5K5ARE Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:09:32 DataFile Name : 022SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	178039	184317	181565	181307	1.74	cps
Lead	207-1	162474	163005	160873	162117	0.68	cps
Lead	208-1	726500	733724	731607	730610	0.51	cps
Lithium	6-1	729836	733148	736770	733252	0.47	cps
Magnesium	24-2	4104118	4039079	4053263	4065487	0.84	cps
Manganese	55-2	9869311	9923682	9925864	9906286	0.32	cps
Molybdenum	94-1	24661	24688	25041	24797	0.86	cps
Molybdenum	95-1	4644	4744	4644	4677	1.23	cps
Molybdenum	96-1	11795	11648	10991	11478	3.73	cps
Molybdenum	97-1	2784	2887	2740	2804	2.69	cps
Molybdenum	98-1	7002	7385	6912	7100	3.54	cps
Neodymium	150-1	49289	50757	50738	50261	1.68	cps
Neodymium	150-2	26896	26779	26879	26851	0.24	cps
Nickel	60-2	55954	56248	56041	56081	0.27	cps
Phosphorus	31-2	8509	8322	7905	8246	3.75	cps
Potassium	39-2	1755487	1754920	1732124	1747511	0.76	cps
Rhodium	103-1	2570658	2619907	2549078	2579881	1.41	cps
Rhodium	103-2	1389337	1370507	1387371	1382405	0.75	cps
Scandium	45-1	1671956	1714219	1692317	1692830	1.25	cps
Scandium	45-2	143909	141211	143529	142883	1.02	cps
Selenium	82-1	855	847	859	854	0.69	cps
Selenium	77-2	133	117	117	122	7.87	cps
Selenium	78-2	373	450	350	391	13.38	cps
Silicon	28-1	12411491	12568796	12543359	12507882	0.68	cps
Silver	107-1	1290	1367	1407	1355	4.38	cps
Silver	109-1	973	1273	1217	1155	13.81	cps
Sodium	23-2	56985	56349	57537	56957	1.04	cps
Strontium	86-1	104286	107069	106229	105861	1.35	cps
Strontium	88-1	926044	926324	924985	925784	0.08	cps
Sulfur	34-1	105014	106847	104581	105481	1.14	cps
Terbium	159-1	4202968	4155350	4132010	4163443	0.87	cps
Terbium	159-2	2639276	2580054	2610845	2610058	1.13	cps
Thallium	203-1	1273	1540	1480	1431	9.78	cps
Thallium	205-1	3250	3410	3447	3369	3.10	cps
Tin	118-1	9073	8976	9046	9032	0.55	cps
Titanium	47-1	478514	486747	478273	481178	1.00	cps

LB Number :LB133015

Operator :Jaswal

Lab Sample ID :P2824-01A

Instrumnet Name :P7

Client Sample ID :MYD5K5ARE

Dilution Factor :1

Date & Time Acquired :2024-10-18 14:09:32

DataFile Name :022SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	11855	12192	12176	12074	1.57	cps
Vanadium	51-2	386263	387106	383779	385716	0.45	cps
Yttrium	89-1	5103471	5103015	5137738	5114742	0.39	cps
Yttrium	89-2	1259667	1236146	1242683	1246165	0.97	cps
Zinc	66-2	79021	78495	78944	78820	0.36	cps
Zirconium	90-1	59490	60829	59898	60072	1.14	cps
Zirconium	91-1	13646	13460	13410	13505	0.92	cps

LB Number :LB133015Operator :Jaswal

Lab Sample ID :P2827-10DLX4Instrumnet Name :P7

Client Sample ID :MYD5Y3Dilution Factor :4

Date & Time Acquired :2024-10-18 14:23:04DataFile Name :023SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1065610	1055381	1044095	1055029	1.02	cps
Antimony	121-1	293	270	260	274	6.23	cps
Arsenic	75-2	1377	1463	1487	1442	4.02	cps
Barium	135-1	156610	155248	155653	155837	0.45	cps
Barium	137-1	269829	270786	268847	269821	0.36	cps
Beryllium	9-1	423	363	473	420	13.11	cps
Bismuth	209-1	2387906	2343011	2378446	2369788	1.00	cps
Bismuth	209-2	1825995	1841002	1604185	1757061	7.55	cps
Bromine	81-1	7795	7609	7582	7662	1.52	cps
Bromine	81-2	177	167	143	162	10.54	cps
Cadmium	108-1	50	20	23	31	52.86	cps
Cadmium	106-1	1210	1280	1253	1248	2.83	cps
Cadmium	111-1	962	975	1004	981	2.21	cps
Calcium	43-1	49199	48487	48377	48688	0.92	cps
Calcium	44-1	787362	781716	776100	781726	0.72	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	20824	21058	20748	20877	0.77	cps
Cobalt	59-2	18588	18258	18178	18341	1.19	cps
Copper	63-2	31914	31466	31066	31482	1.35	cps
Dysprosium	156-1	12989	12779	12853	12874	0.83	cps
Dysprosium	156-2	10317	9880	10080	10093	2.17	cps
Erbium	164-1	10110	10217	10341	10223	1.13	cps
Erbium	164-2	6435	6565	6718	6573	2.16	cps
Gadolinium	160-1	11848	11895	12289	12011	2.01	cps
Gadolinium	160-2	8403	8226	8096	8241	1.87	cps
Holmium	165-1	3918800	3912433	3888909	3906714	0.40	cps
Holmium	165-2	2601820	2592835	2566776	2587144	0.70	cps
Indium	115-1	2892722	2812963	2880030	2861905	1.50	cps
Indium	115-2	886756	896608	883079	888814	0.79	cps
Iron	56-2	22411991	22251749	22169881	22277874	0.55	cps
Iron	57-2	583909	578420	571754	578028	1.05	cps
Iron	54-2	1281935	1275556	1263895	1273795	0.72	cps
Krypton	83-1	80	70	73	74	6.84	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3 Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:23:04 DataFile Name : 023SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	35861	36519	37134	36504	1.74	cps
Lead	207-1	31510	31119	31430	31353	0.66	cps
Lead	208-1	141967	143246	144370	143194	0.84	cps
Lithium	6-1	712119	705085	700847	706017	0.81	cps
Magnesium	24-2	687802	678442	678584	681609	0.79	cps
Manganese	55-2	544930	548277	543464	545557	0.45	cps
Molybdenum	94-1	5848	5725	5684	5752	1.48	cps
Molybdenum	95-1	390	333	407	377	10.21	cps
Molybdenum	96-1	1643	1617	1583	1615	1.86	cps
Molybdenum	97-1	233	270	233	246	8.62	cps
Molybdenum	98-1	537	607	563	569	6.21	cps
Neodymium	150-1	22879	23106	23196	23060	0.71	cps
Neodymium	150-2	12602	12779	12779	12720	0.80	cps
Nickel	60-2	6902	6675	6481	6686	3.15	cps
Phosphorus	31-2	2564	2370	2490	2475	3.94	cps
Potassium	39-2	292342	291271	289451	291021	0.50	cps
Rhodium	103-1	2545972	2551639	2519322	2538978	0.68	cps
Rhodium	103-2	1369961	1384724	1354046	1369577	1.12	cps
Scandium	45-1	1506429	1496062	1478912	1493801	0.93	cps
Scandium	45-2	127686	129283	125366	127445	1.55	cps
Selenium	82-1	116	107	118	114	5.45	cps
Selenium	77-2	17	33	17	22	43.28	cps
Selenium	78-2	113	83	80	92	19.91	cps
Silicon	28-1	4271213	4316942	4292024	4293393	0.53	cps
Silver	107-1	233	290	253	259	11.10	cps
Silver	109-1	187	243	253	228	15.78	cps
Sodium	23-2	21388	20594	20794	20925	1.97	cps
Strontium	86-1	33174	33234	33114	33174	0.18	cps
Strontium	88-1	289639	291722	290176	290512	0.37	cps
Sulfur	34-1	100484	97277	94728	97496	2.96	cps
Terbium	159-1	4020081	4016851	4023029	4019987	0.08	cps
Terbium	159-2	2551099	2524586	2487596	2521094	1.27	cps
Thallium	203-1	443	333	327	368	17.82	cps
Thallium	205-1	837	837	880	851	2.94	cps
Tin	118-1	3084	2980	2987	3017	1.92	cps
Titanium	47-1	54096	55294	54862	54751	1.11	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-10DLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:23:04

DataFile Name :

023SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	3064	3074	2987	3041	1.56	cps
Vanadium	51-2	38148	38475	37811	38145	0.87	cps
Yttrium	89-1	4846731	4835499	4792690	4824974	0.59	cps
Yttrium	89-2	1178504	1186075	1175613	1180064	0.46	cps
Zinc	66-2	11878	11898	11434	11737	2.23	cps
Zirconium	90-1	14861	15041	14814	14905	0.80	cps
Zirconium	91-1	3340	3277	3370	3329	1.43	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-11DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3D Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:26:25 DataFile Name : 024SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1062480	1062393	1056985	1060619	0.30	cps
Antimony	121-1	293	290	277	287	3.08	cps
Arsenic	75-2	1433	1443	1437	1438	0.35	cps
Barium	135-1	154501	156344	154275	155040	0.73	cps
Barium	137-1	267924	268857	267573	268118	0.25	cps
Beryllium	9-1	467	530	447	481	9.04	cps
Bismuth	209-1	2354500	2403634	2304604	2354246	2.10	cps
Bismuth	209-2	1825716	1819462	1832734	1825971	0.36	cps
Bromine	81-1	7355	7292	7435	7361	0.98	cps
Bromine	81-2	163	160	147	157	5.63	cps
Cadmium	108-1	20	13	43	26	61.64	cps
Cadmium	106-1	1430	1373	1273	1359	5.84	cps
Cadmium	111-1	1120	1072	951	1048	8.33	cps
Calcium	43-1	48273	48424	48497	48398	0.24	cps
Calcium	44-1	776487	778012	774832	776444	0.20	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	23662	24580	23432	23891	2.54	cps
Cobalt	59-2	18378	18381	18178	18312	0.64	cps
Copper	63-2	31259	31767	31466	31498	0.81	cps
Dysprosium	156-1	13150	13113	12846	13036	1.27	cps
Dysprosium	156-2	10367	10007	10154	10176	1.78	cps
Erbium	164-1	10394	10554	10167	10372	1.87	cps
Erbium	164-2	6865	6592	6985	6814	2.96	cps
Gadolinium	160-1	12302	11999	11952	12084	1.57	cps
Gadolinium	160-2	8519	8456	8453	8476	0.44	cps
Holmium	165-1	3972882	3954238	3895486	3940869	1.02	cps
Holmium	165-2	2597353	2595122	2592015	2594830	0.10	cps
Indium	115-1	2793795	2807894	2771746	2791145	0.65	cps
Indium	115-2	893081	888968	888894	890314	0.27	cps
Iron	56-2	22500881	22496081	22121763	22372909	0.97	cps
Iron	57-2	583103	576768	576167	578679	0.66	cps
Iron	54-2	1283371	1267952	1272446	1274590	0.62	cps
Krypton	83-1	77	77	87	80	7.22	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-11DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3D Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:26:25 DataFile Name : 024SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	36796	37425	37315	37179	0.90	cps
Lead	207-1	31530	31580	32034	31715	0.88	cps
Lead	208-1	144696	145797	145605	145366	0.40	cps
Lithium	6-1	697725	696634	690444	694934	0.57	cps
Magnesium	24-2	692069	688615	678534	686406	1.02	cps
Manganese	55-2	541313	541573	541114	541334	0.04	cps
Molybdenum	94-1	5898	5968	5891	5919	0.72	cps
Molybdenum	95-1	353	400	370	374	6.32	cps
Molybdenum	96-1	1753	1757	1803	1771	1.58	cps
Molybdenum	97-1	263	270	330	288	12.76	cps
Molybdenum	98-1	657	663	647	656	1.28	cps
Neodymium	150-1	22855	23112	23253	23073	0.87	cps
Neodymium	150-2	12719	12369	12749	12612	1.68	cps
Nickel	60-2	8026	7112	7419	7519	6.19	cps
Phosphorus	31-2	2327	2430	2377	2378	2.17	cps
Potassium	39-2	294081	294068	289379	292510	0.93	cps
Rhodium	103-1	2501198	2510163	2479530	2496963	0.63	cps
Rhodium	103-2	1374284	1365250	1350217	1363251	0.89	cps
Scandium	45-1	1454645	1464988	1439899	1453177	0.87	cps
Scandium	45-2	126809	127145	127548	127167	0.29	cps
Selenium	82-1	142	132	90	121	22.77	cps
Selenium	77-2	27	33	20	27	24.99	cps
Selenium	78-2	70	123	110	101	27.45	cps
Silicon	28-1	4249522	4281545	4222964	4251344	0.69	cps
Silver	107-1	257	217	227	233	8.92	cps
Silver	109-1	230	193	223	216	9.06	cps
Sodium	23-2	22460	21959	22710	22376	1.71	cps
Strontium	86-1	33485	33645	33538	33556	0.24	cps
Strontium	88-1	293536	294217	292833	293529	0.24	cps
Sulfur	34-1	89926	89074	89215	89405	0.51	cps
Terbium	159-1	3989530	4050443	4017790	4019254	0.76	cps
Terbium	159-2	2540472	2504664	2529149	2524762	0.72	cps
Thallium	203-1	310	340	370	340	8.82	cps
Thallium	205-1	763	857	833	818	5.94	cps
Tin	118-1	2917	2924	2820	2887	2.00	cps
Titanium	47-1	53343	53591	53170	53368	0.40	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-11DLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3D

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:26:25

DataFile Name :

024SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	2994	3084	3054	3044	1.51	cps
Vanadium	51-2	38984	38476	38422	38627	0.80	cps
Yttrium	89-1	4755750	4688667	4672854	4705757	0.94	cps
Yttrium	89-2	1182052	1171506	1172573	1175377	0.49	cps
Zinc	66-2	11758	11645	11585	11662	0.76	cps
Zirconium	90-1	14544	15161	15161	14956	2.38	cps
Zirconium	91-1	3357	3197	3290	3282	2.45	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10LDLX20 Instrumnet Name : P7
Client Sample ID : MYD5Y3L Dilution Factor : 20
Date & Time Acquired : 2024-10-18 14:29:50 DataFile Name : 025SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	207749	207391	205578	206906	0.56	cps
Antimony	121-1	77	67	67	70	8.25	cps
Arsenic	75-2	333	293	303	310	6.71	cps
Barium	135-1	29875	29965	30860	30233	1.80	cps
Barium	137-1	52005	52630	52808	52481	0.80	cps
Beryllium	9-1	60	77	153	97	51.50	cps
Bismuth	209-1	2438126	2434328	2411278	2427910	0.60	cps
Bismuth	209-2	1905789	1971835	1864958	1914194	2.82	cps
Bromine	81-1	6708	6982	6855	6848	2.00	cps
Bromine	81-2	70	47	97	71	35.18	cps
Cadmium	108-1	7	10	3	7	50.03	cps
Cadmium	106-1	1140	1363	1257	1253	8.91	cps
Cadmium	111-1	820	981	904	902	8.91	cps
Calcium	43-1	9243	9320	9470	9344	1.23	cps
Calcium	44-1	149713	150861	152921	151165	1.08	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	4404	4441	4404	4416	0.48	cps
Cobalt	59-2	3780	3654	3460	3632	4.44	cps
Copper	63-2	6875	6645	6685	6735	1.83	cps
Dysprosium	156-1	2537	2480	2520	2512	1.16	cps
Dysprosium	156-2	2087	2084	1950	2040	3.82	cps
Erbium	164-1	1950	1853	2010	1938	4.08	cps
Erbium	164-2	1283	1260	1330	1291	2.76	cps
Gadolinium	160-1	2307	2330	2224	2287	2.45	cps
Gadolinium	160-2	1617	1603	1743	1655	4.67	cps
Holmium	165-1	3882923	3903340	3915826	3900696	0.43	cps
Holmium	165-2	2678542	2681687	2579033	2646421	2.21	cps
Indium	115-1	2817727	2877244	2874962	2856644	1.18	cps
Indium	115-2	914520	924944	905605	915023	1.06	cps
Iron	56-2	4483156	4449850	4381782	4438263	1.16	cps
Iron	57-2	116269	116262	113914	115481	1.18	cps
Iron	54-2	251814	250769	252164	251582	0.29	cps
Krypton	83-1	63	77	67	69	10.08	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10LDLX20 Instrumnet Name : P7
Client Sample ID : MYD5Y3L Dilution Factor : 20
Date & Time Acquired : 2024-10-18 14:29:50 DataFile Name : 025SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	7249	7429	7512	7397	1.82	cps
Lead	207-1	6198	6205	6338	6247	1.26	cps
Lead	208-1	28425	28682	29283	28797	1.53	cps
Lithium	6-1	691634	694448	697158	694413	0.40	cps
Magnesium	24-2	135097	135463	135067	135209	0.16	cps
Manganese	55-2	109390	107070	107074	107845	1.24	cps
Molybdenum	94-1	1023	1270	1230	1175	11.27	cps
Molybdenum	95-1	93	83	120	99	19.17	cps
Molybdenum	96-1	343	380	373	366	5.34	cps
Molybdenum	97-1	83	67	67	72	13.32	cps
Molybdenum	98-1	97	133	183	138	31.58	cps
Neodymium	150-1	4611	4424	4521	4519	2.07	cps
Neodymium	150-2	2484	2487	2410	2460	1.76	cps
Nickel	60-2	1467	1573	1477	1506	3.91	cps
Phosphorus	31-2	573	547	553	558	2.49	cps
Potassium	39-2	83349	81929	81302	82193	1.28	cps
Rhodium	103-1	2487232	2514830	2530428	2510830	0.87	cps
Rhodium	103-2	1412120	1433106	1398844	1414690	1.22	cps
Scandium	45-1	1439045	1456936	1456548	1450843	0.70	cps
Scandium	45-2	128338	128959	127097	128132	0.74	cps
Selenium	82-1	103	69	91	88	19.82	cps
Selenium	77-2	0	3	10	4	114.60	cps
Selenium	78-2	120	93	83	99	19.17	cps
Silicon	28-1	934482	951590	960047	948706	1.37	cps
Silver	107-1	87	73	113	91	22.36	cps
Silver	109-1	40	57	50	49	17.16	cps
Sodium	23-2	12468	12275	12218	12321	1.06	cps
Strontium	86-1	6258	6592	6712	6520	3.60	cps
Strontium	88-1	55842	55591	56923	56119	1.26	cps
Sulfur	34-1	85129	85688	86547	85788	0.83	cps
Terbium	159-1	4037449	3990526	4000290	4009422	0.62	cps
Terbium	159-2	2573031	2603487	2523328	2566616	1.58	cps
Thallium	203-1	97	97	137	110	20.99	cps
Thallium	205-1	210	217	297	241	20.00	cps
Tin	118-1	1243	1253	1317	1271	3.13	cps
Titanium	47-1	9940	10497	10510	10316	3.16	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-10LDLX20

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3L

Dilution Factor :

20

Date & Time Acquired :

2024-10-18 14:29:50

DataFile Name :

025SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	610	607	653	623	4.18	cps
Vanadium	51-2	7349	7349	7579	7425	1.79	cps
Yttrium	89-1	4603668	4677256	4669545	4650156	0.87	cps
Yttrium	89-2	1179288	1199072	1163149	1180503	1.52	cps
Zinc	66-2	2447	2434	2467	2449	0.69	cps
Zirconium	90-1	2970	3187	3074	3077	3.52	cps
Zirconium	91-1	680	727	670	692	4.37	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-12DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3S Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:33:11 DataFile Name : 026SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1098608	1088508	1090159	1092425	0.50	cps
Antimony	121-1	66101	66596	66020	66239	0.47	cps
Arsenic	75-2	2760	2850	2730	2780	2.25	cps
Barium	135-1	511293	515863	515614	514257	0.50	cps
Barium	137-1	881303	887225	889568	886032	0.48	cps
Beryllium	9-1	6488	6505	6611	6535	1.02	cps
Bismuth	209-1	2424312	2419533	2403857	2415900	0.44	cps
Bismuth	209-2	1844819	1857884	1899027	1867243	1.52	cps
Bromine	81-1	6765	7078	7212	7018	3.27	cps
Bromine	81-2	113	137	133	128	9.88	cps
Cadmium	108-1	637	543	573	584	8.15	cps
Cadmium	106-1	2084	2180	2150	2138	2.31	cps
Cadmium	111-1	9015	9275	9300	9196	1.71	cps
Calcium	43-1	50564	50520	51691	50925	1.30	cps
Calcium	44-1	824360	832023	822370	826251	0.62	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	69097	69847	68940	69295	0.70	cps
Cobalt	59-2	191680	192015	189986	191227	0.57	cps
Copper	63-2	89811	89411	89136	89453	0.38	cps
Dysprosium	156-1	13537	13974	13984	13831	1.85	cps
Dysprosium	156-2	10998	10931	10808	10912	0.88	cps
Erbium	164-1	11191	10834	10981	11002	1.63	cps
Erbium	164-2	7172	7072	7355	7200	2.00	cps
Gadolinium	160-1	13056	13280	13126	13154	0.87	cps
Gadolinium	160-2	9120	8943	8900	8987	1.30	cps
Holmium	165-1	3905818	3970422	3951569	3942603	0.84	cps
Holmium	165-2	2603777	2630160	2611974	2615304	0.52	cps
Indium	115-1	2872196	2848222	2866052	2862157	0.44	cps
Indium	115-2	901758	898144	902501	900801	0.26	cps
Iron	56-2	23848634	23948195	23653890	23816907	0.63	cps
Iron	57-2	610249	611036	610248	610511	0.07	cps
Iron	54-2	1352602	1348155	1337283	1346013	0.59	cps
Krypton	83-1	73	93	107	91	18.42	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-12DLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3S Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:33:11 DataFile Name : 026SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	50359	51052	51390	50934	1.03	cps
Lead	207-1	43178	43289	42994	43154	0.34	cps
Lead	208-1	197525	197918	198650	198031	0.29	cps
Lithium	6-1	692955	692006	689102	691354	0.29	cps
Magnesium	24-2	712828	714438	714531	713932	0.13	cps
Manganese	55-2	680428	680435	671086	677316	0.80	cps
Molybdenum	94-1	5941	5344	5995	5760	6.27	cps
Molybdenum	95-1	413	490	473	459	8.79	cps
Molybdenum	96-1	1740	1743	1837	1773	3.09	cps
Molybdenum	97-1	320	293	267	293	9.09	cps
Molybdenum	98-1	750	673	783	736	7.67	cps
Neodymium	150-1	24835	25416	25350	25200	1.26	cps
Neodymium	150-2	14000	13687	13653	13780	1.39	cps
Nickel	60-2	50444	50662	50341	50482	0.32	cps
Phosphorus	31-2	2714	2617	2554	2628	3.07	cps
Potassium	39-2	305778	303612	305230	304873	0.37	cps
Rhodium	103-1	2541744	2499127	2519620	2520164	0.85	cps
Rhodium	103-2	1371816	1389118	1391560	1384165	0.78	cps
Scandium	45-1	1455267	1483525	1452459	1463750	1.17	cps
Scandium	45-2	127699	127332	126162	127065	0.63	cps
Selenium	82-1	1082	1120	1043	1082	3.58	cps
Selenium	77-2	137	130	87	118	23.05	cps
Selenium	78-2	393	467	427	429	8.56	cps
Silicon	28-1	4240416	4281225	4205099	4242247	0.90	cps
Silver	107-1	38066	38752	38230	38349	0.93	cps
Silver	109-1	35847	36219	36299	36122	0.67	cps
Sodium	23-2	21375	21775	22443	21864	2.47	cps
Strontium	86-1	35225	34975	35072	35091	0.36	cps
Strontium	88-1	310320	309978	311168	310489	0.20	cps
Sulfur	34-1	86338	84304	84941	85195	1.22	cps
Terbium	159-1	4094381	4051964	4089687	4078677	0.57	cps
Terbium	159-2	2513902	2543720	2546781	2534801	0.72	cps
Thallium	203-1	30952	31206	31533	31230	0.93	cps
Thallium	205-1	74502	74761	74596	74620	0.18	cps
Tin	118-1	3254	3470	3177	3300	4.61	cps
Titanium	47-1	57525	57930	58492	57983	0.84	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-12DLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3S

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:33:11

DataFile Name :

026SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	3367	3354	3214	3312	2.57	cps
Vanadium	51-2	144299	144157	142068	143508	0.87	cps
Yttrium	89-1	4825164	4805438	4737068	4789223	0.97	cps
Yttrium	89-2	1204041	1197737	1204208	1201995	0.31	cps
Zinc	66-2	32368	32318	32549	32412	0.37	cps
Zirconium	90-1	14647	14557	14924	14710	1.30	cps
Zirconium	91-1	3384	3284	3230	3299	2.36	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10MSDLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3SRE Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:36:40 DataFile Name : 027SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	1143768	1141547	1138798	1141371	0.22	cps
Antimony	121-1	314157	316681	315345	315394	0.40	cps
Arsenic	75-2	6795	6732	6948	6825	1.63	cps
Barium	135-1	1964165	1942073	1909449	1938563	1.42	cps
Barium	137-1	3334114	3273201	3360683	3322666	1.35	cps
Beryllium	9-1	29404	29097	28930	29144	0.83	cps
Bismuth	209-1	2314243	2347572	2431195	2364337	2.55	cps
Bismuth	209-2	1872265	1887111	1870393	1876590	0.49	cps
Bromine	81-1	7015	6858	6932	6935	1.13	cps
Bromine	81-2	153	133	163	150	10.18	cps
Cadmium	108-1	2614	2700	2544	2619	3.00	cps
Cadmium	106-1	4707	4684	4694	4695	0.25	cps
Cadmium	111-1	39489	38997	39435	39307	0.69	cps
Calcium	43-1	48651	48273	48792	48572	0.55	cps
Calcium	44-1	780124	787192	785740	784352	0.48	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	240268	241341	241128	240912	0.24	cps
Cobalt	59-2	832244	829564	830633	830814	0.16	cps
Copper	63-2	302623	301748	301996	302122	0.15	cps
Dysprosium	156-1	13063	13580	13193	13279	2.03	cps
Dysprosium	156-2	10227	10371	10437	10345	1.04	cps
Erbium	164-1	10304	10781	10754	10613	2.53	cps
Erbium	164-2	7352	7155	7245	7251	1.36	cps
Gadolinium	160-1	12853	12532	12446	12610	1.70	cps
Gadolinium	160-2	8729	8773	8583	8695	1.15	cps
Holmium	165-1	3841175	3791021	3866243	3832813	1.00	cps
Holmium	165-2	2610411	2630923	2553394	2598242	1.55	cps
Indium	115-1	2795404	2783183	2903659	2827415	2.35	cps
Indium	115-2	896712	900486	891863	896354	0.48	cps
Iron	56-2	23315362	23215062	23356565	23295663	0.31	cps
Iron	57-2	606289	606789	608091	607057	0.15	cps
Iron	54-2	1340032	1339250	1335577	1338287	0.18	cps
Krypton	83-1	93	80	53	76	26.96	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : P2827-10MSDLX4 Instrumnet Name : P7
Client Sample ID : MYD5Y3SRE Dilution Factor : 4
Date & Time Acquired : 2024-10-18 14:36:40 DataFile Name : 027SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	86211	88059	88324	87531	1.32	cps
Lead	207-1	74134	75458	75532	75041	1.05	cps
Lead	208-1	340147	349059	346131	345112	1.32	cps
Lithium	6-1	669119	654688	680580	668129	1.94	cps
Magnesium	24-2	678925	677132	677985	678014	0.13	cps
Manganese	55-2	980723	977853	977420	978665	0.18	cps
Molybdenum	94-1	5328	5661	5638	5542	3.36	cps
Molybdenum	95-1	477	460	470	469	1.79	cps
Molybdenum	96-1	1650	1590	1647	1629	2.07	cps
Molybdenum	97-1	270	300	230	267	13.17	cps
Molybdenum	98-1	687	783	707	726	7.03	cps
Neodymium	150-1	24021	23900	23937	23953	0.26	cps
Neodymium	150-2	13093	12763	13370	13075	2.33	cps
Nickel	60-2	212050	209837	210114	210667	0.57	cps
Phosphorus	31-2	2454	2464	2550	2489	2.14	cps
Potassium	39-2	289339	290419	290504	290087	0.22	cps
Rhodium	103-1	2439720	2392331	2479670	2437240	1.79	cps
Rhodium	103-2	1391838	1381922	1384022	1385927	0.38	cps
Scandium	45-1	1413120	1392632	1428238	1411330	1.27	cps
Scandium	45-2	126486	126630	125169	126095	0.64	cps
Selenium	82-1	4912	4992	4993	4966	0.93	cps
Selenium	77-2	487	470	497	484	2.78	cps
Selenium	78-2	1690	1597	1533	1607	4.91	cps
Silicon	28-1	3981988	4049035	4007292	4012772	0.84	cps
Silver	107-1	180612	183646	182745	182334	0.85	cps
Silver	109-1	174172	175144	175117	174811	0.32	cps
Sodium	23-2	21799	21682	23478	22320	4.50	cps
Strontium	86-1	33902	33515	33268	33562	0.95	cps
Strontium	88-1	294568	295012	297879	295820	0.61	cps
Sulfur	34-1	83902	82421	81734	82685	1.34	cps
Terbium	159-1	3931153	3893017	3952427	3925532	0.77	cps
Terbium	159-2	2571510	2581837	2560225	2571191	0.42	cps
Thallium	203-1	146615	149348	148978	148314	1.00	cps
Thallium	205-1	347482	355446	354008	352312	1.20	cps
Tin	118-1	3330	3327	3287	3315	0.73	cps
Titanium	47-1	54464	55712	55876	55351	1.40	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

P2827-10MSDLX4

Instrumnet Name :

P7

Client Sample ID :

MYD5Y3SRE

Dilution Factor :

4

Date & Time Acquired :

2024-10-18 14:36:40

DataFile Name :

027SMPL.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	2920	3150	3154	3075	4.35	cps
Vanadium	51-2	522704	523786	523990	523493	0.13	cps
Yttrium	89-1	4634968	4592159	4731558	4652895	1.53	cps
Yttrium	89-2	1198040	1197903	1190968	1195637	0.34	cps
Zinc	66-2	105600	105959	107930	106496	1.18	cps
Zirconium	90-1	13670	14197	13767	13878	2.02	cps
Zirconium	91-1	3090	3264	3117	3157	2.96	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCV095 Instrumnet Name : P7
Client Sample ID : CCV095 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:44:37 DataFile Name : 028CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	3959202	4021423	3995113	3991912	0.78	cps
Antimony	121-1	1926832	1960964	1990488	1959428	1.63	cps
Arsenic	75-2	82112	83057	83386	82851	0.80	cps
Barium	135-1	2749894	2790982	2815322	2785399	1.19	cps
Barium	137-1	4785009	4774175	4817695	4792293	0.47	cps
Beryllium	9-1	339097	345340	350671	345036	1.68	cps
Bismuth	209-1	2162861	2118907	2196749	2159506	1.81	cps
Bismuth	209-2	1506409	1501887	1522116	1510137	0.70	cps
Bromine	81-1	6268	6285	6175	6242	0.95	cps
Bromine	81-2	17	43	20	27	54.47	cps
Cadmium	108-1	32553	33833	33833	33406	2.21	cps
Cadmium	106-1	49497	49528	49916	49647	0.47	cps
Cadmium	111-1	439097	446651	452363	446037	1.49	cps
Calcium	43-1	2325753	2391294	2423470	2380173	2.09	cps
Calcium	44-1	37146256	37750601	38118235	37671697	1.30	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	665865	664018	668385	666089	0.33	cps
Cobalt	59-2	1043446	1041241	1050028	1044905	0.44	cps
Copper	63-2	7110194	7044593	7092840	7082542	0.48	cps
Dysprosium	156-1	117	123	120	120	2.78	cps
Dysprosium	156-2	113	137	170	140	20.34	cps
Erbium	164-1	93	100	103	99	5.15	cps
Erbium	164-2	77	77	47	67	25.98	cps
Gadolinium	160-1	97	60	87	81	23.37	cps
Gadolinium	160-2	163	187	173	174	6.71	cps
Holmium	165-1	3708135	3755849	3738197	3734060	0.65	cps
Holmium	165-2	2501918	2531700	2566985	2533534	1.29	cps
Indium	115-1	2543264	2542556	2538386	2541402	0.10	cps
Indium	115-2	811606	807876	817602	812361	0.60	cps
Iron	56-2	146483298	146807731	147490161	146927063	0.35	cps
Iron	57-2	3663369	3605683	3614224	3627759	0.86	cps
Iron	54-2	7980877	8065522	8024447	8023615	0.53	cps
Krypton	83-1	77	73	53	68	18.62	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCV095 Instrumnet Name : P7
Client Sample ID : CCV095 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:44:37 DataFile Name : 028CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	8215247	8279585	8450896	8315242	1.46	cps
Lead	207-1	7170109	7318348	7417096	7301851	1.70	cps
Lead	208-1	32757755	33233826	33748036	33246539	1.49	cps
Lithium	6-1	607561	618126	623113	616267	1.29	cps
Magnesium	24-2	32141449	31783092	32380654	32101731	0.94	cps
Manganese	55-2	5170185	5204218	5251872	5208758	0.79	cps
Molybdenum	94-1	6908839	6929293	6982134	6940089	0.54	cps
Molybdenum	95-1	9954810	10182090	10129118	10088673	1.18	cps
Molybdenum	96-1	11012912	11206212	11180413	11133179	0.94	cps
Molybdenum	97-1	6259701	6405357	6451097	6372052	1.57	cps
Molybdenum	98-1	16257515	16705506	16612325	16525115	1.43	cps
Neodymium	150-1	323	257	297	292	11.48	cps
Neodymium	150-2	87	63	100	83	22.27	cps
Nickel	60-2	253322	252590	254284	253399	0.34	cps
Phosphorus	31-2	38532	39464	38802	38933	1.23	cps
Potassium	39-2	20052357	20031113	20163613	20082361	0.35	cps
Rhodium	103-1	2135263	2138564	2157719	2143849	0.57	cps
Rhodium	103-2	1218401	1213225	1235641	1222422	0.96	cps
Scandium	45-1	1311172	1290442	1320714	1307443	1.18	cps
Scandium	45-2	118267	116943	118381	117864	0.68	cps
Selenium	82-1	24693	24553	24962	24736	0.84	cps
Selenium	77-2	2427	2510	2690	2542	5.29	cps
Selenium	78-2	7919	8366	8039	8108	2.85	cps
Silicon	28-1	1468721	1514286	1535674	1506227	2.27	cps
Silver	107-1	2180546	2193192	2212514	2195417	0.73	cps
Silver	109-1	2048128	2090054	2099550	2079244	1.32	cps
Sodium	23-2	53662969	53800841	54557468	54007093	0.89	cps
Strontium	86-1	14796909	15075105	14967544	14946519	0.94	cps
Strontium	88-1	131047621	131538225	132223655	131603167	0.45	cps
Sulfur	34-1	227305	230824	235301	231143	1.73	cps
Terbium	159-1	3746105	3825307	3805178	3792197	1.09	cps
Terbium	159-2	2486356	2456267	2456463	2466362	0.70	cps
Thallium	203-1	1987126	2041784	2079265	2036058	2.28	cps
Thallium	205-1	4796495	4944563	5006757	4915938	2.20	cps
Tin	118-1	1695118	1715401	1738268	1716263	1.26	cps
Titanium	47-1	2244466	2312855	2347538	2301620	2.28	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCV095

Instrumnet Name :

P7

Client Sample ID :

CCV095

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 14:44:37

DataFile Name :

028CCV.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	6799560	6924991	6961380	6895310	1.23	cps
Vanadium	51-2	607091	613031	615330	611817	0.69	cps
Yttrium	89-1	4235129	4228543	4276569	4246747	0.61	cps
Yttrium	89-2	1085997	1085365	1094876	1088746	0.49	cps
Zinc	66-2	1186217	1198684	1199876	1194926	0.63	cps
Zirconium	90-1	3122949	3128011	3175597	3142186	0.92	cps
Zirconium	91-1	685264	693965	702511	693913	1.24	cps

LB Number :LB133015Operator :Jaswal

Lab Sample ID :CCB095Instrumnet Name :P7

Client Sample ID :CCB095Dilution Factor :1

Date & Time Acquired :2024-10-18 14:50:48DataFile Name :030CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Aluminium	27-2	130	160	150	147	10.41	cps
Antimony	121-1	437	450	517	468	9.16	cps
Arsenic	75-2	10	3	0	4	114.60	cps
Barium	135-1	60	53	43	52	16.07	cps
Barium	137-1	103	33	77	71	49.68	cps
Beryllium	9-1	123	180	133	146	20.78	cps
Bismuth	209-1	2518731	2569695	2524898	2537775	1.10	cps
Bismuth	209-2	1973330	1960886	1996397	1976871	0.91	cps
Bromine	81-1	6708	6652	6568	6643	1.06	cps
Bromine	81-2	37	57	20	38	48.60	cps
Cadmium	108-1	0	7	0	2	173.21	cps
Cadmium	106-1	1387	1460	1267	1371	7.12	cps
Cadmium	111-1	994	1027	917	979	5.80	cps
Calcium	43-1	170	120	133	141	18.35	cps
Calcium	44-1	3030	3217	3124	3124	2.99	cps
Carbon	12-1						cps
Carbon	12-2						cps
Chlorine	35-1						cps
Chlorine	35-2						cps
Chromium	52-2	303	207	230	247	20.45	cps
Cobalt	59-2	37	77	53	56	36.17	cps
Copper	63-2	817	763	777	786	3.53	cps
Dysprosium	156-1	7	0	3	3	100.05	cps
Dysprosium	156-2	0	0	0	0	0.00	cps
Erbium	164-1	40	13	27	27	50.01	cps
Erbium	164-2	13	13	20	16	24.76	cps
Gadolinium	160-1	13	0	7	7	99.98	cps
Gadolinium	160-2	127	80	127	111	24.25	cps
Holmium	165-1	4044559	4091739	4042404	4059567	0.69	cps
Holmium	165-2	2722597	2698520	2718822	2713313	0.48	cps
Indium	115-1	2926944	2969243	2954049	2950078	0.73	cps
Indium	115-2	924153	916679	920368	920400	0.41	cps
Iron	56-2	8432	8479	8389	8434	0.53	cps
Iron	57-2	587	660	617	621	5.94	cps
Iron	54-2	1043	960	910	971	6.94	cps
Krypton	83-1	43	83	63	63	31.58	cps

LB Number : LB133015 Operator : Jaswal
Lab Sample ID : CCB095 Instrumnet Name : P7
Client Sample ID : CCB095 Dilution Factor : 1
Date & Time Acquired : 2024-10-18 14:50:48 DataFile Name : 030CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Lead	206-1	507	430	447	461	8.74	cps
Lead	207-1	393	370	490	418	15.23	cps
Lead	208-1	1933	1840	1993	1922	4.02	cps
Lithium	6-1	721883	722649	715133	719888	0.57	cps
Magnesium	24-2	1870	1817	1683	1790	5.37	cps
Manganese	55-2	160	163	193	172	10.66	cps
Molybdenum	94-1	407	327	290	341	17.49	cps
Molybdenum	95-1	273	233	307	271	13.54	cps
Molybdenum	96-1	267	267	300	278	6.93	cps
Molybdenum	97-1	167	177	153	166	7.07	cps
Molybdenum	98-1	333	370	340	348	5.62	cps
Neodymium	150-1	3	10	3	6	69.34	cps
Neodymium	150-2	7	3	0	3	100.05	cps
Nickel	60-2	87	53	57	66	28.01	cps
Phosphorus	31-2	120	133	117	123	7.15	cps
Potassium	39-2	31269	32077	30758	31368	2.12	cps
Rhodium	103-1	2573957	2600029	2567026	2580337	0.67	cps
Rhodium	103-2	1438758	1424841	1432224	1431941	0.49	cps
Scandium	45-1	1462007	1504972	1456010	1474329	1.81	cps
Scandium	45-2	128284	128126	129064	128491	0.39	cps
Selenium	82-1	110	76	101	96	18.36	cps
Selenium	77-2	0	0	0	0	0.00	cps
Selenium	78-2	97	83	93	91	7.62	cps
Silicon	28-1	166577	169246	168326	168050	0.81	cps
Silver	107-1	87	143	70	100	38.44	cps
Silver	109-1	113	103	100	106	6.58	cps
Sodium	23-2	13686	13583	13132	13467	2.19	cps
Strontium	86-1	290	293	340	308	9.08	cps
Strontium	88-1	1737	1623	1530	1630	6.35	cps
Sulfur	34-1	108750	108596	107709	108352	0.52	cps
Terbium	159-1	4201372	4087922	4134864	4141386	1.38	cps
Terbium	159-2	2609233	2607205	2651725	2622721	0.96	cps
Thallium	203-1	203	167	210	193	12.07	cps
Thallium	205-1	463	417	460	447	5.83	cps
Tin	118-1	837	810	800	816	2.32	cps
Titanium	47-1	67	90	53	70	26.51	cps

LB Number :

LB133015

Operator :

Jaswal

Lab Sample ID :

CCB095

Instrumnet Name :

P7

Client Sample ID :

CCB095

Dilution Factor :

1

Date & Time Acquired :

2024-10-18 14:50:48

DataFile Name :

030CCBE.d

Parameter	Mass	CPS1	CPS2	CPS3	CPSMean	CPSRSD	Units
Uranium	238-1	73	87	83	81	8.56	cps
Vanadium	51-2	13	13	10	12	15.73	cps
Yttrium	89-1	4694389	4722526	4721336	4712751	0.34	cps
Yttrium	89-2	1184986	1174633	1184060	1181226	0.48	cps
Zinc	66-2	110	103	107	107	3.12	cps
Zirconium	90-1	630	547	533	570	9.19	cps
Zirconium	91-1	107	110	113	110	3.03	cps