

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132877

Review By	Sarabjit Jaswal	Review On	10/12/2024 4:20:35 AM
Supervise By	Mohan Bera	Supervise On	10/12/2024 4:21:48 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE014	TUNE	10/07/24 17:11		Jaswal	OK
2	S00	S0	CAL1	10/07/24 17:24		Jaswal	OK
3	S02	S02	CAL3	10/07/24 17:28		Jaswal	OK
4	S03	S03	CAL4	10/07/24 17:31		Jaswal	OK
5	S04	S04	CAL5	10/07/24 17:34		Jaswal	OK
6	S05	S05	CAL6	10/07/24 17:37		Jaswal	OK
7	S06	S06	CAL7	10/07/24 17:40		Jaswal	OK
8	S07	S07	CAL8	10/07/24 17:43		Jaswal	OK
9	S08	S08	CAL9	10/07/24 17:46		Jaswal	OK
10	ICV014	ICV014	ICV	10/07/24 17:56	Inst. pause for calibration review	Jaswal	OK
11	ICB014	ICB014	ICB	10/07/24 17:59		Jaswal	OK
12	ICSA014	ICSA014	ICSA	10/07/24 18:03		Jaswal	OK
13	ICSAB014	ICSAB014	ICSAB	10/07/24 18:09	Inst. pause for review	Jaswal	OK
14	CCV051	CCV051	CCV	10/07/24 18:18	Inst. pause for review	Jaswal	OK
15	CCB051	CCB051	CCB	10/07/24 18:20		Jaswal	OK
16	PB163776BL	PBS776	MB	10/07/24 18:24	Not USe	Jaswal	Not Ok
17	PB163776BS	LCS776	LCS	10/07/24 18:28	Not USe	Jaswal	Not Ok
18	PB163776BSD	LCSD776	LCSD	10/07/24 18:31	Not USe	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132877

Review By	Sarabjit Jaswal	Review On	10/12/2024 4:20:35 AM
Supervise By	Mohan Bera	Supervise On	10/12/2024 4:21:48 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

19	P2817-01	MYD567	SAM	10/07/24 18:34		Jaswal	OK
20	P2817-02	MYD568	SAM	10/07/24 18:38		Jaswal	OK
21	P2817-03	MYD569	SAM	10/07/24 18:41		Jaswal	OK
22	P2817-04	MYD570	SAM	10/07/24 18:44	Not Use - Ba,Co,Ni,V High in MS RE	Jaswal	Not Ok
23	P2817-05	MYD570D	DUP	10/07/24 18:47	Not Use - Ba,Co,Ni,V High in MS RE	Jaswal	Not Ok
24	P2817-04L	MYD570L	SD	10/07/24 18:51	Not Use - Ba,Co,Ni,V High in MS RE	Jaswal	Not Ok
25	P2817-06	MYD570S	MS	10/07/24 18:54	Not Use - Ba,Co,Ni,V High in MS RE	Jaswal	Not Ok
26	P2817-04MS	MYD570SRE	MS	10/07/24 18:57	Not Use - MS Fail For Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Not Ok
27	P2817-07	MYD571	SAM	10/07/24 19:00		Jaswal	OK
28	P2817-08	MYD572	SAM	10/07/24 19:03		Jaswal	OK
29	P2817-09	MYD574	SAM	10/07/24 19:07		Jaswal	OK
30	P2817-10	MYD575	SAM	10/07/24 19:10		Jaswal	OK
31	P2817-11	MYD576	SAM	10/07/24 19:13		Jaswal	OK
32	P2817-12	MYD577	SAM	10/07/24 19:16		Jaswal	OK
33	P2817-13	MYD578	SAM	10/07/24 19:19		Jaswal	OK
34	P2817-14	MYD579	SAM	10/07/24 19:23		Jaswal	OK
35	P2817-15	MYD580	SAM	10/07/24 19:26		Jaswal	OK
36	P2817-16	MYD582	SAM	10/07/24 19:29		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132877

Review By	Sarabjit Jaswal	Review On	10/12/2024 4:20:35 AM
Supervise By	Mohan Bera	Supervise On	10/12/2024 4:21:48 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

37	P2817-17	MYD583	SAM	10/07/24 19:32		Jaswal	OK
38	P4161-08	YE0T1	SAM	10/07/24 19:35		Jaswal	OK
39	P2817-18	MYD584	SAM	10/07/24 19:39		Jaswal	OK
40	CCV052	CCV052	CCV	10/07/24 19:42		Jaswal	OK
41	CCB052	CCB052	CCB	10/07/24 19:45		Jaswal	OK
42	P2817-19	MYD585	SAM	10/07/24 19:48		Jaswal	OK
43	P2817-20	MYD586	SAM	10/07/24 19:51		Jaswal	OK
44	P2817-21	MYD587	SAM	10/07/24 19:54		Jaswal	OK
45	P2817-22	MYD588	SAM	10/07/24 19:58		Jaswal	OK
46	CCV053	CCV053	CCV	10/07/24 20:18		Jaswal	OK
47	CCB053	CCB053	CCB	10/07/24 20:21		Jaswal	OK
48	PB163639BL	PBS639	MB	10/07/24 20:24	Not Use	Jaswal	Not Ok
49	PB163639BS	LCS639	LCS	10/07/24 20:27	Not Use	Jaswal	Not Ok
50	PB163639BSD	LCSD639	LCSD	10/07/24 20:31	Not Use	Jaswal	Not Ok
51	P2792-01	MYD438	SAM	10/07/24 20:34	Not Use	Jaswal	Not Ok
52	P2792-02	MYD438D	DUP	10/07/24 20:37	Not Use	Jaswal	Not Ok
53	P2792-01L	MYD438L	SD	10/07/24 20:41	Not Use	Jaswal	Not Ok
54	P2792-03	MYD438S	MS	10/07/24 20:44	Not Use	Jaswal	Not Ok
55	P2792-01MS	MYD438SRE	MS	10/07/24 20:47	Not Use	Jaswal	Not Ok
56	P2792-04	MYD4K3	SAM	10/07/24 20:50		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132877

Review By	Sarabjit Jaswal	Review On	10/12/2024 4:20:35 AM
Supervise By	Mohan Bera	Supervise On	10/12/2024 4:21:48 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard			
Chk Standard	MP82587,MP82586		

57	P2792-05	MYD4K4	SAM	10/07/24 20:53		Jaswal	OK
58	P2792-06	MYD4K5	SAM	10/07/24 20:56		Jaswal	OK
59	P2792-07	MYD4K6	SAM	10/07/24 21:00		Jaswal	OK
60	P2792-08	MYD4K7	SAM	10/07/24 21:03		Jaswal	OK
61	P2792-09	MYD4K8	SAM	10/07/24 21:06		Jaswal	OK
62	P2792-10	MYD4K9	SAM	10/07/24 21:09		Jaswal	OK
63	CCV054	CCV054	CCV	10/07/24 21:13		Jaswal	OK
64	CCB054	CCB054	CCB	10/07/24 21:16		Jaswal	OK