

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No. : SDG No.: MYD5X5
 Initial Calibration Verification Source : MP82578
 Continuing Calibration Verification Source : MP82579
 Run Batch: LB132882 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	200	99	5	500	520	104	3	480	96	1
Arsenic	200	210	105	2	500	510	103	1	510	102	1
Barium	100	95	95	4	2500	2700	108	3	2500	99	2
Beryllium	100	96	96	3	500	540	107	4	480	96	1
Cadmium	100	100	100	5	500	530	106	4	480	96	1
Chromium	100	96	96	2	500	480	97	1	500	100	1
Cobalt	100	100	101	2	500	500	100	1	520	104	1
Copper	100	98	98	2	5000	4700	95	1	5000	100	0
Lead	200	190	95	5	2500	2700	107	4	2500	102	1
Nickel	110	100	95	2	500	490	98	1	470	95	0
Selenium	200	200	101	5	500	530	106	5	500	100	1
Silver	50.0	53	106	5	500	510	101	4	460	92	1
Thallium	210	190	90	5	500	530	107	4	510	101	1
Vanadium	100	93	93	2	500	500	99	2	510	102	1
Zinc	200	200	100	3	5000	4800	96	1	5000	100	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51495 MA No. : SDG No.: MYD5X5Initial Calibration Verification Source : MP82578Continuing Calibration Verification Source : MP82579Run Batch: LB133015 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV025				ID: CCV094				ID: CCV095		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	190	95	2	500	500	100	1	510	101	1
Arsenic	200	200	100	1	500	500	100	0	510	101	1
Barium	100	93	93	2	2500	2500	101	1	2600	103	1
Beryllium	100	100	102	1	500	520	104	1	520	104	0
Cadmium	100	110	107	1	500	480	96	1	490	98	1
Chromium	100	100	101	1	500	510	101	0	510	102	0
Cobalt	100	100	102	1	500	530	107	0	540	108	0
Copper	100	100	103	0	5000	5100	102	1	5200	105	0
Lead	200	180	91	2	2500	2600	104	1	2600	104	2
Nickel	110	110	98	2	500	530	106	0	540	108	0
Selenium	200	210	103	2	500	490	98	1	500	99	0
Silver	50.0	49	98	2	500	480	96	1	490	98	1
Thallium	210	210	99	1	500	530	105	1	520	105	3
Vanadium	100	97	97	1	500	510	102	0	520	103	1
Zinc	200	200	99	1	5000	5200	104	0	5300	106	1