

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD5K6
 Instrument ID: P7 Start Date: 10/18/2024
 Analytical Method: ICP-MS Run Batch: LB133015
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	2.0	0	50	51	2
Arsenic	0	0.0	0	1	1.1	10	50	51	2
Barium	0	0.0	0	10	9.5	-5	250	250	0
Beryllium	0	0.0	0	1	1.1	10	50	53	6
Cadmium	0	0.0	0	1	1.2	20	50	55	10
Chromium	0	0.0	0	2	2.0	0	50	51	2
Cobalt	0	0.0	0	1	1.1	10	50	52	4
Copper	0	0.0	0	2	2.1	5	500	520	4
Lead	0	0.0	0	1	0.79	-21	250	240	-4
Nickel	0	0.0	0	1	1.2	20	50	54	8
Selenium	0	0.0	0	5	4.9	-2	50	52	4
Silver	0	0.0	0	1	1.0	0	50	54	8
Thallium	0	0.0	0	1	0.86	-14	50	44	-12
Vanadium	0	0.0	0	5	5.0	0	50	50	0
Zinc	0	0.0	0	5	5.1	2	500	520	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD5K6
 Instrument ID: P7 Start Date: 10/18/2024
 Analytical Method: ICP-MS Run Batch: LB133015
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	130	4	250	250	0	500	520	4
Arsenic	125	130	4	250	260	4	500	500	0
Barium	625	630	1	1250	1300	4	2500	2600	4
Beryllium	125	130	4	250	260	4	500	510	2
Cadmium	125	140	12	250	260	4	500	520	4
Chromium	125	130	4	250	260	4	500	510	2
Cobalt	125	130	4	250	260	4	500	510	2
Copper	1250	1300	4	2500	2600	4	5000	5100	2
Lead	625	620	-1	1250	1200	-4	2500	2500	0
Nickel	125	130	4	250	260	4	500	510	2
Selenium	125	130	4	250	260	4	500	510	2
Silver	125	130	4	250	260	4	500	520	4
Thallium	125	110	-12	250	250	0	500	510	2
Vanadium	125	130	4	250	250	0	500	510	2
Zinc	1250	1300	4	2500	2600	4	5000	5100	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: SDG No.: MYD5K6
 Instrument ID: P7 Start Date: 10/18/2024
 Analytical Method: ICP-MS Run Batch: LB133015
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	990	-1						
Arsenic	1000	1000	0						
Barium	5000	4900	-2						
Beryllium	1000	990	-1						
Cadmium	1000	990	-1						
Chromium	1000	990	-1						
Cobalt	1000	990	-1						
Copper	10000	9900	-1						
Lead	5000	5000	0						
Nickel	1000	990	-1						
Selenium	1000	990	-1						
Silver	1000	990	-1						
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	9900	-1						

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD5K6
 Instrument ID: P7 Start Date: 10/20/2024
 Analytical Method: ICP-MS Run Batch: LB133019
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	0	0.0	0	2	2.0	0	50	50	0
Arsenic	0	0.0	0	1	1.1	10	50	51	2
Barium	0	0.0	0	10	9.7	-3	250	250	0
Beryllium	0	0.0	0	1	0.98	-2	50	50	0
Cadmium	0	0.0	0	1	0.98	-2	50	50	0
Chromium	0	0.0	0	2	2.0	0	50	50	0
Cobalt	0	0.0	0	1	1.1	10	50	50	0
Copper	0	0.0	0	2	2.0	0	500	500	0
Lead	0	0.0	0	1	0.92	-8	250	240	-4
Nickel	0	0.0	0	1	0.93	-7	50	50	0
Selenium	0	0.0	0	5	4.8	-4	50	50	0
Silver	0	0.0	0	1	0.95	-5	50	51	2
Thallium	0	0.0	0	1	0.95	-5	50	50	0
Vanadium	0	0.0	0	5	5.0	0	50	49	-2
Zinc	0	0.0	0	5	5.4	8	500	500	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: _____ SDG No.: MYD5K6
 Instrument ID: P7 Start Date: 10/20/2024
 Analytical Method: ICP-MS Run Batch: LB133019
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	125	130	4	250	230	-8	500	500	0
Arsenic	125	120	-4	250	240	-4	500	510	2
Barium	625	620	-1	1250	1100	-12	2500	2500	0
Beryllium	125	120	-4	250	240	-4	500	510	2
Cadmium	125	120	-4	250	240	-4	500	510	2
Chromium	125	130	4	250	240	-4	500	500	0
Cobalt	125	120	-4	250	240	-4	500	500	0
Copper	1250	1200	-4	2500	2500	0	5000	5100	2
Lead	625	630	1	1250	1200	-4	2500	2500	0
Nickel	125	130	4	250	240	-4	500	510	2
Selenium	125	120	-4	250	240	-4	500	510	2
Silver	125	120	-4	250	240	-4	500	510	2
Thallium	125	130	4	250	210	-16	500	500	0
Vanadium	125	130	4	250	240	-4	500	500	0
Zinc	1250	1200	-4	2500	2400	-4	5000	5100	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51495 MA No.: SDG No.: MYD5K6
Instrument ID: P7 Start Date: 10/20/2024
Analytical Method: ICP-MS Run Batch: LB133019
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Antimony	1000	1000	0						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	1000	0						
Cadmium	1000	1000	0						
Chromium	1000	1000	0						
Cobalt	1000	1000	0						
Copper	10000	10000	0						
Lead	5000	5000	0						
Nickel	1000	1000	0						
Selenium	1000	1000	0						
Silver	1000	1000	0						
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	10000	0						