

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/28/24 14:22		Kareem	OK
2	S1	S01	CAL2	06/28/24 14:26		Kareem	OK
3	S2	S02	CAL3	06/28/24 14:31		Kareem	OK
4	S3	S03	CAL4	06/28/24 14:35		Kareem	OK
5	S4	S04	CAL5	06/28/24 14:40		Kareem	OK
6	S5	S05	CAL6	06/28/24 14:44		Kareem	OK
7	S6	S06	CAL7	06/28/24 14:48		Kareem	OK
8	ICV012	ICV012	ICV	06/28/24 15:04	Inst. paused for Calibration review	Kareem	OK
9	ICB012	ICB012	ICB	06/28/24 15:08		Kareem	OK
10	ICSA012	ICSA012	ICSA	06/28/24 15:15	Inst. paused for review	Kareem	OK
11	ICSAB012	ICSAB012	ICSAB	06/28/24 15:20		Kareem	OK
12	CCV080	CCV080	CCV	06/28/24 15:39	Inst. paused for review	Kareem	OK
13	CCB080	CCB080	CCB	06/28/24 15:46	Inst. paused for review	Kareem	OK
14	PB161736BL	PBW736	MB	06/28/24 15:50		Kareem	OK
15	PB161736BS	LCS736	LCS	06/28/24 15:55		Kareem	OK
16	PB161704BL	PBS704	MB	06/28/24 15:59		Kareem	OK
17	PB161704BS	LCS704	LCS	06/28/24 16:04		Kareem	OK
18	PB161705BL	PBS705	MB	06/28/24 16:08		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

19	PB161705BS	LCS705	LCS	06/28/24 16:13		Kareem	OK
20	PB161706BL	PBS706	MB	06/28/24 16:18		Kareem	OK
21	PB161706BS	LCS706	LCS	06/28/24 16:22		Kareem	OK
22	PB161707BL	PBS707	MB	06/28/24 16:27		Kareem	OK
23	PB161707BS	LCS707	LCS	06/28/24 16:31		Kareem	OK
24	PB161708BL	PBS708	MB	06/28/24 16:36		Kareem	OK
25	PB161708BS	LCS708	LCS	06/28/24 16:40		Kareem	OK
26	PB161709BL	PBS709	MB	06/28/24 16:45		Kareem	OK
27	PB161709BS	LCS709	LCS	06/28/24 16:49		Kareem	OK
28	P2975-01	E10R8	SAM	06/28/24 16:54		Kareem	OK
29	P2975-02	E10R9	SAM	06/28/24 16:58		Kareem	OK
30	P2975-03	E10S0	SAM	06/28/24 17:03		Kareem	OK
31	P2975-04	E10S1	SAM	06/28/24 17:07		Kareem	OK
32	P2975-05	E10S2	SAM	06/28/24 17:12		Kareem	OK
33	P2975-06	E10S3	SAM	06/28/24 17:16		Kareem	OK
34	CCV081	CCV081	CCV	06/28/24 17:21		Kareem	OK
35	CCB081	CCB081	CCB	06/28/24 17:25		Kareem	OK
36	P2975-07	E1179	SAM	06/28/24 17:30		Kareem	OK
37	P2975-08	E1180	SAM	06/28/24 17:34		Kareem	OK
38	P3016-18	E11R4	SAM	06/28/24 17:39		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

39	P3016-19	E11R5	SAM	06/28/24 17:43		Kareem	OK
40	P3016-20	E11R6	SAM	06/28/24 17:48		Kareem	OK
41	P2821-01	MYD4L1	SAM	06/28/24 17:52		Kareem	OK
42	P2821-02	MYD4L1D	DUP	06/28/24 17:57		Kareem	OK
43	P2821-01L	MYD4L1L	SD	06/28/24 18:01		Kareem	OK
44	P2821-03	MYD4L1S	MS	06/28/24 18:06	MS fail for Cu(0.3605673PPM), Pb(0.1279198PPM), Se(Below RL), TI(Below RL), Zn(0.7395081PPM)	Kareem	OK
45	P2821-04	MYD5E9	SAM	06/28/24 18:10		Kareem	OK
46	P2821-05	MYD5F0	SAM	06/28/24 18:14		Kareem	OK
47	P2821-06	MYD5F1	SAM	06/28/24 18:19		Kareem	OK
48	P2821-07	MYD5F2	SAM	06/28/24 18:23		Kareem	OK
49	P2821-08	MYD5F3	SAM	06/28/24 18:27		Kareem	OK
50	P2821-09	MYD5F4	SAM	06/28/24 18:32		Kareem	OK
51	P2821-10	MYD5F5	SAM	06/28/24 18:36		Kareem	OK
52	P2821-11	MYD5F6	SAM	06/28/24 18:40		Kareem	OK
53	P2821-12	MYD5F7	SAM	06/28/24 18:45		Kareem	OK
54	P2821-13	MYD5F8	SAM	06/28/24 18:49		Kareem	OK
55	P2821-14	MYD5F9	SAM	06/28/24 18:54		Kareem	OK
56	CCV082	CCV082	CCV	06/28/24 18:58		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

57	CCB082	CCB082	CCB	06/28/24 19:02		Kareem	OK
58	P2821-15	MYD5G0	SAM	06/28/24 19:07		Kareem	OK
59	P2821-16	MYD5G1	SAM	06/28/24 19:11		Kareem	OK
60	P2821-17	MYD5G2	SAM	06/28/24 19:15		Kareem	OK
61	P2821-18	MYD5G3	SAM	06/28/24 19:20		Kareem	OK
62	P2821-19	MYD5G4	SAM	06/28/24 19:24		Kareem	OK
63	P2821-20	MYD5G5	SAM	06/28/24 19:29		Kareem	OK
64	P2821-21	MYD5G6	SAM	06/28/24 19:33		Kareem	OK
65	P2821-22	MYD5G7	SAM	06/28/24 19:37		Kareem	OK
66	P2809-10A	MYD4R7A	PS	06/28/24 19:42	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
67	P2810-01A	MYD4G6A	PS	06/28/24 19:46	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
68	P2811-17A	MYD4Y3A	PS	06/28/24 19:50	PS For Se- 0.007-MP81024,TI-0.005-M P81025,Zn- 0.032 - M5819 - 10ml sample	Kareem	OK
69	P2813-01A	MYD4Y4A	PS	06/28/24 19:55	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
70	P2814-01A	MYD4H3A	PS	06/28/24 19:59	PS For Se- 0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

71	P2815-11A	MYD536A	PS	06/28/24 20:03	PS For Se-0.007-MP81024,TI-0.005-M P81025,Zn-0.027-M5819 - 10ml sample	Kareem	OK
72	P2816-01A	MYD4J8A	PS	06/28/24 20:08	PS For Se-0.007-MP81024,TI-0.005-M P81025,Zn-0.01-M5819 - 10ml sample	Kareem	OK
73	P2817-04A	MYD570A	PS	06/28/24 20:12	PS For As-0.031-MP81248,Cr-0.016-M P81246,Se-0.007-MP81024,TI-0.005-M P81025 - 10ml sample	Kareem	OK
74	P2818-01A	MYD581A	PS	06/28/24 20:17	PS For Se-0.007-MP81024,Zn-0.037-M 5819 - 10ml sample	Kareem	OK
75	P2819-05A	MYD5B3A	PS	06/28/24 20:21	PS For Cr-0.17-MP81246,Cu-0.1-MP81247,Se-0.007-MP81024,Zn-0.012-M 5819 - 10ml sample	Kareem	OK
76	P2820-05A	MYD5D4A	PS	06/28/24 20:25	PS For Se-0.007-MP81024,TI-0.005-M P81025,Zn- 0.016 - M5819 - 10ml sample	Kareem	OK
77	P2814-19DL	MYD522	SAM	06/28/24 20:29	Not required	Kareem	Not Ok
78	CCV083	CCV083	CCV	06/28/24 20:34		Kareem	OK
79	CCB083	CCB083	CCB	06/28/24 20:38		Kareem	OK
80	P2820-19DL	MYD5E7	SAM	06/28/24 20:43	Not required	Kareem	Not Ok
81	P2822-01	MYD5H2	SAM	06/28/24 20:47		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

82	P2822-02	MYD5H2D	DUP	06/28/24 20:52		Kareem	OK
83	P2822-01L	MYD5H2L	SD	06/28/24 20:56		Kareem	OK
84	P2822-03	MYD5H2S	MS	06/28/24 21:01	MS fail for As(0.1722545PPM), Cu(0.3878602PPM), Se(Below RL), TI(Below RL), V(0.5143108PPM)	Kareem	OK
85	P2822-04	MYD5H4	SAM	06/28/24 21:05		Kareem	OK
86	P2822-05	MYD5H5	SAM	06/28/24 21:09		Kareem	OK
87	P2822-06	MYD5H6	SAM	06/28/24 21:13		Kareem	OK
88	P2822-07	MYD5H7	SAM	06/28/24 21:18		Kareem	OK
89	P2822-08	MYD5H8	SAM	06/28/24 21:22		Kareem	OK
90	P2822-09	MYD5H9	SAM	06/28/24 21:27		Kareem	OK
91	P2822-10	MYD5J0	SAM	06/28/24 21:31		Kareem	OK
92	P2822-11	MYD5J1	SAM	06/28/24 21:35		Kareem	OK
93	P2822-12	MYD5J2	SAM	06/28/24 21:40		Kareem	OK
94	P2822-13	MYD5J3	SAM	06/28/24 21:44		Kareem	OK
95	P2822-14	MYD5J4	SAM	06/28/24 21:48	Mn High	Kareem	Dilution
96	P2822-15	MYD5J5	SAM	06/28/24 21:53	Mn High	Kareem	Dilution
97	P2822-16	MYD5J6	SAM	06/28/24 21:57	Mn High	Kareem	Dilution
98	P2822-17	MYD5J7	SAM	06/28/24 22:02	Mn High	Kareem	Dilution
99	P2822-18	MYD5J8	SAM	06/28/24 22:06	Mn High	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

100	CCV084	CCV084	CCV	06/28/24 22:10		Kareem	OK
101	CCB084	CCB084	CCB	06/28/24 22:15		Kareem	OK
102	P2822-19	MYD5K0	SAM	06/28/24 22:19	Mn High	Kareem	Dilution
103	P2822-20	MYD5K1	SAM	06/28/24 22:23	Mn High	Kareem	Dilution
104	P2822-21	MYD5K2	SAM	06/28/24 22:28	Mn High	Kareem	Dilution
105	P2822-22	MYD5K4	SAM	06/28/24 22:32	Mn High	Kareem	Dilution
106	P2823-01	MYD5X1	SAM	06/28/24 22:36		Kareem	OK
107	P2823-02	MYD5X1D	DUP	06/28/24 22:41		Kareem	OK
108	P2823-01L	MYD5X1L	SD	06/28/24 22:45		Kareem	OK
109	P2823-03	MYD5X1S	MS	06/28/24 22:50	MS fail for As(0.0417517PPM),Cr(0.60 59198PPM),Pb(0.0818535P PM),Mn(1.991713PPM),Ni(0 .2536192PPM),Se(Below RL)	Kareem	OK
110	P2823-04	MYD5L1	SAM	06/28/24 22:54	Mn High	Kareem	Dilution
111	P2823-05	MYD5L2	SAM	06/28/24 22:58		Kareem	OK
112	P2823-06	MYD5L3	SAM	06/28/24 23:03	Mn High	Kareem	Dilution
113	P2823-07	MYD5L4	SAM	06/28/24 23:07	Mn High	Kareem	Dilution
114	P2823-08	MYD5L5	SAM	06/28/24 23:11	Mn High	Kareem	Dilution
115	P2823-09	MYD5L6	SAM	06/28/24 23:16	Mn High	Kareem	Dilution
116	P2823-10	MYD5L7	SAM	06/28/24 23:20	Mn High	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

117	P2823-11	MYD5L8	SAM	06/28/24 23:24	Mn High	Kareem	Dilution
118	P2823-12	MYD5L9	SAM	06/28/24 23:29		Kareem	OK
119	P2823-13	MYD5M0	SAM	06/28/24 23:33	Mn High	Kareem	Dilution
120	P2823-14	MYD5M1	SAM	06/28/24 23:37	Mn High	Kareem	Dilution
121	P2823-15	MYD5M2	SAM	06/28/24 23:42	Mn High	Kareem	Dilution
122	CCV085	CCV085	CCV	06/28/24 23:46		Kareem	OK
123	CCB085	CCB085	CCB	06/28/24 23:50		Kareem	OK
124	P2823-16	MYD5M3	SAM	06/28/24 23:55	Mn High	Kareem	Dilution
125	P2823-17	MYD5M4	SAM	06/28/24 23:59	Mn High	Kareem	Dilution
126	P2823-18	MYD5M5	SAM	06/29/24 00:04	Mn High	Kareem	Dilution
127	P2823-19	MYD5M6	SAM	06/29/24 00:08	Mn High	Kareem	Dilution
128	P2823-20	MYD5M7	SAM	06/29/24 00:12	Mn High	Kareem	Dilution
129	P2823-21	MYD5M8	SAM	06/29/24 00:17	Mn High	Kareem	Dilution
130	P2823-22	MYD5M9	SAM	06/29/24 00:21	Mn High	Kareem	Dilution
131	P2824-01	MYD5K5	SAM	06/29/24 00:25	Mn High	Kareem	Dilution
132	P2824-02	MYD5K5D	DUP	06/29/24 00:30	Mn High	Kareem	Dilution
133	P2824-01L	MYD5K5L	SD	06/29/24 00:34	Mn High	Kareem	Dilution
134	P2824-03	MYD5K5S	MS	06/29/24 00:38	Mn high,MS fail for Cr(0.6045513PPM),Se(Below RL),Ti(Below RL)	Kareem	Dilution
135	P2824-04	MYD5N0	SAM	06/29/24 00:43		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

136	P2824-05	MYD5N1	SAM	06/29/24 00:47	Mn High	Kareem	Dilution
137	P2824-06	MYD5N2	SAM	06/29/24 00:51	Mn High	Kareem	Dilution
138	P2824-07	MYD5N3	SAM	06/29/24 00:56	Mn High	Kareem	Dilution
139	P2824-08	MYD5N4	SAM	06/29/24 01:00	Mn High	Kareem	Dilution
140	P2824-09	MYD5N5	SAM	06/29/24 01:04	Mn High	Kareem	Dilution
141	P2824-10	MYD5N6	SAM	06/29/24 01:09	Mn High	Kareem	Dilution
142	P2824-11	MYD5N7	SAM	06/29/24 01:13	Mn High	Kareem	Dilution
143	P2824-12	MYD5N8	SAM	06/29/24 01:17	Mn High	Kareem	Dilution
144	CCV086	CCV086	CCV	06/29/24 01:22		Kareem	OK
145	CCB086	CCB086	CCB	06/29/24 01:26		Kareem	OK
146	P2824-13	MYD5N9	SAM	06/29/24 01:31	Mn High	Kareem	Dilution
147	P2824-14	MYD5P0	SAM	06/29/24 01:35		Kareem	OK
148	P2824-15	MYD5P1	SAM	06/29/24 01:40		Kareem	OK
149	P2824-16	MYD5P2	SAM	06/29/24 01:44		Kareem	OK
150	P2824-17	MYD5P3	SAM	06/29/24 01:48		Kareem	OK
151	P2824-18	MYD5P4	SAM	06/29/24 01:53		Kareem	OK
152	P2824-19	MYD5P5	SAM	06/29/24 01:57		Kareem	OK
153	P2824-20	MYD5P6	SAM	06/29/24 02:01	Mn High	Kareem	Dilution
154	P2824-21	MYD5P7	SAM	06/29/24 02:06		Kareem	OK
155	P2824-22	MYD5P8	SAM	06/29/24 02:10		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

156	P2825-01	MYD542	SAM	06/29/24 02:14		Kareem	OK
157	P2825-02	MYD542D	DUP	06/29/24 02:19		Kareem	OK
158	P2825-01L	MYD542L	SD	06/29/24 02:23		Kareem	OK
159	P2825-03	MYD542S	MS	06/29/24 02:28	MS fail for As(0.2705676PPM), Cu(0.3912561PPM), Se(Below RL), Zn(1.091943PPM)	Kareem	OK
160	P2825-04	MYD5P9	SAM	06/29/24 02:32		Kareem	OK
161	P2825-05	MYD5Q0	SAM	06/29/24 02:36		Kareem	OK
162	P2825-06	MYD5Q1	SAM	06/29/24 02:41		Kareem	OK
163	P2825-07	MYD5Q2	SAM	06/29/24 02:45		Kareem	OK
164	P2825-08	MYD5Q3	SAM	06/29/24 02:49		Kareem	OK
165	P2825-09	MYD5Q4	SAM	06/29/24 02:54		Kareem	OK
166	P2825-10	MYD5Q5	SAM	06/29/24 03:07		Kareem	OK
167	P2825-11	MYD5Q6	SAM	06/29/24 03:11	Mn High	Kareem	Dilution
168	CCV087	CCV087	CCV	06/29/24 03:16		Kareem	OK
169	CCB087	CCB087	CCB	06/29/24 03:25	Inst. paused for review	Kareem	OK
170	P2825-12	MYD5Q7	SAM	06/29/24 03:30		Kareem	OK
171	P2825-13	MYD5Q8	SAM	06/29/24 03:34	Mn High	Kareem	Dilution
172	P2825-14	MYD5Q9	SAM	06/29/24 03:39		Kareem	OK
173	P2825-15	MYD5R0	SAM	06/29/24 03:43		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

174	P2825-16	MYD5R1	SAM	06/29/24 03:47		Kareem	OK
175	P2825-17	MYD5R2	SAM	06/29/24 03:52	Mn High	Kareem	Dilution
176	P2825-18	MYD5G8	SAM	06/29/24 03:56		Kareem	OK
177	P2825-19	MYD5G9	SAM	06/29/24 04:00		Kareem	OK
178	P2825-20	MYD5H0	SAM	06/29/24 04:05		Kareem	OK
179	P2825-21	MYD5H1	SAM	06/29/24 04:09		Kareem	OK
180	P2826-01	MYD546	SAM	06/29/24 04:13		Kareem	OK
181	P2826-01L	MYD546L	SD	06/29/24 04:18		Kareem	OK
182	P2826-02	MYD546D	DUP	06/29/24 04:22		Kareem	OK
183	P2826-03	MYD546S	MS	06/29/24 04:27	MS fail for As(0.2951106PPM),Se(Below RL)	Kareem	OK
184	P2826-04	MYD5R3	SAM	06/29/24 04:31	Mn High	Kareem	Dilution
185	P2826-05	MYD5R4	SAM	06/29/24 04:35		Kareem	OK
186	P2826-06	MYD5R5	SAM	06/29/24 04:40		Kareem	OK
187	P2826-07	MYD5R6	SAM	06/29/24 04:44		Kareem	OK
188	CCV088	CCV088	CCV	06/29/24 04:48		Kareem	OK
189	CCB088	CCB088	CCB	06/29/24 04:53		Kareem	OK
190	P2826-08	MYD5R7	SAM	06/29/24 04:57		Kareem	OK
191	P2826-09	MYD5R8	SAM	06/29/24 05:02		Kareem	OK
192	P2826-10	MYD5R9	SAM	06/29/24 05:06		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

193	P2826-11	MYD5S0	SAM	06/29/24 05:10	Mn High	Kareem	Dilution
194	P2826-12	MYD5S1	SAM	06/29/24 05:15		Kareem	OK
195	P2826-13	MYD5S2	SAM	06/29/24 05:19		Kareem	OK
196	P2826-14	MYD5S3	SAM	06/29/24 05:23		Kareem	OK
197	P2826-15	MYD5S4	SAM	06/29/24 05:28	Mn High	Kareem	Dilution
198	P2826-16	MYD5S5	SAM	06/29/24 05:32		Kareem	OK
199	P2826-17	MYD5S6	SAM	06/29/24 05:37	Mn High	Kareem	Dilution
200	P2826-18	MYD5S7	SAM	06/29/24 05:41		Kareem	OK
201	P2826-19	MYD5S8	SAM	06/29/24 05:45		Kareem	OK
202	P2826-20	MYD5S9	SAM	06/29/24 05:50	Mn High	Kareem	Dilution
203	P2826-21	MYD5T0	SAM	06/29/24 05:54		Kareem	OK
204	P2826-22	MYD5T1	SAM	06/29/24 05:58		Kareem	OK
205	P2828-01	MYD5H3	SAM	06/29/24 06:03		Kareem	OK
206	P2828-01L	MYD5H3L	SD	06/29/24 06:07		Kareem	OK
207	P2828-02	MYD5H3D	DUP	06/29/24 06:11		Kareem	OK
208	P2828-03	MYD5H3S	MS	06/29/24 06:16	MS fail for Cr(0.3328370PPM),Cu(0.65 30863ppm).Se(Below RL),Ti(Below RL),Zn(1.610189PPM)	Kareem	OK
209	P2828-04	MYD5K6	SAM	06/29/24 06:20	Mn High	Kareem	Dilution
210	CCV089	CCV089	CCV	06/29/24 06:24		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM
STD. NAME	STD REF.#		
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935		
ICV Standard	MP80931		
CCV Standard	MP80934		
ICSA Standard	MP80932,MP80933		
CRI Standard	MP81024,M5819		
LCS Standard			
Chk Standard	MP80942,MP80944		

211	CCB089	CCB089	CCB	06/29/24 06:29		Kareem	OK
212	P2828-05	MYD5K7	SAM	06/29/24 06:33	Mn High	Kareem	Dilution
213	P2828-06	MYD5K8	SAM	06/29/24 06:38	Mn High	Kareem	Dilution
214	P2828-07	MYD5K9	SAM	06/29/24 06:42	Mn High	Kareem	Dilution
215	P2828-08	MYD5L0	SAM	06/29/24 06:46	Mn High	Kareem	Dilution
216	P2828-09	MYD4E6	SAM	06/29/24 06:50		Kareem	OK
217	P2828-10	MYD4E7	SAM	06/29/24 06:55		Kareem	OK
218	P2828-11	MYD4E8	SAM	06/29/24 06:59		Kareem	OK
219	P2828-12	MYD4E9	SAM	06/29/24 07:04		Kareem	OK
220	P2828-13	MYD4F0	SAM	06/29/24 07:08		Kareem	OK
221	P2828-14	MYD4F1	SAM	06/29/24 07:12		Kareem	OK
222	P2828-15	MYD4F2	SAM	06/29/24 07:17		Kareem	OK
223	P2828-16	MYD4F3	SAM	06/29/24 07:21		Kareem	OK
224	P2828-17	MYD4F4	SAM	06/29/24 07:25		Kareem	OK
225	P2828-18	MYD4F5	SAM	06/29/24 07:29		Kareem	OK
226	P2828-19	MYD4F6	SAM	06/29/24 07:34		Kareem	OK
227	P2820-19DL2	MYD5E7	SAM	06/29/24 07:38	2X for Mn	Kareem	Confirms
228	P2829-01	MYD5T2	SAM	06/29/24 07:43		Kareem	OK
229	P2829-02	MYD5T3	SAM	06/29/24 07:47		Kareem	OK
230	P2829-03	MYD5T4	SAM	06/29/24 07:51		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

231	P2829-04	MYD5T5	SAM	06/29/24 07:56		Kareem	OK
232	CCV090	CCV090	CCV	06/29/24 08:00		Kareem	OK
233	CCB090	CCB090	CCB	06/29/24 08:04		Kareem	OK
234	PB161710BL	PBS710	MB	06/29/24 08:09		Kareem	OK
235	PB161710BS	LCS710	LCS	06/29/24 08:13		Kareem	OK
236	PB161711BL	PBS711	MB	06/29/24 08:18		Kareem	OK
237	PB161711BS	LCS711	LCS	06/29/24 08:22		Kareem	OK
238	PB161712BL	PBS712	MB	06/29/24 08:27		Kareem	OK
239	PB161712BS	LCS712	LCS	06/29/24 08:31		Kareem	OK
240	PB161713BL	PBS713	MB	06/29/24 08:36		Kareem	OK
241	PB161713BS	LCS713	LCS	06/29/24 08:40		Kareem	OK
242	P2829-05	MYD5T6	SAM	06/29/24 08:45		Kareem	OK
243	P2829-06	MYD5T7	SAM	06/29/24 08:49		Kareem	OK
244	P2829-07	MYD5T8	SAM	06/29/24 08:53		Kareem	OK
245	P2829-08	MYD5T9	SAM	06/29/24 08:58		Kareem	OK
246	P2829-09	MYD5T9D	DUP	06/29/24 09:02		Kareem	OK
247	P2829-08L	MYD5T9L	SD	06/29/24 09:06		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

248	P2829-10	MYD5T9S	MS	06/29/24 09:11	MS fail for Ba(3.039342ppm),Cr(0.5076730ppm),Cu(1.570988ppm),Se(Below RL),Ti(Below RL),V(1.385149ppm),Zn(1.300744ppm)	Kareem	OK
249	P2829-11	MYD5W0	SAM	06/29/24 09:15		Kareem	OK
250	P2829-12	MYD5W1	SAM	06/29/24 09:19	Mn High	Kareem	Dilution
251	P2829-13	MYD5W2	SAM	06/29/24 09:24	Mn High	Kareem	Dilution
252	P2829-14	MYD5W3	SAM	06/29/24 09:28	Mn High	Kareem	Dilution
253	P2829-15	MYD5W4	SAM	06/29/24 09:32		Kareem	OK
254	CCV091	CCV091	CCV	06/29/24 09:37		Kareem	OK
255	CCB091	CCB091	CCB	06/29/24 09:41		Kareem	OK
256	P2829-16	MYD5W5	SAM	06/29/24 09:46		Kareem	OK
257	P2829-17	MYD5W6	SAM	06/29/24 09:50	Mn High	Kareem	Dilution
258	P2829-18	MYD5X0	SAM	06/29/24 09:54		Kareem	OK
259	P2829-19	MYD5X2	SAM	06/29/24 09:59		Kareem	OK
260	P2829-20	MYD5X3	SAM	06/29/24 10:03		Kareem	OK
261	P2830-01	MYD573	SAM	06/29/24 10:07		Kareem	OK
262	P2830-02	MYD573D	DUP	06/29/24 10:12		Kareem	OK
263	P2830-01L	MYD573L	SD	06/29/24 10:16		Kareem	OK
264	P2830-03	MYD573S	MS	06/29/24 10:21	MS fail for Se(Below RL),Ti(Below RL),Zn(1.235142ppm)	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

265	P2831-01	MYD5C5	SAM	06/29/24 10:25		Kareem	OK
266	P2831-02	MYD5C5D	DUP	06/29/24 10:29		Kareem	OK
267	P2831-01L	MYD5C5L	SD	06/29/24 10:34		Kareem	OK
268	P2831-03	MYD5C5S	MS	06/29/24 10:38	MS fail for Cu(0.7983432ppm),Pb(0.10 47544ppm),Se(Below RL),Tl(Below RL)	Kareem	OK
269	P2832-01	MYD5E3	SAM	06/29/24 10:42		Kareem	OK
270	P2832-02	MYD5E3D	DUP	06/29/24 10:47		Kareem	OK
271	P2832-01L	MYD5E3L	SD	06/29/24 10:51		Kareem	OK
272	P2832-03	MYD5E3S	MS	06/29/24 10:55	MS fail for Cu(1.167239ppm),Se(Below RL),Zn(1.444952ppm)	Kareem	OK
273	P2833-01	MYD5J9	SAM	06/29/24 11:00	Mn High	Kareem	Dilution
274	P2833-02	MYD5J9D	DUP	06/29/24 11:04	Mn High	Kareem	Dilution
275	P2833-01L	MYD5J9L	SD	06/29/24 11:08	Mn High	Kareem	Dilution
276	CCV092	CCV092	CCV	06/29/24 11:13		Kareem	OK
277	CCB092	CCB092	CCB	06/29/24 11:17		Kareem	OK
278	P2833-03	MYD5J9S	MS	06/29/24 11:22	Mn high,MS fail for Ba(4.687471ppm),Se(Below RL),Tl(Below RL),Zn(1.234995ppm)	Kareem	Dilution
279	P2834-01	MYD5K3	SAM	06/29/24 11:26	Mn High in Dup,MS	Kareem	Dilution
280	P2834-02	MYD5K3D	DUP	06/29/24 11:30	Mn High	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

281	P2834-01L	MYD5K3L	SD	06/29/24 11:35	Mn High	Kareem	Dilution
282	P2834-03	MYD5K3S	MS	06/29/24 11:39	Mn high,MS fail for Ba(1.897607ppm),Se(Below RL),TI(Below RL)	Kareem	Dilution
283	P3031-01	E11M7	SAM	06/29/24 11:43		Kareem	OK
284	P3031-02	E11M7D	DUP	06/29/24 11:48		Kareem	OK
285	P3031-01L	E11M7L	SD	06/29/24 11:52		Kareem	OK
286	P3031-03	E11M7S	MS	06/29/24 11:56	MS fail for As(0.1258610ppm)	Kareem	OK
287	P3032-01	E11E8	SAM	06/29/24 12:01		Kareem	OK
288	P3032-02	E11E8D	DUP	06/29/24 12:05		Kareem	OK
289	P3032-01L	E11E8L	SD	06/29/24 12:10		Kareem	OK
290	P3032-03	E11E8S	MS	06/29/24 12:14		Kareem	OK
291	P2835-01	MYD4A6	SAM	06/29/24 12:18		Kareem	OK
292	P2835-02	MYD4A7	SAM	06/29/24 12:23		Kareem	OK
293	P2835-03	MYD4A8	SAM	06/29/24 12:27		Kareem	OK
294	P2835-04	MYD4A9	SAM	06/29/24 12:31		Kareem	OK
295	P2835-05	MYD4A9D	DUP	06/29/24 12:36		Kareem	OK
296	P2835-04L	MYD4A9L	SD	06/29/24 12:40		Kareem	OK
297	P2835-06	MYD4A9S	MS	06/29/24 12:45	MS fail for As(0.2225569ppm),Ba(1.530131ppm),Se(Below RL)	Kareem	OK
298	CCV093	CCV093	CCV	06/29/24 12:49		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

299	CCB093	CCB093	CCB	06/29/24 12:53		Kareem	OK
300	P2835-07	MYD4B0	SAM	06/29/24 12:58		Kareem	OK
301	P2835-08	MYD4B1	SAM	06/29/24 13:02		Kareem	OK
302	P2835-09	MYD4B2	SAM	06/29/24 13:07		Kareem	OK
303	P2835-10	MYD4B3	SAM	06/29/24 13:11		Kareem	OK
304	P2835-11	MYD4B4	SAM	06/29/24 13:15		Kareem	OK
305	P2835-12	MYD4B5	SAM	06/29/24 13:20		Kareem	OK
306	P2835-13	MYD4B6	SAM	06/29/24 13:24		Kareem	OK
307	P2835-14	MYD4B7	SAM	06/29/24 13:29		Kareem	OK
308	P2835-15	MYD4B8	SAM	06/29/24 13:33		Kareem	OK
309	P2835-16	MYD4B9	SAM	06/29/24 13:37		Kareem	OK
310	P2835-17	MYD4C0	SAM	06/29/24 13:42		Kareem	OK
311	P2835-18	MYD4C1	SAM	06/29/24 13:46		Kareem	OK
312	P2835-19	MYD4C2	SAM	06/29/24 13:50		Kareem	OK
313	P2835-20	MYD4C3	SAM	06/29/24 13:55		Kareem	OK
314	P2835-21	MYD4C4	SAM	06/29/24 13:59		Kareem	OK
315	P2835-22	MYD4C5	SAM	06/29/24 14:04		Kareem	OK
316	CCV094	CCV094	CCV	06/29/24 14:08		Kareem	OK
317	CCB094	CCB094	CCB	06/29/24 14:12		Kareem	OK
318	CCV095	CCV095	CCV	06/29/24 15:25	Inst. paused for review	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM
STD. NAME	STD REF.#		
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935		
ICV Standard	MP80931		
CCV Standard	MP80934		
ICSA Standard	MP80932,MP80933		
CRI Standard	MP81024,M5819		
LCS Standard			
Chk Standard	MP80942,MP80944		

319	CCB095	CCB095	CCB	06/29/24 15:30		Kareem	OK
320	P2822-14DL	MYD5J4	SAM	06/29/24 15:40	2X For Mn	Kareem	Confirms
321	P2822-15DL	MYD5J5	SAM	06/29/24 15:45	5X For Mn	Kareem	Confirms
322	P2822-16DL	MYD5J6	SAM	06/29/24 15:49	10X For Mn	Kareem	Confirms
323	P2822-17DL	MYD5J7	SAM	06/29/24 15:53	5X For Mn	Kareem	Confirms
324	P2822-18DL	MYD5J8	SAM	06/29/24 15:58	5X For Mn	Kareem	Confirms
325	P2822-19DL	MYD5K0	SAM	06/29/24 16:02	5X For Mn	Kareem	Confirms
326	P2822-20DL	MYD5K1	SAM	06/29/24 16:07	5X For Mn	Kareem	Confirms
327	P2822-21DL	MYD5K2	SAM	06/29/24 16:11	5X For Mn	Kareem	Confirms
328	P2822-22DL	MYD5K4	SAM	06/29/24 16:16	5X For Mn	Kareem	Confirms
329	P2823-04DL	MYD5L1	SAM	06/29/24 16:20	5X For Mn	Kareem	Confirms
330	P2823-06DL	MYD5L3	SAM	06/29/24 16:25	2X For Mn	Kareem	Confirms
331	P2823-07DL	MYD5L4	SAM	06/29/24 16:29	5X For Mn	Kareem	Confirms
332	P2823-08DL	MYD5L5	SAM	06/29/24 16:33	5X For Mn	Kareem	Confirms
333	P2823-09DL	MYD5L6	SAM	06/29/24 16:38	5X For Mn	Kareem	Confirms
334	P2823-10DL	MYD5L7	SAM	06/29/24 16:42	2X For Mn	Kareem	Confirms
335	P2823-11DL	MYD5L8	SAM	06/29/24 16:47	5X For Mn	Kareem	Confirms
336	P2823-13DL	MYD5M0	SAM	06/29/24 16:51	5X For Mn	Kareem	Confirms
337	P2823-14DL	MYD5M1	SAM	06/29/24 16:56	5X For Mn	Kareem	Confirms
338	P2823-15DL	MYD5M2	SAM	06/29/24 17:00	2X For Mn	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

339	P2823-16DL	MYD5M3	SAM	06/29/24 17:04	5X For Mn	Kareem	Confirms
340	CCV096	CCV096	CCV	06/29/24 17:09		Kareem	OK
341	CCB096	CCB096	CCB	06/29/24 17:13		Kareem	OK
342	P2823-17DL	MYD5M4	SAM	06/29/24 17:18	5X For Mn	Kareem	Confirms
343	P2823-18DL	MYD5M5	SAM	06/29/24 17:22	2X For Mn	Kareem	Confirms
344	P2823-19DL	MYD5M6	SAM	06/29/24 17:27	5X For Mn	Kareem	Confirms
345	P2823-20DL	MYD5M7	SAM	06/29/24 17:31	5X For Mn	Kareem	Confirms
346	P2823-21DL	MYD5M8	SAM	06/29/24 17:35	2X For Mn	Kareem	Confirms
347	P2823-22DL	MYD5M9	SAM	06/29/24 17:40	5X For Mn	Kareem	Confirms
348	P2824-04DL	MYD5N0	SAM	06/29/24 17:44	2X For Mn	Kareem	Confirms
349	P2824-05DL	MYD5N1	SAM	06/29/24 17:49	5X For Mn	Kareem	Confirms
350	P2824-06DL	MYD5N2	SAM	06/29/24 17:53	2X For Mn	Kareem	Confirms
351	P2824-07DL	MYD5N3	SAM	06/29/24 17:58	5X For Mn	Kareem	Confirms
352	P2824-08DL	MYD5N4	SAM	06/29/24 18:02	5X For Mn	Kareem	Confirms
353	P2824-09DL	MYD5N5	SAM	06/29/24 18:06	5X For Mn	Kareem	Confirms
354	P2824-10DL	MYD5N6	SAM	06/29/24 18:11	2X For Mn	Kareem	Confirms
355	P2824-11DL	MYD5N7	SAM	06/29/24 18:15	2X For Mn	Kareem	Confirms
356	P2824-12DL	MYD5N8	SAM	06/29/24 18:20	5X For Mn	Kareem	Confirms
357	P2824-13DL	MYD5N9	SAM	06/29/24 18:24	10X For Mn	Kareem	Confirms
358	P2824-20DL	MYD5P6	SAM	06/29/24 18:29	2X For Mn	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

359	P2824-01DL	MYD5K5	SAM	06/29/24 18:33	5X For Mn	Kareem	Confirms
360	P2824-02DL	MYD5K5D	DUP	06/29/24 18:38	5X For Mn	Kareem	Confirms
361	P2824-01LDL	MYD5K5L	SD	06/29/24 18:42	25X For Mn	Kareem	Confirms
362	CCV097	CCV097	CCV	06/29/24 18:46		Kareem	OK
363	CCB097	CCB097	CCB	06/29/24 18:51		Kareem	OK
364	P2824-03DL	MYD5K5S	MS	06/29/24 18:55	5X For Mn	Kareem	Confirms
365	P2825-11DL	MYD5Q6	SAM	06/29/24 19:00	2X For Mn	Kareem	Confirms
366	P2825-13DL	MYD5Q8	SAM	06/29/24 19:04	5X For Mn	Kareem	Confirms
367	P2825-17DL	MYD5R2	SAM	06/29/24 19:09	5X For Mn	Kareem	Confirms
368	P2826-04DL	MYD5R3	SAM	06/29/24 19:13	2X For Mn	Kareem	Confirms
369	P2826-11DL	MYD5S0	SAM	06/29/24 19:17	2X For Mn	Kareem	Confirms
370	P2826-15DL	MYD5S4	SAM	06/29/24 19:22	5X For Mn	Kareem	Confirms
371	P2826-17DL	MYD5S6	SAM	06/29/24 19:26	5X For Mn	Kareem	Confirms
372	P2826-20DL	MYD5S9	SAM	06/29/24 19:31	2X For Mn	Kareem	Confirms
373	P2828-04DL	MYD5K6	SAM	06/29/24 19:35	10X For Mn	Kareem	Confirms
374	P2828-05DL	MYD5K7	SAM	06/29/24 19:40	5X For Mn	Kareem	Confirms
375	P2828-06DL	MYD5K8	SAM	06/29/24 19:44	10X For Mn	Kareem	Confirms
376	P2828-07DL	MYD5K9	SAM	06/29/24 19:49	5X For Mn	Kareem	Confirms
377	P2828-08DL	MYD5L0	SAM	06/29/24 19:53	5X For Mn	Kareem	Confirms
378	P2829-12DL	MYD5W1	SAM	06/29/24 19:57	2X For Mn	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

379	P2829-13DL	MYD5W2	SAM	06/29/24 20:02	2X For Mn	Kareem	Confirms
380	P2829-14DL	MYD5W3	SAM	06/29/24 20:06	5X For Mn	Kareem	Confirms
381	P2829-17DL	MYD5W6	SAM	06/29/24 20:11	5X For Mn	Kareem	Confirms
382	P2833-01DL	MYD5J9	SAM	06/29/24 20:15	5X For Mn	Kareem	Confirms
383	P2833-02DL	MYD5J9D	DUP	06/29/24 20:20	5X For Mn	Kareem	Confirms
384	CCV098	CCV098	CCV	06/29/24 20:24		Kareem	OK
385	CCB098	CCB098	CCB	06/29/24 20:28		Kareem	OK
386	P2833-01LDL	MYD5J9L	SD	06/29/24 20:33	25X For Mn	Kareem	Confirms
387	P2833-03DL	MYD5J9S	MS	06/29/24 20:37	5X For Mn	Kareem	Confirms
388	P2834-01DL	MYD5K3	SAM	06/29/24 20:42	2X For Mn	Kareem	Confirms
389	P2834-02DL	MYD5K3D	DUP	06/29/24 20:46	2X For Mn	Kareem	Confirms
390	P2834-01LDL	MYD5K3L	SD	06/29/24 20:51	10X For Mn	Kareem	Confirms
391	PB161695BL	PBS695	MB	06/29/24 20:59		Kareem	OK
392	PB161695BS	LCS695	LCS	06/29/24 21:04		Kareem	OK
393	P2811-06	MYD4X2	SAM	06/29/24 21:08		Kareem	OK
394	P2821-01A	MYD4L1A	PS	06/29/24 21:13	PS for - Cu-0.0072 ml - M5521 , Pb - 0.026ml - MP81249 , Se - 0.07ml - MP81115 , Tl - 0.05ml - MP81116 , Zn - 0.015ml - M5819 - 10 ML Sample.	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

395	P2822-01A	MYD5H2A	PS	06/29/24 21:17	PS For AS-0.0034ML M5801,CU-0.0078ML M5521,SE-0.07ML MP81115,TL-0.05ML MP81116,V-0.01ML M5227- 10 ML Sample	Kareem	OK
396	P2823-01A	MYD5X1A	PS	06/29/24 21:21	PS For AS-0.008ML MP81248,CR-0.012ML M5658,PB-0.016ML MP81249,MN-0.04ML M5698,NI-0.0051ML M5111,SE-0.07ML MP81115PS for AS-0.0059ML M5801,SE-0.07ML MP81115 - 10 ML Sample	Kareem	OK
397	P2824-01A	MYD5K5A	PS	06/29/24 21:26	PS for CR-0.012ML M5658,SE-0.07ML MPS1115,TL-0.05ML MP81116PS for AS-0.0059ML M5801,SE-0.07ML MP81115 - 10 ML Sample	Kareem	OK
398	P2825-01A	MYD542A	PS	06/29/24 21:30	PS for AS-0.0054ML M5801,CU-0.0078ML M5698,SE-0.07ML MP81115,ZN-0.022ML M5819 -10 ML Sample	Kareem	OK
399	P2826-01A	MYD546A	PS	06/29/24 21:35	PS for AS-0.0059ML M5801,SE-0.07ML MP81115 - 10 ML Sample	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

400	P2828-01A	MYD5H3A	PS	06/29/24 21:39	PS for CR- 0.0067ML M5658,CU- 0.013ML M5521,SE - 0.07ML MP81115,TL-0.05ML MP81116,ZN-0.032ML M5819- 10 ML Sample	Kareem	OK
401	P2829-08A	MYD5T9A	PS	06/29/24 21:43	PS for BA-0.061ML M5387,CR-0.01ML M5658,CU-0.031ML M5521,SE-0.07ML MP81115,TL-0.05ML MP81116,V-0.028ML M5227,ZN-0.026ML M5819-10 ml sample	Kareem	OK
402	P2830-01A	MYD573A	PS	06/29/24 21:48	PS for SE-0.07ML MP81115,TL-0.05ML MP81116,ZN-0.025ML M5819 - 10 ML Sample	Kareem	OK
403	P2831-01A	MYD5C5A	PS	06/29/24 21:52	PS for Cu-0.016ml - M5521 , Pb-0.021ml MP81249 , Se-0.07ml - MP81115 , TI - 0.05ml - MP81116 - in to 10 ml Sample	Kareem	OK
404	P2832-01A	MYD5E3A	PS	06/29/24 21:56	PS for Cu-0.024ml - M5521 , Se-0.07ml MP81115, Zn-0.029 - M5819 - in to 10 ml Sample	Kareem	OK
405	CCV099	CCV099	CCV	06/29/24 22:01		Kareem	OK
406	CCB099	CCB099	CCB	06/29/24 22:05		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB131456

Review By	Sarabjit Jaswal	Review On	6/29/2024 9:25:13 PM
Supervise By	mohammad ahmed	Supervise On	6/30/2024 4:35:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP80924,MP81209,MP80928,MP80927,MP80926,MP80925,MP80935
ICV Standard	MP80931
CCV Standard	MP80934
ICSA Standard	MP80932,MP80933
CRI Standard	MP81024,M5819
LCS Standard	
Chk Standard	MP80942,MP80944

407	P2833-01A	MYD5J9A	PS	06/29/24 22:10	PS for BA-0.094ML M5387,SE-0.07ML MP81115,TL-0.05ML MP81116,ZN-0.025ML M5819 - 10 ML Sample	Kareem	OK
408	P2834-01A	MYD5K3A	PS	06/29/24 22:14	PS for Ba - 0.038ml - M5387 , Se-0.07ml - MP81115 , TI - 0.05ml - MP81116 - in to 10 ml Sample	Kareem	OK
409	P2835-04A	MYD4A9A	PS	06/29/24 22:18	PS for As-0.0044ml M5801, Ba-0.031ml M5387 , Se-0.07ml MP81115 - in to 10 ml Sample.	Kareem	OK
410	P3031-01A	E11M7A	PS	06/29/24 22:23	PS for As-0.025ml MP81248 - in to 10 ml Sample	Kareem	OK
411	P2834-03DL	MYD5K3S	MS	06/29/24 22:27	2X For Mn	Kareem	Confirms
412	CCV001	CCV001	CCV	06/29/24 22:32		Kareem	OK
413	CCB001	CCB001	CCB	06/29/24 22:36		Kareem	OK