

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: SDG No.: MYE440
 Instrument ID: P7 Date: 10/10/2024
 Run Batch: LB132901

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.0	0.7	5	0.93
Co - 59	59.0	0.7	5	0.85
In - 113	113.0	0.7	5	1.2
In - 115	115.0	0.7	5	0.89
Mg - 24	24.0	0.8	5	0.90
Mg - 25	25.0	0.7	5	0.58
Mg - 26	25.9	0.7	5	0.64
Pb - 206	206.1	0.8	5	0.78
Pb - 207	207.0	0.7	5	1.1
Pb - 208	208.0	0.7	5	0.94

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51495 MA No.: SDG No.: MYE440
 Instrument ID: P7 Date: 10/15/2024
 Run Batch: LB132969

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.1	0.7	5	0.57
Co - 59	58.9	0.8	5	0.33
In - 113	113.0	0.8	5	0.50
In - 115	115.0	0.7	5	0.21
Mg - 24	24.0	0.7	5	0.30
Mg - 25	25.0	0.7	5	0.52
Mg - 26	26.0	0.7	5	0.58
Pb - 206	205.9	0.8	5	0.74
Pb - 207	206.9	0.7	5	0.59
Pb - 208	207.9	0.8	5	0.79