

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132969

Review By	Sarabjit Jaswal	Review On	10/17/2024 6:42:29 AM
Supervise By	Mohan Bera	Supervise On	10/17/2024 6:43:13 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE023	TUNE	10/15/24 17:54		Jaswal	OK
2	S00	S0	CAL1	10/15/24 18:55		Jaswal	OK
3	S02	S02	CAL3	10/15/24 18:58		Jaswal	OK
4	S03	S03	CAL4	10/15/24 19:02		Jaswal	OK
5	S04	S04	CAL5	10/15/24 19:05		Jaswal	OK
6	S05	S05	CAL6	10/15/24 19:08		Jaswal	OK
7	S06	S06	CAL7	10/15/24 19:11		Jaswal	OK
8	S07	S07	CAL8	10/15/24 19:17		Jaswal	OK
9	S08	S08	CAL9	10/15/24 19:19		Jaswal	OK
10	ICV023	ICV023	ICV	10/15/24 19:30	Inst. paused for calibration review	Jaswal	OK
11	ICB023	ICB023	ICB	10/15/24 19:33		Jaswal	OK
12	ICSA023	ICSA023	ICSA	10/15/24 19:37		Jaswal	OK
13	ICSAB023	ICSAB023	ICSAB	10/15/24 19:40		Jaswal	OK
14	CCV083	CCV083	CCV	10/15/24 19:43		Jaswal	OK
15	CCB083	CCB083	CCB	10/15/24 19:46		Jaswal	OK
16	PB163820BL	PBS820	MB	10/15/24 19:49		Jaswal	OK
17	PB163820BS	LCS820	LCS	10/15/24 19:53		Jaswal	OK
18	PB163820BSD	LCSD820	LCSD	10/15/24 19:56		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132969

Review By	Sarabjit Jaswal	Review On	10/17/2024 6:42:29 AM
Supervise By	Mohan Bera	Supervise On	10/17/2024 6:43:13 AM
STD. NAME	STD REF.#		
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569		
ICV Standard	MP82578		
CCV Standard	MP82579		
ICSA Standard	MP82580,MP82581		
CRI Standard			
LCS Standard	MP82583		
Chk Standard	MP82587,MP82586		

19	PB163821BL	PBS821	MB	10/15/24 19:59		Jaswal	OK
20	PB163821BS	LCS821	LCS	10/15/24 20:03		Jaswal	OK
21	PB163821BSD	LCSD821	LCSD	10/15/24 20:06		Jaswal	OK
22	P2815-11DL	MYD536	SAM	10/15/24 20:10	4X For Ba,Co,Ni,V	Jaswal	Confirms
23	P2815-12DL	MYD536D	DUP	10/15/24 20:13	4X For Ba,Co,Ni,V	Jaswal	Confirms
24	P2815-11LDL	MYD536L	SD	10/15/24 20:16	20X For Ba,Co,Ni,V	Jaswal	Confirms
25	P2815-13DL	MYD536S	MS	10/15/24 20:20	4X For Ba,Co,Ni,V	Jaswal	Confirms
26	P2815-11MSDL	MYD536SRE	MS	10/15/24 20:23	4X For Ba,Co,Ni,V	Jaswal	Confirms
27	P2821-01DL	MYD4L1	SAM	10/15/24 20:26	4X For Ba,Co,Ni,V	Jaswal	Confirms
28	P2821-02DL	MYD4L1D	DUP	10/15/24 20:30	4X For Ba,Co,Ni,V	Jaswal	Confirms
29	P2821-01LDL	MYD4L1L	SD	10/15/24 20:33	4X For Ba,Co,Ni,V	Jaswal	Confirms
30	P2821-03DL	MYD4L1S	MS	10/15/24 20:36	4X For Ba,Co,Ni,V	Jaswal	Confirms
31	P2821-01MSDL	MYD4L1SRE	MS	10/15/24 20:40	4X For Ba,Co,Ni,V	Jaswal	Confirms
32	P2822-01REDL	MYD5H2	SAM	10/15/24 20:43	Not Use	Jaswal	Not Ok
33	P2822-02REDL	MYD5H2D	DUP	10/15/24 20:46	Not Use	Jaswal	Not Ok
34	P2822-01LREDL	MYD5H2L	SD	10/15/24 20:50	Not Use	Jaswal	Not Ok
35	P2822-03REDL	MYD5H2S	MS	10/15/24 20:53	Not Use	Jaswal	Not Ok
36	P2822-01MSREDL	MYD5H2SRE	MS	10/15/24 20:56	Not Use	Jaswal	Not Ok
37	P2830-01	MYD573	SAM	10/15/24 20:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
38	P2830-02	MYD573D	DUP	10/15/24 21:03	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132969

Review By	Sarabjit Jaswal	Review On	10/17/2024 6:42:29 AM
Supervise By	Mohan Bera	Supervise On	10/17/2024 6:43:13 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

39	P2830-01L	MYD573L	SD	10/15/24 21:06	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
40	P2830-03	MYD573S	MS	10/15/24 21:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
41	P2830-01MS	MYD573SRE	MS	10/15/24 21:12	Ba,Co,Ni,V High	Jaswal	Dilution
42	P2830-01DL	MYD573	SAM	10/15/24 21:16	4X For Ba,Co,Ni,V	Jaswal	Confirms
43	P2830-02DL	MYD573D	DUP	10/15/24 21:19	4X For Ba,Co,Ni,V	Jaswal	Confirms
44	P2830-01LDL	MYD573L	SD	10/15/24 21:22	20X For Ba,Co,Ni,V	Jaswal	Confirms
45	CCV084	CCV084	CCV	10/15/24 21:26		Jaswal	OK
46	CCB084	CCB084	CCB	10/15/24 21:29		Jaswal	OK
47	P2830-03DL	MYD573S	MS	10/15/24 21:32	4X For Ba,Co,Ni,V	Jaswal	Confirms
48	P2830-01MSDL	MYD573SRE	MS	10/15/24 21:35	4X For Ba,Co,Ni,V	Jaswal	Confirms
49	P2831-01DL	MYD5C5	SAM	10/15/24 21:39	4X For Ba,Co,Ni,V	Jaswal	Confirms
50	P2831-02DL	MYD5C5D	DUP	10/15/24 21:42	4X For Ba,Co,Ni,V	Jaswal	Confirms
51	P2831-01LDL	MYD5C5L	SD	10/15/24 21:45	20X For Ba,Co,Ni,V	Jaswal	Confirms
52	P2831-03DL	MYD5C5S	MS	10/15/24 21:49	4X For Ba,Co,Ni,V	Jaswal	Confirms
53	P2831-01MSDL	MYD5C5SRE	MS	10/15/24 21:52	4X For Ba,Co,Ni,V	Jaswal	Confirms
54	P2832-01	MYD5E3	SAM	10/15/24 21:55	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
55	P2832-02	MYD5E3D	DUP	10/15/24 21:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
56	P2832-01L	MYD5E3L	SD	10/15/24 22:02	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
57	P2832-03	MYD5E3S	MS	10/15/24 22:05	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132969

Review By	Sarabjit Jaswal	Review On	10/17/2024 6:42:29 AM
Supervise By	Mohan Bera	Supervise On	10/17/2024 6:43:13 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

58	P2832-01MS	MYD5E3SRE	MS	10/15/24 22:09	MS fail for Pb(107.23ppb), Ba,Co,Ni,V High	Jaswal	Dilution
59	P2832-01DL	MYD5E3	SAM	10/15/24 22:12	4X For Ba,Co,Ni,V	Jaswal	Confirms
60	P2832-02DL	MYD5E3D	DUP	10/15/24 22:15	4X For Ba,Co,Ni,V	Jaswal	Confirms
61	P2832-01LDL	MYD5E3L	SD	10/15/24 22:18	20X For Ba,Co,Ni,V	Jaswal	Confirms
62	P2832-03DL	MYD5E3S	MS	10/15/24 22:22	4X For Ba,Co,Ni,V	Jaswal	Confirms
63	P2832-01MSDL	MYD5E3SRE	MS	10/15/24 22:25	4X For Ba,Co,Ni,V	Jaswal	Confirms
64	P2833-01DL	MYD5J9	SAM	10/15/24 22:29	4X For Ba,Co,Ni,V	Jaswal	Confirms
65	P2833-02DL	MYD5J9D	DUP	10/15/24 22:32	4X For Ba,Co,Ni,V	Jaswal	Confirms
66	P2833-01LDL	MYD5J9L	SD	10/15/24 22:35	20X For Ba,Co,Ni,V	Jaswal	Confirms
67	P2833-03DL	MYD5J9S	MS	10/15/24 22:39	4X For Ba,Co,Ni,V	Jaswal	Confirms
68	P2833-01MSDL	MYD5J9SRE	MS	10/15/24 22:42	4X For Ba,Co,Ni,V	Jaswal	Confirms
69	P2834-01DL	MYD5K3	SAM	10/15/24 22:45	4X For Ba,Co,Ni,V	Jaswal	Confirms
70	P2834-02DL	MYD5K3D	DUP	10/15/24 22:49	4X For Ba,Co,Ni,V	Jaswal	Confirms
71	P2834-01LDL	MYD5K3L	SD	10/15/24 22:52	20X For Ba,Co,Ni,V	Jaswal	Confirms
72	P2834-03DL	MYD5K3S	MS	10/15/24 22:55	4X For Ba,Co,Ni,V	Jaswal	Confirms
73	P2834-01MSDL	MYD5K3SRE	MS	10/15/24 22:59	4X For Ba,Co,Ni,V	Jaswal	Confirms
74	P2815-11	MYD536	SAM	10/15/24 23:08	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
75	P2815-12	MYD536D	DUP	10/15/24 23:12	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
76	P2815-11L	MYD536L	SD	10/15/24 23:15	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
77	P2815-13	MYD536S	MS	10/15/24 23:18	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132969

Review By	Sarabjit Jaswal	Review On	10/17/2024 6:42:29 AM
Supervise By	Mohan Bera	Supervise On	10/17/2024 6:43:13 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

78	CCV085	CCV085	CCV	10/15/24 23:25		Jaswal	OK
79	CCB085	CCB085	CCB	10/15/24 23:27		Jaswal	OK
80	P2815-11MS	MYD536SRE	MS	10/15/24 23:31	MS Fail For Ag(Below RL),Ba,Co,Ni,V High	Jaswal	Dilution
81	P2835-17	MYD4C0	SAM	10/15/24 23:34		Jaswal	OK
82	P2835-18	MYD4C1	SAM	10/15/24 23:37		Jaswal	OK
83	P2835-19	MYD4C2	SAM	10/15/24 23:40		Jaswal	OK
84	P2835-20	MYD4C3	SAM	10/15/24 23:44		Jaswal	OK
85	P2835-04REDL	MYD4A9	SAM	10/15/24 23:47	Not Use	Jaswal	Not Ok
86	P2835-05REDL	MYD4A9D	DUP	10/15/24 23:50	Not Use	Jaswal	Not Ok
87	P2835-04LREDL	MYD4A9L	SD	10/15/24 23:54	Not Use	Jaswal	Not Ok
88	P2835-06REDL	MYD4A9S	MS	10/15/24 23:57	Not Use	Jaswal	Not Ok
89	P2835-04MSREDL	MYD4A9SRE	MS	10/16/24	Not Use	Jaswal	Not Ok
90	P2836-08DL	MYD4D3	SAM	10/16/24 00:03	4X For Ba,Co,Ni,V	Jaswal	Confirms
91	P2836-09DL	MYD4D3D	DUP	10/16/24 00:07	4X For Ba,Co,Ni,V	Jaswal	Confirms
92	P2836-08LDL	MYD4D3L	SD	10/16/24 00:10	20X For Ba,Co,Ni,V	Jaswal	Confirms
93	P2836-10DL	MYD4D3S	MS	10/16/24 00:14	4X For Ba,Co,Ni,V	Jaswal	Confirms
94	P2836-08MSDL	MYD4D3SRE	MS	10/16/24 00:17	4X For Ba,Co,Ni,V	Jaswal	Confirms
95	P2822-01DL	MYD5H2	SAM	10/16/24 00:20	4X For Ba,Co,Ni,V	Jaswal	Confirms
96	P2822-02DL	MYD5H2D	DUP	10/16/24 00:24	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132969

Review By	Sarabjit Jaswal	Review On	10/17/2024 6:42:29 AM
Supervise By	Mohan Bera	Supervise On	10/17/2024 6:43:13 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

97	P2822-01LDL	MYD5H2L	SD	10/16/24 00:27	20X For Ba,Co,Ni,V	Jaswal	Confirms
98	P2822-03DL	MYD5H2S	MS	10/16/24 00:30	4X For Ba,Co,Ni,V	Jaswal	Confirms
99	P2822-01MSDL	MYD5H2SRE	MS	10/16/24 00:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
100	P2829-14	MYD5W3	SAM	10/16/24 00:37		Jaswal	OK
101	CCV086	CCV086	CCV	10/16/24 00:40		Jaswal	OK
102	CCB086	CCB086	CCB	10/16/24 00:43		Jaswal	OK
103	P2829-15	MYD5W4	SAM	10/16/24 00:47		Jaswal	OK
104	P2829-16	MYD5W5	SAM	10/16/24 00:50		Jaswal	OK
105	P2829-17	MYD5W6	SAM	10/16/24 00:53		Jaswal	OK
106	P2829-18	MYD5X0	SAM	10/16/24 00:56		Jaswal	OK
107	P2829-19	MYD5X2	SAM	10/16/24 01:00		Jaswal	OK
108	P2829-20	MYD5X3	SAM	10/16/24 01:03		Jaswal	OK
109	P2835-04DL	MYD4A9	SAM	10/16/24 01:06	4X For Ba,Co,Ni,V	Jaswal	Confirms
110	P2835-05DL	MYD4A9D	DUP	10/16/24 01:10	4X For Ba,Co,Ni,V	Jaswal	Confirms
111	P2835-04LDL	MYD4A9L	SD	10/16/24 01:13	20X For Ba,Co,Ni,V	Jaswal	Confirms
112	P2835-06DL	MYD4A9S	MS	10/16/24 01:16	4X For Ba,Co,Ni,V	Jaswal	Confirms
113	P2835-04MSDL	MYD4A9SRE	MS	10/16/24 01:20	4X For Ba,Co,Ni,V	Jaswal	Confirms
114	P2829-14DL	MYD5W3	SAM	10/16/24 01:23	Not Required	Jaswal	Not Ok
115	P4319-09DL	MC0C97	SAM	10/16/24 01:27	Not use	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB132969

Review By	Sarabjit Jaswal	Review On	10/17/2024 6:42:29 AM
Supervise By	Mohan Bera	Supervise On	10/17/2024 6:43:13 AM

STD. NAME	STD REF.#
ICAL Standard	MP82568,MP82577,MP82576,MP82574,MP82573,MP82572,MP82571,MP82570,MP82569
ICV Standard	MP82578
CCV Standard	MP82579
ICSA Standard	MP82580,MP82581
CRI Standard	
LCS Standard	MP82583
Chk Standard	MP82587,MP82586

116	P4319-13A	MC0CC1A	PS	10/16/24 01:30	PS For Pb(0.02ML MP82819-10ML SAMPLE)	Jaswal	OK
117	P4320-08DL	MC0C96	SAM	10/16/24 01:33	5X For Zn	Jaswal	Confirms
118	P2829-17DL	MYD5W6	SAM	10/16/24 01:37	Not Required	Jaswal	Not Ok
119	P2748-04	MYD3Z2	SAM	10/16/24 01:40	Not Use	Jaswal	Not Ok
120	P4240-02A	MH2B88A	PS	10/16/24 01:43	PS For Pb(0.019ML MP82671-10ML SAMPLE)	Jaswal	OK
121	CCV087	CCV087	CCV	10/16/24 01:53		Jaswal	OK
122	CCB087	CCB087	CCB	10/16/24 01:56		Jaswal	OK