

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB132029

Review By	rubina	Review On	8/16/2024 12:37:28 PM
Supervise By	Iwona	Supervise On	8/16/2024 1:36:13 PM
SubDirectory	LB132029	Test	Phenolics
STD. NAME	STD REF.#		
ICAL Standard	WP109238		
ICV Standard	WP109240		
CCV Standard	WP109239		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP109049		
Chk Standard	WP107255,WP107256,WP109242,WP109241		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.00PPM	0.00PPM	CAL1	08/15/24 11:54		rubina	OK
2	0.05PPM	0.05PPM	CAL2	08/15/24 11:54		rubina	OK
3	0.1PPM	0.1PPM	CAL3	08/15/24 11:54		rubina	OK
4	0.25PPM	0.25PPM	CAL4	08/15/24 11:54		rubina	OK
5	0.50PPM	0.50PPM	CAL5	08/15/24 11:55		rubina	OK
6	1.0PPM	1.0PPM	CAL6	08/15/24 11:55		rubina	OK
7	2.0PPM	2.0PPM	CAL7	08/15/24 11:55		rubina	OK
8	ICV1	ICV1	ICV	08/15/24 12:32		rubina	OK
9	ICB1	ICB1	ICB	08/15/24 12:32		rubina	OK
10	CCV1	CCV1	CCV	08/15/24 12:32		rubina	OK
11	CCB1	CCB1	CCB	08/15/24 12:32		rubina	OK
12	PB162737BL	PB162737BL	MB	08/15/24 12:32		rubina	OK
13	PB162737BS	PB162737BS	LCS	08/15/24 12:32		rubina	OK
14	P3390-10	LOD-MDL-WATER-04	SAM	08/15/24 12:42		rubina	OK
15	P3390-11	LOQ-WATER-05-QT3	LOQ	08/15/24 12:42		rubina	OK
16	P3536-01	MANHOLE	SAM	08/15/24 12:42		rubina	OK
17	P3536-01DUP	MANHOLEDUP	DUP	08/15/24 12:42		rubina	OK
18	P3536-01MS	MANHOLEMS	MS	08/15/24 12:42		rubina	OK

Instrument ID: KONELAB

Daily Analysis Runlog For Sequence/QC Batch ID # LB132029

Review By	rubina	Review On	8/16/2024 12:37:28 PM
Supervise By	Iwona	Supervise On	8/16/2024 1:36:13 PM
SubDirectory	LB132029	Test	Phenolics
STD. NAME	STD REF.#		
ICAL Standard	WP109238		
ICV Standard	WP109240		
CCV Standard	WP109239		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP109049		
Chk Standard	WP107255,WP107256,WP109242,WP109241		

19	P3536-01MSD	MANHOLEMSD	MSD	08/15/24 12:42		rubina	OK
20	CCV2	CCV2	CCV	08/15/24 12:42		rubina	OK
21	CCB2	CCB2	CCB	08/15/24 12:46		rubina	OK