

Report of Analysis

| Client: | Chemtech Consulting Group | | | | Date Collected: | 07/29/24 | |
|---------------------------------|---------------------------|---------------------|-----------|----------|--------------------|------------------|-------------------|
| Project: | LOD-LOQ Study | y | | | Date Received: | 07/29/24 | |
| Client Sample ID: | LOD-MDL-SOIL-01-QT3-2024 | | | | SDG No.: | P3390 | |
| Lab Sample ID: | P3390-01 | | | | Matrix: | SOIL | |
| Analytical Method | : 8015D DRO | | | | % Solid: | 100 De | canted: |
| Sample Wt/Vol: | 30.01 Units | s: g | | | Final Vol: | 1 | mL |
| Soil Aliquot Vol: | | uL | | | Test: | Diesel Range Org | anics |
| Extraction Type: | | | | | Injection Volume : | | |
| GPC Factor : | | PH : | | | | | |
| Prep Method : | SW3541 | | | | | | |
| File ID/Qc Batch: | Dilution: | Dilution: Prep Date | | | Date Analyzed | Prep Batch ID | |
| FG014001.D | 1 | 07/30/24 08:00 | | | 07/31/24 18:16 | PB162368 | |
| CAS Number | Parameter | Conc. | Qualifier | MDL | | LOQ / CRQL | Units(Dry Weight) |
| TARGETS DRO | DRO | 1010 | J | 185 | | 1670 | ug/kg |
| SURROGATES 16416-32-3 | Tetracosane-d50 | 16.4 | | 37 - 130 | | 82% | SPK: 20 |

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates > 25% difference for detected

concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

- J = Estimated Value
- B = Analyte Found in Associated Method Blank
- N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

 $\mathbf{S}=\mathbf{Indicates}$ estimated value where valid five-point calibration

was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit